

Gopal B. Saha

Physics and Radiobiology of Nuclear Medicine

Third Edition

 Springer

Physics and Radiobiology of Nuclear Medicine

Third Edition

Gopal B. Saha, Ph.D.

*Department of Molecular and Functional Imaging
The Cleveland Clinic Foundation*

Physics and Radiobiology of Nuclear Medicine

Third Edition

With 111 Figures

 Springer

Gopal B. Saha, Ph.D.
Department of Molecular and Functional Imaging
The Cleveland Clinic Foundation
Cleveland, OH 44195
USA

Library of Congress Control Number: 2005937015

ISBN-10: 0-387-30754-0 e-ISBN 0-387-30754-0
ISBN-13: 978-0387-30754-1

Printed on acid-free paper.

© 2006, 2001, 1993 Springer Science+Business Media, Inc.

All rights reserved. This work may not be translated or copied in whole or in part without the written permission of the publisher (Springer Science+Business Media, Inc., 233 Spring Street, New York, NY 10013, USA), except for brief excerpts in connection with reviews or scholarly analysis. Use in connection with any form of information storage and retrieval, electronic adaptation, computer software, or by similar or dissimilar methodology now known or hereafter developed is forbidden.

The use in this publication of trade names, trademarks, service marks, and similar terms, even if they are not identified as such, is not to be taken as an expression of opinion as to whether or not they are subject to proprietary rights.

Printed in the United States of America. (BS/EB)

9 8 7 6 5 4 3 2 1

springer.com

*To
All my benefactors*

Preface

A new edition of a book is always warranted when it needs to be updated because of advances in the field over time. Although the basics of physics, instrumentation, and radiobiology have not changed, their technological applications have been changing and improving continually. Nuclear medicine professionals worldwide appreciate the book so much that the previous edition has been published in Japanese. Changes in content and appreciation of the book are the two guiding factors in writing this third edition.

Like the previous editions, the book is aimed at residents taking the American Board of Nuclear Medicine, the American Board of Radiology (Physics part), and the American Board of Radiology with Special Competency in Nuclear Medicine examinations, and for the technologists taking the Nuclear Medicine Technology Certifying Board.

The book contains 16 chapters, and at the end of each chapter, references and suggested readings have been updated and new questions have been added where appropriate. The first 10 chapters have only minor changes because of the basic nature of the contents. A section on the chi-square test and evaluation of diagnostic tests has been added in Chapter 4. Additional radionuclides have been included in Table 5.1. In Chapter 8, the section on scintillation detectors has been rearranged and the section on dead time has been expanded. In Chapter 10, the sections on uniformity, gamma camera tuning, and quality control tests have been revised. A new section on software and DICOM has been added in Chapter 11. Chapter 12 has been revised to include SPECT/CT cameras with only minimal change in the image reconstruction by the filtered backprojection, but with a detailed description of iterative methods. Chapter 13 is a new chapter on positron emission tomography. Table 14.1 has been updated with new dose values. Chapter 15 has been expanded to include more information on cellular damage by radiation, and also a section on dirty bomb and radiation phobia. In Chapter 16, new revised NRC regulations in 10CFR35 have been added along with a new section on European regulations governing radiopharmaceuticals.

I would like to thank the members of our department for their assistance in many ways. I am thankful and grateful to Mrs. Rita Konyves for typing meticulously and conscientiously the major part of the manuscript and to Mrs. Diane Griffis for completing it.

My sincere thanks and gratitude are due to Robert Albano, Senior Clinical Medical Editor of Springer, for his constant support and encouragement, and to others at Springer for their help in the successful completion of the book.

Gopal B. Saha

Contents

Preface	vii
<i>Chapter 1</i> Structure of Matter	1
Matter and Energy	1
Radiation	1
The Atom	3
Electronic Structure of the Atom	3
Structure of the Nucleus	6
Nuclear Binding Energy	7
Nuclear Nomenclature	8
Chart of the Nuclides	8
Questions	10
Suggested Readings	10
<i>Chapter 2</i> Radioactive Decay	11
Spontaneous Fission	11
Isomeric Transition	12
Gamma (γ)-Ray Emission	12
Internal Conversion	13
Alpha (α)-Decay	14
Beta (β^-)-Decay	15
Positron (β^+)-Decay	17
Electron Capture	19
Questions	20
Suggested Readings	20
<i>Chapter 3</i> Kinetics of Radioactive Decay	21
Radioactive Decay Equations	21
General Equation	21
Half-Life	22
Mean Life	25
Effective Half-life	25

Units of Radioactivity	26
Specific Activity	26
Calculation	27
Successive Decay Equations	29
General Equation	29
Transient Equilibrium	30
Secular Equilibrium	32
Questions	32
Suggested Readings	33
<i>Chapter 4</i> Statistics of Radiation Counting	34
Error, Accuracy, and Precision	34
Mean and Standard Deviation	35
Gaussian Distribution	35
Standard Deviation of Count Rates	37
Propagation of Errors	37
Chi-Square Test	39
Minimum Detectable Activity	41
Evaluation of Diagnostic Tests	41
Questions	43
Suggested Readings	43
<i>Chapter 5</i> Production of Radionuclides	44
Cyclotron-Produced Radionuclides	44
Reactor-Produced Radionuclides	46
Fission or (n, f) Reaction	47
Neutron Capture or (n, γ) Reaction	47
Target and Its Processing	49
Equation for Production of Radionuclides	49
Radionuclide Generators	51
⁹⁹ Mo– ^{99m} Tc Generator	53
Questions	54
Suggested Readings	55
<i>Chapter 6</i> Interaction of Radiation with Matter	56
Interaction of Charged Particles with Matter	56
Specific Ionization	57
Linear Energy Transfer	57
Range	58
Bremsstrahlung	60
Annihilation	60
Interaction of γ -Radiations with Matter	60
Mechanism of Interaction of γ -Radiations	60
Attenuation of γ -Radiations	64
Interaction of Neutrons with Matter	68

Questions	68
Suggested Readings	70
<i>Chapter 7 Gas-Filled Detectors</i>	<i>71</i>
Principles of Gas-Filled Detectors	71
Ionization Chambers	74
Cutie Pie Survey Meter	74
Dose Calibrator	74
Pocket Dosimeter	77
Geiger–Müller Counters	77
Questions	79
Suggested Readings	80
<i>Chapter 8 Scintillation and Semiconductor Detectors</i>	<i>81</i>
Scintillation Detectors	81
Solid Scintillation Detectors	82
Semiconductor Detectors	84
Solid Scintillation Counters	86
NaI(Tl) Detector	86
Photomultiplier Tube	86
Preamplifier	87
Linear Amplifier	87
Pulse-Height Analyzer	87
Display or Storage	88
Gamma-Ray Spectrometry	88
Photopeak	89
Compton Valley, Edge, and Plateau	89
Characteristic X-Ray Peak	90
Backscatter Peak	91
Iodine Escape Peak	91
Annihilation Peak	92
Coincidence Peak	92
Liquid Scintillation Counters	93
Characteristics of Counting Systems	95
Energy Resolution	95
Detection Efficiency	96
Dead Time	99
Gamma Well Counters	101
Calibration of Well Counters	101
Counting in Well Counters	102
Effects of Sample Volume	103
Thyroid Probe	104
Thyroid Uptake Measurement	105
Questions	105
Suggested Readings	107

<i>Chapter 9</i>	Gamma Cameras	108
	Gamma Cameras	108
	Principles of Operation	108
	Detector	110
	Collimator	111
	Photomultiplier Tube	112
	X-, Y-Positioning Circuit	112
	Pulse-Height Analyzer	114
	Display and Storage	114
	Digital Cameras	115
	Solid State Digital Cameras	116
	Questions	116
	Suggested Readings	117
<i>Chapter 10</i>	Performance Parameters of Gamma Cameras	118
	Spatial Resolution	118
	Intrinsic Resolution	118
	Collimator Resolution	119
	Scatter Resolution	121
	Evaluation of Spatial Resolution	122
	Bar Phantom	122
	Line-Spread Function	124
	Modular Transfer Function	125
	Sensitivity	127
	Collimator Efficiency	128
	Uniformity	129
	Pulse-Height Variation	129
	Nonlinearity	130
	Edge Packing	131
	Gamma Camera Tuning	131
	Effects of High Counting Rates	132
	Contrast	132
	Quality Control Tests for Gamma Cameras	133
	Daily Checks	134
	Weekly Checks	136
	Annual or As-Needed Checks	136
	Questions	136
	References and Suggested Readings	138
<i>Chapter 11</i>	Digital Computers in Nuclear Medicine	139
	Basics of a Computer	139
	Central Processing Unit	140
	Computer Memory	141
	External Storage Devices	141
	Input/Output Devices	141

Operation of a Computer	142
Digitization of Analog Data	142
Digital-to-Analog Conversion	143
Digital Images	143
Application of Computers in Nuclear Medicine	144
Digital Data Acquisition	144
Static Study	146
Dynamic Study	146
Gated Study	147
Reconstruction of Images	147
Superimposition and Subtraction of Images	147
Display	148
Software and DICOM	149
PACS	150
Questions	152
Suggested Readings	152
<i>Chapter 12</i> Single Photon Emission Computed Tomography	153
Tomographic Imaging	153
Single Photon Emission Computed Tomography	153
Data Acquisition	155
Image Reconstruction	156
SPECT/CT	169
Factors Affecting SPECT	171
Performance of SPECT Cameras	177
Spatial Resolution	177
Sensitivity	177
Other Parameters	177
Quality Control Tests for SPECT Cameras	178
Daily Tests	178
Weekly Tests	178
Quality Control Tests for CT Scanners	179
Questions	180
References and Suggested Readings	180
<i>Chapter 13</i> Position Emission Tomography	182
Positron-Emitting Radionuclides	182
Detector for PET	183
PM Tubes and Pulse-Height Analyzers	183
PET Scanners	184
Block Detectors	184
Coincidence Timing Window	184
PET/CT Scanners	189
Mobile PET or PET/CT	190
Micro-PET	190

Hybrid Gamma Cameras	191
Data Acquisition	191
Two-Dimensional Versus Three-Dimensional Data	
Acquisition	194
Image Reconstruction	196
Factors Affecting PET	196
Normalization	196
Photon Attenuation Correction	197
Random Coincidences	199
Scatter Coincidences	200
Dead Time	200
Radial Elongation	200
Performance of PET Scanners	201
Spatial Resolution	201
Sensitivity	204
Noise Equivalent Count Rate	204
Quality Control Tests for PET Scanners	205
Daily Tests	205
Weekly Tests	205
Questions	206
References and Suggested Readings	206
<i>Chapter 14 Internal Radiation Dosimetry</i>	208
Radiation Units	208
Dose Calculation	211
Radiation Dose Rate	211
Cumulative Radiation Dose	212
Radiation Dose in SI Units	216
Effective Dose Equivalent and Effective Dose	220
Pediatric Dosages	222
Questions	223
Suggested Readings	224
<i>Chapter 15 Radiation Biology</i>	226
The Cell	226
Effects of Radiation	230
DNA Molecule	230
Chromosome	231
Direct and Indirect Actions of Radiation	235
Radiosensitivity of Cells	237
Cell Survival Curves	238
Factors Affecting Radiosensitivity	241
Dose Rate	241
Linear Energy Transfer	242
Chemicals	242

Stage of Cell Cycle	245
Classification of Radiation Damage	245
Stochastic and Deterministic Effects	246
Acute Effects of Total Body Irradiation	247
Hemopoietic Syndrome	247
Gastrointestinal Syndrome	248
Cerebrovascular Syndrome	248
Long-Term Effects of Radiation	249
Somatic Effects	249
Genetic Effects	256
Risk Versus Benefit in Diagnostic Radiology and Nuclear Medicine	258
Risk to Pregnant Women	259
Dirty Bombs	260
Types of Radiation Exposure	261
Protective Measures in Case of Explosion of a Dirty Bomb	262
Verification Card for Radioactive Patients	262
Radiation Phobia	263
Questions	265
References and Suggested Readings	267
 <i>Chapter 16</i> Radiation Regulations and Protection	268
Sources of Radiation Exposure	268
License	270
Radiation Protection	271
Definition of Terms	271
Caution Signs and Labels	272
Occupational Dose Limits	273
ALARA Program	274
Principles of Radiation Protection	274
Personnel Monitoring	277
Dos and Don'ts in Radiation Protection Practice	279
Bioassay	279
Receiving and Monitoring Radioactive Packages	280
Radioactive Waste Disposal	280
Radioactive Spill	282
Recordkeeping	282
Medical Uses of Radioactive Materials	283
Applications, Amendments, and Notifications	283
Authority and Responsibilities of the Licensee	284
Supervision	284
Mobile Nuclear Medicine Service	285
Written Directives	285
Measurement of Dosages	285

Calibration, Transmission, and Reference Sources	286
Requirement for Possession of Sealed Sources	286
Labeling of Vials and Syringes	286
Surveys of Ambient Radiation Exposure Rate	287
Calibration of Survey Instruments	287
Training and Experience Requirements for Medical Uses of By-Product Materials	287
Report and Notification of a Medical Event	288
Report and Notification of a Dose to an Embryo/Fetus or a Nursing Child	289
Release of Patients Administered with Radiopharmaceuticals	290
Recordkeeping	291
Transportation of Radioactive Materials	293
European Regulations Governing Radiopharmaceuticals	295
Questions	297
References and Suggested Readings	298
<i>Appendix A</i> Units and Constants	300
<i>Appendix B</i> Terms Used in Texts	302
<i>Appendix C</i> Answers to Questions	307
Index	308

1 Structure of Matter

Matter and Energy

The existence of the universe is explained by two entities: matter and energy. These two entities are interchangeable and exist in different forms to make up all things visible or invisible in the universe. Whereas matter has a definite size, shape, and form, energy has different forms but no size and shape.

Matter is characterized by its quantity, called the *mass*, and is composed of the smallest unit, the atom. In atomic physics, the unit of mass is the atomic mass unit (amu), which is equal to 1.66×10^{-27} kg.

Energy is the capacity to do work and can exist in several forms: kinetic energy (which is due to the motion of matter); potential energy (which is due to the position and configuration of matter); thermal energy (which is due to the motion of atoms or molecules in matter); electrical energy (which is due to the flow of electrons across an electric potential); chemical energy (which is due to chemical reaction); and radiation (energy in motion). Energy can change from one form to another. Of all these forms, radiation is of great importance in nuclear medicine and, therefore, will be discussed in detail.

Mass and energy are interchangeable, and one is created at the expense of the other. This is predicted by the Einstein's mass–energy relationship:

$$E = mc^2 \quad (1.1)$$

where E is energy in ergs, m is the mass in grams, and c is the velocity of light in a vacuum given as 3×10^{10} cm/sec. This relationship states that everything around us can be classified as matter or energy.

Radiation

Radiation is a form of energy in motion through space. It is emitted by one object and absorbed or scattered by another. Radiations are of two types:

1. *Particulate radiations*: Examples of these radiations are energetic electrons, protons, neutrons, α -particles, and so forth. They have mass and charge, except neutrons, which are neutral particles. The velocity of their motion depends on their kinetic energy. The particulate radiations originate from radioactive decay, cosmic rays, nuclear reactions, and so forth.

2. *Electromagnetic radiations*: These radiations are a form of energy in motion that does not have mass and charge and can propagate as either waves or discrete packets of energy, called the *photons* or *quanta*. These radiations travel with the velocity of light. Various examples of electromagnetic radiations include radio waves, visible light, heat waves, γ -radiations, and so forth, and they differ from each other in wavelength and hence in energy. Note that the sound waves are not electromagnetic radiations.

The energy E of an electromagnetic radiation is given by

$$E = h\nu = \frac{hc}{\lambda} \quad (1.2)$$

where h is the Planck constant given as 6.625×10^{-27} erg·s/cycle, ν is the frequency in hertz (Hz), defined as 1 cycle per second, λ is the wavelength in centimeters, and c is the velocity of light in vacuum, which is equal to nearly 3×10^{10} cm/s.

The energy of an electromagnetic radiation is given in electron volts (eV), which is defined as the energy acquired by an electron when accelerated through a potential difference of 1 volt. Using $1 \text{ eV} = 1.602 \times 10^{-12}$ erg, Eq. (1.2) becomes

$$E(\text{eV}) = \frac{1.24 \times 10^{-4}}{\lambda} \quad (1.3)$$

where λ is given in centimeters. Table 1.1 lists the different electromagnetic radiations along with their frequencies and wavelengths.

TABLE 1.1. Characteristics of different electromagnetic radiations.

Type	Energy (eV)	Frequency (Hz)	Wavelength (cm)
Radio, TV	10^{-10} – 10^{-6}	10^4 – 10^8	10^2 – 10^6
Microwave	10^{-6} – 10^{-2}	10^8 – 10^{12}	10^{-2} – 10^2
Infrared	10^{-2} –1	10^{12} – 10^{14}	10^{-4} – 10^{-2}
Visible	1–2	10^{14} – 10^{15}	10^{-5} – 10^{-4}
Ultraviolet	2–100	10^{15} – 10^{16}	10^{-6} – 10^{-5}
x-Rays and γ -rays	100 – 10^7	10^{16} – 10^{21}	10^{-11} – 10^{-6}

TABLE 1.2. Characteristics of electrons and nucleons.

Particle	Charge	Mass (amu)*	Mass (kg)	Mass (MeV) [†]
Electron	-1	0.000549	0.9108×10^{-30}	0.511
Proton	+1	1.00728	1.6721×10^{-27}	938.78
Neutron	0	1.00867	1.6744×10^{-27}	939.07

* amu = 1 atomic mass unit = 1.66×10^{-27} kg = 1/12 of the mass of ^{12}C .

[†] 1 atomic mass unit = 931 MeV.

The Atom

For the purpose of this book, the atom can be considered as the smallest unit in the composition of matter. The atom is composed of a nucleus at the center and one or more electrons orbiting around the nucleus. The nucleus consists of protons and neutrons, collectively called *nucleons*. The protons are positively charged particles with a mass of 1.00728 amu, and the neutrons are electrically neutral particles with a mass of 1.00867 amu. The electrons are negatively charged particles with a mass of 0.000549 amu. The protons and neutrons are about 1836 times heavier than the electrons but the neutron is heavier than the proton by one electron mass (i.e., by 0.511 MeV). The number of electrons is equal to the number of protons, thus resulting in a neutral atom of an element. The characteristics of these particles are given in Table 1.2. The size of the atom is about 10^{-8} cm (called the angstrom, Å), whereas the nucleus has the size of 10^{-13} cm (termed the fermi, F). The density of the nucleus is of the order of 10^{14} g/cm³. The electronic arrangement determines the chemical properties of an element, whereas the nuclear structure dictates the stability and radioactive transformation of the atom.

Electronic Structure of the Atom

Several theories have been put forward to describe the electronic structure of the atom, among which the theory of Niels Bohr, proposed in 1913, is the most plausible one and still holds today. The Bohr's atomic theory states that electrons rotate around the nucleus in discrete energy shells that are stationary and arranged in increasing order of energy. These shells are designated as the *K* shell, *L* shell, *M* shell, *N* shell, and so forth. When an electron jumps from the upper shell to the lower shell, the difference in energy between the two shells appears as electromagnetic radiations or photons. When an electron is raised from the lower shell to the upper shell, the energy difference is absorbed and must be supplied for the process to occur.

The detailed description of the Bohr's atomic structure is provided by the quantum theory in physics. According to this theory, each shell is designated by a quantum number n , called the *principal quantum number*, and

denoted by integers, for example, 1 for the *K* shell, 2 for the *L* shell, 3 for the *M* shell, 4 for the *N* shell, and 5 for the *O* shell. Each energy shell is subdivided into subshells or orbitals, which are designated as *s*, *p*, *d*, *f*, and so on. For a principal quantum number *n*, there are *n* orbitals in a given shell. These orbitals are assigned the *azimuthal quantum numbers*, *l*, which represent the electron's angular momentum and can assume numerical values of $l = 0, 1, 2, \dots, n - 1$. Thus for the *s* orbital, $l = 0$; the *p* orbital, $l = 1$; the *d* orbital, $l = 2$; the *f* orbital, $l = 3$; and so forth. According to this description, the *K* shell has one orbital, designated as $1s$, the *L* shell has two orbitals, designated as $2s$ and $2p$, and so forth. The orientation of the electron's magnetic moment in a magnetic field is described by the *magnetic quantum number*, *m*. The values of *m* can be $m = -l, -(l - 1), \dots, 0, \dots, (l - 1), l$. Each electron rotates about its own axis clockwise or anticlockwise, and the *spin quantum number*, *s* ($s = -1/2$ or $+1/2$) is assigned to each electron to specify this rotation.

The electron configuration of the atoms of different elements is governed by the following rules:

1. No two electrons can have the same values for all four quantum numbers in a given atom.
2. The orbital of the lowest energy will be filled in first, followed by the next higher energy orbital. The relative energies of the orbitals are $1s < 2s < 2p < 3s < 3p < 4s < 3d < 4p < 5s < 4d < 5p < 6s < 4f < 5d < 6p < 7s$. This order of energy is valid for lighter elements and is somewhat different in heavier elements.
3. There can be a maximum of $2(2l + 1)$ electrons in each orbital.
4. For given values of *n* and *l*, each of the available orbitals is first singly occupied such that no electron pairing occurs. Only when all orbitals are singly occupied does electron pairing take place.
5. Each energy shell contains a maximum of $2n^2$ electrons.

The hydrogen atom has one proton in the nucleus and one electron in the orbit. Its electronic structure is represented as $1s^1$. The helium atom has two electrons, which are accommodated in the $1s$ orbital, and thus has the structure of $1s^2$. Now let us consider the structure of $^{16}_8\text{O}$, which has eight electrons. The first two electrons will fill the $1s$ orbital. The next two electrons will go to the $2s$ orbital. There are three *p* orbitals, designated as p_x , p_y , p_z , which will be occupied by three electrons individually. The eighth electron will occupy the p_x orbital pairing with the electron already in it. Thus, the electronic configuration of $^{16}_8\text{O}$ is given by $1s^2 2s^2 2p^4$.

The electron configurations in different orbitals and shells are illustrated in Table 1.3, and the structure of $^{28}_{28}\text{Ni}$ is shown in Figure 1.1.

The electronic structure of the atom characterizes the chemical properties of elements. The outermost shell in the most stable and chemically inert elements such as neon, argon, krypton, and xenon has the electronic structure of $ns^2 np^6$. Helium, although a noble gas, has the $1s^2$ configuration.

TABLE 1.3. Electron configurations in different energy shells.

Principal shell	Principal quantum number (n)	Orbital (l)	No. of electrons = $2(2l + 1)$ in each orbital	$2n^2$
<i>K</i>	1	$s(0)$	2	2
<i>L</i>	2	$s(0)$	2	8
		$p(1)$	6	
<i>M</i>	3	$s(0)$	2	
		$p(1)$	6	18
		$d(2)$	10	
<i>N</i>	4	$s(0)$	2	
		$p(1)$	6	32
		$d(2)$	10	
		$f(3)$	14	
<i>O</i>	5	$s(0)$	2	50
		$p(1)$	6	
		$d(2)$	10	
		$f(3)$	14	
		$g(4)$	18	

Elements having electronic configurations different from that of the noble gases either lose or gain electrons to achieve the structure ns^2np^6 of the nearest noble gas atom. The electrons in these shells are called the *valence electrons* and are primarily responsible for the chemical bond formation.

Electrons in different shells are held by *binding energy* in different shells of the atom. The binding energy of an electron is defined as the energy that is required to be supplied to remove it completely from a shell. The binding energy of the electron is the greatest in the *K* shell and decreases with higher shells such as *L*, *M*, and so on. The binding energy also increases with increasing atomic number of the elements. Thus, the *K*-shell binding energy (21.05 keV) of technetium, with atomic number 43, is higher than the *K*-shell binding energy (1.08 keV) of sodium, with atomic number 11. The *K*-shell binding energy of electrons in several elements are: carbon, 0.28 keV, gallium, 10.37 keV, technetium, 21.05 keV; indium, 27.93 keV; iodine, 33.16 keV; lead, 88.00 keV.

When an electron is removed completely from an atom, the process is called *ionization*. The atom is said to be ionized and becomes an ion. On




FIG. 1.1. The electronic configuration of ${}_{28}\text{Ni}$. The *K* shell has 2 electrons, the *L* shell has 8 electrons, and the *M* shell has 18 electrons.

the other hand, when the electron is raised from a lower energy shell to an upper energy shell, the process is called *excitation*. Both ionization and excitation processes require a supply of energy from outside the atom such as heating, applying an electric field, and so forth. In the excited atoms, electrons jump from the upper energy shell to the lower energy shell to achieve stability. The difference in energy appears as electromagnetic radiations or photons. Thus, if the binding energy of *K*-shell electrons in, say, bromine is 13.5 keV and the *L*-shell binding energy is 1.8 keV, the transition of electrons from the *L* shell to the *K* shell will occur with the emission of 11.7 keV ($13.5 - 1.8 = 11.7$ keV) photons. As we shall see later, these radiations are called the *characteristic x-rays* of the product atom.

Structure of the Nucleus

As already stated, the nucleus of an atom is composed of protons and neutrons. The number of protons is called the *atomic number* of the element and denoted by Z . The number of neutrons is denoted by N , and the sum of the protons and neutrons, $Z + N$, is called the *mass number*, denoted by A . The symbolic representation of an element, X , is given by A_ZX_N . For example, sodium has 11 protons and 12 neutrons with a total of 23 nucleons. Thus, it is represented as ${}^{23}_{11}\text{Na}_{12}$. However, the atomic number Z of an element is known, and N can be calculated as $A - Z$; therefore, it suffices to simply write ${}^{23}\text{Na}$ (or Na-23).

To explain the various physical observations related to the nucleus of an atom, two models for the nuclear structure have been proposed: the liquid drop model and the shell model. The liquid drop model was introduced by Niels Bohr and assumes a spherical nucleus composed of closely packed nucleons. This model explains various phenomena, such as nuclear density, energetics of particle emission in nuclear reactions, and fission of heavy nuclei.

In the shell model, both protons and neutrons are arranged in discrete energy shells in a manner similar to the electron shells of the atom in the Bohr atomic theory. Similar to the electronic configuration of the noble gas atoms, nuclei with 2, 8, 20, 28, 50, 82, or 126 protons or neutrons are found to be very stable. These nucleon numbers are called the *magic numbers*.

It is observed that atomic nuclei containing an odd number of protons or neutrons are normally less stable than those with an even number of protons or neutrons. Thus, nuclei with even numbers of protons and neutrons are more stable, whereas those with odd numbers of protons and neutrons are less stable. For example, ${}^{12}\text{C}$ with six protons and six neutrons is more stable than ${}^{13}\text{C}$ containing six protons and seven neutrons.

There are about 270 stable atoms of naturally occurring elements. The stability of these elements is dictated by the configuration of protons and neutrons. The ratio of the number of neutrons to the number of protons (N/Z) is an approximate indicator of the stability of a nucleus. The N/Z ratio




FIG. 1.2. The plot of atomic number (Z) versus the number of neutrons (N) for all nuclides. The proton-rich nuclides fall on the left (dotted) and the neutron-rich nuclides fall on the right (cross-hatched) of the line of stability, indicated by the dark-shaded area. The solid line represents nuclides with $Z = N$.

is 1 in low- Z elements such as $^{12}_6\text{C}$, $^{14}_7\text{N}$, and $^{16}_8\text{O}$, but it increases with increasing atomic number of elements. For example, it is 1.40 for $^{127}_{53}\text{I}$ and 1.54 for $^{208}_{82}\text{Pb}$. The plot of the atomic number versus the neutron number of all nuclides is shown in Figure 1.2. All stable nuclear species fall on or around what is called the *line of stability*. The nuclear species on the left side of the line have fewer neutrons and more protons; that is, they are proton-rich. On the other hand, those on the right side of the line have fewer protons and more neutrons; that is, they are neutron-rich. The nuclides away from the line of stability are unstable and disintegrate to achieve stability.

Nuclear Binding Energy

According to the classical electrostatic theory, the nucleus of an atom cannot exist as a single entity, because of the electrostatic repulsive force among the protons in the nucleus. The stability of the nucleus is explained by the existence of a strong binding force called the *nuclear force*, which overcomes the repulsive force of the protons. The nuclear force is effective equally among all nucleons and exists only in the nucleus, having no influence outside the nucleus. The short range of the nuclear force leads to a very small size ($\sim 10^{-13}$ cm) and very high density ($\sim 10^{14}$ g/cm³) of the nucleus.

The mass M of a nucleus is always less than the combined masses of the nucleons A in the nucleus. The difference in mass ($M - A$) is termed the *mass defect*, which has been used as binding energy for all nucleons in the nucleus. The average binding energy of a nucleon is equal to the total binding energy (calculated from the mass defect) divided by the number of nucleons. It is of the order of 6–9 MeV, although the binding energy of an individual nucleon has a definite value, depending on the shell it occupies. The binding energy of a nucleon must be supplied to completely remove it from the nucleus. Note that whereas the binding energy of the nucleons is in the megaelectron volt (MeV) range, the electron binding energy in the atomic orbital is of the order of kiloelectron volts (keV), a factor of 1000 lower.

Nuclear Nomenclature

A *nuclide* is an atomic species with a definite number of protons and neutrons arranged in a definite order in the nucleus.

Radionuclides are those nuclides that are unstable and thus decay by emission of particles or electromagnetic radiations or by spontaneous fission.

Isotopes are the nuclides having the same atomic number Z but different mass number A . Isotopes exhibit the same chemical properties. Examples of carbon isotopes are ${}^{11}_6\text{C}$, ${}^{12}_6\text{C}$, and ${}^{13}_6\text{C}$.

Isotones are the nuclides having the same number of neutrons N but different numbers of protons. Examples of isotones are: ${}^{134}_{55}\text{Cs}$, ${}^{133}_{54}\text{Xe}$, and ${}^{132}_{53}\text{I}$, each having 79 neutrons.

Isobars are the nuclides with the same number of nucleons, that is, the same mass number A , but a different combination of protons and neutrons. For example: ${}^{82}\text{Y}$, ${}^{82}\text{Sr}$, ${}^{82}\text{Rb}$, and ${}^{82}\text{Kr}$ are all isobars having the mass number 82.

Isomers are the nuclides with the same number of protons and neutrons, but having different energy states and spins. ${}^{99}\text{Tc}$ and ${}^{99m}\text{Tc}$ are isomers of the same nuclide. Individual nuclides can exist in different energy states above the ground state due to excitation. These excited states are called the *isomeric states*, which can have a lifetime varying from picoseconds to years. When the isomeric states are long-lived, they are referred to as *metastable states*. These states are denoted by “m” as in ${}^{99m}\text{Tc}$.

Chart of the Nuclides

Nearly 3000 nuclides, both stable and unstable, are arranged in the form of a chart, called the *chart of the nuclides*, a section of which is presented in Figure 1.3. Each square in the chart represents a specific nuclide,

containing various information such as the half-life, type and energy of radiations, and so forth of the nuclide, and neutron capture cross section of the stable nuclide. The nuclides are arranged in increasing neutron number N horizontally and in increasing proton number Z vertically. Each horizontal group of squares contains all isotopes of the same element, whereas the vertical group contains all isotones with the same number of neutrons. For isomers, the square is subdivided into sections representing each isomer.

Questions

1. If a mass of matter (m) is converted to electromagnetic radiation, what should be the energy of this radiation?
2. Describe the Bohr's atomic theory in terms of the electronic configuration of the atom.
3. What is the difference between the orbital electron binding energy and the nuclear binding energy of an atom?
4. Define the mass defect and mass number of an atom. What does the mass defect account for?
5. Write the electronic configuration of ^{99m}Tc and ^{131}I .
6. How many electrons can the $3d$ orbital contain?
7. The electron binding energy of the K shell in an atom is higher than that of the L shell. True or false?
8. What is the difference between ionization and excitation of an atom?
9. What is a metastable state of a nuclide? How is it designated?

Suggested Readings

- Evans RD. *The Atomic Nucleus*. Malabar, FL: Kreiger; 1982.
- Friedlander G, Kennedy TW, Miller JM. *Nuclear and Radiochemistry*. 3rd ed. New York: Wiley; 1981.
- Turner JE. *Atoms, Radiation, and Radiation Protection*. 2nd ed. New York: Wiley; 1995.

2

Radioactive Decay

In 1896, Henri Becquerel first discovered natural radioactivity in potassium uranyl sulfate. Artificial radioactivity was not produced until 1934, when I. Curie and F. Joliot made boron, aluminum, and magnesium radioactive by bombarding them with α -particles from polonium. This introduction of artificial radioactivity prompted the invention of cyclotrons and reactors in which many radionuclides are now produced. So far, more than 2700 radionuclides have been artificially produced and characterized in terms of their physical properties.

Radionuclides are unstable and decay by emission of particle or γ radiation to achieve stable configuration of protons and neutrons in the nucleus. As already mentioned, the stability of a nuclide in most cases is determined by the N/Z ratio of the nucleus. Thus, as will be seen later, whether a nuclide will decay by a particular particle emission or γ -ray emission is determined by the N/Z and/or excitation energy of the nucleus. Radionuclides can decay by one or more of the six modes: *spontaneous fission, isomeric transition (IT), alpha (α) decay, beta (β^-) decay, positron (β^+) decay, and electron capture (EC) decay*. In all decay modes, energy, charge, and mass are conserved. Different decay modes of radionuclides are described later in detail.

Spontaneous Fission

Fission is a process in which a heavy nucleus breaks into two fragments accompanied by the emission of two or three neutrons. The neutrons carry a mean energy of 1.5 MeV and the process releases about 200 MeV energy that appears mostly as heat.

Spontaneous fission occurs in heavy nuclei, but its probability is low and increases with mass number of the nuclei. The half-life for spontaneous fission is 2×10^{17} years for ^{235}U and only 55 days for ^{254}Cf . As an alternative to the spontaneous fission, the heavy nuclei can decay by α -particle or γ -ray emission.

Isomeric Transition

As previously mentioned, a nucleus can exist in different energy or excited states above the ground state, which is considered as the state involving the arrangement of protons and neutrons with the least amount of energy. These excited states are called the *isomeric states* and have lifetimes of fractions of picoseconds to many years. When isomeric states are long-lived, they are referred to as *metastable states* and denoted by “m” as in ^{99m}Tc . An excited nucleus decays to a lower energy state by giving off its energy, and such transitions are called isomeric transitions (ITs). Several isomeric transitions may occur from intermediate excited states prior to reaching the ground state. As will be seen later, a parent radionuclide may decay to an upper isomeric state of the product nucleus by α -particle or β -particle emission, in which case the isomeric state returns to the ground state by one or more isomeric transitions. A typical isomeric transition of ^{99m}Tc is illustrated in Figure 2.1. Isomeric transitions can occur in two ways: gamma (γ)-ray emission and internal conversion.

Gamma (γ)-Ray Emission

The common mode of an isomeric transition from an upper energy state of a nucleus to a lower energy state is by emission of an electromagnetic radiation, called the γ -ray. The energy of the γ -ray emitted is the difference between the two isomeric states. For example, a decay of a 525-keV isomeric state to a 210-keV isomeric state will result in the emission of a 315-keV γ -ray.




FIG. 2.1. Isomeric transition of ^{99m}Tc . Ten percent of the decay follows internal conversion.




FIG. 2.2. Internal conversion process. The excitation energy of the nucleus is transferred to a K -shell electron, which is then ejected, and the K -shell vacancy is filled by an electron from the L shell. The energy difference between the L shell and K shell appears as the characteristic K x-ray. Alternatively, the characteristic K x-ray may transfer its energy to an L -shell electron, called the Auger electron, which is then ejected.

Internal Conversion

An alternative to the γ -ray emission is the *internal conversion* process. The excited nucleus transfers the excitation energy to an orbital electron—preferably the K -shell electron—of its own atom, which is then ejected from the shell, provided the excitation energy is greater than the binding energy of the electron (Fig. 2.2). The ejected electron is called the *conversion electron* and carries the kinetic energy equal to $E_\gamma - E_B$, where E_γ is the excitation energy and E_B is the binding energy of the electron. Even though the K -shell electrons are more likely to be ejected because of the proximity to the nucleus, the electrons from the L shell, M shell, and so forth also may undergo the internal conversion process. The ratio of the number of conversion electrons (N_e) to the number of observed γ -radiations (N_γ) is referred to as the *conversion coefficient*, given as $\alpha = N_e/N_\gamma$. The conversion coefficients are subscripted as $\alpha_K, \alpha_L, \alpha_M \dots$ depending on which shell the electron is ejected from. The total conversion coefficient α_T is then given by

$$\alpha_T = \alpha_K + \alpha_L + \alpha_M + \dots$$

Problem 2.1

If the total conversion coefficient (α_T) is 0.11 for the 140-keV γ -rays of ^{99m}Tc , calculate the percentage of 140-keV γ -radiations available for imaging.

Answer

$$\alpha_T = \frac{N_e}{N_\gamma} = 0.11$$

$$N_e = 0.11 N_\gamma$$

Total number of disintegrations

$$\begin{aligned}
 &= N_e + N_\gamma \\
 &= 0.11N_\gamma + N_\gamma \\
 &= 1.11N_\gamma
 \end{aligned}$$

Thus, the percentage of γ -radiations


$$\begin{aligned}
 &= \frac{N_\gamma}{1.11N_\gamma} \times 100 \\
 &= \frac{1}{1.11} \times 100 \\
 &= 90\%
 \end{aligned}$$

An internal conversion process leaves an atom with a vacancy in one of its shells, which is filled by an electron from the next higher shell. Such situations may also occur in nuclides decaying by electron capture (see later). When an L electron fills in a K -shell vacancy, the energy difference between the K shell and the L shell appears as a *characteristic K x-ray*. Alternatively, this transition energy may be transferred to an orbital electron, which is emitted with a kinetic energy equal to the characteristic x-ray energy minus its binding energy. These electrons are called *Auger electrons*, and the process is termed the *Auger process*, analogous to internal conversion. The Auger electrons are monoenergetic. Because the characteristic x-ray energy (energy difference between the two shells) is always less than the binding energy of the K -shell electron, the latter cannot undergo the Auger process and cannot be emitted as an Auger electron.

The vacancy in the shell resulting from an Auger process is filled by the transition of an electron from the next upper shell, followed by emission of similar characteristic x-rays and/or Auger electrons. The fraction of vacancies in a given shell that are filled by emitting characteristic x-ray emissions is called the *fluorescence yield*, and the fraction that is filled by the Auger processes is the *Auger yield*. The Auger process increases with the increasing atomic number of the atom.

Alpha (α)-Decay


The α -decay occurs mostly in heavy nuclides such as uranium, radon, plutonium, and so forth. Beryllium-8 is the only lightest nuclide that decays by breaking up into two α -particles. The α -particles are basically helium ions with two protons and two neutrons in the nucleus and two electrons removed from the helium atom. After α -decay, the atomic number of the nucleus is reduced by 2 and the mass number by 4.



The α -particles from a given radionuclide all have discrete energies corresponding to the decay of the initial nuclide to a particular energy level of the product (including, of course, its ground state). The energy of the α -particles is, as a rule, equal to the energy difference between the two levels and ranges from 1 to 10 MeV. The high-energy α -particles normally originate from the short-lived radionuclides and vice versa. The range of the α -particles is very short in matter and is approximately 0.03 mm in body tissue. The α -particles can be stopped by a piece of paper, a few centimeters of air, and gloves.

Beta (β^-)-Decay

When a radionuclide is neutron rich—that is, the N/Z ratio is greater than that of the nearest stable nuclide—it decays by the emission of a β^- -particle (note that it is an electron*) and an antineutrino, $\bar{\nu}$. In the β^- -decay process, a neutron is converted to a proton, thus raising the atomic number Z of the product by 1. Thus:



The difference in energy between the parent and daughter nuclides is called the *transition or decay energy*, denoted by E_{max} . The β^- -particles carry E_{max} or part of it, exhibiting a spectrum of energy as shown in Figure 2.3. The average energy of the β^- -particles is about one-third of E_{max} . This observation indicates that β^- -particles often carry only a part of the transition energy, and energy is not apparently conserved in β^- -decay. To satisfy the law of energy conservation, a particle called the *antineutrino*, $\bar{\nu}$, with no charge and a negligible mass has been postulated, which carries the remainder of E_{max} in each β^- -decay. The existence of antineutrinos has been proven experimentally.

After β^- -decay, the daughter nuclide may exist in an excited state, in which case, one or more γ -ray emissions or internal conversion will occur to dispose of the excitation energy. In other words, β^- -decay is followed by isomeric transition if energetically permitted.

The decay process of a radionuclide is normally represented by what is called the *decay scheme*. Typical decay schemes of ${}^{131}\text{I}$ and ${}^{99}\text{Mo}$ are shown in Figures 2.4 and 2.5, respectively. The β^- -decay is shown by a left-to-right arrow from the parent nuclide to the daughter nuclide, whereas the isomeric transition is displayed by a vertical arrow between the two states.

* The difference between a β^- -particle and an electron is that a β^- -particle originates from the nucleus, and an electron originates from the extranuclear electron orbitals.




FIG. 2.3. A typical energy spectrum of the β^- -particles of ^{32}P .

(Note: The β^+ -decay is shown by a two-step right-to-left arrow between the two states, the electron capture decay by a right-to-left arrow, and the α -decay by a down arrow). Although it is often said that ^{131}I emits 364-keV γ -rays, it should be understood that the 364-keV γ -ray belongs to ^{131}Xe as




FIG. 2.4. Decay scheme of ^{131}I . Eighty-one percent of the total ^{131}I radionuclides decay by 364-keV γ -ray emission. The 8.0-day half-life of ^{131}I is shown in parentheses.





FIG. 2.5. Decay scheme of ^{99}Mo . Approximately 87% of the total ^{99}Mo ultimately decays to $^{99\text{m}}\text{Tc}$, and the remaining 13% decays to ^{99}Tc . A 2-keV transition occurs from the 142-keV level to the 140-keV level. All the 2-keV γ -rays are internally converted. (The energy levels are not shown in scale.)

an isomeric state. This is true for all β^- , β^+ , or electron capture decays that are followed by γ -ray emission.

Some examples of β^- -decay follow:




It should be noted that in β^- -decay, the atomic number of the daughter nuclide is increased by 1 and the mass number remains the same.

Positron (β^+)-Decay

When a radionuclide is proton rich—that is, the N/Z ratio is low relative to that of the nearest stable nuclide—it can decay by positron (β^+) emission accompanied by the emission of a neutrino (ν), which is an opposite entity of the antineutrino. Positron emission takes place only when the energy difference (transition energy) between the parent and daughter nuclides is

greater than 1.02 MeV. In β^+ -decay, essentially a proton is converted to a neutron plus a positron, thus, decreasing the atomic number Z of the daughter nuclide by 1. Thus,



The requirement of 1.02 MeV for β^+ -decay arises from the fact that one electron mass has to be added to a proton to produce a neutron and one positron is created. Since each electron or positron mass is equal to 0.511 MeV, one electron and one positron are equal to 1.02 MeV, which is required as a minimum for β^+ -decay.

Some examples of β^+ -decay follow:



The energetic β^+ -particle loses energy while passing through matter. The range of positrons is short in matter. When it loses almost all of its energy, it combines with an atomic electron of the medium and is annihilated, giving rise to two photons of 511 keV emitted in opposite directions. These photons are called *annihilation radiations*.

The decay scheme of ${}^{68}\text{Ga}$ is presented in Figure 2.6. Note that the β^+ -decay is represented by a two-step right-to-left arrow.




Fig. 2.6. Decay scheme of ${}^{68}\text{Ga}$. The positrons are annihilated in medium to give rise to two 511-keV γ -rays emitted in opposite directions.










FIG. 2.7. Decay scheme of ^{111}In illustrating the electron capture process. The abundances of 171-keV and 245-keV γ -rays are 90% and 94%, respectively.

Electron Capture

Decay by electron capture (EC) is an alternative to the β^+ -decay for proton-rich radionuclides with N/Z lower than that of the stable nuclide. In EC decay, an electron from an extranuclear shell, particularly the K shell because of its proximity, is captured by a proton in the nucleus, forming a neutron accompanied by the emission of a neutrino for conservation of energy. Thus,



In this process, the atomic number of the daughter nuclide is lowered by 1. The EC process occurs usually in nuclides having excitation energy less than 1.02 MeV. In nuclides having excitation energy greater than 1.02 MeV, both EC and β^+ -decay can occur, although the probability of β^+ -decay increases with higher excitation energy. The decay scheme of ^{111}In is shown in Figure 2.7. The EC decay is indicated by a right-to-left arrow. Some examples of EC decay follow:



In EC decay, analogous to the situation in internal conversion, a vacancy is created in the shell from which the electron is captured. It is filled in by

the transition of an electron from the next upper shell, in which case the difference in energy between the two shells appears as a characteristic x-ray of the daughter nuclide. Also, as described earlier, instead of characteristic x-ray emission, the Auger process can occur, whereby an Auger electron is emitted.

Questions

1. What are the primary criteria for β^+ and β^- -decay?
2. If the energy difference between the proton-rich parent nuclide and the daughter nuclide is 1.2 MeV, could the parent radionuclide decay by β^+ -decay and/or electron capture? If the energy difference is 0.8 MeV, what should be the mode of decay?
3. If the total conversion coefficient (α_T) of 195-keV γ -rays of a radionuclide is 0.23, calculate the percentage of 195-keV photons available for imaging.
4. Can a K -shell electron be emitted as an Auger electron? Explain.
5. Explain how characteristic x-rays and Auger electrons are emitted.
6. Why is an antineutrino emitted in β^- -decay?
7. A K -shell electron is ejected by the internal conversion of a 155-keV γ -ray photon. If the binding energy of the K -shell electron is 25 keV, what is the kinetic energy of the electron?
8. What is the average energy of the β^- -particles emitted from a radionuclide?
9. Explain the production of annihilation radiations.

Suggested Readings

- Evans RD. *The Atomic Nucleus*. Malabar, FL: Kreiger; 1982.
- Friedlander G, Kennedy JW, Miller JM. *Nuclear and Radiochemistry*. 3rd ed. New York: Wiley; 1981.
- Turner JE. *Atoms, Radiation, and Radiation Protection*. 2nd ed. New York: Wiley; 1995.

3

Kinetics of Radioactive Decay

Radioactive Decay Equations

General Equation

As mentioned in Chapter 2, radionuclides decay by spontaneous fission, α -, β^- -, and β^+ -particle emissions, electron capture, or isomeric transition. The radioactive decay is a random process, and it is not possible to tell which atom from a group of atoms disintegrates at a specific time. Thus, one can only talk about the average number of radionuclides disintegrating during a period of time. This gives the disintegration rate of a particular radionuclide.

The disintegration rate of a radionuclide, that is, the number of disintegrations per unit time, given as $-dN/dt$, is proportional to the total number of radioactive atoms present at that time. Mathematically,

$$-dN/dt = \lambda N \quad (3.1)$$

where N is the number of radioactive atoms present, and λ is referred to as the *decay constant* of the radionuclide. As can be seen from Eq. (3.1), it is a small fraction of the radioactive atoms that decays in a very short period of time. The unit of λ is $(\text{time})^{-1}$. Thus, if λ is 0.2 sec^{-1} for a radionuclide, then 20% of the radioactive atoms present will disappear per second.

The disintegration rate $-dN/dt$ is referred to as the *radioactivity* or simply the *activity* of the radionuclide and denoted by A . It should be understood from Eq. (3.1) that the same amount of radioactivity means the same disintegration rate for any radionuclide, but the total number of atoms present and the decay constants differ for different radionuclides. For example, a radioactive sample A containing 10^6 atoms and with $\lambda = 0.01 \text{ min}^{-1}$ would give the same disintegration rate (10,000 disintegrations per minute) as that by a radioactive sample B containing 2×10^6 atoms and with a decay constant 0.005 min^{-1} .

Now from the preceding discussion, the following equation can be written:

$$A = \lambda N \quad (3.2)$$

From a knowledge of the decay constant and radioactivity of a radionuclide, one can calculate the total number of atoms or mass of the radionuclides present (using Avogadro's number $1 \text{ g} \cdot \text{atom} = 6.02 \times 10^{23} \text{ atoms}$).

Because Eq. (3.1) is a first-order differential equation, the solution of this equation by integration leads to

$$N_t = N_0 e^{-\lambda t} \quad (3.3)$$

where N_0 and N_t are the number of radioactive atoms at $t = 0$ and time t , respectively. Equation (3.3) is an exponential equation indicating that the radioactivity decays exponentially. By multiplying both sides of Eq. (3.3) by λ , one obtains

$$A_t = A_0 e^{-\lambda t} \quad (3.4)$$

The factor $e^{-\lambda t}$ is called the *decay factor*. The decay factor becomes $e^{+\lambda t}$ if the activity at time t before $t = 0$ is to be determined. The plot of activity versus time on a linear graph gives an exponential curve, as shown in Figure 3.1. However, if the activity is plotted against time on semilogarithmic paper, a straight line results, as shown in Figure 3.2.

Half-Life

Every radionuclide is characterized by a *half-life*, which is defined as the time required to reduce its initial activity to one half. It is usually denoted





FIG. 3.1. Plot of radioactivity versus time on a linear graph indicating an exponential curve.

FIG. 3.2. Plot of radioactivity against time on a semilogarithmic graph indicating a straight line. The half-life of the radionuclide can be determined from the slope of the line, which is given as the decay constant λ . Alternatively, an activity and half its value and their corresponding times are read from the plot. The difference in the two time readings gives the half-life.



by $t_{1/2}$ and is unique for a radionuclide. It is related to the decay constant λ of a radionuclide by

$$\lambda = \frac{0.693}{t_{1/2}} \quad (3.5)$$

From the definition of half-life, it is understood that A_0 is reduced to $A_0/2$ in one half-life; to $A_0/4$, that is, to $A_0/2^2$ in two half-lives; to $A_0/8$, that is, to $A_0/2^3$ in three half-lives; and so forth. In n half-lives of decay, it is reduced to $A_0/2^n$. Thus, the radioactivity A_t at time t can be calculated from the initial radioactivity A_0 by

$$A_t = \frac{A_0}{2^n} = \frac{A_0}{2^{(t/t_{1/2})}} = A_0(0.5)^{t/t_{1/2}} \quad (3.6)$$

where t is the time of decay. Here, $t/t_{1/2}$ can be an integer or a fraction depending on t and $t_{1/2}$. For example, a radioactive sample with $t_{1/2} = 3.2$ days decaying at a rate of 10,000 disintegrations per minute would give, after 7 days of decay, $10,000/2^{(7/3.2)} = 10,000/2^{2.2} = 10,000/4.59 = 2178$ disintegrations per minute.

It should be noted that ten half-lives of decay reduce the radioactivity by a factor of about 1000 ($2^{10} = 1024$), or to 0.1% of the initial activity.

The half-life of a radionuclide is determined by measuring the radioactivity at different time intervals and plotting them on semilogarithmic paper, as shown in Figure 3.2. An initial activity and half its value are read from the line, and the corresponding times are noted. The difference in time between the two readings gives the half-life of the radionuclide. For a very long-lived radionuclide, the half-life is determined by Eq. (3.2) from a knowledge of its activity and the number of atoms present. The number of atoms N can be calculated from the weight W of the radionuclide with




FIG. 3.3. A composite radioactive decay curve for a sample containing two radionuclides of different half-lives. The long-lived component (a) has a half-life of 27 hr and the short-lived component (b) has a half-life of 5.8 hr.

atomic weight A and Avogadro's number 6.02×10^{23} atoms per g-atom as follows:

$$N = \frac{W}{A} \times 6.02 \times 10^{23} \quad (3.7)$$

When two or more radionuclides are present in a sample, the measured count of such a sample comprises counts of all individual radionuclides. A semilogarithmic plot of the activity of a two-component sample versus time is shown in Figure 3.3. The half-life of each of the two radionuclides can be determined by what is called the *peeling or stripping method*. In this method, first, the tail part (second component) of the curve is extrapolated as a straight line up to the ordinate, and its half-life can be determined as mentioned previously (e.g., 27 hr). Second, the activity values on this line are subtracted from those on the composite line to obtain the activity values for the first component. A straight line is drawn through these points, and the half-life of the first component is determined (e.g., 5.8 hr).

The stripping method can be applied to more than two components in the similar manner.

Mean Life

Another relevant quantity of a radionuclide is its *mean life*, which is the average lifetime of a group of radionuclides. It is denoted by τ and is related to the decay constant λ and half-life $t_{1/2}$ as follows:

$$\tau = 1/\lambda \quad (3.8)$$

$$\tau = t_{1/2}/0.693 = 1.44 t_{1/2} \quad (3.9)$$

In one mean life, the activity of a radionuclide is reduced to 37% of its initial value.

Effective Half-Life

As already mentioned, a radionuclide decays exponentially with a definite half-life, which is called the *physical half-life*, denoted by T_p (or $t_{1/2}$). The physical half-life of a radionuclide is independent of its physicochemical conditions. Analogous to physical decay, radiopharmaceuticals administered to humans disappear exponentially from the biological system through fecal excretion, urinary excretion, perspiration, or other routes. Thus, after in vivo administration every radiopharmaceutical has a *biological half-life* (T_b), which is defined as the time needed for half of the radiopharmaceutical to disappear from the biologic system. It is related to decay constant λ_b by $\lambda_b = 0.693/T_b$.

Obviously, in any biologic system, the loss of a radiopharmaceutical is due to both the physical decay of the radionuclide and the biologic elimination of the radiopharmaceutical. The net or effective rate (λ_e) of loss of radioactivity is then related to λ_p and λ_b by

$$\lambda_e = \lambda_p + \lambda_b \quad (3.10)$$

Because $\lambda = 0.693/t_{1/2}$, it follows that

$$\frac{1}{T_e} = \frac{1}{T_p} + \frac{1}{T_b} \quad (3.11)$$

or,

$$T_e = \frac{T_p \times T_b}{T_p + T_b} \quad (3.12)$$

The effective half-life, T_e , is always less than the shorter of T_p or T_b . For a very long T_p and a short T_b , T_e is almost equal to T_b . Similarly, for a very long T_b and short T_p , T_e is almost equal to T_p .

Units of Radioactivity

The unit of radioactivity is a curie. It is defined as

$$\begin{aligned} 1 \text{ curie (Ci)} &= 3.7 \times 10^{10} \text{ disintegrations per second (dps)} \\ &= 2.22 \times 10^{12} \text{ disintegrations per minute (dpm)} \\ 1 \text{ millicurie (mCi)} &= 3.7 \times 10^7 \text{ dps} \\ &= 2.22 \times 10^9 \text{ dpm} \\ 1 \text{ microcurie } (\mu\text{Ci}) &= 3.7 \times 10^4 \text{ dps} \\ &= 2.22 \times 10^6 \text{ dpm} \end{aligned}$$

The System Internationale (SI) unit for radioactivity is the becquerel (Bq), which is defined as 1 dps. Thus,

$$\begin{aligned} 1 \text{ becquerel (Bq)} &= 1 \text{ dps} = 2.7 \times 10^{-11} \text{ Ci} \\ 1 \text{ kilobecquerel (kBq)} &= 10^3 \text{ dps} = 2.7 \times 10^{-8} \text{ Ci} \\ 2 \text{ megabecquerel (MBq)} &= 10^6 \text{ dps} = 2.7 \times 10^{-5} \text{ Ci} \\ 1 \text{ gigabecquerel (GBq)} &= 10^9 \text{ dps} = 2.7 \times 10^{-2} \text{ Ci} \\ 1 \text{ terabecquerel (TBq)} &= 10^{12} \text{ dps} = 27 \text{ Ci} \end{aligned}$$

Similarly,

$$\begin{aligned} 1 \text{ Ci} &= 3.7 \times 10^{10} \text{ Bq} = 37 \text{ GBq} \\ 1 \text{ mCi} &= 3.7 \times 10^7 \text{ Bq} = 37 \text{ MBq} \\ 1 \mu\text{Ci} &= 3.7 \times 10^4 \text{ Bq} = 37 \text{ kBq} \end{aligned}$$

Specific Activity

The presence of “cold,” or nonradioactive, atoms in a radioactive sample always induces competition between them in their chemical reactions or localization in a body organ, thereby compromising the concentration of the radioactive atoms in the organs. Thus, each radionuclide or radioactive sample is characterized by *specific activity*, which is defined as the radioactivity per unit mass of a radionuclide or a radioactive sample. For example, suppose that a 200-mg ^{125}I -labeled monoclonal antibody sample contains 350-mCi (12.95-GBq) ^{125}I radioactivity. Its specific activity would be $350/200 = 1.75 \text{ mCi/mg}$ or 64.75 MBq/mg . Sometimes, it is confused with concentration, which is defined as the radioactivity per unit volume of a sample. If a 10-ml radioactive sample contains 50 mCi (1.85 GBq), it will have a concentration of $50/10 = 5 \text{ mCi/ml}$ or 185 MBq/ml .

Specific activity is at times expressed as radioactivity per mole of a labeled compound, for example, mCi/mole (MBq/mole) or mCi/ μ mole (MBq/ μ mole) for ^3H -, ^{14}C -, and ^{35}S -labeled compounds.

The specific activity of a carrier-free (see Chapter 5) radionuclide sample is related to its half-life: the shorter the half-life, the higher the specific activity. The specific activity of a carrier-free radionuclide with mass number A and half-life $t_{1/2}$ in hours can be calculated as follows:

Suppose 1 mg of a carrier-free radionuclide is present in the sample.

$$\begin{aligned} \text{Number of atoms in the sample} &= \frac{1 \times 10^{-3}}{A} \times 6.02 \times 10^{23} \\ &= \frac{6.02 \times 10^{20}}{A} \end{aligned}$$

$$\text{Decay constant } \lambda = \frac{0.693}{t_{1/2} \times 60 \times 60} \text{ sec}^{-1}$$

Thus, disintegration rate $D = \lambda N$

$$\begin{aligned} &= \frac{0.693 \times 6.02 \times 10^{20}}{t_{1/2} \times A \times 60 \times 60} \\ &= \frac{1.1589 \times 10^{17}}{A \times t_{1/2}} \text{ dps} \end{aligned}$$

$$\begin{aligned} \text{Thus, specific activity (mCi/mg)} &= \frac{1.1589 \times 10^{17}}{A \times t_{1/2} \times 3.7 \times 10^7} \\ &= \frac{3.13 \times 10^9}{A \times t_{1/2}} \end{aligned} \quad (3.13)$$

where A is the mass number of the radionuclide, and $t_{1/2}$ is the half-life of the radionuclide in hours.

From Eq. (3.13), specific activities of carrier-free $^{99\text{m}}\text{Tc}$ and ^{131}I can be calculated as 5.27×10^6 mCi/mg (1.95×10^5 GBq/mg) and 1.25×10^5 mCi/mg (4.6×10^3 GBq/mg), respectively.

Calculation

Some examples related to the calculation of radioactivity and its decay follow:

Problem 3.1

Calculate the total number of atoms and total mass of ^{201}Tl present in 10 mCi (370 MBq) of ^{201}Tl ($t_{1/2} = 3.04$ d).

*Answer*For ^{201}Tl ,

$$\lambda = \frac{0.693}{3.04 \times 24 \times 60 \times 60} = 2.638 \times 10^{-6} \text{ sec}^{-1}$$

$$A = 10 \times 3.7 \times 10^7 = 3.7 \times 10^8 \text{ dps}$$

Using Eq. (3.2),

$$N = \frac{A}{\lambda} = \frac{3.7 \times 10^8}{2.638 \times 10^{-6}} = 1.40 \times 10^{14} \text{ atoms}$$

Because $1 \text{ g} \cdot \text{atom } ^{201}\text{Tl} = 201 \text{ g } ^{201}\text{Tl} = 6.02 \times 10^{23} \text{ atoms of } ^{201}\text{Tl}$ (Avogadro's number),

$$\begin{aligned} \text{Mass of } ^{201}\text{Tl in 10 mCi (370 MBq)} &= \frac{1.40 \times 10^{14} \times 201}{6.02 \times 10^{23}} \\ &= 46.7 \times 10^{-9} \text{ g} \\ &= 46.7 \text{ ng} \end{aligned}$$

Therefore, 10 mCi of ^{201}Tl contains 1.4×10^{14} atoms and 46.7 ng.

Problem 3.2

At 10:00 a.m., the $^{99\text{m}}\text{Tc}$ radioactivity was measured as 150 mCi (5.55 GBq) on Wednesday. What was the activity at 6 a.m. and 3 p.m. on the same day ($t_{1/2}$ of $^{99\text{m}}\text{Tc} = 6 \text{ hr}$)?

Answer

Time from 6 a.m. to 10 a.m. is 4 hr:

$$\lambda \text{ for } ^{99\text{m}}\text{Tc} = \frac{0.693}{6} = 0.1155 \text{ hr}^{-1}$$

$$A_t = 150 \text{ mCi (5.55 GBq)}$$

$$A_0 = ?$$

Using Eq. (3.4)

$$150 = A_0 e^{+0.1155 \times 4}$$

$$A_0 = 150 \times e^{0.462}$$

$$= 150 \times 1.5872$$

$$= 238.1 \text{ mCi (8.81 GBq) at 6 a.m.}$$

Time from 10 a.m. to 3 p.m. is 5 hr:

$$A_0 = 150 \text{ mCi}$$

$$A_t = ?$$

Using Eq. (3.4)

$$\begin{aligned}
 A_t &= 150 \times e^{-0.1155 \times 5} \\
 &= 150 \times e^{-0.5775} \\
 &= 150 \times 0.5613 \\
 &= 84.2 \text{ mCi (3.1 GBq) at 3 p.m.}
 \end{aligned}$$

Problem 3.3

If a radionuclide decays at a rate of 30%/hr, what is its half-life?

Answer

$$\begin{aligned}
 \lambda &= 0.3 \text{ hr}^{-1} \\
 \lambda &= \frac{0.693}{t_{1/2}} \\
 t_{1/2} &= \frac{0.693}{\lambda} = \frac{0.693}{0.3} \text{ hr} = 2.31 \text{ hr}
 \end{aligned}$$

Problem 3.4

If 11% of ^{99m}Tc -labeled diisopropyliminodiacetic acid (DISIDA) is eliminated via renal excretion, 35% by fecal excretion, and 3.5% by perspiration in 5 hr from the human body, what is the effective half-life of the radio-pharmaceutical ($T_p = 6 \text{ hr}$ for ^{99m}Tc)?

Answer

$$\begin{aligned}
 \text{Total biological elimination} &= 11\% + 35\% + 3.5\% \\
 &= 49.5\% \text{ in 5 hr}
 \end{aligned}$$

Therefore, $T_b \approx 5 \text{ hr}$

$$T_p = 6 \text{ hr}$$

$$T_e = \frac{T_b \times T_p}{T_b + T_p} = \frac{5 \times 6}{5 + 6} = \frac{30}{11} = 2.7 \text{ hr}$$

Successive Decay Equations

General Equation

In the preceding section, we derived equations for the activity of any radionuclide that is decaying. Here we shall derive equations for the activity of a radionuclide that is growing from another radionuclide and at the same time is itself decaying.

If a parent radionuclide p decays to a daughter radionuclide d , which in turn decays to another radionuclide (i.e., $p \rightarrow d \rightarrow$), then the rate of growth of d becomes

$$\frac{dN_d}{dt} = \lambda_p N_p - \lambda_d N_d \quad (3.14)$$

By integration, Eq. (3.14) becomes

$$(A_d)_t = \lambda_d N_d = \frac{\lambda_d (A_p)_0}{\lambda_d - \lambda_p} (e^{-\lambda_p t} - e^{-\lambda_d t}) \quad (3.15)$$

Equation (3.15) gives the activity of the daughter nuclide d at time t as a result of growth from the parent nuclide p and also due to the decay of the daughter itself.

Transient Equilibrium

If $\lambda_d > \lambda_p$, that is, $(t_{1/2})_d < (t_{1/2})_p$, then $e^{-\lambda_d t}$ in Eq. (3.15) is negligible compared to $e^{-\lambda_p t}$ when t is sufficiently long. Then Eq. (3.15) becomes

$$\begin{aligned} (A_d)_t &= \frac{\lambda_d (A_p)_0}{\lambda_d - \lambda_p} e^{-\lambda_p t} \\ &= \frac{\lambda_d (A_p)_t}{\lambda_d - \lambda_p} \end{aligned} \quad (3.16)$$

$$= \frac{(t_{1/2})_p (A_p)_t}{(t_{1/2})_p - (t_{1/2})_d} \quad (3.17)$$

This relationship is called the *transient equilibrium*. This equilibrium holds good when $(t_{1/2})_p$ and $(t_{1/2})_d$ differ by a factor of about 10 to 50. The semilogarithmic plot of this equilibrium equation is shown in Figure 3.4. The daughter nuclide initially builds up as a result of the decay of the parent nuclide, reaches a maximum, and then achieves the transient equilibrium decaying with an apparent half-life of the parent nuclide. At equilibrium, the ratio of the daughter to parent activity is constant. It can be seen from Eq. (3.17) that the daughter activity is always greater than the parent activity, because $(t_{1/2})_p / [(t_{1/2})_p - (t_{1/2})_d]$ is always greater than 1. The time to reach maximum daughter activity is given by the formula:

$$t_{\max} = \frac{1.44 \times (t_{1/2})_p \times (t_{1/2})_d \times \ln[(t_{1/2})_p / (t_{1/2})_d]}{[(t_{1/2})_p - (t_{1/2})_d]} \quad (3.18)$$

A typical example of transient equilibrium is ^{99}Mo ($t_{1/2} = 66$ hr) decaying to $^{99\text{m}}\text{Tc}$ ($t_{1/2} = 6$ hr). Because 87% of ^{99}Mo decays to $^{99\text{m}}\text{Tc}$, and the remaining 13% to the ground state, Eqs. (3.15), (3.16), and (3.17) must be multiplied by a factor of 0.87. Therefore, in the time–activity plot, the $^{99\text{m}}\text{Tc}$ daughter activity will be lower than the ^{99}Mo parent activity (Fig. 3.5). Also, the $^{99\text{m}}\text{Tc}$ activity reaches maximum in about 23 hr, i.e., $4(t_{1/2})_d$ [Eq. (3.18)].




FIG. 3.4. Plot of activity versus time on a semilogarithmic graph illustrating the transient equilibrium. Note that the daughter activity reaches a maximum, then transient equilibrium, and follows an apparent half-life of the parent. The daughter activity is higher than the parent activity at equilibrium.




FIG. 3.5. Plot of logarithm of ^{99}Mo and $^{99\text{m}}\text{Tc}$ activities versus time showing transient equilibrium. The activity of the daughter $^{99\text{m}}\text{Tc}$ is less than that of the parent ^{99}Mo , because only 87% of ^{99}Mo decays to $^{99\text{m}}\text{Tc}$. If 100% of the parent were to decay to the daughter, then the daughter activity would be higher than the parent activity after reaching equilibrium, as recognized from Eq. (3.17), and Figure 3.4.




FIG. 3.6. Plot of activity versus time illustrating secular equilibrium. In equilibrium, the daughter activity becomes equal to that of the parent.

Secular Equilibrium

When $\lambda_d \gg \lambda_p$, that is, when the parent half-life is much longer than that of the daughter nuclide, in Eq. (3.16), we can neglect λ_p compared to λ_d . Then Eq. (3.16) becomes

$$(A_d)_t = (A_p)_t \quad (3.19)$$

Equation (3.19) is called the *secular equilibrium*. This equilibrium holds when the half-life of the parent is much longer than that of the daughter nuclide by more than a factor of 100 or so. In secular equilibrium, both parent and daughter activities are equal, and both decay with the half-life of the parent nuclide. A semilogarithmic plot of activity versus time representing secular equilibrium is shown in Figure 3.6. Typical examples of secular equilibrium are ^{113}Sn ($t_{1/2} = 117$ days) decaying to $^{113\text{m}}\text{In}$ ($t_{1/2} = 100$ min), and ^{68}Ge ($t_{1/2} = 280$ days) decaying to ^{68}Ga ($t_{1/2} = 68$ min).

Questions

1. Calculate (a) the total number of atoms and (b) the total mass of ^{131}I present in a 30-mCi (1.11-GBq) ^{131}I sample ($t_{1/2} = 8.0$ days).
2. Calculate (a) the disintegration rate per minute and (b) the activity in curies and becquerels present in 1 mg of ^{201}Tl ($t_{1/2} = 73$ hr).

3. A radiopharmaceutical has a biologic half-life of 10hr in humans and a physical half-life of 23hr. What is the effective half-life of the radiopharmaceutical?
4. If the radioactivity of ^{111}In ($t_{1/2} = 2.8$ days) is 200mCi (7.4GBq) on Monday noon, what is the activity (a) at 10:00 a.m. the Friday before and (b) at 1:00 p.m. the Wednesday after?
5. How long will it take for a 10-mCi (370-MBq) sample of ^{123}I ($t_{1/2} = 13.2$ hr) and a 50-mCi (1.85-MBq) sample of $^{99\text{m}}\text{Tc}$ ($t_{1/2} = 6$ hr) to possess the same activity?
6. What is the time interval during which ^{111}In ($t_{1/2} = 2.8$ days) decays to 37% of the original activity?
7. For antibody labeling, 5mCi of $^{111}\text{InCl}_3$ is required. What amount of $^{111}\text{InCl}_3$ should be shipped if transportation takes 1 day?
8. What are the specific conditions of transient equilibrium and secular equilibrium?
9. How long will it take for the decay of 7/8 of an ^{18}F ($t_{1/2} = 110$ min) sample?
10. What fraction of the original activity of a radionuclide has decayed in a period equal to the mean life of the radionuclide?
11. A radioactive sample initially gives 8564cpm and 2 hr later gives 3000cpm. Calculate the half-life of the radionuclide.
12. If N atoms of a sample decay in one half-life, how many atoms would decay in the next half-life?
13. The ^{99}Mo ($t_{1/2} = 66$ hr) and $^{99\text{m}}\text{Tc}$ ($t_{1/2} = 6$ hr) are in transient equilibrium in a Moly generator. If 600mCi (22.2GBq) of ^{99}Mo is present in the generator, what would be the activity of $^{99\text{m}}\text{Tc}$ 132 hr later, assuming that 87% of ^{99}Mo decays to $^{99\text{m}}\text{Tc}$?
14. A radionuclide decays with a half-life of 10 days to a radionuclide whose half-life is 1.5hr. Approximately how long will it take for the daughter activity to reach a maximum?

Suggested Readings

- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia: W. B. Saunders; 2003.
- Friedlander G, Kennedy JW, Miller JM. *Nuclear Chemistry and Radiochemistry*. 3rd ed. New York: Wiley; 1981.
- Saha GB. *Fundamentals of Nuclear Pharmacy*. 5th ed. New York: Springer-Verlag; 2004.

4

Statistics of Radiation Counting

As mentioned in previous chapters, radioactive decay is a random process, and therefore one can expect fluctuations in the measurement of radioactivity. The detailed discussion of the statistical treatment of radioactive measurements is beyond the scope of this book. Only the salient points of statistics related to radiation counting are given here.

Error, Accuracy, and Precision

In the measurement of any quantity, an error in or deviation from the true value is likely to occur. Errors can be two types: *systematic* and *random*. Systematic errors appear as constant deviations and arise from malfunctioning instruments and inappropriate experimental conditions. These errors can be eliminated by rectifying the incorrect situations. Random errors are variable deviations and arise from the fluctuations in experimental conditions such as high-voltage fluctuations or statistical fluctuations in a quantity such as radioactive decay.

The *accuracy* of a measurement of a quantity indicates how closely it agrees with the “true” value. The *precision* of a series of measurements describes the reproducibility of the measurement, although the measurements may differ from the “average” or “mean” value. The closer the measurement is to the average value, the higher the precision, whereas the closer the measurement is to the true value, the more accurate the measurement. Remember that a series of measurements may be quite precise, but their average value may be far from the true value (i.e., less accurate). Precision can be improved by eliminating the random errors, whereas better accuracy is obtained by removing both the random and systematic errors.

Mean and Standard Deviation

When a series of measurements is made on a radioactive sample, the most likely value of these measurements is the *average*, or *mean* value, which is obtained by adding the values of all measurements divided by the number of measurements. The mean value is denoted by \bar{n} .

The standard deviation of a series of measurements indicates the deviation from the mean value and is a measure of the precision of the measurements. Radioactive decay follows the Poisson distribution law, from which one can show that if a radioactive sample gives an average count of \bar{n} , then its standard deviation σ is given by

$$\sigma = \sqrt{\bar{n}} \quad (4.1)$$

The mean of measurements is then expressed as

$$\bar{n} \pm \sigma$$

Gaussian Distribution

If a series of measurements are made repeatedly on a radioactive sample giving a mean count \bar{n} , then the distribution of counts would normally follow a Poisson distribution. If the number of measurements is large, the distribution can be approximated by a Gaussian distribution, illustrated in Figure 4.1. It can be seen that 68% of all measurements fall within one




FIG. 4.1. A Gaussian distribution of radioactive measurements. Note the 68% confidence level at $\pm 1\sigma$, 95% confidence level at $\pm 2\sigma$, and 99% confidence level at $\pm 3\sigma$.

standard deviation on either side of the mean, that is, within the range $\bar{n} \pm \sigma$; 95% of all measurements fall within the range $\bar{n} \pm 2\sigma$; and 99% fall within the range $\bar{n} \pm 3\sigma$. Also the Gaussian curve shows that half of the measurements are below the mean value, and the other half are above it.

The standard deviations in radioactive measurements indicate the statistical fluctuations of radioactive decay. For practical reasons, only single counts are obtained on radioactive samples instead of multiple repeat counts to determine the mean value. In this situation, if a single count n of a radioactive sample is large, then n can be estimated as close to \bar{n} ; that is, $\bar{n} = n$ and $\sigma = \sqrt{n}$. It can then be said that there is a 68% chance that the true value of the count falls within $n \pm \sigma$ or that the count n falls within one standard deviation of the true value (Fig. 4.1). This is called the 68% *confidence level*. That is, one is 68% confident that the count n is within one standard deviation of the true value. Similarly, 95% and 99% *confidence levels* can be set at two standard deviations (2σ) and three standard deviations (3σ), respectively, of any single radioactive count.

Another useful quantity in the statistical analysis of the counting data is the percent standard deviation, which is given as

$$\% \sigma = \frac{\sigma}{n} \times 100 = \frac{100\sqrt{n}}{n} = \frac{100}{\sqrt{n}} \quad (4.2)$$

Equation (4.2) indicates that as n increases, the $\% \sigma$ decreases, and hence, precision of the measurement increases. Thus, the precision of a count of a radioactive sample can be increased by accumulating a large number of counts in a single measurement. For example, for a count of 10,000, $\% \sigma$ is 1%, whereas for 1,000,000, $\% \sigma$ is 0.1%.

Problem 4.1

How many counts should be collected for a radioactive sample to have a 2% error at a 95% confidence level?

Answer

95% confidence level is 2σ , that is, $2\sqrt{n}$

$$2\% = \frac{2\sigma \times 100}{n} = \frac{2\sqrt{n} \times 100}{n}$$

Therefore,

$$\begin{aligned} 2 &= \frac{200}{\sqrt{n}} \\ \sqrt{n} &= 100 \\ n &= 10,000 \text{ counts} \end{aligned}$$

Standard Deviation of Count Rates

The standard deviation of a count rate is

$$\sigma_c = \frac{\sigma}{t}$$

where σ is the standard deviation of the total count n of a radioactive sample obtained in time t . Because n is equal to the count rate c times the counting time t ,

$$\sigma_c = \sqrt{n}/t = \sqrt{ct}/t = \sqrt{\frac{c}{t}} \quad (4.3)$$

Problem 4.2

A radioactive sample is counted for 12 min and gives 8640 counts. Calculate the count rate and its standard deviation for the sample.

Answer

$$\text{Count rate } c = \frac{8640}{12} = 720 \text{ counts per minute (cpm)}$$

Standard deviation,

$$\sigma_c = \sqrt{c/t} = \sqrt{720/12} \approx 8$$

Therefore, the count rate is 720 ± 8 cpm.

Propagation of Errors

Situations arise in which two quantities, say x and y , with their respective standard deviations, σ_x and σ_y , are either added, subtracted, multiplied, or divided. The standard deviations of the results of these arithmetic operations are given by the following expressions:

Addition:

$$\sigma_{x+y} = \sqrt{\sigma_x^2 + \sigma_y^2} \quad (4.4)$$

Subtraction:

$$\sigma_{x-y} = \sqrt{\sigma_x^2 + \sigma_y^2} \quad (4.5)$$

Multiplication:

$$\sigma_{(x \times y)} = (x \times y) \sqrt{(\sigma_x/x)^2 + (\sigma_y/y)^2} \quad (4.6)$$

Division:

$$\sigma_{(x/y)} = (x/y) \sqrt{(\sigma_x/x)^2 + (\sigma_y/y)^2} \quad (4.7)$$

Problem 4.3

A radioactive sample and the background were counted each for 5 min and found to give 8000 counts and 3000 counts, respectively. Calculate the net count rate, its standard deviation, and percent standard deviation.

Answer

$$\text{Gross sample count rate} = \frac{8000}{5} = 1600 \text{ cpm}$$

$$\text{Background count rate} = \frac{3000}{5} = 600 \text{ cpm}$$

$$\text{Net count rate} = 1600 - 600 = 1000 \text{ cpm}$$

Using Eqs. (4.3) and (4.4)

$$\begin{aligned} \sigma_c &= \sqrt{\left(\sqrt{\frac{1600}{5}}\right)^2 + \left(\sqrt{\frac{600}{5}}\right)^2} \\ &= \sqrt{\frac{1600}{5} + \frac{600}{5}} \\ &= \frac{1}{\sqrt{5}} \sqrt{1600 + 600} \\ &= \frac{1}{2.24} \sqrt{2200} \\ &= \frac{46.9}{2.24} \\ &\approx 21 \text{ cpm} \end{aligned}$$

Thus, the count rate of the sample is 1000 ± 21 cpm

$$\% \text{ standard deviation of the count rate} = \frac{21}{1000} \times 100 = 2.1\%$$

Problem 4.4

A thyroid patient is given a ^{131}I -NaI capsule to measure the 24-hr thyroid uptake. The 2-min counts are: standard, 90,000; room background, 1000; thyroid, 40,000; and thigh, 2000. Calculate the thyroid uptake and its percent standard deviation.

Answer

$$\text{Net standard count} = 90,000 - 1000 = 89,000$$

$$\begin{aligned} \sigma_s &= \sqrt{90,000 + 1000} \\ &= 302 \end{aligned}$$

$$\text{Net thyroid count} = 40,000 - 2000 = 38,000$$

$$\begin{aligned}\sigma_t &= \sqrt{40,000 + 2000} \\ &= 205\end{aligned}$$

Percent thyroid uptake

$$= \frac{38,000}{89,000} \times 100 = 42.7\%$$

Standard deviation in uptake [using Eq. (4.7)]

$$\begin{aligned}&= \frac{38,000}{89,000} \times \sqrt{\left(\frac{302}{89,000}\right)^2 + \left(\frac{205}{38,000}\right)^2} \\ &= 0.427 \times \sqrt{0.000011514 + 0.000029103} \\ &= 0.427 \times 0.006373 \\ &= 0.0027\end{aligned}$$

$$\begin{aligned}\% \text{ standard deviation in uptake} &= \frac{0.0027}{0.427} \times 100 \\ &= 0.63\%\end{aligned}$$

Thus, the thyroid uptake = $42.7 \pm 0.63\%$.

It should be noted that although all counts were taken for 2 min, count rates (cpm) were not used in the calculations. One can do similar calculations using count rates and obtain the same results.

Chi-Square Test

The chi-square (χ^2) test is a useful test for verifying if the variations in a set of measurements are due to statistical randomness of the data or due to variations in entities, such as equipment, patients, and the like, used in the measurements. The latter variations may be systematic, such as a fixed voltage drop throughout the measurement or random, such as fluctuations in voltage supply to the equipment. If there are N measurements made of a parameter, then for Gaussian distribution of the data, which is true in radioactive measurement, the χ^2 is given by

$$\chi^2 = \sum_i^N \frac{(X_i - \bar{X})^2}{\bar{X}} \quad (4.8)$$

where X_i is the value of the i th measurement and \bar{X} is the average of N measurements. The value of N should be 10 or larger. The χ^2 values are given in Table 4.1 for various probability (p) values as a function of degree of freedom, which is equal to the number of measurements minus one, $N - 1$.

TABLE 4.1. Critical chi-square values.

Degrees of Freedom ($N - 1$)	Probability						
	0.99	0.95	0.90	0.50	0.10	0.05	0.01
	That the calculated chi-square value will be equal to or greater than						
2	0.02	0.10	0.21	1.39	4.61	5.99	9.21
3	0.13	0.35	0.58	2.37	6.25	7.82	11.35
4	0.30	0.71	1.06	3.36	7.78	9.49	13.28
5	0.55	1.15	1.61	4.35	9.24	11.07	15.09
6	0.87	1.64	2.20	5.35	10.65	12.59	16.81
7	1.24	2.17	2.83	6.35	12.02	14.07	18.48
8	1.65	2.73	3.49	7.34	13.36	15.51	20.09
9	2.09	3.33	4.17	8.34	14.68	16.92	21.67
10	2.56	3.94	4.87	9.34	15.99	18.31	23.21
11	3.05	4.58	5.58	10.34	17.28	19.68	24.73
12	2.57	5.23	6.30	11.34	18.55	21.03	26.22
13	4.11	5.89	7.04	12.34	19.81	22.36	27.69
14	4.66	6.57	7.79	13.34	21.06	23.69	29.14
15	5.23	7.26	8.55	14.34	22.31	25.00	30.58
16	5.81	7.96	9.31	15.34	23.54	26.30	32.00
17	6.41	8.67	10.09	16.34	24.77	27.59	33.41
18	7.02	9.39	10.87	17.34	25.99	28.87	34.81
19	7.63	10.12	11.65	18.34	27.20	30.14	36.19
20	8.26	10.85	12.44	19.34	28.41	31.41	37.57
21	8.90	11.59	13.24	20.34	29.62	32.67	38.93
22	9.54	12.34	14.04	21.34	30.81	33.92	40.29
23	10.20	13.09	14.85	22.34	32.01	35.17	41.64
24	10.86	13.85	15.66	23.34	33.20	36.42	42.98
25	11.53	14.61	16.47	24.34	34.38	37.38	44.31
26	12.20	15.38	17.29	25.34	35.56	38.89	45.64
27	12.88	16.15	18.11	26.34	36.74	40.11	46.96
28	13.57	16.93	18.94	27.34	37.92	41.34	48.28
29	14.26	17.71	19.77	28.34	39.09	42.56	49.59

In Table 4.1, for a series of eight measurements, (degrees of freedom 7), there is a 50% chance of the value of chi-square being greater than 6.35 and equal probability of it's being smaller. Similarly, the probability of chi-square being as large as 12.02 is only 10%, and being as large as 18.48 is only 1%. Typically, χ^2 values that fall within 0.1 to 0.95 are acceptable. If the observed χ^2 value falls outside this range, it is an indication that the variation is beyond the statistical randomness of the data and something is wrong with the experimental set-up, for example, measuring equipment, measurement technique, and so on.

In performing the χ^2 test, a number of measurements (a minimum of 10) are made of the quantity, and the mean and χ^2 of the measured values are calculated by Eq. (4.8). The χ^2 value is then compared with the value in Table 4.1 corresponding to the actual degree of freedom and for a particular p value.

Problem 4.5

The following repeat counts of a radioactive sample were obtained in a well-counter. Use the χ^2 test to see if the variations in counts are due to statistical variations of radioactivity or the counter is not working properly.

4580	4263
4635	4481
4625	4356
4578	4699
4525	4344
4668	4483
4391	4529

Answer

The average value of 14 measurements is $\bar{N} = 4511$.

Using Eq. (4.8),

$$\chi^2 = \frac{227637}{4511} = 50.5$$

From Table 4.1, for degree of freedom 13 and the p -value of 0.01, $\chi^2 = 27.69$. The computed χ^2 far exceeds the theoretical value, so something in addition to the statistical fluctuations of the counts is operating. Most likely, the well-counter is not functioning properly.

Minimum Detectable Activity

The efficiency of different detectors is limited by the dead time at high count rates and by statistical fluctuations at low count rates of the backgrounds. In the latter situation, the minimum detectable activity (MDA) that gives a statistically significant count is given by

$$MDA = 3\sigma_R = 3\sqrt{R_B/t} \quad (4.9)$$

where σ_R is the standard deviation of the background count rate R_B obtained by counting over a period of time t . Equation (4.9) requires that the sample count rate must be at least three standard deviations of the background to be significant.

Evaluation of Diagnostic Tests

It is often required to evaluate the usefulness of a new diagnostic test to determine the presence or absence of a particular disease. This aspect of the test is commonly described by two entities: sensitivity and specificity. The sensitivity of a test is the probability of being able to identify correctly

TABLE 4.2. Distribution of data obtained by a new diagnostic test.

Test result	Disease present	Disease absent	Totals
+	TP	FP	N ⁺
-	FN	TN	N ⁻
TOTALS	N _p	N _A	N

those having the disease in a diseased population (true positive, TP). The specificity of the test is the probability of being able to identify correctly those who do not have the disease in a healthy population (true negative, TN). By these definitions, it is obvious that a given test may not identify all patients correctly whether or not they have the disease. This results in false-positives (FP) in the healthy group and false-negatives (FN) in the diseased group.

It should be noted that when sensitivity is assessed for a diseased population or specificity for a healthy group, the disease or healthy status of the group must be assessed by an established standard diagnostic test. This test is called the “gold standard” and is considered the best method available for comparison. As determined by the gold standard in a group of N patients, if N_p is the total number of persons with disease, and N_A is the total number of persons without disease, then the results of the new test can be summarized in Table 4.2. The following parameters can be obtained from data in Table 4.2.

$$\text{Sensitivity} = \frac{TP}{TP + FN} = \frac{TP}{N_p} \quad (4.10)$$

$$\text{Specificity} = \frac{TN}{TN + FP} = \frac{TN}{N_A} \quad (4.11)$$

$$\text{Accuracy} = \frac{TP + TN}{TP + TN + FP + FN} \quad (4.12)$$

$$\text{Positive predictive value} = \frac{TP}{TP + FP} \quad (4.13)$$

$$\text{Negative predictive value} = \frac{TN}{TN + FN} \quad (4.14)$$

Problem 4.6

In a group of 1000 patients, 840 patients (Group A) had brain tumor and 160 patients (Group B) did not have tumor by biopsies. The SPECT study diagnosed 780 patients having tumor in Group A and 15 patients having tumor in Group B. Calculate the sensitivity, specificity, accuracy, positive predictive value, and negative predictive value of the SPECT study.

Answer

True positive = 780

True negative = 160 - 15 = 145

False negative = 840 - 780 = 60

False positive = 15

$$\text{Sensitivity} = \frac{780}{780 + 60} = \frac{780}{840} \times 100 = 92.6\%$$

$$\text{Specificity} = \frac{145}{145 + 15} = \frac{145}{160} \times 100 = 90.6\%$$

$$\text{Accuracy} = \frac{780 + 145}{780 + 145 + 60 + 15} = \frac{925}{1000} \times 100 = 92.5\%$$

$$\text{Positive predictive value} = \frac{780}{780 + 15} = \frac{780}{795} \times 100 = 98.1\%$$

$$\text{Negative predictive value} = \frac{145}{145 + 60} = \frac{145}{205} \times 100 = 70.7\%$$

Questions

1. Define accuracy and precision.
2. Do systematic errors give an accurate measurement? Can systematic errors give a precise measurement?
3. A radioactive sample gives 15,360 counts in 9 min:
 - (a) What are the count rate of the sample and its standard deviation?
 - (b) If the sample contained a background count rate of 60 cpm obtained from a 2-min count, what would be the net count rate of the sample and its standard deviation?
4. How many counts of a sample are to be collected to have a 1% error at the 95% confidence level?
5. Within how many standard deviations of a mean count of 62,001 is 730?
6. To achieve an estimated percent standard error of 3%, how many counts must be collected?
7. The χ^2 value of 11 measurements of a quantity is 4.2. What is the probability that the variations of measurements are due to statistical variations of the quantity?

Suggested Readings

- Bahn AK. *Basic Medical Statistics*. New York: Grune & Stratton; 1972.
- Martin PM. Nuclear medicine statistics. In: Rollo FD, ed. *Nuclear Medicine Physics, Instrumentation and Agents*. St. Louis: Mosby; 1977:479-512.

5

Production of Radionuclides


More than 3000 nuclides are known, of which approximately 2700 are radioactive, and the rest are stable. The majority of radionuclides are artificially produced in the cyclotron and reactor. Some short-lived radionuclides are available from the so-called radionuclide generators in which long-lived parents are loaded and decay to short-lived daughters. The following is a brief description of the sources of different radionuclides.

Cyclotron-Produced Radionuclides

In a cyclotron (Fig. 5.1), charged particles (S) such as protons, deuterons, α -particles, ^3He -particles, and so forth are accelerated in circular paths within the dees (A and B) under vacuum by means of an electromagnetic field. These accelerated particles can possess a few kiloelectron volts (keV) to several billion electron volts (BeV) of kinetic energy depending on the design of the cyclotron. Because charged particles move along the circular paths under the magnetic field with gradually increasing energy, the larger the radius of the particle trajectory, the higher the kinetic energy of the particle. The charged particles are deflected by a deflector (D) through a window (W) outside the cyclotron to form an external beam.

When targets of stable elements are irradiated by placing them in the external beam of the accelerated particles or in the internal beam at a given radius inside a cyclotron, the accelerated particles interact with the target nuclei, and nuclear reactions take place. In a nuclear reaction, the incident particle may leave the nucleus after interaction with a nucleon, leaving some of its energy in it, or it may be completely absorbed by the nucleus, depending on the energy of the incident particle. In either case, a nucleus with excitation energy is formed and the excitation energy is disposed of by the emission of nucleons (i.e., protons and neutrons). Particle emission is followed by γ -ray emission when the former is no longer energetically feasible. Depending on the energy deposited by the incident particle, several nucleons are emitted randomly from the irradiated target nucleus,


FIG. 5.1. Schematics of a cyclotron. A and B, dees with vacuum; D, deflector; S, ion source; V, alternating voltage; W, window.



leading to the formation of different nuclides. As the energy of the irradiating particle is increased, more nucleons are emitted, and therefore a much wider variety of nuclides is produced.

Medical cyclotrons are compact cyclotrons that are used to produce routinely short-lived radionuclides, particularly those used in positron emission tomography. In these cyclotrons, protons, deuterons, and α -particles of low-to-medium energy are available. These units are available commercially and can be installed in a relatively small space.

An example of a typical cyclotron-produced radionuclide is ^{111}In , which is produced by irradiating ^{111}Cd with 12-MeV protons in a cyclotron. The nuclear reaction is written as follows:



where ^{111}Cd is the target nuclide, the proton p is the irradiating particle, the neutron n is the emitted particle, and ^{111}In is the product radionuclide. In this case, a second nucleon may not be emitted, because there is not enough energy left after the emission of the first neutron. The excitation energy that is not sufficient to emit any more nucleons will be dissipated by γ -ray emission.

As can be understood, radionuclides produced with atomic numbers different from those of the target isotopes do not contain any stable (“cold,” or “carrier”) isotope detectable by ordinary analytical methods, and such preparations are called *carrier-free*. In practice, however, it is impossible to have these preparations without the presence of any stable isotopes. Another term for these preparations is *no-carrier-added* (NCA), meaning that no stable isotope has been added purposely to the preparations.

The target material for irradiation must be pure and preferably monoisotopic or at least enriched isotopically to avoid the production of extraneous radionuclides. Because various isotopes of different elements may be produced in a target, it is necessary to isolate isotopes of a single element; this can be accomplished by appropriate chemical methods such as solvent extraction, precipitation, ion exchange, and distillation. Cyclotron-produced

radionuclides are usually neutron deficient and therefore decay by β^+ -emission or electron capture.

Reactor-Produced Radionuclides

A variety of radionuclides is produced in nuclear reactors. A nuclear reactor is constructed with fuel rods made of fissile materials such as enriched ^{235}U and ^{239}Pu . These fuel nuclei undergo spontaneous fission with extremely low probability. *Fission* is defined as the breakup of a heavy nucleus into two fragments of approximately equal mass, accompanied by the emission of two to three neutrons with mean energies of about 1.5 MeV. In each fission, there is a concomitant energy release of ~ 200 MeV that appears as heat and is usually removed by heat exchangers to produce electricity in the nuclear power plant.

Neutrons emitted in each fission can cause further fission of other fissionable nuclei in the fuel rod, provided the right conditions exist. This obviously will initiate a chain reaction, ultimately leading to a possible meltdown of the reactor core. This chain reaction must be controlled, which is in part accomplished by the proper size, shape, and mass of the fuel material and other complicated and ingenious engineering techniques. To maintain a self-sustained chain reaction, only one fission neutron is needed and excess neutrons (more than one) are removed by positioning cadmium rods, called *control rods*, in the reactor core (cadmium has a high probability of absorbing a thermal neutron).


The fuel rods of fissile materials are interspersed in the reactor core with spaces in between. Neutrons emitted with a mean energy of 1.5 MeV from the surface of the fuel rod have a low probability of interacting with other nuclei and therefore do not serve any useful purpose. It has been found, however, that neutrons with thermal energy (0.025 eV) interact with many other stable nuclei efficiently, producing various radionuclides. To make the high-energy neutrons, or so-called fast neutrons, more useful, they are thermalized or slowed down by interaction with low molecular weight materials, such as water, heavy water (D_2O), beryllium, and graphite (C), which are distributed in the spaces between the fuel rods. These materials are called *moderators*. The flux, or intensity, of the thermal neutrons so obtained ranges from 10^{11} to 10^{14} neutrons/cm²·sec, and they are useful in the production of many radionuclides. When a target element is inserted in the reactor core, a thermal neutron will interact with the target nucleus, with a definite probability of producing another nuclide. The probability of formation of a radionuclide by thermal neutrons varies from element to element.

In the reactor, two types of interaction with thermal neutrons occur to produce various radionuclides: fission of heavy elements and neutron capture or (n, γ) reaction. These two nuclear reactions are described next.

Fission or (n, f) Reaction

When a target of heavy elements is inserted in the reactor core, heavy nuclei absorb thermal neutrons and undergo fission. Fissionable heavy elements are ^{235}U , ^{239}Pu , ^{237}Np , ^{233}U , ^{232}Th , and many others having atomic numbers greater than 92. Fission of heavy elements may also be induced in a cyclotron by irradiation with high-energy charged particles, but fission probability depends on the type and energy of the irradiating particle. Nuclides produced by fission may range in atomic number from about 28 to nearly 65. These isotopes of different elements are separated by appropriate chemical procedures that involve precipitation, solvent extraction, ion exchange, chromatography, and distillation. The fission radionuclides are normally carrier-free or NCA, and therefore radionuclides of high specific activity are available from fission. The fission products are usually neutron rich and decay by β^- -emission.

Many clinically useful radionuclides such as ^{131}I , ^{99}Mo , ^{133}Xe , and ^{137}Cs are produced by fission of ^{235}U . An example of thermal fission of ^{235}U follows, showing a few representative radionuclides:



Many other nuclides besides those mentioned in the example are also produced.

Neutron Capture or (n, γ) Reaction

In neutron capture reaction, the target nucleus captures one thermal neutron and emits γ -rays to produce an isotope of the same element. The radionuclide so produced is therefore not carrier-free, and its specific activity is relatively low. This reaction takes place in almost all elements with varying probability. Some examples of neutron capture reactions are $^{98}\text{Mo}(n, \gamma)^{99}\text{Mo}$, $^{196}\text{Hg}(n, \gamma)^{197}\text{Hg}$, and $^{50}\text{Cr}(n, \gamma)^{51}\text{Cr}$. Molybdenum-99 so produced is called the *irradiated molybdenum* as opposed to the *fission molybdenum* described earlier. This method is commonly used in the analysis of trace elements in various samples.

The method of production and various characteristics of radionuclides commonly used in nuclear medicine are presented in Table 5.1.

TABLE 5.1. Characteristics of commonly used radionuclides.

Nuclide	Physical half-life	Mode of delay (%)	γ -ray energy* (MeV)	γ -ray abundance (%)	Common production method
${}^3_1\text{H}$	12.3 yr	β^- (100)	—	—	${}^6\text{Li}(n, \alpha){}^3\text{H}$
${}^{11}_6\text{C}$	20.4 min	β^+ (100)	0.511 (annihilation)	200	${}^{10}\text{B}(d, n){}^{11}\text{C}$ ${}^{14}\text{N}(p, \alpha){}^{11}\text{C}$
${}^{13}_7\text{N}$	10 min	β^+ (100)	0.511 (annihilation)	200	${}^{12}\text{C}(d, n){}^{13}\text{N}$ ${}^{16}\text{O}(p, \alpha){}^{13}\text{N}$ ${}^{13}\text{C}(p, n){}^{13}\text{N}$
${}^{14}_6\text{C}$	5730 yr	β^- (100)	—	—	${}^{14}\text{N}(n, p){}^{14}\text{C}$
${}^{15}_8\text{O}$	2 min	β^+ (100)	0.511 (annihilation)	200	${}^{14}\text{N}(d, n){}^{15}\text{O}$ ${}^{15}\text{N}(p, n){}^{15}\text{O}$
${}^{18}_9\text{F}$	110 min	β^+ (97) EC (3)	0.511 (annihilation)	194	${}^{18}\text{O}(p, n){}^{18}\text{F}$
${}^{32}_{15}\text{P}$	4.3 day	β^- (100)	—	—	${}^{32}\text{S}(n, p){}^{32}\text{P}$
${}^{57}_{27}\text{Co}$	271 days	EC (100)	0.014 0.122 0.136	9 86 11	${}^{56}\text{Fe}(d, n){}^{57}\text{Co}$
${}^{67}_{31}\text{Ga}$	78 hr	EC (100)	0.093 0.184 0.300 0.393	40 20 17 5	${}^{68}\text{Zn}(p, 2n){}^{67}\text{Ga}$
${}^{68}_{31}\text{Ga}$	68 min	β^+ (89) EC (11)	0.511 (annihilation)	178	${}^{68}\text{Zn}(p, n){}^{68}\text{Ga}$
${}^{82}_{37}\text{Rb}$	75 s	β^+ (95) EC(5)	0.511 0.777	190 13.4	${}^{98}\text{Mo} \xrightarrow{\text{spallation}} {}^{82}\text{Sr}$ $\downarrow 25.6 \text{ d}$ ${}^{82}\text{Rb}$
${}^{82}_{38}\text{Sr}$	25.6 days	EC (100)	—	—	${}^{98}\text{Mo} \xrightarrow{\text{spallation}} {}^{82}\text{Sr}$
${}^{99}_{42}\text{Mo}$	66 hr	β^- (100)	0.181 0.740 0.780	6 12 4	${}^{98}\text{Mo}(n, \gamma){}^{99}\text{Mo}$ ${}^{235}\text{U}(n, f){}^{99}\text{Mo}$
${}^{99m}_{43}\text{Tc}$	6.0 hr	IT (100)	0.140	90	${}^{99}\text{Mo} \xrightarrow{\beta^-}_{67\text{h}} {}^{99m}\text{Tc}$
${}^{111}_{49}\text{In}$	2.8 days	EC (100)	0.171 0.245	90 94	${}^{111}\text{Cd}(p, n){}^{111}\text{In}$
${}^{123}_{53}\text{I}$	13.2 hr	EC (100)	0.159	83	${}^{121}\text{Sb}(\alpha, 2n){}^{123}\text{I}$ ${}^{127}\text{I}(p, 5n){}^{123}\text{Xe}$ $\downarrow 2.1 \text{ hr}$ ${}^{123}\text{I}$ ${}^{124}\text{Xe}(p, 2n){}^{123}\text{Cs}$ $\downarrow 5.9 \text{ min}$ ${}^{123}\text{Xe}$ $\downarrow 2.1 \text{ hr}$ ${}^{123}\text{I}$
${}^{124}_{53}\text{I}$	4.2 days	β^+ (23) EC (77)	511 603	46 61	${}^{124}\text{Te}(p, n){}^{124}\text{I}$
${}^{125}_{53}\text{I}$	60 days	EC (100)	0.035 X-ray (0.027–0.032)	7 140	${}^{124}\text{Xe}(n, \gamma){}^{125}\text{Xe}$ ${}^{125}\text{Xe} \xrightarrow{\text{EC}}_{17\text{hr}} {}^{125}\text{I}$
${}^{131}_{53}\text{I}$	8.0 days	β^- (100)	0.284 0.364 0.637	6 81 7	${}^{130}\text{Te}(n, \gamma){}^{131}\text{Te}$ ${}^{235}\text{U}(n, f){}^{131}\text{Te}$ ${}^{131}\text{Te} \xrightarrow{\beta^-}_{25\text{min}} {}^{131}\text{I}$ ${}^{235}\text{U}(n, f){}^{131}\text{I}$

TABLE 5.1. *Continued*

Nuclide	Physical half-life	Mode of delay (%)	γ -ray energy* (MeV)	γ -ray abundance (%)	Common production method
$^{133}_{54}\text{Xe}$	5.3 days	β^- (100)	0.081	37	$^{235}\text{U}(\text{n}, \text{f})^{133}\text{Xe}$
$^{137}_{55}\text{Cs}$	30.0 yr	β^- (100)	0.662	85	$^{235}\text{U}(\text{n}, \text{f})^{137}\text{Cs}$
$^{201}_{81}\text{Tl}$	73 hr	EC (100)	0.167	9.4	$^{203}\text{Tl}(\text{p}, 3\text{n})^{201}\text{Pb}$
			X-ray (0.069–0.083)	93	$^{201}\text{Pb} \xrightarrow[9.3\text{hr}]{\text{EC}} ^{201}\text{Tl}$

* γ -rays with abundance less than 4% have not been cited.

d, deuteron; EC, electron capture; f, fission; IT, isomeric transition; n, neutron; p, proton. Data from Browne E, Finestone RB. *Table of Radioactive Isotopes*. New York: Wiley; 1986.

Target and Its Processing

Various types of targets have been designed and used for both reactor and cyclotron irradiation. In the design of targets, primary consideration is given to heat deposition in the target by irradiation with neutrons in the reactor or charged particles in the cyclotron. In both cases, the temperature can rise to 1000°C, and if proper material is not used or a method of heat dissipation is not properly designed, the target is likely to be burned or melted. For this reason, water cooling of the cyclotron probe to which the target is attached is commonly adopted. In the case of the reactor, the core cooling with heavy water is sufficient to cool the target. Most often, the targets are designed in the form of a foil to maximize heat dissipation.

The common form of the target is metallic foil, for example, copper, aluminum, uranium, vanadium, and so on. Other forms of targets are oxides, carbonates, nitrates, and chlorides contained in an aluminum tubing, which is then flattened to maximize the heat loss. Aluminum tubing is used because of its high melting point. In some cases, compounds are deposited on the appropriate metallic foil by vacuum distillation or by electrodeposition, and the plated foils are then used as targets. In specific cases, high-pressure gases (e.g., ^{124}Xe for ^{123}I production) and liquid targets (e.g., H_2^{18}O for ^{18}F production) are also used.

Equation for Production of Radionuclides

While irradiating a target for the production of a radionuclide, it is essential to know various parameters affecting its production, preferably in a mathematical form, to estimate how much of it would be produced for a given set of parameters. These parameters are therefore discussed below.

The activity of a radionuclide produced by irradiation of a target material with charged particles in a cyclotron or with neutrons in a nuclear reactor is given by

$$A = IN\sigma(1 - e^{-\lambda t}) \quad (5.1)$$

where

A = activity in disintegrations per second of the radionuclide produced

I = intensity or flux of the irradiating particles [number of particles/
($\text{cm}^2 \cdot \text{sec}$)]

N = number of target atoms

σ = formation cross section (probability) of the radionuclide (cm^2); it is given in units of "barn," which is equal to 10^{-24}cm^2

λ = decay constant given by $0.693/t_{1/2}$ (sec^{-1})

t = duration of irradiation (sec)

Equation (5.1) indicates that the amount of radioactivity produced depends on the intensity and energy (related to the cross-section σ) of the incident particles, the amount of the target material, the half-life of the radionuclide produced, and the duration of irradiation. The term $(1 - e^{-\lambda t})$ is called the *saturation factor* and approaches unity when t is approximately 5 to 6 half-lives of the radionuclide in question. At that time, the yield of the product nuclide becomes maximum, and its rates of production and decay become equal. For a period of irradiation of 5 to 6 half-lives, Eq. (5.1) becomes

$$A = IN\sigma \quad (5.2)$$

A graphic representation of Eqs. (5.1) and (5.2) is given in Figure 5.2.

The intensity of the irradiating particles is measured by various physical techniques, the description of which is beyond the scope of this book; however, the values are available from the operator of the cyclotron or the reactor. The formation cross sections of various nuclides are determined by experimental methods using Eq. (5.1), and they have been compiled and published by many investigators. The number of atoms N of the target is calculated from the weight W of the material irradiated, the atomic weight A_w and natural abundance K of the target isotope, and Avogadro's number (6.02×10^{23}) as follows:

$$N = \frac{W \times K}{A_w} \times 6.02 \times 10^{23} \quad (5.3)$$

After irradiation, isotopes of different elements may be produced and therefore should be separated by the appropriate chemical methods. These radionuclides are identified and quantitated by detecting their radiations and measuring their half-lives by the use of the NaI(Tl) or




FIG. 5.2. Production of radionuclides in a reactor or a cyclotron. The activity produced reaches a maximum (saturation) in 5 to 6 half-lives of the radionuclide.

Ge(Li) detectors coupled to a multichannel pulse height analyzer. They may also be assayed in an ionization chamber if the amount of radioactivity is high.

Radionuclide Generators

Radionuclide generators provide the convenient sources of short-lived radionuclides that are very useful clinically. The basic requirements for a generator are that *a parent radionuclide has a longer half-life than that of the daughter*, and the daughter can be easily separated from the parent. In a generator, a long-lived parent radionuclide is allowed to decay to its short-lived daughter radionuclide, and the latter is then chemically separated. The importance of radionuclide generators lies in the fact that they are easily transportable and serve as sources of short-lived radionuclides in institutions without cyclotron or reactor facilities.

A radionuclide generator consists of a glass or plastic column fitted at the bottom with a fretted disk. The column is filled with adsorbent material such as ion exchange resin, alumina, and so forth, on which the parent nuclide is adsorbed. The parent decays to the daughter until transient or secular equilibrium is established [Eqs. (3.17) and (3.18)] in several half-




FIG. 5.3. Typical radionuclide generator system. Vacuum in vial B draws the eluant from vial A through adsorbent material, and the daughter is collected in vial B.

lives of the daughter. After equilibrium, the daughter appears to decay with the same half-life as the parent. Because of the differences in chemical properties, the daughter activity is eluted with an appropriate solvent, leaving the parent on the column. After elution, the daughter activity builds up again and can be eluted repeatedly.

A schematic diagram of a radionuclide generator is shown in Figure 5.3. The vial containing the eluant is first inverted onto needle A, and an evacuated vial is inverted on the other needle B. The vacuum in the vial on needle B draws the eluant from the vial A through the column and elutes the daughter nuclide, leaving the parent nuclide on the column. In some commercial generators, a bottle of eluant is placed inside the housing, and aliquots of eluant are used up in each elution by the evacuated vial.

An ideal radionuclide generator should be simple and sturdy for transportation. The generator eluate should be free of the parent nuclide and the adsorbent material.

Several radionuclide generators are available for ready supply of short-lived radionuclides: ^{99}Mo (66 hr)– $^{99\text{m}}\text{Tc}$ (6 hr); ^{113}Sn (117 days)– $^{113\text{m}}\text{In}$ (100 min); ^{68}Ge (271 days)– ^{68}Ga (68 min); ^{82}Sr (25.6 days)– ^{82}Rb (75 sec); ^{81}Rb (4.6 hr)– $^{81\text{m}}\text{Kr}$ (13 sec). Of these, the ^{99}Mo – $^{99\text{m}}\text{Tc}$ generator is the work-horse of nuclear pharmacy in nuclear medicine.

^{99}Mo – $^{99\text{m}}\text{Tc}$ Generator

The ^{99}Mo – $^{99\text{m}}\text{Tc}$ generator is constructed with alumina (Al_2O_3) loaded in a plastic or glass column. The ^{99}Mo activity is adsorbed on alumina in the chemical form MoO_4^{2-} (molybdate) and in various amounts. The amount of alumina used is about 5–10 g depending on the ^{99}Mo activity. Currently, all generators use fission-produced ^{99}Mo . The growth and decay of $^{99\text{m}}\text{Tc}$ along with the decay of ^{99}Mo is shown in Figure 3.5 in Chapter 3. The $^{99\text{m}}\text{Tc}$ activity is eluted with 0.9% NaCl solution (isotonic saline) and obtained in the chemical form of $\text{Na}^{99\text{m}}\text{TcO}_4$.

Considering that only 87% of ^{99}Mo decays to $^{99\text{m}}\text{Tc}$, the $^{99\text{m}}\text{Tc}$ activity A_{Tc} can be calculated from Eq. (3.15) as follows:

$$A_{\text{Tc}} = 0.957(A_{\text{Mo}})_0(e^{-0.0105t} - e^{-0.1155t}) \quad (5.4)$$

where $(A_{\text{Mo}})_0$ is the ^{99}Mo activity at $t = 0$, $\lambda_{\text{Mo}} = 0.0105 \text{ hr}^{-1}$, and $\lambda_{\text{Tc}} = 0.1155 \text{ hr}^{-1}$. The time t has the unit of hour. At transient equilibrium [from Eq. (3.16)],

$$\begin{aligned} A_{\text{Tc}} &= 0.957(A_{\text{Mo}})_0 e^{-0.0105t} \\ &= 0.957(A_{\text{Mo}})_t \end{aligned} \quad (5.5)$$

Upon elution with saline, approximately 75% to 85% of the total activity is eluted from the column. After about 4 half-lives, the $^{99\text{m}}\text{Tc}$ activity reaches maximum.

Problem 5.1

A 2.5-Ci ^{99}Mo – $^{99\text{m}}\text{Tc}$ generator calibrated for Thursday noon was received on Wednesday afternoon. What would be the $^{99\text{m}}\text{Tc}$ activity, eluted at 8:00 a.m. on Friday, assuming a transient equilibrium between ^{99}Mo and $^{99\text{m}}\text{Tc}$ at the time of elution, and 85% elution yield?

Answer

^{99}Mo activity on Thursday noon = 2500 mCi

The time from Thursday noon to 8:00 a.m. on Friday = 20 hrs

Therefore, ^{99}Mo activity at 8:00 a.m. Friday = $2500 \times \exp(-0.0105 \times 20)$
= 2026 mCi

Using Eq. (5.5) at transient equilibrium,

$$\begin{aligned} ^{99\text{m}}\text{Tc activity} &= 0.956 \times 2026 \times 0.85 \\ &= 1646 \text{ mCi} \end{aligned}$$

The ^{99}Mo activity is likely to be eluted in trace quantities along with $^{99\text{m}}\text{Tc}$ activity. This is called the ^{99}Mo or *Moly breakthrough*. The presence of ^{99}Mo gives unnecessary radiation dose to the patient. According to the Nuclear

Regulatory Commission (NRC) regulations, *the permissible limit for ^{99}Mo breakthrough is $0.15\ \mu\text{Ci}$ ($5.5\ \text{kBq}$) per millicurie ($37\ \text{MBq}$) of $^{99\text{m}}\text{Tc}$ at the time of injection.* The ^{99}Mo breakthrough is determined by the detection of high-energy photons ($740\ \text{keV}$ and $780\ \text{keV}$) of ^{99}Mo in a dose calibrator after stopping 140-keV photons of $^{99\text{m}}\text{Tc}$ in a lead container (6-mm thick).

Aluminum is also likely to be eluted with $^{99\text{m}}\text{Tc}$ activity and must be checked. The presence of aluminum interferes with the preparation of $^{99\text{m}}\text{Tc}$ -labeled sulfur colloid by forming larger particles, which are trapped in the lungs. It also agglutinates the red blood cells during labeling. Its acceptable limit is $10\ \mu\text{g/ml}$ of $^{99\text{m}}\text{Tc}$ solution. Aluminum ion (Al^{3+}) is checked by the colorimetric test using a paper strip impregnated with a coloring agent and comparing the intensity of the color developed by the sample solution with that by a standard test solution ($10\ \mu\text{g/ml}$).

Questions

- Describe the different methods of production of radionuclides. Which method gives relatively more proton-rich and neutron-rich radionuclides?
- What is the average number of neutrons emitted in fission?
- What is the difference between carrier-free and no-carrier-added radionuclides?
- Why are cadmium rods and graphite used in the reactor?
- If ^{68}Zn is bombarded with protons in a cyclotron and three neutrons are emitted from the nucleus, what is the product of the nuclear reaction? Write the nuclear reaction.
- (a) What are the primary requirements for a radionuclide generator?
(b) How long does it take to reach transient equilibrium in a ^{99}Mo – $^{99\text{m}}\text{Tc}$ generator?
(c) What is the permissible limit of ^{99}Mo breakthrough in the $^{99\text{m}}\text{Tc}$ eluate?
(d) A 20-mCi (740-MBq) $^{99\text{m}}\text{Tc}$ eluate is found to contain $10\ \mu\text{Ci}$ ($0.37\ \text{MBq}$) of ^{99}Mo . Can this preparation be injected into a patient?
- Why is Al^{3+} undesirable in the $^{99\text{m}}\text{Tc}$ eluate? What is the permissible limit of Al^{3+} in the $^{99\text{m}}\text{Tc}$ eluate?
- A 2-Ci (74-GBq) ^{99}Mo – $^{99\text{m}}\text{Tc}$ generator calibrated for Thursday noon was eluted at 10 a.m. on the following Monday. Calculate the $^{99\text{m}}\text{Tc}$ activity assuming 80% yield and 87% of ^{99}Mo decays to $^{99\text{m}}\text{Tc}$.
- Calculate the activity in millicuries (MBq) of ^{111}In produced by irradiation of 1 g of pure ^{111}Cd for 3 hr with 12-MeV protons having a beam intensity of 10^{13} particles/($\text{cm}^2 \cdot \text{sec}$). The cross section for formation of ^{111}In ($t_{1/2} = 2.8$ days) is 200 millibarns (1 millibarn = $10^{-27}\ \text{cm}^2$).

Suggested Readings

- Colombetti LG. Radionuclide generators. In: Rayudu GVS, ed. *Radiotracers for Medical Applications*. Boca Raton, Fla: CRC Press; 1983;II:133–168.
- Friedlander G, Kennedy JW, Miller JM. *Nuclear and Radiochemistry*. 3rd ed. New York: Wiley; 1981.
- Gelbard AS, Hara T, Tilbury RS, Laughlin JS. Recent aspects of cyclotron production of medically useful radionuclides. In: *Radiopharmaceuticals and Labelled Compounds*. Vienna: IAEA; 1973:239–247.
- Noronha OPD, Sewatkar AB, Ganatra RD, et al. Fission-produced ^{99}Mo – $^{99\text{m}}\text{Tc}$ generator system for medical use. *J Nucl Med Biol*. 1976;20:32–36.
- Poggenburg JK. The nuclear reactor and its products. *Semin Nucl Med*. 1974; 4:229–243.
- Saha GB. Miscellaneous tracers for imaging. In: Rayudu GVS, ed. *Radiotracers for Medical Applications*. Boca Raton, Fla: CRC Press; 1983;II:119–132.
- Saha GB. *Fundamentals of Nuclear Pharmacy*. 5th ed. New York: Springer; 2004.
- Saha GB. *Basics of PET Imaging*. New York: Springer; 2005.
- Saha GB, MacIntyre WJ, Go RT. Cyclotron and positron emission tomography radiopharmaceuticals for clinical imaging. *Semin Nucl Med*. 1992;22:150–161.

6

Interaction of Radiation with Matter

All particulate and electromagnetic radiations can interact with the atoms of an absorber during their passage through it, producing ionization and excitation of the absorber atoms. These radiations are called *ionizing radiations*. Because particulate radiations have mass and electromagnetic radiations do not, the latter travel through matter longer distance before losing all energy than the former of the same energy. Electromagnetic radiations are therefore called *penetrating* radiations and particulate radiations *non-penetrating* radiations. The mechanisms of interaction with matter, however, differ for the two types of radiation, and therefore they are discussed separately.

Interaction of Charged Particles with Matter

The energetic charged particles such as α -particles, protons, deuterons, and β -particles (electrons) interact with the absorber atoms, while passing through it. The interaction occurs primarily with the orbital electrons of the atoms and rarely with the nucleus. During the interaction, both *ionization* and *excitation* as well as the breakdown of the molecule may occur. In excitation, the charged particle transfers all or part of its energy to the orbital electrons, raising them to higher energy shells. In ionization, the energy transfer may be sufficient to overcome the binding energy of the orbital electrons, ultimately ejecting them from the atom. Electrons ejected from the atoms by the incident charged particles are called *primary electrons*, which may have sufficient kinetic energy to produce further excitation or ionization in the absorber. The high-energy secondary electrons from secondary ionizations are referred to as *delta* (δ -) rays. The process of excitation and ionization will continue until the incident particle and all electrons come to rest. Both these processes may rupture chemical bonds in the molecules of the absorber, forming various chemical entities.

In ionization, an average energy of W is required to produce an ion pair in the absorber and varies somewhat with the type of absorber. The value

of W is about 35 eV in air and less in oxygen and xenon gases but falls in the range of 25–45 eV for most gases. The process of ionization, that is, the formation of ion pairs, is often used as a means of the detection of charged particles in ion chambers and Geiger–Müller counters described in Chapter 7.

Three important quantities associated with the passage of charged particles through matter are specific ionization, linear energy transfer, and range of the particle in the absorber, and these are described next.

Specific Ionization

Specific ionization (SI) is the total number of ion pairs produced per unit length of the path of the incident radiation. The SI values of α -particles are slightly greater than those of protons and deuterons, which in turn are larger than those of electrons.

Specific ionization increases with decreasing energy of the charged particle because of the increased probability of interaction at low energies. Therefore, toward the end of the travel, the charged particle shows a sharp increase in ionization. This peak ionization is called *Bragg ionization*. This phenomenon is predominant for heavy charged particles and is negligible for electrons.

Linear Energy Transfer

The linear energy transfer (LET) is the amount of energy deposited per unit length of the path by the radiation. From the preceding, it is clear that

$$\text{LET} = \text{SI} \times W \quad (6.1)$$

The LET is expressed in units of keV/ μm and is very useful in concepts of radiation protection. Electromagnetic radiations and β -particles interact with matter, losing only little energy per interaction and therefore have low LETs. In contrast, heavy particles (α -particles, neutrons, and protons) lose energy very rapidly, producing many ionizations in a short distance, and thus have high LETs. Some comparative approximate LET values in keV/ μm in tissue are given in Table 6.1.

TABLE 6.1. LET values of some radiations in tissue.

Radiation	LET (keV/ μm)
3 MV x-rays	0.5
250 KV x-rays	3.0
5-MeV α -particles	100.0
1-MeV electrons	0.25
14-MeV neutrons	20.0

Problem 6.1

If a particulate radiation produces 45,000 ion pairs per centimeter in air, calculate the LET of the radiation.

Answer

$$W = 35 \text{ eV per ion pair}$$

Using Eq. (6.1),

$$\begin{aligned} \text{LET} &= SI \times W \\ &= 45,000 \times 35 \\ &= 1,575,000 \text{ eV/cm} \\ &= 157.5 \text{ eV}/\mu\text{m} \\ &= 0.1575 \text{ keV}/\mu\text{m} \end{aligned}$$

Range

The *range* (R) of a charged particle in an absorber is the straight-line distance traversed by the particle in the direction of the particle. The range of a particle depends on the mass, charge, and kinetic energy of the particle and also on the density of the absorber. The heavier and more highly charged particles have shorter ranges than lighter and lower charged particles. The range of charged particles increases with the energy of the particle. Thus, a 10-MeV particle will have a longer range than a 1-MeV particle. The range of the particle depends on the density of the absorber, in that the denser the absorber, the shorter the range. The unit of range is given in mg/cm^2 of the absorber.

Depending on the type of the charged particle, the entire path of travel may be unidirectional along the initial direction of motion, or tortuous (Fig. 6.1). Because the α -particle loses only a small fraction of energy in a single collision with an electron because of its heavier mass and is not appreciably deflected in the collision, the α -particle path is nearly a straight line along its initial direction (Fig. 6.1a). Many collisions in a short distance create many ion pairs in a small volume. In contrast, β -particles or electrons interact with extranuclear orbital electrons of the same mass and are deflected considerably. This leads to tortuous paths of these particles (Fig. 6.1b). In this situation, the true range is less than the total path traveled by the particle.

It is seen that the ranges of all identical particles in a given absorber are not exactly the same but show a spread of 3% to 4% near the end of their path. (Fig. 6.2). This phenomenon, referred to as the *straggling of the ranges*, results from the statistical fluctuations in the number of collisions and in the energy loss per collision. The range straggling is less prominent with α -particles but is severe with electrons because it is mostly related to the




FIG. 6.1. Concept of passage of α particles and electrons through an absorber: (a) heavy α particles move in almost a straight line; (b) light electrons move in zigzag paths.

mass of the particle. The light mass electrons are considerably deflected during collisions and hence exhibit more straggling. If the transmission of a beam of charged particles through absorbers of different thicknesses is measured, the beam intensity will remain constant until the region of range straggling is encountered, where the beam intensity falls sharply from its initial value to zero. The absorber thickness that reduces the beam intensity by one half is called the *mean range*. The mean range of heavier particles such as α -particles is more well defined than that of electrons. Because β^- -particles are emitted with a continuous energy spectrum, their absorption, and hence their ranges, become quite complicated.




FIG. 6.2. Mean range and straggling of charged particles in an absorber.

Bremsstrahlung

When energetic charged particles, particularly electrons, pass through matter and come close to the nucleus of the atom, they lose energy as a result of deceleration in the Coulomb field of atomic nuclei. The loss in energy appears as an x-ray that is called *bremsstrahlung* (German for “braking” or “slowing down” radiation). These bremsstrahlung radiations are commonly used in radiographic procedures and are generated by striking a tungsten target with a highly accelerated electron beam.

Bremsstrahlung production increases with the kinetic energy of the particle and the atomic number (Z) of the absorber. For example, a 10-MeV electron loses about 50% of its energy by bremsstrahlung, whereas a 90-MeV electron loses almost 90% of its energy by this process. The bremsstrahlung production is proportional to Z^2 of the absorber atom. Therefore, bremsstrahlung is unimportant in lighter metals such as air, aluminum, and so forth, whereas it is very significant in heavy metals such as lead and tungsten. High-energy β^- -particles from radionuclides such as ^{32}P can produce bremsstrahlung in heavy metals such as lead and tungsten. For this reason, these radionuclides are stored in low- Z materials such as plastic containers rather than in lead containers.

Bremsstrahlung is inversely proportional to the mass of the charged particles and therefore is insignificant for heavy particles, namely α -particles and protons, because the probability of penetrating close to the nuclei is relatively low due to their heavier masses.

Annihilation

When energetic β^+ -particles pass through an absorber, they lose energy via interaction with orbital electrons of the atoms of the absorber. When the β^+ -particle comes to almost rest after losing all energy, it combines with an orbital electron of the absorber atom and produces two 511-keV annihilation radiations that are emitted in opposite directions (180°). These annihilation radiations are the basis of positron emission tomography (PET) in which two photons are detected in coincidence, which is discussed in Chapter 13.

Interaction of γ -Radiations with Matter

Mechanism of Interaction of γ -Radiations

When penetrating γ -rays pass through matter, they lose energy by interaction with the orbital electrons or the nucleus of the absorber atom. The γ ray photons may lose all of their energy, or a fraction of it, in a single encounter. The specific ionization of γ -rays is one-tenth to one-hundredth

of that caused by a non-penetrating electron of the same energy. There is no quantity equivalent to a range of particles for γ -rays, but they travel a long path in the absorber before losing all energy. The average energy loss per ion pair produced by the photons is the same as for electrons, that is, 35 keV in air.

There are three mechanisms by which γ -rays interact with absorber atoms during their passage through matter, and they are described below.

Photoelectric Effect

In the photoelectric effect, the incident γ -ray transfers all its energy to an orbital electron of the absorber atom whereby the electron, called the *photoelectron*, is ejected with kinetic energy equal to $E_\gamma - E_B$, where E_γ and E_B are the energy of the γ -ray and the binding energy of the electron, respectively (Fig. 6.3). The photoelectron loses its energy by ionization and excitation in the absorber, as discussed previously. The photoelectric effect occurs primarily in the low-energy range and decreases sharply with increasing photon energy. It also increases very rapidly with increasing atomic number Z of the absorber atom. Roughly, the photoelectric effect is proportional to Z^5/E_γ^3 . The photoelectric contribution from the 0.15-MeV γ -rays in aluminum ($Z = 13$) is about the same ($\sim 5\%$) as that from the 4.7-MeV γ -rays in lead ($Z = 82$).

The photoelectric effect occurs primarily with the K -shell electrons, with about 20% contribution from the L -shell electrons and even less from higher shells. There are sharp increases (discontinuities) in photoelectric effects at energies exactly equal to binding energies of K -, L - (etc.) shell electrons. These are called K -, L - (etc.) absorption edges. The vacancy created by the ejection of an orbital electron is filled in by the transition of an electron from the upper energy shell. It is then followed by emission of




FIG. 6.3. The photoelectric effect in which a γ -ray with energy E_γ transfers all its energy to a K -shell electron, and the electron is ejected with $E_\gamma - E_B$, where E_B is the binding energy of the K -shell electron.




FIG. 6.4. The Compton scattering, in which a γ ray interacts with an outer orbital electron of an absorber atom. Only a part of the photon energy is transferred to the electron, and the photon itself is scattered at an angle. The scattered photon may undergo subsequent photoelectric effect or Compton scattering in the absorber or may escape the absorber.

a characteristic x-ray or Auger electron, analogous to the situations in internal conversion or electron capture decay.

Compton Scattering

In Compton scattering, the γ ray photon transfers only a part of its energy to an electron in the outer shell of the absorber atom, and the electron is ejected. The photon, itself with reduced energy, is deflected from its original direction (Fig. 6.4). This process is called the *Compton scattering*. The scattered photon of lower energy may then undergo further photoelectric or Compton interaction, and the Compton electron may cause ionization or excitation, as discussed previously.

At low energies, only a small fraction of the photon energy is transferred to the Compton electron, and the photon and the Compton electron are scattered at an angle θ . Using the law of conservation of momentum and energy, the scattered photon energy is given by

$$E_{sc} = E_{\gamma} / [1 + (E_{\gamma}/0.511)(1 - \cos \theta)] \quad (6.2)$$

where E_{γ} and E_{sc} are the energies in MeV of the initial and scattered photons. The scattered photon energy varies from a maximum in a collision at 0° (forward) to a minimum at $\theta = 180^{\circ}$ in a backscattering collision. Conversely, the Compton electron carries a minimum energy in the forward collision to a maximum energy in the backscattering collision. At higher energies, both the scattered photon and the Compton electron are predominantly scattered in the forward direction.

If the photon is backscattered, that is, scattered at 180° , then the backscattered photon has the energy E_{sc} given by the expression ($\cos 180^{\circ} = -1$):

$$E_{sc} = E_{\gamma} / (1 + E_{\gamma}/0.256) \quad (6.3)$$

In backscattering of a 140-keV photon, the scattered photon and the Compton electron would have 91 keV and 49 keV, respectively, whereas for

a 1330-keV photon these values are 215 keV and 1115 keV, respectively. It can be seen that as the photon energy increases, the scattered photon energy approaches the minimum limit of 256 keV, and the Compton electron receives the maximum energy.

Compton scattering is almost independent of the atomic number Z of the absorber. Compton scattering contributes primarily in the energy range of 0.1 to 10 MeV, depending on the type of absorber.

Pair Production

When the γ ray photon energy is greater than 1.02 MeV, the photon can interact with the nucleus of the absorber atom during its passage through it, and a positive electron and a negative electron are produced at the expense of the photon (Fig. 6.5). The energy in excess of 1.02 MeV appears as the kinetic energy of the two particles. This process is called *pair production*. It varies almost linearly with Z^2 of the absorber and increases slowly with the energy of the photon. In soft tissue, pair production is insignificant at energies up to 10 MeV above 1.02 MeV. Positive electrons created by pair production are annihilated to produce two 0.511-MeV photons identical to those produced by positrons from radioactive decay.

The relative importance of photoelectric, Compton, and pair production interactions with absorbers of different atomic numbers is shown in Figure 6.6, as a function of the energy of the incident photons. It is seen that the photoelectric effect is predominant in high Z absorbers at lower energies (<0.1 MeV), whereas the Compton scattering is predominant in inter-




FIG. 6.5. Illustration of the pair production process. An energetic γ ray with energy greater than 1.02 MeV interacts with the nucleus, and one positive electron (e^+) and one negative electron (e^-) are produced at the expense of the photon. The photon energy in excess of 1.02 MeV appears as the kinetic energy of the two particles. The positive electron eventually undergoes annihilation to produce two 511-keV photons emitted in opposite directions.




FIG. 6.6. Relative contributions of the photoelectric effect, Compton scattering, and pair production as a function of photon energy in absorbers of different atomic numbers. (Adapted with permission from Hendee WR. *Medical Radiation Physics*. 1st ed. Chicago: Year Book Medical Publishers, Inc; 1970:141.)

mediate Z absorbers at medium energies (~ 1 MeV). At higher energies (>10 MeV), pair production predominates in all Z absorbers.

Photodisintegration

When the γ -ray photon energy is very high (>10 MeV), the photon may interact with the nucleus of the absorber atom and transfer sufficient energy to the nucleus such that one or more nucleons may be emitted. This process is called the *photodisintegration reaction*, or *photonuclear reaction* and produces new nuclides. The (γ, n) reactions on targets such as ^{12}C and ^{14}N have been used to produce ^{11}C and ^{13}N radionuclides but now are rarely used to produce radionuclides.

Attenuation of γ -Radiations

Linear and Mass Attenuation Coefficients

γ -ray and x-ray photons are either attenuated or transmitted as they travel through an absorber. Attenuation results from absorption by the photoelectric effect, Compton scattering, and pair production at higher energies. Depending on the photon energy and the density and thickness of the absorber, some of the photons may pass through the absorber without any interaction leading to the transmission of the photons (Fig. 6.7). Attenuation of γ -radiations is an important factor in radiation protection.

As shown in Figure 6.7, if a photon beam of initial intensity I_0 passes through an absorber of thickness x , then the transmitted beam I_t is given by the exponential equation

$$I_t = I_0 e^{-\mu x} \quad (6.4)$$

where μ is the *linear attenuation coefficient* of the absorber for the photons of interest and has the unit of cm^{-1} . The factor $e^{-\mu x}$ represents the fraction of the photons transmitted. Because attenuation is primarily due to photoelectric, Compton, and pair production interactions, the linear attenuation coefficient μ is the sum of photoelectric coefficient (τ), Compton coefficient (σ), and pair production coefficient (κ). Thus,

$$\mu = \tau + \sigma + \kappa \quad (6.5)$$

Linear attenuation coefficients normally decrease with the energy of the γ -ray or x-ray photons and increase with the atomic number and density of the absorber. The relative contributions of photoelectric effect, Compton scattering, and pair production in water (equivalent to body tissue) at different energies are illustrated in Figure 6.8.

An important quantity, μ_m , called the *mass attenuation coefficient*, is given by the linear attenuation coefficient divided by the density ρ of the absorber

$$\mu_m = \frac{\mu}{\rho} \quad (6.6)$$

The mass attenuation coefficient μ_m has the unit of cm^2/g or cm^2/mg . The mass attenuation coefficients for fat, bone, muscle, iodine, and lead are given in Figure 6.9.

Half-Value Layer

The concept of *half-value layer* (HVL) of an absorbing material for γ or x-radiations is important in the design of shielding for radiation protection.




FIG. 6.7. Illustration of attenuation of a photon beam (I_0) in an absorber of thickness x . Attenuation comprises a photoelectric effect (τ), Compton scattering (σ), and pair production (κ). Photons passing through the absorber without interaction constitute the transmitted beam (I_t).




FIG. 6.8. Plot of linear attenuation coefficient of γ -ray interaction in water (equivalent to body tissue) as a function of photon energy. The relative contributions of photoelectric, Compton, and pair production processes are illustrated.




FIG. 6.9. Attenuation coefficients for fat, muscle, bone, iodine, and lead as a function of photon energy. (Adapted with permission from Hendee WR. *Medical Radiation Physics*. 1st ed. Chicago: Year Book Medical Publishers, Inc; 1970:221.)

TABLE 6.2. Half-value layer values (HVLs) of lead for commonly used radionuclides.

Radionuclides	HVL, Lead (cm)*	HVL, Water (cm)†
^{137}Cs	0.65	—
$^{99\text{m}}\text{Tc}$	0.03	4.6
^{201}Tl	0.02	—
^{99}Mo	0.70	—
^{67}Ga	0.10	—
^{123}I	0.04	—
^{111}In	0.10	—
^{125}I	0.003	1.7
^{57}Co	0.02	—
^{131}I	0.30	6.3
^{18}F	0.39	11.2

* Adapted from Goodwin PN. Radiation safety for patients and personnel. In: Freeman LM, ed. *Freeman and Johnson's Clinical Radionuclide Imaging*. 3rd ed. Philadelphia: WB Saunders; 1984: 320.

† HVL in water is considered equivalent to HVL in tissue.

It is defined as the thickness of the absorber that reduces the intensity of a photon beam by one-half. Thus, an HVL of an absorber around a source of γ -radiations with an exposure rate of 150 mR/hr will reduce the exposure rate to 75 mR/hr. The HVL depends on the energy of the radiation and the atomic number of the absorber. It is greater for high-energy photons and smaller for high- Z materials.

For monoenergetic photons, the HVL of an absorber is related to its linear attenuation coefficient as follows:

$$\text{HVL} = \frac{0.693}{\mu} \quad (6.7)$$

Because μ has the unit of cm^{-1} , the HVL has the unit of cm. The HVLs of lead for different radionuclides are given in Table 6.2.

Another important quantity, tenth-value layer (TVL), is the thickness of an absorber that reduces the initial beam by a factor of 10. It is given by

$$\begin{aligned} \text{TVL} &= -\frac{\ln(0.1)}{\mu} \\ &= \frac{2.30}{\mu} \end{aligned} \quad (6.8)$$

$$= 3.32 \text{ HVL} \quad (6.9)$$

Problem 6.2

If the HVL of lead for the 140-keV photons of $^{99\text{m}}\text{Tc}$ is 0.03 cm of lead, calculate the linear attenuation coefficient of lead for the 140-keV photons

and the amount of lead needed to reduce the exposure of a point source of radiation by 70%.

Answer

$$\mu = \frac{0.693}{\text{HVL}} = \frac{0.693}{0.03} = 23.1 \text{ cm}^{-1}$$

Because the initial beam is reduced by 70%, the remaining beam is 30%.

$$\begin{aligned} 0.3 &= 1 \times e^{-23.1x} \\ \ln(0.3) &= -23.1 \times x \\ 1.20 &= 23.1 \times x \\ x &= 0.052 \text{ cm} \\ &= 0.52 \text{ mm} \end{aligned}$$

Thus, 0.52 mm of lead will reduce a beam of 140-keV photons by 70%.

Interaction of Neutrons with Matter

Because neutrons are neutral particles, their interactions in the absorber differ from those of the charged particles. They interact primarily with the nucleus of the absorber atom and very little with the orbital electrons. The neutrons can interact with the atomic nuclei in three ways: elastic scattering, inelastic scattering, and neutron capture. If the sum of the kinetic energies of the neutron and the nucleus before collision is equal to the sum of these quantities after collision, then the interaction is called *elastic*. If a part of the initial energy is used for the excitation of the struck nucleus, the collision is termed *inelastic*. In neutron capture, a neutron is captured by the absorber nucleus, and a new excited nuclide is formed. Depending on the energy deposited, an α -particle, a proton, a neutron, or γ -rays can be emitted from the excited nucleus, and a new product nuclide (usually radioactive) is produced.

Questions

- What is the difference between excitation and ionization?
 - How are δ -rays produced?
 - How much energy is required on the average to produce an ion pair in air by charged particles?
- Define specific ionization (SI), linear energy transfer (LET), and range (R).

3. Electromagnetic radiations and electrons have low LETs compared to heavy particles (e.g., α -particles), which have high LETs. Explain.
4. The range of an α -particle is almost equal to the total path traveled, whereas the range of an electron is less than the total path traveled by the particle. Explain.
5. Indicate how the range of a charged particle is affected by the following conditions:
 - (a) As the mass increases, the range increases or decreases.
 - (b) As the energy of the particle increases, the range increases or decreases.
 - (c) As the charge of the particle increases, the range increases or decreases.
6. Define Bragg ionization and straggling of ranges. Which has more straggling, an α -particle or an electron? Explain.
7. How is bremsstrahlung produced? Does its production increase or decrease with increasing kinetic energy of the electron and the atomic number of the absorber? Explain why ^{32}P is stored in plastic containers.
8. Discuss the mechanism of the photoelectric effect. Does this process increase or decrease with increasing energy of the γ -ray and with increasing atomic number of the absorber?
9. A 0.495-MeV γ -ray interacts with a K -shell electron by the photoelectric process. If the binding energy of the K -shell electron is 28 keV, what happens to the rest of the photon energy?
10. (a) Explain the Compton scattering of electromagnetic radiations in the absorber.
 - (b) Does it depend on the atomic number of the absorber?
 - (c) How is it affected by the γ -ray energy?
11. If a relatively high-energy γ -ray is scattered at 180° (backscattered) by the Compton scattering, what is the maximum energy of the scattered photon?
12. (a) How does pair production occur?
 - (b) Why does pair production require a minimum of 1.02 MeV for γ -ray energy?
 - (c) Is this process affected by the atomic number of the absorber and the photon energy?
13. Which electrons of the absorber atom are involved in the photoelectric and Compton interactions of electromagnetic radiations?
14. (a) Discuss the attenuation of a photon beam passing through an absorber.
 - (b) Does it depend on the density and the atomic number of the absorber?
 - (c) Define the half-value layer (HVL) of an absorbing material for a γ -ray energy.

15. If 1 mCi of a radionuclide is adequately shielded by 5 HVLs of a shielding material, how many HVLs are needed to provide equal shielding for (a) 5 mCi and (b) 8 mCi?
16. A 1-mm lead apron will afford approximately twice as much protection as a 0.5-mm apron, or does this shielding depend on the energy of the radiation?
17. How many HVLs are approximately equivalent to three tenth-value layers?
18. Suppose 5% of the 364-keV photons of ^{131}I are transmitted after passing through a lead brick of 10-cm thickness. Calculate the HVL of lead for ^{131}I .
19. There is a 75% chance that a monoenergetic photon beam will be attenuated by 4 mm of lead. What is the HVL of lead for the photon?
20. Which of the following radiations has the highest LET?
 - (a) 120-keV x-ray
 - (b) 100-keV electron
 - (c) 5-MeV α -particle
 - (d) 10-MeV proton
 - (e) 14-MeV neutron

Suggested Readings

- Cherry SR, Sorensen JA, Philips ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia: W.B. Saunders; 2003.
- Friedlander G, Kennedy JW, Miller JM. *Nuclear and Radiochemistry*. 3rd ed. New York: Wiley; 1981.
- Johns HE, Cunningham JR. *The Physics of Radiology*. 4th ed. Springfield, Ill: Charles C Thomas; 1983.
- Knoll GF. *Radiation Protection and Measurement*. 3rd ed. New York: Wiley; 2003.
- Lapp RE, Andrews HL. *Nuclear Radiation Physics*. 4th ed. Englewood Cliffs, NJ: Prentice-Hall; 1972.

7

Gas-Filled Detectors

Principles of Gas-Filled Detectors

The operation of a gas-filled detector is based on the ionization of gas molecules by radiation, followed by collection of the ion pairs as charge or current with the application of a voltage between two electrodes. The measured charge or current is proportional to the applied voltage and the amount and energy of radiation, and depends on the type and pressure of the gas.

A schematic diagram of a gas-filled detector is shown in Figure 7.1. When an ionizing radiation beam passes through the gas, it will cause ionization of the gas molecules and ion pairs will be produced depending on the type and pressure of the gas. When a voltage is applied between the two electrodes, the negative electrons will move to the anode and the positive ions to the cathode, thus producing a current that can be measured on a meter.

At very low voltages, the ion pairs do not receive enough acceleration to reach the electrodes and therefore may combine together to form the original molecule instead of being collected by the electrodes. This region is called the *region of recombination* (Fig. 7.2). As the applied voltage is gradually increased, a *region of saturation* is encountered, where the current measured remains almost the same over the range of applied voltages. In this region, only the primary ion pairs formed by the initial radiations are collected. Individual events cannot be detected; only the total current passing through the chamber is measured. Because specific ionization differs for α -, β -, and γ -radiations, the amount of current produced by these radiations differs in this region. The voltage in this region is of the order of 50–300 V. Ionization chambers such as dose calibrators are operated in this region.

When the applied voltage is further increased, the electrons and positive ions gain such high velocities and energies during their acceleration toward the electrodes that they cause secondary ionization. The latter will increase the measured current. This process is referred to as the *gas amplification*. This factor can be as high as 10^6 per individual primary event depending on




FIG. 7.1. A schematic diagram of a gas-filled detector illustrating the principles of operation.




FIG. 7.2. A composite curve illustrating the current output as a result of increasing voltages for different radiations. (A) Region of recombination, (B) region of saturation, (C) proportional region, (D) region of limited proportionality, (E) Geiger region, and (F) continuous discharge.

the design of the gas detector and the applied voltage. In this region, the total current measured is equal to the number of ionizations caused by the primary radiation multiplied by the gas amplification factor. In this region, the current increases with the applied voltage in proportion to the initial number of ion pairs produced by the incident radiation. Therefore, as in the case of the region of saturation, the current amplification is relatively proportional to the types of radiations, e.g., α -, β -, and γ -radiations. This region is referred to as the *proportional region* (see Fig. 7.2). Proportional counters are usually filled with 90% argon and 10% methane at atmospheric pressure. These counters can be used to count individual counts and to discriminate radiations of different energies. These counters, however, are not commonly used for γ and x-ray counting because of poor counting efficiency (<1%).

As the applied voltage is increased further, the current produced by different types of radiation tends to become identical. The voltage range over which the current tends to converge is referred to as the *region of limited proportionality*. This region is not practically used for detecting any radiation in nuclear medicine.

With additional increase in voltage beyond the region of limited proportionality, the current becomes identical, regardless of how many ion pairs are produced by the incident radiations. This region is referred to as the *Geiger region* (see Fig. 7.2). In the Geiger voltage region, the current is produced by an avalanche of interactions. When highly accelerated electrons strike the anode with a great force, ultraviolet (UV) light is emitted, which causes further emission of photoelectrons by gas ionization and from the chamber walls. The photoelectrons will again strike the anode to produce more UV, and hence an avalanche spreads along the entire length of the anode. The amplification factor can be as high as 10^{10} . During the avalanche, however, the lightweight electrons are quickly attracted to the anode, whereas a sheath of slow-moving heavy positive ions builds up around the anode. As a result, the voltage gradient falls below the value necessary for ion multiplication, and therefore the avalanche is terminated. All this occurs in less than 0.5 microsecond, and the counter is left insensitive and must recover before another event can be counted.

Recovery begins with the migration of the positive ions toward the cathode (i.e., chamber wall) and takes about 200 microseconds at a gas pressure of 0.1 atmosphere, which is equal to the dead time of the counter that varies with gas pressure. As the positive ions approach the cathode, secondary electrons may be emitted from the surface of the cathode, which then set another discharge just about 200 microseconds after the previous one. Such repetitive discharges that are due to secondary electrons are independent of the types and energy of radiation that the counter is intended to measure. The emission of secondary electrons is suppressed by a technique known as *quenching* to eliminate repetitive counter discharges (see later).

As the applied voltage is increased beyond the Geiger region, a single ionizing event produces a series of repetitive discharges leading to what is called *spontaneous discharge*. This region is called the *region of continuous discharge* because the gas may be ionized in the absence of radiation at this high voltage (see Fig. 7.2). Operation of a detector in this region may cause damage to the detector.

Ionization Chambers

Ionization chambers are operated at voltages in the saturation region that spans 50–300 V. The detector is a cylindrical chamber filled with air or a gas, sometimes at high pressure. A central wire and the chamber act as the electrodes and the current is measured by an electrometer. The detection efficiency of the ionization chambers for x-rays and γ -rays is very low (<1%) and depends on the energy of these radiations. Ionization chambers are primarily used for measuring high-intensity radiation such as x-ray beams and high activity of radiopharmaceuticals. Cutie Pie meters, dose calibrators, and pocket dosimeters are the common ionization chambers used in nuclear medicine.

Cutie Pie Survey Meter

The Cutie Pie survey meter is made of an outer metallic cylindrical electrode and a central wire. It uses air for ionization and is operated with a battery. It is used to measure the radiation exposure in the range of mR/hr to R/hr. It is primarily used to monitor the exposure at high radiation levels such as those from x-ray beams and ^{99}Mo – $^{99\text{m}}\text{Tc}$ generators. It does not have a dead time.

Dose Calibrator

The dose calibrator is an ionization chamber and one of the most essential instruments in nuclear medicine for measuring the activity of radionuclides and radiopharmaceuticals. Since it measures the current produced by activity, it does not have deadtime effects. It is a cylindrically shaped, sealed chamber with a central well and is filled with argon and traces of halogen at high pressure (~5–12 atmospheres). Its operating voltage is about 150 V. A typical dose calibrator is shown in Figure 7.3.

Because radiations of different types and energies produce different amounts of ionization (hence current), equal activities of different radionuclides generate different quantities of current. For example, the amount of the current produced by 1 mCi (37 MBq) of $^{99\text{m}}\text{Tc}$ differs from that produced by 1 mCi (37 MBq) of ^{131}I . Isotope selectors provided on the dose calibrator are the feedback resistors to compensate for the differences in ioniza-



FIG. 7.3. A typical dose calibrator. (Courtesy of Capintec, Inc., Ramsey, NJ.)

tion (current) produced by different radionuclides so that equal activities produce the same reading. In most dose calibrators, isotope selectors for common radionuclides are push-button type, whereas those for other radionuclides are set by a continuous dial. An activity range selector is a variable resistor that adjusts the range of activity (μCi , mCi , Ci , or kBq , MBq , GBq) for display.

In the past, the NRC required the calibration of dose calibrators for constancy, accuracy, and linearity of their operation and geometry of samples and accordingly prescribed specific recommendations for these tests. However, current NRC regulations (10CFR35) require only to have these calibrations performed according to nationally recognized standards or the manufacturers' instructions. In the absence of specific recommendations, the earlier frequency and other related requirements of these calibration tests have been given as follows:

1. Constancy (daily)
2. Accuracy (at installation, annually, and after adjustment or repairs)
3. Linearity (at installation, quarterly, and after adjustment or repairs)
4. Geometry (at installation and after adjustment or repairs)

Constancy

Daily constancy check is performed by measuring a long-lived radioactivity (e.g., ^{137}Cs) in the dose calibrator and observing the variation not to exceed $\pm 10\%$ relative to the previous day reading. If the variation exceeds $\pm 10\%$, the unit must be repaired or replaced.

Accuracy

Accuracy of the dose calibrator is determined by measuring the activity of at least two long-lived radionuclides (e.g., ^{137}Cs and ^{57}Co) certified by the National Institute of Standards and Technology (NIST) in the dose calibrator and comparing the measured activity with the activity reported by the NIST. The measured value should not differ from the standard value by more than $\pm 10\%$. If it exceeds $\pm 10\%$, the unit must be repaired or replaced.

Linearity

Decay Method

The linearity test indicates the dose calibrator's ability to measure the activity accurately over a range of values. It is performed by measuring a radioactive source (e.g., $^{99\text{m}}\text{Tc}$), containing the highest activity normally used in the clinical setting, in the dose calibrator at different time intervals until the source decays down to less than $30\ \mu\text{Ci}$ ($1.1\ \text{MBq}$). The measured activities are plotted against time on a semilog paper and the "best fit" line is drawn. If the deviation of any point from the line exceeds $\pm 10\%$, the dose calibrator needs to be replaced, or a correction factor must be applied to the data in the nonlinear region.

Shielding Method

The advantage of this method is that it is less time-consuming and is easy to perform. The method utilizes a commercial kit, called Calicheck, that contains seven concentric tubes or "sleeves." All sleeves except the innermost one are lead-lined with increasing thickness simulating the various times of decay. When an activity source is measured by using first the inner sleeve, followed sequentially by increasingly thick sleeves, the data represent the activities at different decay times. Calibration factors are calculated by dividing the innermost tube reading by each outer tube reading. For subsequent linearity tests, identical measurements are made using the sleeves, and each measurement is multiplied by the corresponding calibration factors. Each corrected sleeve reading should give an identical value, and the average of all values is calculated. If an individual reading exceeds the average value by $\pm 10\%$, then the calibrator needs replacement, or a correction factor needs to be applied.

It should be noted that before the shielding method can be instituted, the linearity test must be first performed by the decay method for a new dose calibrator.

Geometry

Variations in sample volumes or in geometric configurations of the container can affect the accuracy of measurements in a dose calibrator, particularly for low-energy radiations. Thus, 1 mCi (37 MBq) in 1-ml or 30-ml volume, or 1 mCi (37 MBq) in 1-cc syringe, 10-cc syringe, or 10-cc vial, or in containers of different materials (plastic or glass) may give different readings in the dose calibrator. Correction factors must be determined for these geometric variations and applied to the measured activities, if the error exceeds $\pm 10\%$.

Pocket Dosimeter

The pocket dosimeter operates on the principle of a charged electroscope equipped with a scale inside. It consists of a quartz fiber electroscope inside the chamber. Initially, the dosimeter is fully charged by means of an external power supply (a dosimeter charger), and the scale then reads zero. After exposure to radiation, charge is lost, and the loss of charge is proportional to the amount of radiation exposure, which is read on the inside scale in mR. This reading can be seen through a viewing window at the end of the dosimeter. After complete discharge of the dosimeter, it can be charged and used again. It is primarily used to determine personnel exposure while working with radiation and has the advantage of giving immediate readings. These dosimeters are available in full-scale readings of 200 mR, 500 mR, and 1 R. Discharge due to leakage is the major disadvantage of these dosimeters.

Geiger–Müller Counters

The Geiger–Müller (GM) counter operates in the Geiger region of the voltage, as shown in Figure 7.2. As already mentioned, in this region, an avalanche of ionizations occurs as a result of high voltage. Once an ionization is initiated, the avalanche of ionizations can lead to repetitive discharges unless the process is interrupted by the quenching technique. An electronic technique of quenching can be applied in which the voltage applied to the GM tube is temporarily reduced below the Geiger region until all ion pairs return to their de-excited states. This happens in a few tenths of a millisecond. The original voltage is then restored for the detection of the next event. This technique is no longer in use.

The common technique of quenching is to add a small quantity of a quenching gas to the counting gas. Either organic solvent vapors (e.g., ethyl alcohol, xylene, or isobutane) or halogen gases (chlorine or bromine) are commonly used as the quenching gas. These molecules transfer electrons to the “positive” ion cloud and become themselves ionized. Ionized molecules of the quenching gas migrate to and dislodge electrons from the cathode.



FIG. 7.4. An end-window-type Geiger-Müller counter. (Courtesy of Nuclear Associates, Carle Place, NY.)

When these electrons neutralize the ionized molecules of the quenching gas, energy is released, which causes the dissociation of the molecules of the gas but with no UV emissions to prolong the avalanche. This prevents the continuous discharge of the GM counter. Organic molecules are more effective quenchers but dissociate irreversibly and therefore give a limited lifetime for the GM tube ($\sim 10^8$ – 10^{10} pulses). In contrast, dissociated inorganic molecules recombine to form the original molecules, and therefore halogen-quenched GM tubes have infinite useful lifetime.

The probes of GM counters can be either end-window type or side-window type. An end-window-type GM counter is shown in Figure 7.4. The window is made of thin mica (0.01 mm thick), and gases such as argon, methane, helium, and neon mixed with halogen are commonly used as the counting gas. The typical gas pressure in GM probes is about 0.8 atmosphere. Different shapes of GM probes are available, such as cylindrical and pancake types. Some GM probes are provided with a metal cover that stops all β -particles and very low-energy γ -radiations so that only high-energy photons are detected. Without the cover, both β -particles and γ -rays are detected. The GM counter is usually battery operated at a voltage of 500–900 V. Lower voltages are used for smaller tubes, and some special tubes are operated even at 1300 V. The meter connected to the GM probe gives readings in mR/hr or counts per minute. Some counters are equipped with audible alarms or flashing light alarms that are triggered by radiation above a preset intensity. The latter counter is often used to monitor the radiation level in work areas and is called an *area monitor*.

The exposure rate (mR/hr or counts/min) given by the GM tube for x-rays or γ -rays depends on the energy of the photons, because they primarily interact with the walls of the tube rather than with the gas volume. The GM tubes are made of aluminum or steel with atomic number Z higher than that of air. Since at low energies the photons primarily interact via the photoelectric process, which is proportional to Z^3 , the exposure rate measured by the GM counter will be overestimated. On the other hand, at medium energies, the Compton interaction predominates, which is independent of Z and will therefore give correct reading of exposure rate. Since most GM counters are commonly calibrated for 662-keV photons of ^{137}Cs , they give only crude estimations of exposure rates for energies less than 150 keV and must be calibrated at these energies. However, the so-called energy-compensated detectors have been devised in which a thin layer of high Z material such as tin is placed around the inside of the detector. The increased photoelectric absorption of low energy photons in tin significantly flattens the energy response of the detector. A disadvantage of this detector is that the low energy photon sensitivity is greatly reduced.

The GM survey meters are more sensitive than ionization chambers by a factor of about 10. Because voltage pulses generated in GM tubes are independent of the energy deposited, they cannot discriminate between energies and types of radiations. These counters are almost 100% efficient for counting β -particles but have only 1% to 2% efficiency for counting γ and x-rays. The dead time, or resolving time (Chapter 8), of the GM counters is about 100 to 500 microseconds. This limits the count rates to about 15,000–20,000 counts per minute (cpm) for these counters, and at higher activities they tend to saturate, thus losing counts. The GM counters are normally used for area survey for contamination with low-level activity. According to the NRC regulations, these survey meters must be calibrated annually with standard calibrated sources such as ^{137}Cs .

Questions

1. Describe the principles of gas-filled detectors.
2. What are the differences between an ionization chamber and a Geiger–Müller counter?
3. What is the function of a push-button isotope selector on a dose calibrator?
4. Can you discriminate between 140-keV γ -rays, 364-keV γ -rays, and 5-MeV α -particles using a GM counter?
5. What type of instruments would you use for:
 - (a) Survey of the laboratory?
 - (b) X-ray beam exposure?
 - (c) Area survey around x-ray room?
 - (d) Spill of $50\ \mu\text{Ci}$ (1.85 MBq) of ^{201}Tl ?

- (e) Background radiation?
 - (f) Radiation survey of a diagnostic x-ray installation?
6. (a) Why are halogen gases added to GM counters?
 - (b) What is the typical dead time for GM counters?
 - (c) How often do the GM counters need to be calibrated?
 - (d) Why cannot the GM counters be used for detecting high-activity samples?
7. What are the typical voltages applied to the ionization chambers and GM counters?
 8. Describe the various tests of the dose calibrator and mention the frequency of each test.

Suggested Readings

- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia: W.B. Saunders; 2003.
- Hendee WR, Ritenour R. *Medical Imaging Physics*. New York: Wiley-Liss; 2002.
- Ouseph PJ. *Introduction to Nuclear Radiation Detectors*. New York: Plenum Press; 1975.
- Robinson CV. Geiger-Müller and proportional counters. In: Hine GJ, ed. *Instrumentation in Nuclear Medicine*. New York: Academic Press, 1967:57-72.
- Rollo FD, ed. *Nuclear Medicine Physics, Instrumentation and Agents*. St. Louis: Mosby; 1977.

8

Scintillation and Semiconductor Detectors

Scintillation Detectors

As stated in Chapter 7, the detection efficiency of γ and x-rays in gas detectors is very low, because these penetrating radiations travel through the low-density gas with little interaction. To improve counting efficiency for these radiations, solid and liquid scintillation detectors with high density are used. These detectors have the unique property of emitting scintillations or flashes of light after absorbing γ or x-radiations. The γ or x-rays interact with scintillation detectors via photoelectric, Compton, and/or pair production mechanisms, whereby the detector molecules are raised to higher energy states through ionization or excitation. These high-energy states return to ground states by emitting light photons. The time to reach the ground state is called the *scintillation decay time*. The light photons produced are converted to an electrical pulse by means of a photomultiplier (PM) tube (described later). The pulse is then amplified by a linear amplifier, sorted by a pulse-height analyzer (PHA), and then registered as a count. Different solid or liquid detectors are used for different types of radiation. For example, sodium iodide detectors containing a trace of thallium (NaI[Tl]) are used for γ and x-ray detection, whereas organic detectors such as anthracene and plastic fluors are used for β^- particle detection.

In liquid scintillation counting, a β^- emitting radioactive sample and an organic scintillator are dissolved in a solvent. The β^- particle interacts with solvent molecules emitting electrons. The latter interact with the organic scintillator, whereby light photons are produced, which are then directed to two PM tubes coupled in coincidence. A pulse is generated by the PM tube, which is registered as a count, as in the solid scintillation counting.

Organic scintillators usually have a lower density and, hence, a lower counting efficiency than inorganic scintillators. The decay time also limits the efficiency of a detector at high-count rates. The faster decay time allows high-count rate capability. The decay time for organic scintillators is much shorter than that for inorganic scintillators. For example, the decay time for

NaI(Tl) is $0.25\ \mu\text{sec}$ and that for anthracene is $0.026\ \mu\text{sec}$. The faster decay time permits the use of organic scintillators at higher count rates.

Solid Scintillation Detectors

See Table 8.1 for a summary of the various characteristics of the following detectors.

NaI (Tl) Detector

Pure sodium iodide does not produce any scintillation after interaction with γ -radiations at room temperature. However, if it is doped with a trace amount (0.1–0.4%) of thallium as an activator, NaI(Tl) becomes quite efficient in producing light photons after γ -radiations interact with it. NaI(Tl) molecules are excited or ionized by interaction with γ -rays or x-rays, and the high-energy states return to ground states by emitting light photons. Approximately 20–30 light photons are produced per 1 keV of energy.

The choice of NaI(Tl) crystals for γ -ray detection is primarily due to the high density ($3.67\ \text{g/cm}^3$) of the detector and the high atomic number of iodine ($Z = 53$), compared to organic scintillators. However, NaI(Tl) crystals are hygroscopic, and absorbed water causes color changes that reduce light transmission to the PM tubes. Therefore, the NaI(Tl) crystals are hermetically sealed in aluminum containers. Also, the entrance and sides of the crystals are coated with a reflective substance (e.g., magnesium oxide) so that light photons are reflected toward the photocathode of the PM tube (see later). These crystals are fragile, and must be handled with care. Room temperature should not be changed abruptly, because such changes in temperature can cause cracks in the crystal.

Bismuth Germanate Detector

The bismuth germinate ($\text{Bi}_4\text{Ge}_3\text{O}_{12}$ or BGO) detector has a higher density and effective atomic number and so higher attenuation coefficient (hence, higher stopping power) for 511 keV photons than NaI(Tl). But it has a slightly longer scintillation decay time (300ns) compared to NaI(Tl) (250ns) and its light output is relatively small causing poor energy resolution. However, energy resolution has minimal effect on the spatial resolution of PET, which is mainly determined by the size of the detectors. Moreover, BGO crystals are not hygroscopic. Because of these factors, BGO is preferred to NaI(Tl) for most positron emission tomography (PET) cameras.

Barium Fluoride Detector

Barium fluoride (BaF_2) is an inorganic crystal that has a very fast decay time (0.8ns) and offers a suitable detector for time-of flight PET. The photon yield of this crystal is relatively small and it is slightly hygroscopic.

TABLE 8.1. Characteristics of different solid scintillation detectors.*

Detectors	Effective atomic no (Z)	Density (g/cm ³)	Scintillation decay time (ns)	Photon yield (per keV)	Linear atten. coeff (cm ⁻¹) of 511 keV	Energy resolution (% at 511 keV)
NaI(Tl)	51	3.67	250	38	0.34 [†]	10
BGO	74	7.13	300	6	0.92	20
BaF ₂	54	4.89	0.6	2	0.44	11.4
GSO	59	6.71	60	10	0.70	8.5
Anthracene	—	—	26	30	—	—
LSO	66	7.40	40	29	0.87	10
YSO	34	4.53	70	46	0.39	12.5
CsI(Tl) [‡]	54	4.51	1000	52	0.48 [§]	—

* BGO, bismuth germanate; BaF₂, barium fluoride; LSO, lutetium oxyorthosilicate; YSO, yttrium oxyorthosilicate; GSO, gadolinium oxyorthosilicate.

[†] The linear attenuation coefficient of 140 keV photons of ^{99m}Tc in NaI(Tl) is 0.155 cm⁻¹.

[‡] Early P, Private communication, 2005.

[§] Of 660 keV.

Lutetium Oxyorthosilicate Detector

Lutetium oxyorthosilicate ($\text{Lu}_2[\text{SiO}_4]\text{O}(\text{Ce})$ or LSO) doped with cerium is another solid detector that is used for scintillation counting in PET imaging. LSO has a shorter scintillation decay time (40 ns) than BGO that favors the use of a narrow pulse window to cut down random coincidences in PET. Also, its higher light output gives a better energy resolution than BGO. These detectors have high efficiency for photon detection and can be fabricated in the size of a few millimeters. Many commercial manufacturers use LSO detectors in place of BGO detectors in clinical PET scanners and in micro-PET scanners for scanning small animals.

Gadolinium Oxyorthosilicate Detector

Gadolinium oxyorthosilicate (GSO) is another detector that can be used for coincidence counting in PET imaging. Even though it has lower light output and stopping power than LSO, its better energy resolution has prompted some commercial manufacturers to use it in PET scanners. GSO crystals are fragile and great care is warranted in their fabrication.

Yttrium Oxyorthosilicate Detector

Yttrium oxyorthosilicate (YSO) is another inorganic crystal similar to LSO introduced for scintillation counting. The scintillation decay time of YSO is 70 ns and it gives high light output. A combination detector of YSO/LSO has been reported for potential use in simultaneous single photon and coincidence imaging. YSO detects low-energy photons and LSO detects 511 keV photons, and the two pulses are readily separated by pulse shape discriminators.

Semiconductor Detectors

Germanium and Silicon Detectors

Semiconductor detectors or solid-state detectors are made of germanium or silicon materials commonly doped with lithium. These detectors are designated as Ge(Li) or Si(Li) detectors, of which the former are commonly used for high-energy γ -ray detection and the latter for α -particle and low-energy radiation detection. Currently, high purity germanium (HPGe) alone is commonly used. The basic principle of operation of these detectors involves ionization of the semiconductor atoms, as in gas detectors. Ionizations produced in the detector by radiation are collected as current and converted to voltage pulses through a resistor by the application of a voltage. The pulses are then amplified and counted. The size of the pulse is proportional to the radiation energy absorbed in the detector, but does not depend on the type of radiation.

Because semiconductors are much denser than gases, they are more efficient for x- and γ -ray detection than gas detectors. Also in semiconductor detectors, each ionization requires only about 3 eV compared to 35 eV in gas detectors. Thus, almost ten times more ions are produced in semiconductor detectors than in gas detectors for a given γ -ray energy, thus yielding a better spectral resolution of γ -ray photons of closer energies. Fabrication of Ge(Li) and Si(Li) detectors is quite time-consuming and expensive. The size of the detectors is also small, which prevents their use in gamma cameras.

Thermal noise at room temperature introduces a high background that can obscure the sample counts, but is reduced at low temperature. Therefore, these detectors are operated at low temperature usually employing liquid nitrogen (-196°C or 77°K). A disadvantage of these detectors is that liquid nitrogen evaporates over time and needs to be replenished periodically, typically weekly. Nowadays HPGGe detectors can be kept at room temperature when not used, and cooled when used for counting by means of helium-based cryoelectric and freon-based coolers.

Semiconductor detectors are most useful in differentiating photon energies because of the high-energy resolution, particularly in detecting radionuclidic contamination. These detectors are not in common use in nuclear medicine.

Cadmium–Zinc–Tellurium Detector

Cadmium–Zinc–Tellurium (CZT) detectors are another type of semiconductor made of Cd, Zn, and Te metals, and provide very high efficiency for γ -ray detection because of their high atomic numbers. For reasons of high detection efficiency, these detectors can be made as small as 2 mm thick and 2 mm diameter with almost 100% efficiency for 100 keV photons. The energy resolution of these detectors is very good for a wide range of γ -ray energies. These detectors are operated at room temperature. The electronics used are similar to those of other scintillation detectors. Different types of handheld probes have been devised for various purposes. One probe, called the Neoprobe 1000, is used for the detection of metastatic sites containing radioactivity (e.g., ^{125}I -labeled monoclonal antibody) during surgery for their removal by incision.

Cesium Iodide (CsI(Tl)) Detector

The CsI(Tl) detector has higher density and hence greater stopping power than the NaI(Tl) detector and also yields more light photons per keV. But its scintillation decay time is very long (1000 ns) resulting in longer dead time for the counting system. The crystal is weakly hygroscopic and does not require hermetic sealing. Unlike NaI(Tl), it can withstand a wide variation in temperature.




FIG. 8.1. A basic scintillation counter consisting of a NaI(Tl) detector, a photomultiplier (PM) tube, a preamplifier, a linear amplifier, a pulse-height analyzer, and a recording device. The high voltage applied to the PM tube is typically 1000V.

Solid Scintillation Counters

A basic solid scintillation counter consists of a scintillation detector, a PM tube, a preamplifier, a linear amplifier, a PHA, and a recording device (Fig. 8.1). The most commonly used scintillation detector in γ -ray counting is NaI(Tl), although BGO and LSO are commonly used in PET, discussed later. Each of these components is described in detail next.

NaI(Tl) Detector

The NaI(Tl) detectors are made of various sizes for different types of equipment. They vary from 3.8 to 50cm in diameter and 0.63 to 23cm in thickness. In thyroid probes and well counters, smaller and thicker crystals are used, whereas larger and thinner crystals are employed in scintillation cameras.

Photomultiplier Tube

A PM tube consists of a light-sensitive *photocathode* at one end, a series (usually 10) of metallic electrodes known as *dynodes* in the middle, and an *anode* at the other end—all enclosed in a vacuum glass tube (see Fig. 8.1).

The photocathode is usually an alloy of cesium and antimony that releases electrons after absorption of light photons. The PM tube is fixed on to the NaI(Tl) crystal with the photocathode facing the crystal by a special optical grease or connected to the crystal using light pipes.

A high voltage of ~ 1000 V is applied between the photocathode and the anode of the PM tube in steps of 50–150 V between dynodes (see Fig. 8.1). When light photons from the NaI(Tl) crystal strike the photocathode, photoelectrons are emitted, which are accelerated toward the next closest (i.e., first) dynode by the voltage difference between the electrodes. Approximately one to three photoelectrons are produced from the photocathode per 7 to 10 light photons. Each of these photoelectrons is accelerated to the second dynode and emits two to four electrons upon impingement. The accelerated electrons strike the successive dynodes, and more electrons are emitted. The process of multiplication continues until the last dynode is reached, where a pulse of 10^5 to 10^8 electrons is produced. The pulse is then attracted to the anode and finally delivered to the preamplifier. The amplitude of the pulse is proportional to the number of light photons received by the photocathode and in turn to the energy of the γ ray photon absorbed in the detector. The applied voltage must be very stable, because slight changes in dynode voltage cause a great variation in electron multiplication factor.

Preamplifier

The pulse from the PM tube is small in amplitude and is initially amplified by a preamplifier. The preamplifier adjusts the voltage of the pulse (pulse shaping) and matches impedance level between the detector and the subsequent circuits so that the pulse is appropriately processed by the system.

Linear Amplifier

A linear amplifier amplifies further the signal from the preamplifier and delivers it to the pulse height analyzer for analysis of its amplitude. The amplification of the pulse is given by the amplifier gain expressed as the ratio of the amplitude of the outgoing pulse to that of the initial pulse from the PM tube. The amplifier gains are given in the range of 1 to 1000 by gain control knobs provided on the amplifier. The output pulses normally have amplitudes of up to 10 V.

Pulse-Height Analyzer

γ -rays of different energies can arise from a source of the same or different radionuclides or can be due to scattering of γ -rays in the source and the detector. Thus, in counting a radioactive source, the pulses coming out of the amplifier may differ in magnitude. A PHA is a device that selects for

counting only those pulses falling within preselected voltage intervals or “channels” and rejects all others (see Fig. 8.1). Pulses corresponding to γ -ray energies of interest are selected by energy discriminator knobs, known as the *lower level* or *upper level*, or the *baseline* and *window*, provided on the PHA, and are ultimately delivered to the recording devices such as scalars, computers, films, and so on.

There are two modes of counting using PHAs: *differential* and *integral*. In differential counting, only pulses of preselected energies are counted by appropriate selection of lower and upper level knobs (discriminators) or the baseline and window. In scintillation cameras, however, differential counting is achieved by a peak voltage knob and a percent window knob. The peak voltage knob sets the energy of the desired γ -ray, and the percent window knob sets the window width in percentage of the γ -ray energy, which is normally placed symmetrically on each side of the peak voltage.

In integral counting, γ -rays of all energies or all γ -rays of energies above a certain energy are counted by setting the appropriate lower level or baseline and bypassing the upper level or window mechanism.

A PHA normally selects only one range of pulses corresponding to only one γ -ray energy by means of differential counting. Such a PHA is called a single-channel analyzer (SCA). A multichannel analyzer (MCA) is a device that can simultaneously sort pulses into many predetermined voltage ranges or channels, corresponding to different photon energies. By using an MCA, one obtains a simultaneous spectrum of different γ -ray energies from a radioactive source.

Display or Storage

Pulses processed by the PHA can be displayed on a cathode ray tube (CRT) or can be counted for a preset count or time by a scaler-timer device. A rate meter can be used to display the pulses in terms of counts per minute (cpm) or counts per second (cps). In scintillation cameras, pulses are used to form the image on a CRT and polaroid or x-ray films. These pulses can also be stored in a computer or on a magnetic tape or laser disc for processing later.

Gamma-Ray Spectrometry

Pulses are generated by the PM tube and associated electronics after the γ -ray energy is absorbed in the NaI(Tl) detector. Because γ -rays interact with the NaI(Tl) detector by photoelectric, Compton, and pair production mechanisms, and also because various scattered radiations from outside the detector may interact with the detector, a distribution of pulse heights will be obtained depicting a spectrum of γ -ray energies. Such a γ -ray spectrum may result from a single γ -ray or from many γ -rays in a sample. Different features of this spectrum are discussed here.




FIG. 8.2. γ -ray spectra. (A) An ideal spectrum would represent the different γ -rays as lines. (B) An actual spectrum showing the spread of the photopeak that is due to statistical fluctuations in the pulse formation.

Photopeak

In an ideal situation, if the γ -ray photon energy is absorbed by the photoelectric mechanism and each γ -ray photon yields a pulse of the same height, then each γ -ray would be seen as a line on the γ -ray spectrum (Fig. 8.2A). In reality, the photopeak is broader, which is due to various statistical variations in the process of forming the pulses. These random fluctuations arise from the following conditions:

1. Because 20–30 light photons are produced for every keV of γ -ray energy absorbed, there is a statistical variation in the number of light photons produced by the absorption of a given γ -ray energy in the detector. Also, statistically all light photons produced may not strike the photocathode.
2. As already stated, 7 to 10 light photons are required to release 1 to 3 photoelectrons from the photocathode. Therefore, the number of photoelectrons that one γ -ray will produce may vary from one event to another.
3. The number of electrons released from the successive dynodes by impingement of each electron from the previous dynode varies from 2 to 4, and therefore pulse heights from the PM tube will vary from one γ -ray to the next of the same energy.

All of the preceding statistical fluctuations in generating a pulse cause a spread in the photopeak (see Fig. 8.2B). A typical spectrum of the 662-keV γ -ray of ^{137}Cs is shown in Figure 8.3.

Compton Valley, Edge, and Plateau

When γ -rays interact with the NaI(Tl) detector via Compton scattering and scattered photons escape from the detector, the Compton electrons result




FIG. 8.3. A typical spectrum of the 662-keV γ -ray of ^{137}Cs illustrating the photopeak, Compton plateau, Compton edge, Compton valley, backscatter, characteristic lead K x-ray, and barium K x-ray peaks.

in pulse heights that are smaller than that of the photopeak. The Compton electrons, however, can have variable energies from zero to E_{max} , where E_{max} is the kinetic energy of those electrons that are produced by the 180° Compton backscattering of the γ -ray photons in the detector. At relatively high photon energy, E_{max} is given by the photon energy minus 256 keV (Eq. 6.3). Thus, the γ -ray spectrum will show a continuum of pulses corresponding to Compton electron energies between zero and E_{max} . The peak at E_{max} is called the *Compton edge*, and the portion of the spectrum below the Compton edge down to about zero energy is called the *Compton plateau* (see Fig. 8.3). The portion of the spectrum between the photopeak and the Compton edge is called the *Compton valley*, which results from multiple Compton scattering of a γ -ray in the detector yielding a narrow range of pulses in this region.

The relative heights of the photopeak and the Compton edge depend on the photon energy as well as the size of the NaI(Tl) detector. At low energies, photoelectric effect predominates over Compton scattering, whereas at higher energies the latter becomes predominant. In larger detectors, γ -rays may undergo multiple Compton scattering, which can add up to the absorption of the total photon energy identical to the photoelectric effect. This increases the contribution to the photopeak and decreases to the Compton plateau.

Characteristic X-Ray Peak

Photoelectric interactions of the γ -ray photons in the lead shield around the detector can lead to the ejection of the K -shell electrons, followed by tran-

sition of electrons from the upper shells, mainly the L shell, to the K shell. The difference in binding energy between the K -shell electron ($\sim 88\text{keV}$) and the L -shell electron ($\sim 16\text{keV}$) appears as lead K x-ray of $\sim 72\text{keV}$. These characteristic x-ray photons may be directed toward the detector and absorbed in it and may appear as a peak in the γ -ray spectrum (see Fig. 8.3). These photons can be reduced by increasing the distance between the detector and the shielding material.

Backscatter Peak

When γ -ray photons, before striking the detector, are scattered at 180° by Compton scattering in lead shielding and housing, and the scattered photons are absorbed in the detector, then a peak, called the *backscatter peak*, appears in the γ -ray spectrum (see Fig. 8.3). For high-energy photons, the backscattered peak appears at 256keV [see Eq. (6.3)]. This peak can be mostly eliminated by increasing the distance between the shield and the detector.

Iodine Escape Peak

Photoelectric interaction of γ -ray photons with iodine atoms of the NaI(Tl) detector usually results in the emission of characteristic K x-rays. These x-ray photons may escape the detector, resulting in a peak equivalent to photon energy minus 28keV (binding energy of the K -shell electron of iodine). This is called the *iodine escape peak*, which appears about 28keV below the photopeak (Fig. 8.4). This peak becomes prominent when the energy of the photon is less than about 200keV , because, at energies above




FIG. 8.4. A spectrum of 81-keV γ -ray of ^{133}Xe showing an iodine escape peak.




FIG. 8.5. A spectrum of ^{111}In with 171- and 245-keV photons showing a coincidence (sum) peak at 416 keV.

200 keV, the iodine escape peak would fall within the width of the photopeak, because of the small differences between the two peaks.

Annihilation Peak

γ -rays with energy greater than 1.02 MeV may undergo pair production in the detector in which a positive-negative electron pair is produced. The β^+ -particles are annihilated to produce two 511-keV photons, which appear as photopeaks in the γ -ray spectrum. If, however, one of the 511-keV photons escapes from the detector, then a peak, called the *single-escape peak*, corresponding to the primary photon energy minus 511 keV, will appear in the spectrum. If both annihilation photons escape, then a *double-escape peak* results, corresponding to the primary photon energy minus 1.02 MeV. Larger detectors can prevent the escape of the annihilation radiations.

Coincidence Peak

A *coincidence* or *sum peak* results when more than one photon is absorbed simultaneously in the detector to be considered as a single event. The peak equals the sum of the energies of the photons. Such situations occur with radionuclides that have short-lived isomeric states and thus emit γ -rays in cascade. For example, ^{111}In emits 171- and 245-keV photons, which can result in a sum peak of 416 keV (Fig. 8.5). Sum peaks are also caused by counting high-activity samples in which two photons may strike the detector at the same time. These peaks can be reduced by counting the samples at larger distances between the source and the detector or by using smaller

detectors so that the likelihood of two photons striking the detector at the same time is reduced. In the case of high-activity samples, the level of activity has to be reduced either by dilution or allowing to decay, in order to reduce the sum peak.

Liquid Scintillation Counters

Low-energy β^- -particles are normally absorbed within the source and in the window and walls of the detectors, and therefore β^- -emitters are difficult to detect in gas or solid detectors. For this reason, β^- -emitting radionuclides are counted using the liquid scintillation technique in which the radioactive sample is mixed with a scintillating material. A sample vial containing the liquid scintillator and the radioactive sample of interest is placed between two PM tubes connected in coincidence (Fig. 8.6). Each PM tube receives the light photons emitted by the interaction of the by β^- -particle with the scintillator and converts them into a pulse, which is further amplified by an amplifier. The two amplified pulses are then delivered to the coincidence circuit that contains a PHA to analyze the pulse height for acceptance. A count is registered in the scaler if an event is recorded in both PM tubes simultaneously. Such coincidence counting reduces the background counts due to noise, including terrestrial and cosmic radiations, radioactive patients, etc.

The liquid scintillation solution is prepared by dissolving a primary scintillating solute or fluor and often a secondary fluor in a solvent. The radioactive sample is added to and thoroughly mixed with the scintillating solution




FIG. 8.6. A schematic diagram of a liquid scintillation counting system. Light photons emitted from the sample strike the two photomultiplier tubes to produce pulses. Only coincident pulses are counted.

for counting. The primary fluors include 2,5-diphenyloxazole (PPO), 2,5-bis-2-(5-T-butyl-benzoxazolyl)-thiophene (BBOT), and *p*-terphenyl, of which PPO is most commonly used in a concentration of 5 g/L.

Toluene, xylene, and dioxane are the most common solvents that easily dissolve the primary fluor and often the radioactive sample, which is a requirement for a good solvent. These solvents, however, are poorly miscible in water, and therefore their disposal in the sewer system is restricted. For this reason, biodegradable solvents such as linear alkylbenzene and phenylxylylethane are widely used. Counting vials are usually glass or plastic, but the latter is not used when toluene or xylene is used as a solvent because the solvent tends to dissolve plastic.

When radiations pass through the solvent, electrons are released from the solvent molecules after absorption of radiation energies. These electrons transfer energy to primary fluor molecules, which then emit light photons for further processing by PM tubes and associated electronics. The wavelength of these light photons may be somewhat shorter than required for the spectral sensitivity of the photocathode of the PM tube. This mismatch is rectified by adding a secondary fluor or solute, called the *wavelength shifter*, to the scintillating solution. The wavelength shifter absorbs the light photons emitted by the primary fluor and reemits them with a longer wavelength, which is more suitable for the photocathode of the PM tube. The compound 1,4-bis-2-(5-phenyloxazolyl)-benzene (POPOP) is most commonly used as a secondary solute in a concentration of about 0.1%.

An attempt is always made to keep the radioactive sample in solution in the liquid scintillator. Solubilizing agents are added to improve dissolution of specific samples, and the common example is the hydroxide of Hyamine 10-X used in counting tissue samples.

In liquid scintillation counting, quenching is a problem caused by interference with the production and transmission of light, which ultimately reduces the detection efficiency of the system. Quenching can be of the following types:

1. *Chemical type*, resulting from interference in energy transfer by substances such as samples or extraneous materials (e.g., dissolved O₂)
2. *Color type*, resulting from absorption of light photons by colored substances, such as hemoglobin, before striking the PM tube
3. *Dilution type*, resulting from relatively large dilution of the scintillation mixture, in which case many light photons may be absorbed by the diluted sample.
4. *Optical type*, resulting from absorption of light by a dirty vial containing frost or fingerprints.

Quenching must be corrected to obtain accurate counting of samples, and three methods have been adopted for this purpose, namely, internal standard method, channel ratio method, and external standard method.

The readers are referred to reference physics books for details of these methods.

A problem with liquid scintillation counting is the noise due to spontaneous thermal emission of electrons from the photocathode of the PM tube. Background noise also arises from the interaction of light with scintillation solution. Thermal emission of electrons is reduced by refrigeration of the counting chamber to keep the PM tubes at low temperature. But the coincidence counting is the most effective method to reduce the noise.

The liquid scintillation counting systems are provided with automatic sample changers for counting as many as 500 samples. Also, one to five PHAs are available on a liquid scintillation counter, so that β^- -particles of different energies can be counted simultaneously by using different baselines and windows on each PHA. The β^- -emitters, ^3H , ^{14}C , ^{32}P , and ^{35}S , are commonly detected by liquid scintillation counting. Whereas the counting efficiencies of ^3H ($E_{\max} = 0.018\text{ MeV}$) and ^{32}P ($E_{\max} = 1.71\text{ MeV}$) are $\sim 60\text{--}70\%$ and $\sim 100\%$ respectively, they are negligible for γ and x-rays.

Characteristics of Counting Systems

Detection of radiation and therefore counting of radioactive samples is affected by different characteristics of the detector and the associated electronics. The following is a discussion of these properties.

Energy Resolution

As already mentioned, even though γ -rays of the same energy are absorbed in the NaI(Tl) detector by the photoelectric effect, pulses of different amplitudes are produced because of the statistical variations in the production of light photons in the detector and photoelectrons and electrons in the PM tube. This results in the broadening of the photopeak. The width of the peak or the sharpness of the peak (i.e., the energy resolution of the detector) predicts the ability of the NaI(Tl) spectrometer to discriminate between the γ -ray photons of dissimilar energies. A similar situation exists for semiconductor detectors where the number of ionizations may vary from one γ -ray to another of the same energy, leading to the broadening of the peak.

The energy resolution of a system is given by the full width at half-maximum (FWHM) amplitude of the photopeak and is expressed as a percentage of the photon energy as follows:

$$\text{Energy resolution (\%)} = \frac{\text{FWHM}}{E_\gamma} \times 100 \quad (8.1)$$

where E_γ is the energy of the γ -ray photon. In Figure 8.7, FWHM is 55 keV for the 662-keV peak of ^{137}Cs ; therefore,




FIG. 8.7. The full width at half maximum (FWHM) of the 662-keV γ -ray of ^{137}Cs in a NaI(Tl) detector.

$$\begin{aligned} \text{Energy resolution (\%)} &= \frac{55}{662} \times 100 \\ &= 8.3\% \end{aligned}$$

The energy resolution depends on the photon energy. The higher the photon energy, the better the energy resolution (i.e., smaller FWHM), because of the decrease in the percentage of statistical variations in the pulse production. The FWHM (%) of NaI(Tl) detectors is about 7% to 10% for the 662-keV γ -ray of ^{137}Cs and 10% to 14% for the 140-keV γ -ray of $^{99\text{m}}\text{Tc}$. In contrast, the FWHM (%) in Ge(Li) detectors is about 0.42% for 140-keV γ -rays and about 0.2% for photons of more than 1 MeV.

Detection Efficiency

The detection efficiency of a counter is given by the observed count rate divided by the disintegration rate of a radioactive sample. The count rate of a sample differs from the disintegration rate because of several factors. Radiations from a source are emitted isotropically around 4π steradians, but only a fraction of all photons emitted strikes the detector, depending on the solid angle subtended by the detector on the source. Only a fraction of all photons striking the detector may interact in the detector and produce pulses. Only a fraction of all pulses produces a single photopeak. Furthermore, the count rate is affected by the abundance of a particular radiation

from a radionuclide. Considering these factors, the overall counting efficiency of a counter for a radiation is given by the following expression:

$$\text{Efficiency} = f_i \times f_p \times f_g \times N_i \quad (8.2)$$

where f_i is the intrinsic efficiency; f_p is the photopeak efficiency, or photofraction; f_g is the geometric efficiency; and N_i is the abundance of the radiation in question. N_i is available in literature on Tables of Isotopes.

Intrinsic Efficiency

The fraction of all radiations of a given type and energy impinging on the detector that interacts with it to produce pulses is called the *intrinsic efficiency*, f_i , of the detector:

$$\begin{aligned} f_i &= \frac{\text{No. of radiations detected by the detector}}{\text{No. of radiations impinging on the detector}} \\ &= \frac{\text{All counts under the entire spectrum}}{\text{No. of radiations impinging on the detector}} \end{aligned} \quad (8.3)$$

It includes all photons undergoing both photoelectric absorption and Compton scattering. Intrinsic efficiency depends on the type and energy of the radiation and the linear attenuation coefficient (μ) and thickness of the detector. The dependence of intrinsic efficiency on the photon energy and the detector thickness is illustrated in Figure 8.8. The value of f_i is almost 1 for low-energy γ -rays and thicker detectors. The f_i tends to 0 for high-energy γ -rays and thinner detectors. These conditions apply to all solid scintillation detectors. The intrinsic efficiency of gas detectors is almost unity for α - and β -particles but is about 0.01 (1%) for γ and x-rays.

Photopeak Efficiency or Photofraction

The fraction of all detected γ -rays that contributes only to the photopeak is called the *photopeak efficiency*, or *photofraction* (f_p). It is given as the total photopeak area divided by the total area under the entire spectrum:

$$f_p = \frac{\text{All counts under the photopeak}}{\text{All photons detected by the detector}} \quad (8.4)$$

This value is affected by all factors that influence photoelectric effect, such as the size and composition of the detector and γ -ray energy, but is primarily determined by the discriminator settings on the PHA. The f_p increases with increasing window width of the PHA.

Geometric Efficiency

Radiations from a source are emitted uniformly with equal intensity in all directions. If a source of radiation is placed at a distance from a detector,




FIG. 8.8. Dependence of intrinsic efficiency on photon energy and detector thickness.

then only a fraction of all radiations emitted from the source will be detected by the detector. This fraction is determined by the solid angle subtended by the detector on the source. The geometric efficiency, f_g , is equal to the number of radiations striking the detector divided by the total number of radiations emitted by the source. Thus,

$$f_g = \frac{\text{No. of radiations striking the detector}}{\text{Total number of radiations emitted by the source}} \quad (8.5)$$

For a circular detector with radius r , it is equal to the area πr^2 of the detector divided by the total spherical area $4\pi R^2$, where R is the distance between a point source S and the detector D (Fig. 8.9).

$$f_g = \frac{\pi r^2}{4\pi R^2} \quad (8.6)$$

As the distance R between the source and the detector increases, the f_g decreases, according to the inverse square law, that is, $f_g \propto 1/R^2$ (Fig. 8.9A). Thus the f_g at $2R$ is one fourth of the f_g at R . The value of f_g increases with the size of the detector. Also, the finite size of the radiation source affects the f_g values.

When the source and the detector are in close contact, the f_g tends to be about 50% (Fig. 8.9B). In the case of gamma well counters and liquid scintillation counters, the f_g approaches 100% (Fig. 8.9C).

Dead Time

Each counting system takes a certain amount of time to process a radiation event, starting from interaction of radiation with the detector all the way up to forming a pulse and ultimately recording it. The counter remains insensitive to a second event for this period of time, that is, if a second radiation arrives during this time, the counter cannot process it. This period is called the *dead time*. When the counter recovers after this period, only then can a second radiation be detected. Thus the second event arriving during the dead time is lost. Counts lost during the dead time are called the *dead-time loss*. In scintillation detectors two events may be processed simultaneously to form a single event of amplitude that is equal to the sum of the amplitudes of both events. This is referred to as *pulse pileup*. If one or both of the events were photopeaks originally, then the combined peak will fall outside the PHA window setting and be lost. Dead time loss at high count rates is a serious problem for any counting system and is more so for scintillation cameras due to pulse pileup (see Chapter 11).

The dead time of a counting system may arise from different components of the entire counting system: detector, PHA, PM tube, scaler, computer interface, and so on. While Geiger-Müller (GM) detectors have a longer dead time of 100 to 500 microseconds, the typical values for NaI(Tl) and semiconductors are of the order of 0.5 to 10 microseconds and for liquid scintillators, ~0.1 to 1 microsecond (Cherry et al., 2003).




FIG. 8.9. Illustration of geometric efficiency, f_g , of a detector D with a circular area, πr^2 , where r is the radius of the detector. (A) The detector D placed at a distance R from the point source S has an f_g four times greater than the f_g when the detector is placed at a distance $2R$. (B) When the source and the detector are in close contact, the f_g is about 50%. (C) When the source is well inside the detector as in a well counter, the f_g approaches 100%.




FIG. 8.10. Plot of observed count rates versus true count rates indicating the dead time loss in paralyzable and nonparalyzable systems.

Based on how successive pulses are processed owing to the dead time, the counting systems fall into two categories: *paralyzable* and *nonparalyzable*. In paralyzable systems, each event sets its own dead time, even if it arrives within the dead time of the previous event and is not counted. Each event prolongs the dead time induced by the previous event, and thus adds to the total dead time of the system, whereby a paralyzable system can become totally unresponsive to process events if the count rate of the source is very high. On the other hand, in nonparalyzable systems, the instrument remains insensitive to successive events for a period of time equal to the dead time, and these events are lost. But unlike paralyzable systems, the dead time is not changed or lengthened. When the system recovers after the detection of the first event, only then is the second event processed and detected. The two types of dead time losses are illustrated in Figure 8.10.

The paralyzable and nonparalyzable systems can be represented by mathematical relationships among the observed count rate R_o , true count rate R_t , and dead time τ . For nonparalyzable systems,

$$R_t = R_o / (1 - R_o \tau) \quad (8.7)$$

and for paralyzable systems,

$$R_o = R_t e^{-R_t \tau} \quad (8.8)$$

Different components of a radiation detection system can have either paralyzable or nonparalyzable dead time. Scalers and pulse-height analyzers are nonparalyzable systems, whereas radiation detectors themselves

are paralyzable systems. Scintillation cameras have both paralyzable and nonparalyzable components of dead time. As can be seen from Eq. (8.8), the radiation detectors become totally paralyzed at very high count rates giving no reading.

Dead time loss is a serious problem for a counting system at high-count rates. Therefore, either count rates must be lowered or corrections must be made to the observed count rates. There are several methods to determine or correct for dead time. An empirical method is to plot the observed count rates as a function of increasing concentrations of known activity. From the plot and Eqs. (8.7) and (8.8), the dead time is calculated for the nonparalyzable or paralyzable system. For subsequent measurements of unknown samples, correction is made to compensate for the dead time loss giving true count rates. Another method uses two radioactive sources, which are counted in the counter individually and together. From these three measurements, one can calculate dead time using appropriate equations (see Cherry et al., 2003). Various techniques, such as use of buffers, in which overlapping events are held off during the dead time, use of pulse pileup rejection circuits, and use of high-speed electronics have been employed to improve the dead time correction.

Gamma Well Counters

The gamma well counter consists of a NaI(Tl) detector with a hole in the center for a sample to be placed and associated electronics such as a PM tube, preamplifier, amplifier, PHA, and scaler-timer. Placing a radioactive sample in the central hole of the detector increases the geometric efficiency (almost 99%) and hence the counting efficiency of the counter. The NaI(Tl) detectors have dimensions in the range of 5-cm diameter \times 5-cm thick to 23-cm diameter \times 23-cm thick. Smaller detectors are used for low-energy γ -rays (less than 200 keV), and larger detectors are used for high-energy γ -rays. Most well counters are shielded with about 8.5-cm thick circular lead ring to reduce background from cosmic rays, natural radioactivity such as ^{40}K , or background activity in the work area. A typical well counter detector is shown in Figure 8.11.

Calibration of Well Counters

It is essential that the dial settings of the discriminators on the PHA are calibrated so that the dial readings correspond directly to the pulse height (i.e., the energy of the γ -ray photon); that is, the dial readings can be read in units of keV. This calibration is called the *high-voltage* or *energy calibration* of the well counter. The energy calibration is carried out by using the 662-keV photons of ^{137}Cs . After placing a ^{137}Cs source in the well counter, the lower and upper discriminator levels are set at 640 divisions




FIG. 8.11. A schematic diagram of a NaI(Tl) well counter with a PM, photomultiplier tube.

and 684 divisions, respectively, thus assigning the center of the photopeak at 662 divisions corresponding to the 662-keV γ -ray. Starting from low values, the high voltage is increased in small increments for a given amplifier gain until the observed count rate reaches a maximum. The high voltage at the maximum count rate is kept as the operating voltage for subsequent counting of photons of different energies. The discriminator dials are then said to be energy calibrated; for example, each dial unit corresponds to 1 keV at an amplifier gain of 1. Thus, the center of the 140-keV photopeak of ^{99m}Tc can be set at 140 divisions of the discriminator setting, with lower and upper values set as desired. After calibration, well counters should be checked regularly for any voltage drift using a long-lived source, such as ^{137}Cs .

Counting in Well Counters

For relative comparison of count rates between samples, the well counter does not need to be calibrated, provided all samples for comparison have the same volume. In radioimmunoassays, ferrokinetics, blood volume, red cell mass measurements, a standard of the same geometry (volume) and with relatively the same activity is counted along with all samples, and then a comparison is made between each sample and the standard. However, when the absolute activity of a radioactive sample needs to be determined, then the detection efficiency of the counter must be measured for the γ -ray energy of interest using a standard of the radioactive sample of known activity. The photopeak efficiency is determined from the count rate of the standard at the appropriate PHA setting divided by the disintegration rate from the known activity of the standard. The efficiency correction can then be applied to the count rates of samples of unknown radioactivity when counted at the same setting as the standard to give the absolute activity. For

absolute activity, the photopeak efficiency must be determined for each photon energy.

When multiple γ -rays, either from a single radionuclide or from many radionuclides, are present in a radioactive sample, then the energy spectrum becomes complicated by the overlapping of different photopeaks and also by Compton contributions from the high-energy photons to the low-energy photopeaks. The latter contributions are termed the *spillover*, or *crosstalk contributions*.

Figure 8.12 illustrates an energy spectrum of the 140-keV peak of ^{99m}Tc and 364-keV peak of ^{131}I , in which the Compton contribution from the 364-keV peak to the 140-keV peak is shown. Corrections must be made for this spillover to the 140-keV peak. This is accomplished by counting a sample of pure ^{131}I in both 140-keV and 364-keV discriminator settings and determining the percentage of spillover from the ratio of the counts in the 140-keV photopeak to those in the 364-keV photopeak.

Effects of Sample Volume

The sample volume affects the counting efficiency of well counters. As the sample volume for a given activity is increased, more radiations are lost through the opening of the well without interacting in the detector, and hence, the counting efficiency drops. Therefore, correction factors should be determined for different sample volumes and applied to the measured activity.




FIG. 8.12. A combined spectrum of the 140-keV γ -ray of ^{99m}Tc and 364-keV γ -ray of ^{131}I . The dotted line under the 140-keV photopeak is the spillover, or crosstalk, contribution from the 364-keV photon.

Well counters are available with automatic sample changers having provisions of counting as many as 500 samples. Most counters are programmable with computers and provide printouts with various information on counting. The major advantage of the well counter is its high detection efficiency due to increased geometric efficiency, which approaches almost 100% depending on the volume of the sample. The detection efficiency of a well counter decreases with increasing photon energy and decreasing detector size. Typically, the overall detection efficiency is close to 100% for 140-keV photons of ^{99m}Tc and 30% to 90% for 364-keV photons of ^{131}I , depending on the detector size.

Thyroid Probe

The thyroid probe is a counter commonly employed to measure the uptake of ^{131}I or ^{123}I in the thyroid gland after the oral administration of a ^{131}I -NaI or ^{123}I -NaI capsule. It consists of a NaI(Tl) detector, 5 cm in diameter by 5 cm in thickness, and other associated electronics, as in a well counter. The operation of the probe is similar to that of a well counter.

One of the differences between the well counter and the thyroid probe is that the latter requires a collimator, which limits the field of view on the thyroid. The collimator is a 20- to 25-cm long cylindrical barrel made of lead and covers the detector as well as the PM tube (Fig. 8.13). This reduces the background activity from the γ -radiations from areas outside the thyroid reaching the detector.

The efficiency of a thyroid probe varies inversely with the square of the distance between the detector and the thyroid. The probe is initially cali-




FIG. 8.13. A schematic diagram of a thyroid probe. PHA, pulse-height analyzer; PM, photomultiplier.

brated for photon energies in the same manner as the well counter using the 662-keV γ -ray energy of ^{137}Cs , and then discriminator settings are set for the 364-keV γ -ray of ^{131}I . Attenuation of photons in the thyroid tissues reduces the overall detection efficiency of the probe.

Photons scattered in the thyroid gland by Compton scattering may reach and interact in the detector because they originate in the field of view and are not stopped by the collimator thickness. These scattered photons, however, are excluded from the total measured counts by selecting the appropriate lower and upper discriminator settings on the PHA for the 364-keV γ -ray of ^{131}I .

Thyroid Uptake Measurement

In the thyroid uptake test, a ^{131}I -NaI capsule containing about 10 to 15 μCi (0.37–0.55 MBq) of ^{131}I is measured in a lucite thyroid phantom at a fixed distance using the thyroid probe and the settings for 364-keV photons of ^{131}I . The thickness and composition of the lucite phantom are equivalent to those of the patient's neck. This count is considered as the standard count. The capsule is then administered to the patient orally, and the thyroid count is obtained at the same distance as the standard count 24 hr after administration. The room background count is taken to subtract from the standard count, and the thigh count is taken as background to subtract from the thyroid count. The thyroid uptake is then calculated as follows:

$$\% \text{ uptake} = \frac{(A - B) \times 100}{(C - D)} \quad (8.9)$$

where A is the thyroid count, B is the thigh count, C is the standard count corrected for 24-hr decay, and D is the room background. Identical procedures are employed with ^{123}I -NaI using $\sim 300 \mu\text{Ci}$ (11.1 MBq). At times, the 6-hr thyroid uptake also is determined depending upon the clinical judgement of the physicians.

Questions

- Describe the mechanism of γ -ray interaction in the NaI(Tl) detector. In γ -ray counting, why is NaI(Tl) commonly chosen as the detector?
- Describe the operation of a photomultiplier (PM) tube.
 - What is the typical high voltage applied to the PM tube?
 - What are the photocathodes commonly made of?
 - How many photoelectrons are emitted from the photocathode for each keV of photon energy?
- Ideally, a photopeak should appear as a line in a γ -ray spectrum. Indicate different factors that contribute to the broadening of the photopeak.

- (b) A photopeak is due to only photoelectric effect of γ -rays, or due to all γ -rays that deposit full energy in the detector. True or false?
4. (a) Describe the function of a pulse-height analyzer (PHA).
 (b) Do the following factors affect the size (pulse height) of the photopeak pulses?
 - i. Gain of the amplifier
 - ii. High voltage of PM tubes
 - iii. Distance between the source and the detector
 - iv. Light photons produced in the detector
 5. In a γ -ray spectrum, describe the origins of:
 - (a) Backscatter peak
 - (b) Compton valley
 - (c) Characteristic K x-ray peak
 - (d) Iodine escape peak
 - (e) Sum peak
 6. (a) Describe the principles of a liquid scintillation counter.
 (b) What is a scintillation solution, and how does it work?
 (c) What is the purpose of using a secondary solute to the scintillation solution?
 (d) What are the most common solvents for liquid scintillation counting?
 (e) Can you count ^3H (β^- energy = 0.018 MeV) and ^{14}C (β^- energy = 0.156 MeV) in the same sample using a liquid scintillation counter equipped with three PHAs?
 7. (a) Define the energy resolution of a detector.
 (b) For a given detector, the energy resolution of low-energy photons is poorer than that of high-energy photons. True or false?
 (c) For NaI(Tl) detectors, the energy resolution should be less than 10% for the 662-keV photon of ^{137}Cs . True or false?
 8. (a) A point source of $^{99\text{m}}\text{Tc}$ is placed 10 cm away from a NaI(Tl) detector that has a diameter of 20 cm. Calculate the geometric efficiency.
 (b) What would be the geometric efficiency if the source were placed on the surface of the detector?
 9. (a) Explain the dead time and pulse pileup of a counter.
 (b) What is the distinction between the paralyzable and nonparalyzable systems?
 (c) What are the typical dead times for Geiger-Müller (GM) counters and NaI(Tl) counters?
 10. (a) Describe the energy calibration of a NaI(Tl) well counter.
 (b) Why does the count rate differ from the disintegration rate of a sample of a radionuclide?
 (c) How does the spillover volume affect the count rate?
 11. What are the spillover or crosstalk contributions in a spectrum of several γ -rays? How would you correct for them?

12. A radioactive sample has two γ -ray photons of 130- and 120-keV energies. If a NaI(Tl) crystal has an energy resolution of 10% at 125 keV, could the two photons be detected as separate photopeaks?
13. Both gas-filled detectors and semiconductor detectors operate by ionization of atoms by radiation. Why do semiconductor detectors give better energy resolution than gas-filled detectors?
14. A patient is given orally a $10\text{-}\mu\text{Ci}$ ^{131}I -NaI capsule. Before administration, the count rate of the capsule in a thyroid phantom is 297,000 cpm. The 24-hour count rate of the patient's thyroid is 168,000 cpm. If the room background is 200 cpm and the patient's thigh count rate is 1000 cpm, calculate the thyroid uptake.
15. High-activity sources such as radiopharmaceutical dosages and x-ray exposure outputs are better measured with ionization chambers than GM counters and NaI(Tl) well counters. Why?
16. Which of the following counters can detect individual events of the radiation interacting with the detector?
 - (a) Ionization chamber
 - (b) GM counter
 - (c) NaI(Tl) well counter
17. What type of Compton scattering causes the Compton edge of a γ -ray spectrum?
18. Discuss the properties of newer detectors. Explain why BGO is preferably used in PET cameras.

Suggested Readings

- Bushberg JT, Seibert JA, Leidholdt EM Jr, Boone JM. *The Essential Physics of Medical Imaging*. 2nd ed. Philadelphia: Lippincott Williams & Wilkins; 2002.
- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia: W.B. Saunders; 2003.
- Craddock TD. Fundamentals of scintillation counting. *Semin Nucl Med*. 1973; 3:205–223.
- Hendee WR, Ritenour ER. *Medical Imaging Physics*. 4th ed. New York: Wiley-Liss; 2002.
- Hine GJ. Sodium iodide scintillators. In: Hine GJ, eds. *Instrumentation in Nuclear Medicine*. New York: Academic Press; 1967;I:95–117.
- Peng CT, Horrocks DL, Alpen EL, eds. *Liquid Scintillation Counting*. New York: Academic Press; 1980; I, II.
- Rollo FD, ed. *Nuclear Medicine Physics, Instrumentation and Agents*. St. Louis: Mosby; 1977.

9

Gamma Cameras

The principles of nuclear medicine studies are based on the assessment of radionuclidic distribution in different parts of a given organ after in vivo administration of a radiopharmaceutical to distinguish between the normal and abnormal tissues. Such assessment of radionuclide distribution is performed by gamma cameras that primarily comprise NaI(Tl) detectors and the associated electronics described in the previous chapter. The gamma cameras permit the dynamic acquisition of the images with better spatial resolution, and can be oriented in any direction around the patient. Various aspects of gamma cameras are discussed below.

Gamma Cameras

The gamma or scintillation camera is an imaging device that is most commonly used in nuclear medicine. It is also called the Anger camera in honor of Hal O. Anger, who invented it in the late 1950s. Gamma cameras detect radiation from the entire field of view simultaneously and therefore are capable of recording dynamic as well as static images of the area of interest in the patient. Various designs of gamma cameras have been proposed and made available, but the Anger camera with a single crystal is by far the most widely used. Although many sophisticated improvements have been made on the gamma cameras over the years, the basic principles of the operation have essentially remained the same.

Principles of Operation

The gamma camera usually consists of several components: a detector, a collimator, PM tubes, a preamplifier, an amplifier, a pulsed-height analyzer (PHA), an X -, Y -positioning circuit, and a display or recording device. A schematic diagram of a gamma camera is illustrated in Figure 9.1, and a commercial gamma camera is shown in Figure 9.2. The detector, PM tubes, and amplifiers are housed in a unit called the *detector head*, which is




FIG. 9.1. A schematic electronics diagram of a gamma camera. CRT, cathode ray tube; PHA, pulse-height analyzer; PM, photomultiplier.




FIG. 9.2. A typical single-head gamma camera. (Courtesy of Siemens Medical Solutions USA, Inc.)

mounted on a stand. The head can be moved up or down and rotated with electrical switches to position it in the field of view on the patient. The X -, Y -positioning circuits, PHA, and some recording devices are mounted on a console. In the past, the cameras were operated by switches and dials on the console. Currently, much of the operation of the camera is performed by a computer built in it. The computer is run by appropriate software in conjunction with a keyboard, a mouse, and a video monitor. High voltage, window, and photopeaks are all set by the operator's choice of parameters. Acquisition of the data and processing of the data are carried out by the computer. Whereas stationary cameras are permanently installed at desired locations, portable gamma cameras are mounted on wheels for use in situations requiring movement of the camera from room to room, such as to the patient's bedside. Mobile cameras are installed in wheeled vans such that they can be moved to places where gamma cameras are not available for nuclear medicine studies.

The operational principles of a gamma camera are identical to those of solid scintillation counters described in Chapter 8. Basically, γ -rays from a source interact with the NaI(Tl) detector, and light photons are emitted. The latter strike the photocathode of PM tubes, and a pulse is generated, which is then amplified by an amplifier and sorted out by a PHA. Finally, the pulse is positioned by an X -, Y -positioning circuit on the recording device or stored in the computer, corresponding to the location of γ -ray interaction in the detector.

The functions of PM tubes, preamplifier, amplifier, PHA, and recording devices are the same as described in Chapter 8, and therefore only essential features pertaining to the use of gamma cameras will be highlighted.

Detector

Detectors used in gamma cameras are typically circular NaI(Tl) detectors, which have dimensions of 25–50 cm in diameter and 0.64–1.84 cm in thickness. The most common thickness is 0.95 cm. The 0.64-cm thick detectors are used in mobile gamma cameras and are useful for ^{201}Tl , $^{99\text{m}}\text{Tc}$, and ^{123}I radionuclides. Larger detectors (>40 cm in diameter) are used in large field of view (LFOV) cameras. Rectangular NaI(Tl) detectors (38.7×61 cm or 45×66 cm) are also available in gamma cameras.

Increasing the thickness of a detector increases the probability of complete absorption of γ -rays and hence the sensitivity (defined in Chapter 10) of the camera. However, the probability of multiple Compton scattering also increases in thicker detectors, and therefore the X , Y coordinates of the point of γ -ray interaction can be misplaced. This results in poor resolution of the image of the area of interest. For this reason, thin NaI(Tl) detectors are used in gamma cameras, but this decreases the sensitivity of the camera, because many γ -rays may escape from the detector without interaction.

Collimator

In gamma cameras, a collimator is attached to the face of the NaI(Tl) detector to limit the field of view so that γ -radiations from outside the field of view are prevented from reaching the detector. Collimators are normally made of material with high atomic number and stopping power, such as tungsten, lead, and platinum, among which lead is the material of economic choice in nuclear medicine. They are designed in different sizes and shapes and contain one or many holes to view the area of interest.

Collimators are primarily classified by the type of focusing, although other classifications are also made based on septal thickness and the number of holes. Depending on the type of focusing, collimators are classified as parallel-hole, pinhole, converging, and diverging types; these are illustrated in Figure 9.3. *Pinhole* collimators are made in conical shape with a single hole and are used in imaging small organs such as the thyroid glands to provide magnified images. *Converging* collimators are made with tapered holes converging to an outside point and are employed to provide magnified images when the organ of interest is smaller than the size of the detector. Images are magnified by converging collimators. *Diverging* collimators are constructed with tapered holes that are divergent from the detector face and are used in imaging organs such as lungs that are larger than the size of the detector. The images are minified with these collimators. *Parallel-hole* collimators are made with holes that are parallel to each other and perpendicular to the detector face and have between 4000 and 46,000 holes depending on the collimator design. These collimators are most commonly used in nuclear medicine procedures and furnish a one-to-one projected image. Because pinhole and converging collimators magnify and the diverging collimators minify the image of the object, some distortion occurs in images obtained with these collimators. Because LFOV cameras are readily available now, diverging collimators are not used in routine nuclear medicine studies.

COLLIMATOR DESIGNS




FIG. 9.3. Different designs of collimators.

Parallel-hole collimators are classified as high-resolution, all-purpose, and high-sensitivity type, or low-energy, medium-energy, or high-energy type, depending on the resolution and sensitivity they provide in imaging. High-sensitivity collimators are made with smaller thickness than all-purpose collimators, whereas high-resolution collimators are thickest of all. These characteristics are discussed in detail in Chapter 10.

Currently, ultra-high-energy collimators useful for 511-keV photons are commercially available and used for single photon emission computed tomography (SPECT) imaging with ^{18}F -fluorodeoxyglucose.

Several collimators are available that are designed for some specific purposes. *Fan-beam* collimators are designed with holes that converge in one dimension but are parallel to each other in the other dimension. These collimators are primarily used for imaging smaller objects and hence magnify the images. *Cone-beam* collimators are similar to fan-beam collimators and magnify the images except that the holes are designed such that they converge in two dimensions.

In earlier collimators, the holes were originally circular, but current designs have square, hexagonal, or even triangular holes with uniform thickness of lead around the opening. These collimators provide better spatial resolution than the circular-hole ones.

Photomultiplier Tube

As in scintillation counters, PM tubes are essential in gamma cameras for converting the light photons in the NaI(Tl) detector to a pulse. Instead of one PM tube, an array of PM tubes (19 to 107) are mounted in a hexagonal fashion to the back of the detector with optical grease, or in some instances, using lucite light pipes between the detector and the PM tubes. In modern gamma cameras, square or hexagonal PM tubes are used for better packing. The output of each PM tube is used to define the X , Y coordinates of the point of interaction of the γ -ray in the detector by the use of an X -, Y -positioning circuit (see later) and also is summed up by a summing circuit to form a pulse known as the Z pulse. The Z pulse is then subjected to pulse-height analysis and is accepted if it falls within the range of selected energies.

X-, Y-Positioning Circuit

Each pulse arising out of the γ -ray interaction in the NaI(Tl) detector is projected at an X , Y location on the image corresponding to the X , Y location of the point of interaction of the γ -ray. This is accomplished by an X -, Y -positioning circuit in conjunction with the PM tubes and a summing circuit. Figure 9.4 illustrates the principles of X , Y positioning of pulses arising from γ -ray interactions in the detector employing seven PM tubes. All PM tubes are connected through capacitors to four output leads rep-




FIG. 9.4. Arrangement of seven photomultiplier (PM) tubes to produce X and Y pulses for the X, Y location of the γ -ray interaction in the detector. X^+ , X^- , Y^+ , and Y^- pulses are obtained by summing the output of all PM tubes weighted by capacitors for the location of each PM tube in relation to the site of γ -ray interaction. (Adapted from Anger HO. Scintillation camera. Rev Sci Instr. 1958; 29:27.)

resenting four directional signals, X^+ , X^- , Y^+ , and Y^- . The capacitance values are assigned in direct proportion to the location of the PM tube relative to the four signals. Suppose a γ -ray interacts at a location (*) near tube 7. The largest amount of light is received by tube 7, and other tubes receive light in proportion to their distances from the point of interaction. The output signals of PM tubes are weighted by the appropriate capacitance values and then summed to form each of the X^+ , X^- , Y^+ , and Y^- signals individually. In this case, X^- will be greater than X^+ , and Y^+ will be greater than Y^- , because the interaction occurred in the upper left quadrant. The X^- , Y^- -designating pulses, X and Y , and the Z pulse are then obtained as follows:

$$Z = X^+ + X^- + Y^+ + Y^- \quad (9.1)$$

$$X = \frac{k}{Z}(X^+ - X^-) \quad (9.2)$$

$$Y = \frac{k}{Z}(Y^+ - Y^-) \quad (9.3)$$

where k is a constant and k/Z is the amplifier gain. The X and Y pulses are then projected on a CRT to depict the X , Y coordinates of the point of γ -ray interaction, which in turn corresponds to the coordinates of the location in the field of view from which the γ -ray originated. Similarly, these pulses can be stored in the computer in a square matrix so that the data can be processed later to reproduce an image. Details of data acquisition and storage in the computer are given in Chapter 11. Or, they can be projected on an x-ray film. Nowadays, resistors and microprocessors have been used in place of capacitors.

The larger the number of PM tubes, the better the accuracy of the X , Y locations of all pulses on the image; that is, the better the spatial resolution of the image (see Chapter 10).

Pulse-Height Analyzer

After the Z pulses are formed by the summing circuit, the PHA analyzes their amplitude and selects only those of desired energy by the use of appropriate peak and window settings. In many gamma cameras, the energy selection is made automatically by push-button-type isotope selectors designated for different radionuclides such as ^{99m}Tc , ^{131}I , and so on. In modern cameras, isotope peak and window settings are selected by the mouse-driven menu on a computer monitor interfaced with the camera. In some gamma cameras, two or three PHAs are used to select simultaneously two or three γ -rays of different energies. These types of cameras are useful in imaging with ^{111}In and ^{67}Ga that possess two or three predominant γ -rays. The window settings are provided in percentages of the peak energy by a control knob. For most studies, a 15% to 20% window centered symmetrically on the photopeak is employed.

It should be noted that X and Y pulses are accepted if the Z pulse is within the energy range selected by the PHA. If the Z pulse is outside this range, then X and Y pulses are discarded.

Display and Storage

In a typical nuclear medicine study, data are collected normally for preset counts (e.g., 500,000 counts) or a preset time (e.g., 10 min). Until the mid-90s, image data were captured on x-ray film or Polaroid film or stored on magnetic tapes, laser disks, and the like. Nowadays, all camera systems use computer memories for storage of image data. The details of storage in computers are given in Chapter 11. In older systems, images were mostly displayed on cathode ray tube (CRT) monitors and at present, all systems commonly use LCD (liquid crystal display) video monitors for better

display of images. The computer manipulation of image contrast on LCD monitors provides a better view of images leading to more accurate diagnosis of diseases. The details of display and storage are given in Chapter 11.

Digital Cameras

It is seen from the above description that the X - and Y -pulses are obtained in analog form and are projected on different display and recording systems. Such analog processing inherently includes instability in pulse formation and results in image nonlinearity and nonuniformity. These are caused by fluctuations in PM tube output due to high voltage (HV) variations, drift in preamplifier output, and variations in PH and X -, Y -positioning analyses. To correct for these effects and also for the manipulation of data at a later time, analog data are digitized to be stored in a matrix map in a computer. Digitization of the analog signal is performed by an electronic circuit, called the analog-to-digital converter (ADC). The digitized data are later retrieved for further processing to display on video monitors.

In modern cameras each PM tube output is digitized by the ADC before PH and X -, Y -positioning analyses. These cameras are called "all-digital" cameras. In these cameras, the gains of all PM tubes are initially optimized by placing a narrow beam of a radioactive source in front of each PM tube and determining the center of the photopeak by adjusting the high voltage of the PM tube with a digital computer. Next, the camera is calibrated, in which a source of interest is positioned in front of each PM tube, and output from each PM tube is sampled, integrated, and digitized by a high-speed ADC in the computer. Each signal is then normalized by dividing it with the sum of all digital signals arising from the same scintillation event. In a two-dimensional array of PM tubes, the normalized digital output $Z_i(X, Y)$ corresponds to the X, Y location of the PM tube i . To determine the location of each signal Z_i , a weighting factor is calculated from the inverse of the uncertainties of X and Y positions, i.e., $1/\Delta X$ and $1/\Delta Y$, that are related to the spatial distribution of Z_i values around the center of the PM tube. The X, Y locations and weighting factors are mapped and stored in reference tables as functions of Z_i values for all PM tubes for positional and Z -pulse analyses of a scintillation event in later imaging studies.

In subsequent patient imaging studies, the output signal of each PM tube from a scintillation event is sampled, integrated, digitized and finally normalized to give Z_i . The location (X, Y) of the scintillation event is then calculated by using the appropriate values of locations and weighting factors in the reference tables in the memory. The digitized $Z_i(X, Y)$ is stored in the X, Y location of the image matrix, if the pulse discrimination does not reject the signal. Since the location of each event is determined by digitizing and analyzing the individual signal from each PM tube, the accuracy of positioning of the signals is greatly improved. For these reasons, the digital

cameras provide excellent intrinsic linearity and hence superior spatial resolution in image formation.

Solid-State Digital Cameras

The Digirad Corporation has made commercially available several gamma cameras using solid-state detectors. Initially CZT detectors were used, but later the company replaced them with CsI(Tl) detectors. The detectors are fabricated in modules made of 128 3-mm \times 3-mm detector elements. The cameras are single-head, two-head, or three-head. Each head is comprised of 32 modules consisting of a total of 4096 elements, resulting in a detector area of 8 in. \times 8 in. in a 64 \times 64 crystal array. The uniqueness of these cameras is that they do not use PM tubes for pulse formation and use silicon diodes instead. No X -, Y -positioning circuit is used, because each CsI(Tl)/silicon diode element functions as an individual detection system, independent of other elements, and each event of photon interaction in the crystal is positioned on the image matrix corresponding to the location of the element (Early, 2005). This provides an excellent spatial resolution and quality of the images in the energy range of 60–300 keV.

Various Digirad camera models include Digirad 2020tc, SPECTpack, Cardius-1 (single-head), Cardius-2 (two-head), and Cardius-3 (three-head). Appropriate collimators are required for imaging different organs and for photons of different energies. Many units are small and portable. These cameras are capable of SPECT studies with the use of a rotating chair for the patient, which rotates with the patient on it at incremental angles with respect to the detector.

Questions

- Describe the operational principles of a gamma camera.
 - The main purpose of a collimator is to limit the field of view of an imaging device for imaging. True or false?
 - The purpose of a photomultiplier tube is to convert light photons to an electron pulse. True or false?
 - The scattered photons are excluded by the proper choice of a collimator. True or false?
 - Scattered photons are excluded by the proper choice of discriminator settings (windows). True or false?
- What are the different categories of collimators?
 - Which collimator is most used in nuclear medicine?
 - Which types of collimator give image distortion and why?
- Describe the function of the X , Y circuit in the gamma camera system.

4. A pulse-height analyzer:
 - (a) Reduces the background. True or false?
 - (b) Rejects γ -rays that undergo Compton scattering in the patient and the detector. True or false?
 - (c) Rejects γ -rays undergoing photoelectric effect in patients. True or false?
 - (d) Increases the signal-to-noise ratio. True or false?
5.
 - (a) The counting efficiency of a gamma camera increases with the thickness of the detector. True or false?
 - (b) What are the most common thicknesses of the NaI(Tl) detector used?
 - (c) A gamma camera detector with a 20-cm field of view is used to image the lungs, which fill 75% of the image. The camera is set to accumulate 450,000 counts. Calculate the information density.
6. In pulse-height analysis, a 20% window means 10% on either side of the photopeak. True or false?
7. Describe how the digital camera works.
8. What is the advantage of a digital camera over an analog camera (Anger type)?

Suggested Readings

- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia: W.B. Saunders; 2003.
- Early P. Private communication, 2005.
- Erickson J. Imaging systems. In: Harbert J, da Rocha AFG, eds. *Textbook of Nuclear Medicine, Volume I: Basic Science*. Philadelphia: Lea & Febiger; 1984.
- Rollo FD, ed. *Nuclear Physics, Instrumentation, and Agents*. St Louis: CV Mosby; 1977.

10

Performance Parameters of Gamma Cameras

The quality and detail of an image obtained by gamma cameras are affected by several parameters associated with these imaging systems. These parameters include spatial resolution, sensitivity, uniformity, and contrast, and they are described here in detail. A brief description of the quality control tests for gamma cameras is also included.

Spatial Resolution

The spatial resolution of a gamma camera is a measure of the ability of the device to faithfully reproduce the image of an object, thus clearly depicting the variations in the distribution of radioactivity in the object (Erickson, 1984). The *spatial resolution* of a gamma camera is empirically defined as the minimum distance between two points in an image that can be detected by the system. The overall spatial resolution (R_o) of a gamma camera comprises three components, namely, intrinsic resolution (R_i) of the detection system, collimator resolution (R_g), and scatter resolution (R_s), and is given by

$$R_o = \sqrt{R_i^2 + R_g^2 + R_s^2} \quad (10.1)$$

The smaller numerical values of R_o indicate better resolution and vice versa.

Intrinsic Resolution

Intrinsic resolution, R_i , is the component of spatial resolution contributed by the detector and associated electronics, and is a measure of how well an imaging device can localize an event on the image. Intrinsic resolution arises primarily from the statistical fluctuations in pulse formation that have been discussed in the section entitled Gamma Ray Spectrometry in Chapter 8. The statistical variations in the production of light photons after γ -ray interaction in the detector and variations in the number of electrons emitted

from the photocathode and dynodes in the photomultiplier (PM) tubes have significant effects on the intrinsic resolution. In gamma cameras, the X, Y positioning of the pulses is improved by increasing the number of PM tubes, thus improving the intrinsic resolution. Also, PM tubes with greater quantum efficiency and their improved optical coupling to the detector for greater light collection provide better intrinsic resolution.

Intrinsic resolution improves with higher γ -ray energy and deteriorates with lower energy because greater statistical fluctuations occur in the production of light photons by lower energy photons and vice versa. For example, the 140-keV photons of ^{99m}Tc produce almost twice as many light photons in the detector as the 69- to 80-keV photons of ^{201}Tl and thus result in better intrinsic resolution. However, there is little improvement in intrinsic resolution with photon energy above 250 keV because of multiple scattering of photons within the detector that can result in photoelectric absorption (see below). Intrinsic resolution improves with narrow PHA window settings, because scattered radiations are avoided.

Multiple Compton scattering of a γ -ray photon followed by absorption of all scattered photons in the detector causes uncertainty in the X, Y location of the original γ -ray interaction and makes the intrinsic resolution, and hence spatial resolution, worse. This effect is worse with thicker detectors and high-energy photons (>250 keV) because of the increased chances of multiple scattering. For this reason, only thinner detectors (0.63–1.84 cm) are used in gamma cameras.

Most modern cameras have intrinsic resolution of the order of 4-mm full width at half maximum (FWHM) for 140-keV photons of ^{99m}Tc .

Collimator Resolution

Collimator resolution, also termed the *geometric resolution* (R_g), constitutes the major part of the overall spatial resolution and primarily arises from the collimator design. In general, collimator resolution is worse than intrinsic resolution. As already mentioned in Chapter 9, there are four major collimators: parallel-hole, pinhole, converging, and diverging. Of these, parallelhole collimators are most commonly used in nuclear medicine.

The different parameters of a typical parallel-hole collimator are shown in Figure 10.1. The spatial resolution for this collimator is given by the geometric radius of acceptance, R_g :

$$R_g = \frac{d(t_e + b + c)}{t_e} \quad (10.2)$$

where d is the hole diameter of the collimator, b is the distance between the collimator face and the source of radiation, c is the distance between the back face of the collimator and the midplane of the detector, and t_e is the effective length of the collimator holes. The t_e is empirically given by $t_e = t - 2\mu^{-1}$, where μ is the linear attenuation coefficient of the photons in




FIG. 10.1. A parallel-hole collimator with thickness t , hole diameter d , septal thickness a , and source-to-collimator distance b . The collimator is attached to a detector whose midplane is at a distance c from the back surface of the collimator. R_g is the collimator resolution.

the collimator material (e.g., lead), and t is the length or thickness of the collimator hole. This corrects for the penetration of the two corners of the holes by the photons.

As seen from Eq. (10.2), the collimator resolution is improved by increasing the length, t , of the collimator holes or by decreasing the diameter, d , of the holes. Thus, long narrow holes provide best spatial resolution. Also, the collimator resolution deteriorates with increasing source-to-collimator distance, b , and is best at the collimator face. Therefore, in nuclear medicine studies, patients should be placed as close to the collimator as possible to provide the best resolution.

The thickness of lead, a , between holes is called the septum. *Septal penetration* of γ -rays plays an important role in the collimator resolution and depends on the γ -ray energy. High-energy photons from outside the field of view can cross the septum and yet interact in the detector, thus blurring the image. Because of this, γ -rays of only ~ 50 – 300 keV are suitable for com-

monly used collimators, the most preferable photon energy being 150 keV. At energies below ~ 50 keV, photons are absorbed in the body tissue, whereas at energies above ~ 300 keV septal penetration of the photons can occur. Current collimators are made with appropriate septal thickness for specific photon energies to limit septal penetration. Parallel-hole collimators are classified as low-energy collimators with a few tenths of a millimeter septal thickness (for up to 150-keV γ -rays) and medium-energy collimators with a few millimeter thickness (up to 400-keV photons) (Cherry et al., 2003). Currently very high energy collimators are available for counting 511-keV photons. It is understandable that for a given diameter collimator, the number of holes are greater in low-energy collimators than in high-energy collimators. Normally, high-energy collimators have poorer efficiency and resolution than low-energy collimators.

In another classification, collimators are termed high-sensitivity and high-resolution collimators. Often, these collimators are made with an identical number of holes with identical diameters but with different thicknesses. Thus, the collimator with longer holes is called the high-resolution collimator and that with shorter holes is called the high-sensitivity collimator. The spatial resolution for the high-sensitivity collimator deteriorates sharply with the source-to-collimator distance. All-purpose, or general-purpose, collimators are designed with intermediate values of resolution and sensitivity. Typical values of resolution (FWHM at 10 cm) are about 8 mm for low-energy high-resolution and all-purpose parallel-hole collimators and about 13 mm for high-sensitivity collimators.

The collimator resolution for pinhole, diverging, and converging collimators is expressed by similar but somewhat complex equations, and their details are available in reference books on nuclear physics and instrumentation. For pinhole and converging collimators, best resolution is obtained when the object is at the focal plane. The overall system resolutions of different collimators are illustrated in Figure 10.2. Fan-beam collimators give better spatial resolution but poorer sensitivity than parallel-hole collimators.

Scatter Resolution

Radiations are scattered by interaction with tissue in patients and with the detector. It is possible that some of these radiations are scattered without much loss of energy and fall within the field of view, resulting in pulses of acceptable amplitude set by pulse-height analyzer (PHA). This degrades the overall spatial resolution. This component is called the *scatter resolution* (R_s) and depends on the composition of the scattering medium, the source configuration, and PHA discriminator settings. Scatter increases in heavy patients, and decreasing the PHA window reduces the scatter contribution. The effect of scatter resolution is essentially the same for all collimators (Rollo and Harris, 1977).




FIG. 10.2. Effect of source-to-collimator distance on overall system resolution for various types of collimators. (A) High sensitivity parallel hole. (B) Diverging. (C) All purpose parallel hole. (D) Converging. (E) High resolution parallel hole. (F) Pinhole. (From Rollo FD, Harris CC. Factors affecting image formation. In: Rollo FD, ed. *Nuclear Medicine Physics. Instrumentation and Agents*. St. Louis: Mosby; 1977:407. Modified from Moyer RA. *J Nucl Med* 1974, 15:59.)

Evaluation of Spatial Resolution

Bar Phantom

Qualitative evaluation of the spatial resolution of an imaging device can be made by visual inspection of the images obtained using bar phantoms. Bar phantoms consist of four sets of parallel lead bar strips arranged perpendicular to each other in four quadrants in a lucite holder (Fig. 10.3A). The widths and spacings of the strips are the same within each quadrant but differ in different quadrants. The Hine-Duley phantom consists of five groups of lead strips of different thicknesses and spacings arranged in parallel fashion in a lucite holder (Fig. 10.3B). In all bar phantoms, the thickness of lead should be sufficient to stop photons of a given energy for which spatial resolution is being estimated.

The bar phantom is placed over the detector of a gamma camera. A ^{57}Co flood source (described later under Quality Control) is placed on the top of it and an image is taken. For evaluation of spatial resolution for differ-




FIG. 10.3. (A) Bar phantom. (B) Hine-Duley phantom. (Courtesy of Nuclear Associates, Carle Place, NY.)

ent photon energies, flood sources of radiations of respective energies should be used. The image of the bar phantom obtained is visually inspected, and spatial resolution is estimated from the smallest strips or spacings distinguishable on the image (Fig. 10.4). Obviously, this technique is qualitative and does not give a quantitative measure of the spatial resolution.




FIG. 10.4. Bar phantom images at different distances from a parallel-hole collimator.

Line-Spread Function

An improved method of estimating spatial resolution of a gamma camera is based on the use of a line-spread function (LSF). A long plastic tubing filled with a radioactive solution is placed in the field of view of the detector. The gamma camera interfaced with a computer collects and stores counts from the line source in a single view, and then the computer generates the LSF. The counts (response) obtained at incremental distances are plotted against the distance from the center axis of the collimator to give a bell-shaped LSF (Fig. 10.5). The FWHM of the LSF curve gives the spatial resolution of the imaging device.

Spatial resolution by the LSF method varies with the design of the collimator and is, therefore, different for parallel-hole, converging, and diverging collimators. Also, the FWHM values of the LSF may not represent the true spatial resolution, because the scatter and septal penetration components fall in the tail part of the LSF (i.e., below 50%) and therefore are not accounted for.




FIG. 10.5. Line spread function of a gamma camera equipped with a low-energy all-purpose parallel hole collimator obtained in (a) air and (b) water at different distances using a ^{99m}Tc -line source. Note that the resolution deteriorates in water due to attenuation.




FIG. 10.6. An illustration of the principles of modulation transfer function (MTF) of an imaging system. See text for details.

Modulation Transfer Function

A more complete and quantitative assessment of spatial resolution of imaging devices is made by using modulation transfer function (MTF) (Rollo, 1977). The concept of MTF is illustrated in Figure 10.6. Suppose a source of activity (e.g., a patient) has a sinusoidal distribution with peaks (maximum activity, A_{max}) and valleys (minimum activity, A_{min}), as illustrated in Figure 10.6. Such a distribution gives a spatial frequency (ν) in cycles per centimeter or cycles per millimeter. The contrast, or *modulation* (M_s), in the source activity is given by

$$M_s = \frac{A_{max} - A_{min}}{A_{max} + A_{min}} \quad (10.3)$$

If a perfect imaging device were to image the source faithfully, it would depict the same distribution of activity in the image with A_{max} and A_{min} as in the source activity. Because the imaging devices are not absolutely perfect, it will portray the distribution of activity in the image with C_{max} for the peak and C_{min} for the valley, which are smaller in magnitude than A_{max} and A_{min} . The modulation in the image (M_i) is then expressed by

$$M_i = \frac{C_{max} - C_{min}}{C_{max} + C_{min}} \quad (10.4)$$

The MTF at a spatial frequency ν is then calculated as the ratio of M_i to M_s :

$$\text{MTF}(\nu) = \frac{M_i}{M_s} \quad (10.5)$$

When $M_s = M_i$, $\text{MTF} = 1$. This is true if the sinusoidal cycles are well separated and if the imaging device reproduces the image of each cycle faithfully. Thus, the system with $\text{MTF} = 1$ gives the best overall spatial resolution.

When the distribution of activity is such that spatial frequency increases, the peaks and valleys come closer. When the peaks and valleys are too close, the imaging device cannot delineate them, and the MTF tends to 0, yielding the poorest spatial resolution of the system. The values of the MTF between 0 and 1 give intermediate spatial resolutions. It is important to note that small objects are better imaged at higher frequencies and large objects at lower frequencies.

It has been demonstrated that the MTF is a normalized Fourier transform of the LSF discussed previously. In practice, the source activity distribution is assumed to be composed of line sources separated by infinitesimal distances, and the MTFs are then calculated from the LSFs of all line sources. The mathematical expression of these functions is quite complex and can be found in reference physics books.

Plots of the MTFs against spatial frequencies are useful in determining the overall spatial resolution of imaging devices and are presented in Figure 10.7 for three imaging systems. It is seen that, at very low frequencies (i.e., larger separation of sinusoidal cycles), the MTFs are almost unity for all three systems; that is, they reproduce equally good images of the radiation source. At higher frequencies (i.e., sinusoidal cycles are too close), system A in Figure 10.7 gives the best resolution, followed in order by system B and system C.




FIG. 10.7. Plot of modulation transfer function (MTF) against spatial frequency. System A gives better spatial resolution than systems B and C, and system B provides better resolution than system C.

It is appropriate to mention that whereas the FWHM of the LSF does not account for the scatter and septal penetration of γ -radiations, the MTF takes these two factors into consideration and provides a complete description of the spatial resolution of a system. Furthermore, individual components of an imaging system may have separate MTFs, and they are combined to give the overall MTF, as follows:

$$\text{MTF} = \text{MTF}_1 \times \text{MTF}_2 \times \text{MTF}_3 \dots \quad (10.6)$$

Problem 10.1

The MTFs at a certain spatial frequency of the detector, PM tubes, and PHA of a gamma camera are 0.8, 0.6, and 0.7, respectively. What is the overall MTF of the camera?

Answer

$$\begin{aligned} \text{MTF} &= 0.8 \times 0.6 \times 0.7 \\ &= 0.34 \end{aligned}$$

Sensitivity

Sensitivity of a gamma camera is defined as the number of counts per unit time detected by the device for each unit of activity present in a source. It is normally expressed in counts per second per microcurie (cps/ μ Ci). Sensitivity depends on the geometric efficiency of the collimator, the intrinsic photopeak efficiency of the detector, PHA discriminator settings, and the dead time of the system. Intrinsic photopeak efficiency of a detector (Chapter 8), PHA discriminator settings (Chapters 8 and 9), and the dead time (Chapter 8) have been discussed previously. Briefly,

- (a) the intrinsic photopeak efficiency of a detector decreases with increasing photon energy and with increasing source-to-detector distance (see Fig. 8.9A) but increases with the thickness of the detector (see Fig. 8.8). Most modern cameras have 0.95 cm thick NaI(Tl) crystals. The photopeak efficiency of these crystals is about 90% for 140-keV photons of ^{99m}Tc and about 30% for 364-keV photons of ^{131}I .
- (b) A narrow window setting on the PHA reduces the counts measured and therefore compromises the counting efficiency.
- (c) Counts are lost when counting a high-activity sample using a device with a long dead time, and hence the counting efficiency is reduced.

Sensitivity of a gamma camera is most affected by the collimator efficiency, which is described next.




FIG. 10.8. Variation of geometric efficiency with source-to-collimator distance for various collimators. (A) High sensitivity parallel hole. (B) Converging. (C) All purpose parallel hole. (D) Diverging. (E) High resolution parallel hole. (F) Pinhole. (From Rollo FD, Harris CC. Factors affecting image formation. In: Rollo FD, ed. *Nuclear Medicine Physics. Instrumentation and Agents*. St. Louis: Mosby; 1977:407. Modified from Moyer RA. *J Nucl Med* 1974, 15:59.)

Collimator Efficiency

Collimator efficiency, or geometric efficiency (E_g), is defined as the number of γ -ray photons passing through the collimator holes per unit activity present in a source. For parallel-hole collimators (see Fig. 10.1), it is given by

$$E_g = K \cdot \frac{d^4}{t_c^2(d+a)^2} \quad (10.7)$$

where d is the hole diameter, t_c is the effective length of the collimator hole defined before, and a is the septal thickness. The constant K is a function of the shape and arrangement of holes in the collimator and varies between 0.24 for round holes in a hexagonal array to 0.28 for square holes in a square array.

The collimator efficiency for parallel-hole collimators increases with increasing diameter of the collimator holes and decreases with increasing collimator thickness (t) and septal thickness (a), which is quite opposite to

spatial resolution [see Eq. (10.2)]. Thus, *for a given collimator, as the spatial resolution of a system increases, its sensitivity decreases, and vice versa*. Note that collimator efficiency, E_g , for parallel-hole collimators is not affected by the source-to-detector distance for an extended planar source; that is, it essentially remains the same at different distances from the detector. The low-energy all-purpose (LEAP) parallel-hole collimators have efficiencies of about 2×10^{-4} . Collimator efficiency varies with different types of collimators, and the values are shown as a function of source-to-collimator distance in Figure 10.8.

Uniformity

It is always expected that a gamma camera should yield a uniform response throughout the field of view. That is, a point source counted at different locations in the field of view should give the same count rate by the detector at all locations. However, even properly tuned and adjusted gamma cameras produce nonuniform images with count density variations of as much as 10%. Such nonuniformity adds to the degradation of the spatial resolution of the system.

The nonuniformity in detector response arises from several factors: (a) variations in PM tube response, (b) nonlinearity in X , Y -positioning of pulses across the field of view, and (c) edge packing. Although factor (c) is preventable as discussed below, factors (a) and (b) are the leading causes of deterioration in uniformity and special attention is needed to remedy them.

Pulse-Height Variation

As already mentioned in Chapter 8, there are variations in the light production from γ -ray interaction in the detector, light transmission to PM tubes, and in the pulse formation in PM tubes. These variations result in photopeaks of different amplitude. Because there are many PM tubes across the detector, the amplitude of the photopeaks will vary from location to location (i.e., from PM tube to PM tube) across the detector even with a fixed energy window. These variations in PM tube response contribute significantly to the nonuniformity of the detector response.

This part of nonuniformity is corrected by acquiring an image using a well-tuned gamma camera and a ^{99m}Tc point source that is placed at a distance of five times the detector's FOV. An intrinsic image (without the collimator) is acquired using the appropriate window and stored in a 128×128 matrix. The pulse height in each (X , Y) pixel is determined and stored in a 128×128 look-up table. In subsequent patient studies, a microprocessor compares the pulse height in each pixel of the patient image with the corresponding value in the look-up table, and then either moves the energy

window or adjusts the pulse height to compensate for the variations. This is performed in real-time during the data collection in patient studies and should be carried out for each radionuclide.

Nonlinearity

The spatial nonlinearities are systematic errors in the positioning of X -, Y -coordinates of pulses in the image and result from local count compression or expansion. For example, when a radioactive source is moved across from the edge to the center of a PM tube, more counts are found at the center (pincushion distortion) giving a hot spot. Counts are less toward the edges of the PM tubes, causing a cold spot (barrel distortion) in the area.

The spatial distortions due to nonlinearity are corrected by microprocessors built into modern cameras that use correction matrices. Nonlinearity correction factors are generated by calculating the spatial shift of the observed position of an event from its actual position. An intrinsic image (without collimator) is taken using a line or orthogonal hole test pattern and a source of radioactivity of interest placed at a distance five times the detector's FOV. The test pattern is placed directly on the detector, and an image is stored in a 128×128 matrix. The actual location (X , Y) of each pixel is known and the corresponding location on the image is measured. These variations in X , Y for all pixels are calculated as correction factors and are normally supplied by the manufacturer in the form of a look-up table. These correction factors are subsequently applied in real-time to each detected event to move it to the actual position during patient imaging. This is performed for all events throughout the field.

As stated above, modern cameras include two look-up matrices, the pulse-height correction and the linearity correction, to compensate for variations in the overall uniformity of images. As the camera slowly drifts over time, the correction tables have to be updated for proper correction of the patient scan. The exact frequency of reacquiring the correction tables depends on stability of the camera and varies with the manufacturer. The pulse-height correction tables require more frequent acquisitions, whereas the linearity correction tables are typically performed by a service engineer. Different manufacturers recommend monthly to quarterly acquisition of these correction factor maps.

It should be pointed out that uniformity corrections are mainly for minor variations in PM tube response across the field of view and are not a replacement for proper tuning of the gamma camera. Even though the uniformity corrections at times can correct for large nonuniformities, frequent retuning of the gamma camera is essential as these corrections affect linearity, resolution, and overall sensitivity of the camera. However, in modern cameras, uniformity has improved considerably with better attachment of the PM tubes to the detector and the use of PM tubes with higher quantum efficiency.

Edge Packing

Edge packing is seen around the edge of an image as a bright ring and results in nonuniformity of the image. This results from the fact that more light photons are reflected near the edge of the detector to the PM tube. Normally a 5-cm wide lead ring is attached around the edge of the collimator to mask this effect. In modern cameras, electronic masking is employed.

Gamma Camera Tuning

Nonlinearity and pulse-height correction factors are generated with properly balanced PM tubes, and for them to work accurately, the PM tubes must be maintained properly balanced or “tuned.” Changes can occur over time in the gain of the PM tubes and other electronic circuits that can invalidate these factors. Most modern cameras have circuits to monitor the PM tube gains and automatically can make minor adjustments using uniformity corrections. Still they need routine maintenance of tuning monthly or quarterly.

The methods of tuning of the gamma camera vary from manufacturer to manufacturer. One method of tuning involves the use of a ^{99m}Tc source in front of the camera and then obtaining the pulse-height spectrum for each PM tube. If a photopeak shift in a PM tube is seen from the original value in a PM tube, the preamplifier gain is adjusted. This technique was employed in earlier models of the gamma camera.

Some manufacturers use a split-window technique for each PM tube with two narrow tuning windows set on the high side of the photopeak to minimize the scattered radiation. The source of radiation can be either the radiation from the patient or an external radioactive source, for example, ^{99m}Tc . Counts are collected in the two windows for each PM tube and the ratios are calculated, which are stored in look-up tables and should be constant over time for each well-tuned PM tube. Deviation of the ratio from the constant value in subsequent patient studies indicates a drift in the gain of the PM tube, which is then adjusted electronically.

In another method, once a week each PM tube of the camera is tuned up with a radioactive source followed by illumination of the crystal with a light-emitting diode (LED) attached to each PM tube. The LED is pulsed at a high frequency and the response of each PM tube is recorded as a reference value. Subsequently, the illumination of the detector is carried out daily by the LED optical system, and the PM tube response is measured and compared with the reference value. If there is any drift, the gain is adjusted. Tuning should be performed separately for each radionuclide.

Effects of High Counting Rates

As discussed in Chapter 8, the scintillation cameras suffer count losses at high counting rates due to pulse pileup. Pulse pileup results from the detection by the camera of two events simultaneously as one event with amplitude different from that of either original event. If one or both of the events are photopeaks, then the amplitude of the new event will be outside the pulse-height window setting and so the event will be rejected resulting in a loss of counts. If, however, two Compton scattered photons are processed together to produce an event equivalent to the photopeak in amplitude, then the event will be counted within the window setting. But the X, Y position of the event will be misplaced on the image somewhere between the locations of the two events. This causes image distortion. Both count rate loss and image distortion at high count rates must be taken into consideration in evaluating the performance of different cameras.

Several techniques are employed to improve the high count rate performance of a gamma camera. In modern cameras, buffers (or derandomizers) are used in which pulses are processed one at a time, and overlapping events are kept on “hold” until the processing of the preceding event is completed. In some cameras, the dead time is shortened by shortening the time to integrate the signals from the PM tubes. Other cameras use pulse pileup rejection circuits to minimize the count loss and image distortion and thus to improve images, although they tend to increase the dead time of the camera. Recent developments include high-speed electronics that reduce the number of misplaced events and improve the image quality significantly.

Contrast

Contrast of an image is the relative variations in count densities between adjacent areas in the image of an object. Contrast (C) gives a measure of detectability of an abnormality relative to normal tissue and is expressed as

$$C = \frac{A - B}{A} \quad (10.8)$$

where A and B are the count densities recorded over the normal and abnormal tissues, respectively.

Lesions on the image are seen as either “hot” or “cold” spots indicating increased or decreased uptakes of radioactivity in the corresponding areas in the object. Several factors affect the contrast of the image, namely, count density, scattered radiation, pulse pileup, size of the lesion, and patient motion, and each contributes to the contrast to a varying degree.

Statistical variations of the count rates give rise to noise that increases with decreasing information density or count density (counts/cm²) and is given by $(1/\sqrt{N}) \times 100$, where N is the count density. For a given imaging setting, a minimum number of counts need to be collected for reasonable image contrast. Even with adequate spatial resolution from the imaging device, lack of sufficient counts may give rise to poor contrast due to increased noise, so much so that lesions may be missed. This count density depends on the amount of activity administered and the uptake in the organ of interest. Contrast is improved with increasing administered activity and also with the differential uptake between the normal and abnormal tissues. However, due consideration should be given to the radiation dose to the patient from a large amount of administered activity. Sometimes, high count density is achieved by counting for a longer period of time in the case of low administered activity. It should be emphasized that spatial resolution is not affected by the increased count density from increased administered activity or longer counting.

Background in the image increases with scattered radiations and thus degrades the image contrast. Maximum scatter radiations originate from the patient. Narrow PHA window settings can reduce the scatter radiations, but sensitivity, that is, counting efficiency, is reduced by narrow window settings. A 15–20% PHA window centered on the photopeak of interest is most commonly used in routine imaging.

As discussed above, at high count rates, pulse pileup can degrade the image contrast.

Image contrast to distinguish a lesion depends on its size relative to system resolution and its surrounding background. Unless a minimum size of a lesion larger than system resolution develops, contrast may not be sufficient to appreciate the lesion, even at higher count density. The lesion size factor depends on the background activity surrounding it and on whether it is a “cold” or “hot” lesion. A relatively small-size “hot” lesion can be well contrasted against a lower background, whereas a small “cold” lesion may be missed against surrounding tissues of increased activities.

Patient motion during imaging reduces the image contrast. This primarily results from the overlapping of normal and abnormal areas by the movement of the organ. It is somewhat alleviated by restraining the patients or by having them in a comfortable position.

Quality Control Tests for Gamma Cameras

To ensure high quality of images produced by imaging devices, several quality control tests must be performed routinely on gamma cameras. The frequency of tests is daily, weekly, and, for some tests, monthly or even quarterly. The most common tests are the positioning of the photopeak

(peaking), uniformity, and spatial resolution of the camera. These tests can be carried out with the collimator attached to the camera (*extrinsic*) or without the collimator (*intrinsic*), and should be performed for each radionuclide used in a specific clinical study.

In the intrinsic method, the source of a particular radionuclide containing approximately 100 to 200 μCi (3.7 to 7.4 MBq) in a syringe is normally placed at a distance of four to five times the detector field of view to ensure uniform irradiation of the detector. Because the collimator is removed, the integrity of the collimator cannot be assessed by this method.

In the extrinsic method, a sheet source is used made of plastic containing the radionuclide of interest. Because $^{99\text{m}}\text{Tc}$ is most commonly used in nuclear medicine studies, a $^{99\text{m}}\text{Tc}$ sheet source is prepared by adding several millicuries of $^{99\text{m}}\text{Tc}$ activity to a water-filled plastic sheet container. The source should be thoroughly mixed and free of air bubbles. Due to the inconvenience of daily preparation of the $^{99\text{m}}\text{Tc}$ sheet source and radiation exposure to the technologist during preparation, an alternative solid ^{57}Co sheet source is used, which is commercially available in rectangular or circular forms. ^{57}Co has a longer half-life (~ 270 d) and emits photons of 122 keV and 136 keV, which are equivalent to the 140 keV photons of $^{99\text{m}}\text{Tc}$. These sources, typically made with 10 mCi (370 MBq) of ^{57}Co , are also called *flood* sources and most commonly used for over a period of one to two years. The source is placed on the collimator and an image is taken. The use of these sources in the extrinsic method provides information on PM tubes as well as any structural imperfections in the collimator. Because ^{57}Co activity decays over time, counting time increases with time to accumulate sufficient counts for the image.

Daily Checks

Positioning of Photopeak

Positioning of the photopeak must be done daily or as needed for each radionuclide used in the clinical study to center the PHA window on the center of the photopeak. In older analog cameras, a source of radioactivity of interest is placed on the collimator attached to the detector (extrinsic) and the high voltage on the PM tube is adjusted to center the energy window on the photopeak. For $^{99\text{m}}\text{Tc}$, typically 1 mCi (37 MBq) of the activity in a syringe is used as a source for peaking and a 20% window is set around 140 keV. Peaking for ^{111}In , ^{67}Ga , ^{123}I , ^{201}Tl , and so on must be done separately, as needed.

In modern cameras, peaking is performed automatically by menu-driven protocol-based software provided by the manufacturer. Initially at the time of the camera set-up, the photopeak window is set with a $^{99\text{m}}\text{Tc}$ source using the intrinsic method. Subsequently the daily check of the position of the photopeak is performed with a ^{57}Co flood source by the extrinsic method

using a low-energy high-resolution collimator. If the peak shift is more than 10%, the camera must be tuned. Tuning is performed by the computer program by repeaking of the camera with a ^{99m}Tc source placed at least 30 cm away from the detector and without a collimator (intrinsic method). The same method is applied for other radionuclides.

Uniformity

The uniformity of the detector response is checked daily by using a ^{57}Co flood source. The flood source is placed on the detector with a low-energy high-resolution collimator attached (extrinsic) and an energy window of 20% is used. An image is acquired containing about 1 million counts (Fig. 10.9) and then assessed for uniformity, linearity (visibility of tube pattern), integrity of PM tubes (energy calibration), artifacts, and so on by visual inspection. For SPECT camera heads, 3–5 million counts are acquired. Nonuniformity exceeding 5 to 10% is detectable by human eyes. Visual inspection of the image should reveal any nonuniformity that warrants the tuning of the camera, particularly the PM tubes.

According to the NEMA protocol, the intrinsic method is used with a 100 to 200 μCi (3.7 to 7.4 MBq) ^{99m}Tc source, and two parameters are evaluated for uniformity, which are defined as follows.

$$\text{integral uniformity} = \frac{C_{\max} - C_{\min}}{C_{\max} + C_{\min}} \times 100$$

where C_{\max} and C_{\min} are the maximum and minimum count rates across the field of view in a nine-point smoothed image.




FIG. 10.9. Images of a ^{57}Co flood source showing the uniformity (a) and nonuniformity (b) of the response of a gamma camera.

$$\text{differential uniformity} = \frac{\text{high} - \text{low}}{\text{high} + \text{low}} \times 100$$

where “high” and “low” are the maximum and minimum differences in counts over five contiguous pixels in all rows and columns of the matrix. The computer program provides these values for both central field of view (CFOV) and the useful field of view (UFOV). These values should be in the range of 1 to 2%, otherwise, the camera needs to be tuned. The UFOV is a circular area with the largest diameter inscribed within the collimated field of view. The CFOV is the circular area with the diameter equal to 75% of the diameter of the UFOV. These definitions may not apply to rectangular FOV scanners.

Weekly Checks

Spatial Resolution and Linearity

The spatial resolution and linearity of the gamma camera is checked weekly by using a bar phantom (Fig. 10.3). The bar phantom is placed on the detector head with a low-energy high-resolution collimator attached (extrinsic), and a flood source of ~10 mCi (370 MBq) of ^{57}Co is placed on the top of the bar phantom. An image is taken with approximately 10 million counts and visually inspected to check the linearity and separation of the smallest bars. This is a qualitative method. For a quantitative method, a line spread function must be determined and then FWHM measured as discussed earlier. Although extrinsic tests are done for convenience, intrinsic tests are preferable for better accuracy.

Annual or As-Needed Checks

Tuning of the camera is performed monthly or quarterly by the protocols described earlier. Other essential parameters such as energy resolution, high count rate response, multiwindow registration (e.g., ^{111}In and ^{67}Ga with multiple photons) and sensitivity should be evaluated at least annually or as needed after extensive modification or repair of the camera.

In addition, tests on accessories such as computers, multifunction cameras, scanning tables, rotation of gantry, and so on should be performed periodically. Furthermore, all tests must be documented in a record book with pertinent information, such as the date, time, total counts, window settings, the type of radioactive source, the type of camera, and initials of the technologist performing the tests.

Questions

- (a) Define the spatial resolution of a gamma camera.
(b) What are the different components of the spatial resolution?

- (c) A system with a spatial resolution of 5 mm is better than a system with a spatial resolution of 8 mm. True or false?
2. The intrinsic resolution of a gamma camera depends on:
- (a) The thickness of the NaI(Tl) detector. True or false?
 - (b) The energy of the γ -ray. True or false?
 - (c) The width of the pulse-height window. True or false?
 - (d) The number of counts collected. True or false?
3. (a) What is the best photon energy for imaging with a gamma camera?
- (b) Why is a thinner NaI(Tl) detector used in a gamma camera?
 - (c) Intrinsic resolution improves with higher γ -ray energy. True or false?
 - (d) Spatial resolution of a gamma camera improves as the number of photomultiplier tubes is increased. True or false?
4. For a gamma camera with a parallel-hole collimator,
- (a) The spatial resolution increases with decreasing detector thickness. True or false?
 - (b) The collimator efficiency decreases with increasing collimator length. True or false?
 - (c) The spatial resolution increases with decreasing collimator length. True or false?
 - (d) High-energy collimators have higher efficiency and resolution than low-energy collimators. True or false?
 - (e) The best resolution is obtained at the face of the parallel-hole collimator. True or false?
5. What are the effects of the following factors on the spatial resolution and sensitivity of a gamma camera?
- (a) Photomultiplier (PM) tubes with higher quantum efficiency
 - (b) A wider "window" on the pulse-height analyzer (PHA)
 - (c) Increasing the activity of ^{99m}Tc from 5 mCi (185 MBq) to 15 mCi (555 MBq)
 - (d) Increasing the diameter of the collimator hole
 - (e) Adding more tissue between the collimator face and the patient's organ
 - (f) Using a diverging collimator
 - (g) Increasing the source-to-collimator distance for a parallel-hole collimator
 - (h) Using a γ -ray of higher energy, which penetrates the septum of the collimator
6. (a) In routine practice, how is the spatial resolution of a gamma camera checked?
- (b) The full width at half maximum of the line spread function of a gamma camera does not give a true picture of spatial resolution. Why?
 - (c) What is the modulation transfer function (MTF) of a system?
 - (d) A system gives the best spatial resolution when its MTF is equal to 1. True or false?

- (e) If PM tubes and the PHA of a gamma camera have MTFs of 0.5 and 0.7 at a certain spatial frequency, what is the overall MTF of the camera?
 - (f) As the sensitivity of a gamma camera increases, its spatial resolution decreases. True or false?
 - (g) The collimator efficiency of a parallel-hole collimator is not affected by the source-to-detector distance. True or false?
7. (a) What is the primary cause of nonuniformity in an image?
(b) What is edge packing?
(c) How is the nonuniformity in an image corrected?
8. (a) What is the contrast of an image?
(b) What are the different factors that affect the contrast of an image?
(c) How does pulse pileup affect the contrast?
(d) Is contrast or spatial resolution affected by increasing the administered activity?
9. (a) What are the daily and weekly tests performed for gamma cameras?
(b) What is meant by extrinsic and intrinsic tests?

References and Suggested Readings

- Bushberg JT, Seibert JA, Leidholdt EM Jr, Boone JM. *The Essential Physics of Medical Imaging*. 2nd ed. Philadelphia: Lippincott Williams & Wilkins; 2002.
- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia: W.B. Saunders; 2003
- Erickson J. Imaging systems. In: Harbert J, da Rocha AFG, ed. *Textbook of Nuclear Medicine. Volume I. Basic Science*. Philadelphia: Lea & Febiger; 1984:105.
- Groch MW, Erwin WD. Single-photon emission computed tomography in the year 2001: Instrumentation and quality control. *J Nucl Med Technol*. 2001;29:12.
- Murphy PH. Acceptance testing and quality control of gamma cameras, including SPECT. *J Nucl Med* 1987;28:1221.
- Rollo FD. Evaluating imaging devices. In: Rollo FD, ed. *Nuclear Physics, Instrumentation and Agents*. St. Louis: Mosby; 1977:436.
- Rollo FD, Harris CC. Factors affecting image formation. In: Rollo FD, ed. *Nuclear Physics, Instrumentation and Agents*. St. Louis: Mosby; 1977:387.

11

Digital Computers in Nuclear Medicine

Digital computers were introduced in nuclear medicine practice in the mid-1960s, but did not become an integral part in both imaging and nonimaging applications until the mid-1970s. In imaging modalities, the computers are used to quantitate the distribution of radiopharmaceuticals in an object both spatially and temporally. Both data acquisition and image processing in scintigraphy are accomplished by digital computers. In nonimaging applications, patient scheduling, archiving, inventory of supplies, management of budget, record keeping, and health physics are just a few examples of what is accomplished with the help of digital computers. Computational capabilities have advanced tremendously over the years and are still evolving, and the utility of a computer is limited only by the limitations of hardware and software.

Basics of a Computer

The basic elements of a computer are a *central processing unit (CPU)*, *main memory*, *external storage* and *input/output (I/O)* devices, which are connected to one another by pathways called *buses*. The main memory stores all program instructions and acquired data, while the CPU executes all instructions given in a program. External storage includes floppy disks, CD-ROMs, DVD-ROMs, and hard drives. I/O devices include peripherals such as keyboards, mouse, video monitors, and printers, whose functions are to communicate with the computer for input of the acquired data and output of the processed data. A typical setup of a computer is illustrated in Figure 11.1.

The voltage signals, i.e., electrical pulses from a scintillation camera, are obtained in analog form and are digitized by the digital computers for further processing and storage. Digital computers operate with *binary* numbers using only two digits, 0 and 1, as opposed to 10 digits 0 to 9 in the decimal system. The basic unit of the binary system is a *bit* (binary digit) that is either 0 or 1. The binary numbers are expressed by placing 0's and




FIG. 11.1. Basic components of a computer.

1's in a row, e.g., 10101, which are equal to a sum of a series of powers of two, as opposed to decimal numbers that are expressed in powers of 10. Thus, the binary number 10101 is equal to the decimal number $(1 \times 2^4 + 0 \times 2^3 + 1 \times 2^2 + 0 \times 2^1 + 1 \times 2^0) = 21$, which is given as $2 \times 10^1 + 1 \times 10^0$ in the decimal system.

The bits, 1 and 0, are represented by the "on" or "off" states of many transistor components present in the computer memory. A two-bit number can be expressed in 2^2 , or four, ways (00, 01, 10, 11) corresponding to decimal numbers, 0, 1, 2, 3; a three-bit number can be expressed in 2^3 , or eight, ways (000, 001, 010, 011, 100, 101, 110, 111) corresponding to decimal numbers 0, 1, 2, . . . 7, and an n -bit number can be expressed in 2^n ways corresponding to decimal numbers from 0 to $2^n - 1$. In computer nomenclature, a *byte* of memory is equal to eight bits that can store up to 2^8 , i.e., 0 to 255, units of information. Similarly, a *word* of memory consists of 16 bits or two bytes and can store up to 2^{16} , i.e., 0 to 65,535, units of information. In newer computers, a word can consist of 32 or 64 bits, allowing more counts to be stored in memory.

Central Processing Unit

The CPU, also called the microprocessor, performs all control, logic, and arithmetic operations in a computer. A computer program is a set of sequential instructions for the computer to perform with essential data inserted whenever appropriate. The CPU retrieves the instructions and data

from memory storage, executes the instructions sequentially, and displays or stores the results in appropriate locations. Transfer of data from one location to another is performed by the CPU using buses, which are essentially a set of electrical connections.

How fast a program is executed depends on the speed of the computer operation, which increases with the faster electrical components of the CPU. The efficiency of the computer operation is further increased by using parallel transfer of data (where many transfers are performed simultaneously) rather than serial transfer (where only one transfer is carried out at a time).

Computer Memory

The memory of a computer is a section assigned for temporary storage of data during the operation of a program. The program instructions and processed data are all stored in the computer memory with the help of the CPU. When the CPU sends data to the memory, it writes the data into the memory. If the CPU retrieves data from the memory, it is then said to read the data from the memory. The memory can be of two types: random access memory (RAM) and read-only memory (ROM). RAM has the advantage of both write and read capacity; however, the data stored in it is lost when the computer is shutdown or the electrical power is lost. With larger RAMs, computation time becomes shorter. On the other hand, data stored in ROMs such as CD-ROMs, DVD-ROMs, etc. cannot be erased by electrical failure or computer shutdown.

External Storage Devices

Floppy disks, hard disk drives, CD-ROMs, DVD-ROMs, magnetic tapes, and optical (laser) disks are varieties of external storage devices that are commonly used for storage of programs and data. Each of them has variable storage space. Hard drives are installed virtually in all computers for internal storage of the programs and data. Floppy disks are commonly used for storing data externally as backup copies, although in some applications programs and data can be stored for input into the computer for execution. While hard drives have the storage capacity of hundreds of gigabytes, floppy disks can only store up to a few megabytes and are getting out of use. Currently, CD-ROMs are available with capacities of 650–700 megabytes and DVD-ROMs with 4.7–9 gigabytes. Magnetic tapes and laser optical disks have large storage space in compact form and can be utilized primarily for archiving of patient data that can be retrieved for future reference.

Input/Output Devices

Input/output (I/O) devices are essential for input of the initial data and for output of the processed data. Input and output of data are carried out by

the use of acquisition and video interfaces via serial or parallel buses. The keyboard and the mouse are the most common input devices used in computers, although joysticks, light pens, and trackballs are occasionally used as input devices. While the keyboard is essential for the input of alphanumeric data such as patient identification, date, time, and operator's name, the mouse and trackballs are used to select items from the menu. Light pens, a mouse, and touch screens are often used for the selection of regions of interest (ROI) in an image.

Common output devices include display screens (video monitor) for texts, images or graphics, and printers for printing. Display screens normally have a capability of a gray scale or a color scale for comparison between the intensities or amplitudes of different regions of the image.

Operation of a Computer

A computer operates according to instructions provided by an operator. These instructions are given in the form of one or more programs. A collection of programs is called the software, which is developed by specialists according to the specific need for a project. The most essential program for the operation of a computer is the so-called operating system such as DOS (disk operating system), Windows 98, Unix, and Linux. The utility of this program is to facilitate communication between the computer and operator's instructions. The operating system transfers the program instructions from the input device to the memory, commands the CPU to carry out the specific instruction and returns the data to the output devices. Other utilities of this system include file transfer from one location to another, storing data in the external storage device, and display of the data.

Data must be provided as input to the computer for processing, and in nuclear medicine they are available in the form of counts or voltage pulses obtained from scintigraphic studies. Data are processed according to instructions in the software program, and the processed data are then stored in computer memory or external storage spaces or displayed on video monitors. The time to complete a task by the computer depends on a number of factors such as the speed of the CPU, the size of the main memory, serial or parallel processing of data, and the data transfer rates of the I/O devices. The faster CPU, the larger size of memory and the parallel buses provide speedy computation.

Digitization of Analog Data

In nuclear medicine, signals from a gamma camera are acquired in analog form, which are digitized before storing and further processing by the computer. Conversion of analog signals to discrete digital values is performed by the so-called *analog-to-digital converters* (ADCs) and the process is called digitization.

While analog signals are continuous in time, digital signals consist of a fixed number of bits produced by the ADC by sampling a selected number of time points in the analog signal. ADCs are available as 8-, 10-, 12-, or 16-bit depending on the number of bits they produce in the digital signal from the analog signal. While the analog signals can be distorted by the electronic noise, there is some inherent loss of signal information as a result of digitization, i.e., due to different time-point selections during the analog-to-digital conversion. This arises from the fact that there is a likelihood of a small fraction of the signal being lost during the conversion of a continuous analog signal to discrete digital values. However, higher bit ADCs minimize this loss by producing a large number of bits from each analog signal. The faster ADCs can handle higher count rates. The slower ADCs increase the dead time of the system and hence are good for low count rates only.

Digital-to-Analog Conversion

For video displays, data must be in the analog form, and therefore digitized data must be converted back to analog data. This is performed by units, called the digital-to-analog converters (DACs), similar to the ADCs. DACs are connected to the computer via video interface cards, and the speed of digital-to-analog conversion depends on the speed of various electrical components included in its operation.

Digital Images

Digital images are characterized by two quantities: matrix size and pixel depth. The computer memory approximates the area of the detector in a gamma camera as a square matrix of a definite size that can range from 32×32 to 1024×1024 with 1024 (1 K) to 1,048,576 (1 M) picture elements, called *pixels*, respectively. The size of a matrix is selected by the operator, depending on the type of task to be performed and is approximated to the field of view (FOV). Each pixel corresponds to a specific location in the detector. As discussed in Chapter 9, the *X*- and *Y*-pulses are obtained in the analog form from the photomultiplier (PM) tube, which originate from the interaction of γ -rays in the detector. The *X*- and *Y*-analog pulses are digitized by the ADC and stored in the appropriate pixel of the matrix. How many counts can be stored in a pixel depends on the depth of the pixel, which is represented by a byte or a word. Thus, a 1-byte pixel could record up to 2^8 , or 256, events, whereas a 1-word pixel could store up to 2^{16} , or 65,536, events.

The pixel size, which depends on the choice of the matrix size for a study, is an important factor that affects the spatial resolution of a digital image. The field of view is approximated to the matrix size; therefore the pixel size is calculated by dividing the FOV by the number of pixels across the matrix.

Thus if an image of 250×250 mm FOV is obtained in a matrix of 128×128 mm, the pixel size would be $250/128 \approx 2$ mm. If the matrix size is changed to 64×64 , then the pixel size would be ~ 4 mm. Often, a *zoom factor* is applied during data acquisition to improve spatial resolution because it reduces the pixel size. Overall, the pixel size d can be calculated as

$$d = FOV/(z \times N) \quad (11.1)$$

where z is the zoom factor (1.2, 2.0, etc.,) and N is the number of pixels across the matrix. The use of a zoom factor of, say, 2, reduces the pixel size by half, improving the spatial resolution, but counts per pixel are reduced thus increasing the noise on the image (see later).

The choice of pixel size and zoom factor is limited by the spatial resolution of the imaging device, particularly in tomographic systems. Ideally, the pixel size should be less than $1/3$ of the expected spatial resolution of the SPECT system, measured at the center of rotation. That is,

$$d \leq FWHM/3 \quad (11.2)$$

where $FWHM$ is the full width at half maximum of the line spread function of the imaging system. If the expected system resolution is 18 mm, then the pixel size in the matrix should be less than 6 mm. Pixel size larger than this limit would degrade the image.

For a typical SPECT gamma camera, the FOV size is 400 mm across and the spatial resolution is of the order of 18 to 25 mm. Thus, the pixel size in a 64×64 matrix is $400/64 = 6.25$ mm, which is nearly equal to or less than the $1/3$ of the spatial resolution of the SPECT system. Thus, a 64×64 matrix should be good enough in most SPECT imaging. Using a 128×128 matrix (pixel size is 3.13 mm, which is much less than $1/3$ of the system resolution), would improve the spatial resolution significantly. However, as mentioned before, the counts in each pixel would be reduced by $1/4$, as the total counts are distributed over four times the pixels, compared to a 64×64 matrix. Thus the noise increases in the image and so the signal-to-noise ratio decreases causing degradation in image contrast.

Application of Computers in Nuclear Medicine

Digital Data Acquisition

The X - and Y -signals obtained in scintigraphic studies in nuclear medicine are digitized by ADCs in the computer and stored in one of two ways: (a) *frame mode* and (b) *list mode*. In both modes, a technique of magnification or zooming can be applied, whereby the pixel size is decreased by a zoom factor. Zoom factors typically vary from 1 to 4 in increments of 0.25.

Data acquisition in the frame mode is the most common practice in nuclear medicine and widely used in static, gated, dynamic, and single




FIG. 11.2. Data acquisition in the frame mode and the list mode.

photon emission computed tomography (SPECT) studies. In this mode, a matrix is chosen that approximates the entire area of the detector so that a position (X, Y) in the detector corresponds to a pixel position in the matrix. Digitized signals (X, Y) are stored in the corresponding (X, Y) positions (pixel) of the matrix of choice in the computer. Every time a new X, Y signal arrives, it is added to the (X, Y) pixel (Fig. 11.2A). In this mode, one must specify the size and depth of the matrix, the number of frames per study, and the time of collection of data per frame or total counts to be collected. Data acquisition continues until a preselected time or total count is reached. This mode provides instant images for storage and display.

In the list mode, digitized X - and Y -signals are coded with "time marks" as they are received in sequence in time, and are stored as individual events in the order they occur (Fig. 11.2B). After the data acquisition is completed, the data can be sorted to form images in a variety of ways to suit a specific need. Data can be manipulated by changing the matrix size and the time of acquisition per frame. Also, physiologic markers, such as the start of a cardiac cycle (e.g., the start of the R-wave) in the gated cardiac studies, can be incorporated in the list mode acquisition. Since the data are listed sequentially without overlapping each other, the bad signals from an arrhythmic cardiac cycle can be discarded, as found appropriate, in the

postacquisition reformatting. Although the list mode acquisition provides wide flexibility, its major disadvantages are larger memory space and longer processing time required and unavailability of images during or immediately after the completion of the study.

Static Study

A static study is the collection of data in one view of a region of interest in an object for a preset time or preset total counts. Data are acquired in the frame mode, and normally the matrix size is specified prior to starting the study. The choice of a matrix size depends on the field of view of the imaging system and the pixel size to give desired image resolution. For all practical purposes, a pixel size of 2 to 3 mm is considered appropriate for good image resolution. Thus, for large FOV scintillation cameras (>400 mm), one would need a 256×256 matrix to obtain the above pixel size. Because of the high count densities in static views, data acquisition in byte mode may overflow in individual pixels and, therefore, the word mode is usually employed.

Digital images essentially represent the count density in regions of interest in an object. How many counts should be acquired in an image? It depends on how small a region in an image is to be identified and its apparent contrast with the surrounding background. Large and high-contrast objects are easily detectable at low count densities, whereas small and low-contrast objects are difficult to delineate from the statistical noise. Count density should be optimum for desirable contrast.

Dynamic Study

In dynamic studies, a series of images are collected and each image (frame) is acquired over a certain period of time selected by the operator. While the patient's position cannot be changed during the image acquisition, the matrix size and the frame rate (time of acquisition) can be changed. The frame rate can vary from many frames per second to a single frame per hour. The acquisition of image data is buffered such that while one frame is being collected, the previous frame is stored in the external storage device (e.g., disk). Data in dynamic studies can be collected in a sequence of several phases, e.g., 1 frame per second for 1 minute, then 1 frame/minute for 5 minutes, followed by 1 frame/10 minutes for 2 hours. The choice of frame rate for a given study depends on the kinetics of the radiotracer through the organ of interest.

The common matrix size used in dynamic studies is 64×64 or 128×128 , although some loss of spatial resolution is expected with these matrices. However, 256×256 or larger matrices require larger memory size. Since counts collected per frame are low in number, the data are collected in byte mode, which obviates the need for a large memory space, and normally does not allow pixel counts to exceed 255 with little chance of counts overflow.

Gated Study

The gated study was introduced in the mid-1970s to determine the ejection fraction of the heart by acquiring two images, one at end diastole and the other at end systole. It was later substituted by continuous acquisition of data in multiple sequential images (multiple gated acquisition, MUGA) in each cardiac cycle by gating between successive cycles.

In the MUGA study, the data are acquired in synchronization with the R-wave of the cardiac cycle. The normal heart beat is about 1 beat/second, and the R-R interval is therefore about 1 second, i.e., 1000 msec. First, the R-R interval is divided into several segments or frames (16–32 segments) depending on the number of frames one chooses to obtain. For example, with a choice of 20 frames in the R-R interval, each frame will be 50 msec long. In actual data collection, first the counts are acquired in frame 1 for 50 msec, followed by the collection of counts in frame 2 for another 50 msec, and so on. After completion of counting in all 20 frames, a new R-wave is detected, and the above sequence of counting continues until sufficient counts have been accumulated in each frame. Assuming a count rate of 10,000 to 20,000 counts/s in a typical cardiac study, each 50 msec frame would accumulate counts of the order of 500 to 1000. Normally, 64×64 or larger matrices are used for the gated study.

The heart beat must be regular for the above method to work well. If the heart beat is irregular such as in cardiac arrhythmia, the R-R interval is sufficiently altered and the data become corrupted from R-wave to R-wave. Modern acquisition programs have been devised to reject the bad heart beat cycle. Using the list mode acquisition, bad heart beat data can be sorted out and rejected in postacquisition reformatting.

SPECT is routinely used in nuclear medicine for various organ imaging, particularly cardiac imaging. The gated SPECT study is also employed for the cardiac studies using the typical 20 frames in each R-R interval.

Reconstruction of Images

In planar imaging, the acquired data are displayed in a two-dimensional images without further processing. In tomographic imaging, data are acquired in different angular projections around the patient. The data of each projection are processed further using the methods described in Chapter 12 to reconstruct the images at different depths of the patient's organ in 3-D directions. All reconstruction methods are accomplished by the use of modern computers.

Superimposition and Subtraction of Images

It has been a common practice to superimpose image data from one modality onto another for better interpretation of the images. For example, com-

puted tomography (CT) or magnetic resonance imaging (MRI) anatomical images of an organ are superimposed onto the corresponding functional images obtained by positron emission tomography or other nuclear medicine studies to match the functional abnormalities with the anatomical defects. Computers are well utilized to perform these superimpositions of images.

Another important utility of the computer is the subtraction of background activity from an image or one set of images from another set. An example of the latter is to subtract the interictal images obtained in epilepsy patients using ^{99m}Tc -ethyl cysteinyl dimer (ECD) from those obtained in ictal period using the same radiotracer. Resultant difference images provide better delineation of epileptogenic foci in these patients.

Display

Digital images are displayed on video monitors which are either cathode ray tubes (CRT) or flat-panel type liquid crystal display (LCD) monitors. These monitors are characterized by parameters such as spatial resolution, contrast, aspect ratio, luminance, persistence, refresh rate, and dynamic range. The spatial resolution and luminance of LCD monitors are far superior to those of CRTs. These monitors are placed in what is called the workstation where nuclear physicians view, manipulate, and interpret the images using the computer.

Display can be in either grayscale (black and white) or colorscale. In either case, grading of scale is achieved by variations in counts in the pixels in the digital image. In grayscale, the number of counts in the pixel defines the brightness level of a pixel. Thus, the black and white contrast in a digital image is obtained by applying the grayscale.

Color hues are assigned to different pixels corresponding to counts stored in the individual pixels in order to provide contrast between areas on the image. In a gradient colorscale, blue, green, yellow, and red are assigned in order to pixels with increasing counts: blue to the lowest count and red to the highest count. Edges of color bands are blended to produce a gradual change over the full range of the color scale.

Often a grayscale or colorscale bar is shown on the side of the image in order to help the interpreter differentiate the image contrast. Images can be displayed in transaxial (transverse), coronal (horizontal long axis), or sagittal (vertical long axis) views individually or simultaneously on the video monitor. On the simultaneous display of SPECT images, a point on the image is chosen using the cursor and three images that pass through the point are displayed. New sets of images are obtained by choosing a different point on the image. Such sequential screening of images is helpful in delineating the abnormal areas on images of the patient.

Angular projections around an object computed from the 3-D tomographic data can be displayed in continuous rotation. This presents the

image data in a movie or cinematographic (cine) mode, whereby a rotating 3-D image is seen on the monitor screen. This type of presentation identifies the location of a lesion in an organ in relation to other organs in the body.

In cardiac, brain, and respiratory studies, a popular technique called the *bull's eye*, or *polar map*, method is employed in which the activities in each transverse slice are displayed on a circumferential profile. The circumferential profile of each slice is projected on a bull's-eye format where the intensity of a point in the slice represents the magnitude of the activity, and the location of the point represents the radial location of the slice (Fig. 11.3). In polar images, the activity distribution in an object is essentially unfolded from inside out, and three-dimensional data are presented in a two-dimensional format. The major advantage of this technique is that one can identify the location of the defect in relation to adjacent areas on a single image.

Software and DICOM

As already mentioned, software is a collection of instructions for the computer to perform in carrying out a particular imaging study. Different vendors develop software programs, which are proprietary to them to operate their own equipment, and it is difficult to use one vendor's software for another's equipment. Also, there are third-party companies who develop software specific for equipment of a particular vendor. To partially circumvent such situations, one may stick to one vendor all the time using the same software. However, the American College of Radiology and the National Engineering Manufacturing Association (NEMA) jointly sponsored a standard format for the software, called Digital Imaging and Com-




FIG. 11.3. A typical illustration of a bull's eye or polar map image.

munications in Medicine (DICOM), which all vendors are recommended to adopt for compatibility among different software. Some of the standards of DICOM formats include image storage, protocols for intertransfer of data between the workstation and PACS (see later), query and retrieval of image data, print, and scheduling of data acquisition. DICOM formats are encoded in binary form. NEMA upgrades DICOM formats from time to time to meet the requirements of advancing technology and the medical community.

Essentially, vendors conform to the DICOM standard in developing their software, although compliance is voluntary. It provides a common format for imaging systems recognized by the hardware and software components of various manufacturers. This allows interoperability in the transfer of images and associated information among multiple vendors' devices. DICOM is very useful in the implementation of PACS (see below).

PACS

The modern networking of computers has offered a great advantage for exchange of information among individuals and organizations. It has been particularly useful for healthcare facilities in exchanging patient information among the physicians and hospitals. One type of network systems implemented in healthcare facilities is called the Picture Archiving and Communication System (PACS) and is solely used for the archiving and exchanging of patient information among health professionals. A PACS consists of devices to produce and store digital images electronically, workstations to view and interpret images, and a networking of these devices at different sites. Appropriate PACS software allows the interpreter to retrieve images from other locations and manipulate and interpret them as needed at his own location, and then return them with a report back to the original locations. In the absence of PACS, one can read the images only at the local facility and cannot transport them electronically to and from other facilities, if needed. PACS has improved the workflow profoundly by facilitating and expediting the transfer of information through network connections among various facilities.

In a radiology department, a small network system called the Radiology Information System (RIS) is normally implemented to maintain all types of workflow, such as image storage, patient scheduling, study type and its time of completion, image reporting, all the billing codes, and so on, within the department. Similarly, hospitals also have the Hospital Information System (HIS) that maintains similar information on patients including their demographic data, laboratory data, clinical history, and medication, and again, scheduling, tracking, reporting, and billing. A PACS can integrate both RIS and HIS for a broader exchange of information among healthcare personnel that will save time and money in healthcare operations. In such an integrated system, a referring physician can retrieve an image of a

patient on his/her computer from other locations within the PACS, rather than waiting for the hardcopy from the imaging department. He/She can then correlate the images with the clinical findings with a considerable saving of time. A typical integrated PACS is shown in Figure 11.4.

A PACS can be run by software on different operating systems such as Windows, MAC OS, Linux, or UNIX, although most PACS workstations are PC-based. PACS software must preserve confidentiality of patient information as mandated by the U.S. Hospital Insurance Portability and Accountability Act (HIPAA). The system must be reliable so that its downtime is nil. Also, the integrity of the system should be intact to avoid any medical errors in the patients' information. It should be always and easily accessible to all concerned to avoid delay in patient care. PACS software is constantly evolving to meet the new demands of healthcare professionals, and it is usually upgraded every 6 to 12 months. Numerous vendors (CTI/Siemens Medical, GE Medical, Phillips Medical, Kodak, Stentor, Spectra, etc.) have made commercially available their copyrighted PACS software, which are claimed to be robust, reliable, and user-friendly, with an uptime of more than 99.9%. A drawback is that there is a lack of a uniform standard among PACS software. It is desirable that the medical community, and perhaps the federal government, come up with a consensus policy similar to DICOM to make PACS uniform among different vendors.

An important application of PACS is in teleradiology which is being implemented throughout the country, and even worldwide between coun-




FIG. 11.4. A PACS integrating the networks of HIS and RIS.

tries connecting through PACS' different healthcare institutions for exchange of patient care information. By virtue of teleradiology, a radiologist or a nuclear physician can retrieve and interpret diagnostic images from a distant hospital and send back the report to the original hospital. This type of practice has resulted in outsourcing practitioners at a lower cost from one country to interpret imaging scans performed in another country, where the practitioner's pay is high.

Questions

1. What is a binary number? Express the decimal number 23 in binary form.
2. What is the difference between RAM and ROM memory?
3. The speed of a computer depends on the size of RAM memory and faster electrical components in the computer. True or false?
4. Why are parallel buses more efficient than serial buses in the computer?
5. Resolution of digital images are poorer than analog images. True or false? Explain why.
6. Which of the two matrices gives better resolution— 64×64 or 128×128 ?
7. Describe the method and advantages and disadvantages of the list mode acquisition and the frame mode acquisition.
8. Which mode would you use—byte mode or word mode—in static studies versus dynamic studies? Explain.
9. What is the essential difference between the Anger type analog camera and the “all-digital” camera?

Suggested Readings

- Bushberg JT, Seibert JA, Leidholdt, Jr EM, Boone JM. *The Essential Physics of Medical Imaging*. 2nd ed. Philadelphia: Lippincott, Williams & Wilkins; 2002.
- Lee K. *Computers in Nuclear Medicine: A Practical Approach*. New York: Society of Nuclear Medicine; 1992.
- Royal HD, Parker JA, Holman BL. Basic principles of computers. In: Sandler MP, Coleman RE, Wackers FJT, et al., eds. *Diagnostic Nuclear Medicine*. 3rd ed. Baltimore: Williams and Wilkins; 1995:93.

12

Single Photon Emission Computed Tomography

Tomographic Imaging

Conventional gamma cameras provide two-dimensional planar images of three-dimensional objects. Structural information in the third dimension, depth, is obscured by superimposition of all data along this direction. Although imaging of the object in different projections (posterior, anterior, lateral, and oblique) gives some information about the depth of a structure, precise assessment of the depth of a structure in an object is made by tomographic scanners. The prime objective of these scanners is to display the images of the activity distribution in different sections of the object at different depths.

The principle of tomographic imaging in nuclear medicine is based on the detection of radiations from the patient at different angles around the patient. It is called *emission computed tomography* (ECT), which is based on mathematical algorithms, and provides images at distinct depths (slices) of the object (Fig. 12.1). In contrast, in transmission tomography, a radiation source (x-rays or a radioactive source) projects an intense beam of radiation photons through the patient's body, and the transmitted beam is detected by the detector and further processed for image formation.

In nuclear medicine, two types of ECT have been in practice based on the type of radionuclides used: *single photon emission computed tomography* (SPECT), which uses γ -emitting radionuclides such as ^{99m}Tc , ^{123}I , ^{67}Ga , and ^{111}In , and *positron emission tomography* (PET), which uses β^+ -emitting radionuclides such as ^{11}C , ^{13}N , ^{15}O , ^{18}F , ^{68}Ga , and ^{82}Rb . SPECT is described in detail in this chapter and PET in Chapter 13.

Single Photon Emission Computed Tomography

The most common SPECT system consists of a typical gamma camera with one to three NaI(Tl) detector heads mounted on a gantry, an online computer for acquisition and processing of data, and a display system (Fig. 12.2).




Fig. 12.1. Four slices of the heart in the short axis.




Fig. 12.2. A dual-head SPECT camera, e-CAM model (Courtesy of Siemens Medical Solutions, Hoffman Estates, IL).

The detector head rotates around the long axis of the patient at small angle increments (3° to 10°) for collection of data over 180° or 360° . The data are collected in the form of pulses at each angular position and normally stored in a 64×64 or 128×128 matrix in the computer for later reconstruction of the images of the planes of interest. Note that the pulses are formed by the PM tubes from the light photons produced by the interaction of γ -ray photons from the object, which are then amplified, verified by X , Y position and PHA analyses, and finally stored. Transverse (short axis), sagittal (vertical long axis), and coronal (horizontal long axis) images can be generated from the collected data. Multihead gamma cameras collect data in several projections simultaneously and thus reduce the time of imaging. For example, a three-head camera collects a set of data in about one third of the time required by a single-head camera for 360° data acquisition.

Data Acquisition

The details of data collection and storage such as digitization of pulses, use of frame mode or list mode, choice of matrix size, etc., have been given in Chapter 11.

Data are acquired by rotating the detector head around the long axis of the patient over 180° or 360° . Although 180° data collection is commonly used (particularly in cardiac studies), 360° data acquisition is preferred by some investigators, because it minimizes the effects of attenuation and variation of resolution with depth. In 180° acquisition using a dual-head camera with heads mounted in opposition (i.e., 180°), only one detector head is needed for data collection and the data acquired by the other head essentially can be discarded. In some situations, the arithmetic mean $(A_1 + A_2)/2$ or the geometric mean $(A_1 \times A_2)^{1/2}$ of the counts, A_1 and A_2 , of the two heads are calculated to correct for attenuation of photons in tissue. However, in 180° collection, a dual-head camera with heads mounted at 90° angles to each other has the advantage of shortening the imaging time required to sample 180° by half (Table 12.1). Dual-head cameras with heads mounted at 90° or 180° angles to each other and triple-head cameras with heads oriented at 120° to each other are commonly used for 360° data acquisition and offer shorter imaging time than a one-head camera for this type of angular sampling.

The sensitivity of a multihead system increases with the number of heads depending on the orientation of the heads and whether 180° or 360° acquisition is made. Table 12.1 illustrates the relationship among sensitivity, time of imaging, and acquisition arc (180° or 360°) for different camera head configurations.

Older cameras were initially designed to rotate in circular orbits around the body. Such cameras are satisfactory for SPECT imaging of symmetric organs such as the brain, but because the body contour is not uniform, such a circular orbit places the camera heads far from the other parts of the body

TABLE 12.1. Relationship of sensitivity and time of imaging for 180° and 360° acquisitions for different camera head configurations.

Camera Type	180° Acquisition		360° Acquisition	
	Acquisition time (min)	Relative sensitivity	Acquisition time (min)	Relative sensitivity
Single-head	15	1	15	1
Dual-head (heads at 180°)	15	1	7.5	2
Dual-head (heads at 90°)	7.5	2	7.5	2
Triple-head (heads at 120°)	10	1.5	5	3

in the anterior and posterior positions. This causes loss of data and hence loss of spatial resolution in these projections. To circumvent this problem, many modern cameras are designed to include a feature called *noncircular orbit* (NCO) (i.e., to follow the body contour) that moves the camera heads such that the detector remains at the same distance close to the body contour at all angles.

Data collection can be made in either continuous motion or “step-and-shoot” mode. In continuous acquisition, the detector rotates continuously at a constant speed around the patient, and the acquired data are later binned into the number of segments equal to the number of projections desired. In the step-and-shoot mode, the detector moves around the patient at selected incremental angles (e.g., 3°) and collects the data for the projection at each angle.

Image Reconstruction

Data collected in two-dimensional projections give planar images of the object at each projection angle. To obtain information along the depth of the object, tomographic images are reconstructed using these projections. Two common methods of image reconstruction using the acquired data are the backprojection method and the iterative method, of which the former is the more popular, although lately the latter is gaining more attention. Both methods are described below.

Simple Backprojection

The principle of simple backprojection in image reconstruction is illustrated in Figure 12.3, in which three projection views are obtained by a gamma camera at three equidistant angles around an object containing two sources of activity designated by black dots. In the two-dimensional data acquisition, each pixel count in a projection represents the sum of all counts along the straight-line path through the depth of the object (Fig. 12.3A). Reconstruction is performed by assigning each pixel count of a given projection in the acquisition matrix to all pixels along the line of collection (perpendicular to the detector face) in the reconstruction matrix (Fig. 12.3B). This

is called *simple backprojection*. When many projections are backprojected, a final image is produced as shown in Figure 12.3C.

Backprojection can be better explained in terms of data acquisition in the computer matrix. Suppose the data are collected in a 4×4 acquisition matrix, as shown in Figure 12.4A. In this matrix, each row represents a slice, projection, or profile of a certain thickness and is backprojected individually. Each row consists of four pixels. For example, the first row has pixels A_1 , B_1 , C_1 , and D_1 . Counts in each pixel are considered to be the sum of all counts along the depth of the view. In the backprojection technique, a new reconstruction matrix of the same size (i.e., 4×4) is designed so that counts in pixel A_1 of the acquisition matrix are added to each pixel of the first column of the reconstruction matrix (Fig. 12.4B). Similarly, counts from pixels B_1 , C_1 , and D_1 are added to each pixel of the second, third, and fourth columns of the reconstruction matrix, respectively.

Next, suppose a lateral view (90°) of the same object is taken, and the data are again stored in a 4×4 acquisition matrix. The first row of pixels (A_2 , B_2 , C_2 , and D_2) in the 90° acquisition matrix is shown in Figure 12.4C. Counts from pixel A_2 are added to each pixel of the first row of the same reconstruction matrix, counts from pixel B_2 to the second row, counts from pixel C_2 to the third row, and so on. If more views are taken at angles between 0° and 90° , or any other angle greater than 90° and stored in 4×4 acquisition matrices, then the first row data of all these views can be




Fig. 12.3. Basic principle of reconstruction of an image by the backprojection technique. (A) An object with two “hot” spots (solid spheres) is viewed at three projections (at 120° angles). Each pixel count in a projection represents the sum of all counts along the straight-line path through the depth of the object. (B) Collected data are used to reconstruct the image by backprojection. (C) When many views are obtained, the reconstructed image represents the activity distribution with “hot” spots. (D) Blurring effect described by $1/r$ function where r is the distance away from the central point.




Fig. 12.4. An illustration of the backprojection technique using the data from an acquisition matrix into a reconstruction matrix.

similarly backprojected into the reconstruction matrix. This type of backprojection results in superimposition of data in each projection, thereby forming the final transverse image with areas of increased or decreased activity (Fig. 12.3C).

Similarly, backprojecting data from the other three rows of the 4×4 matrix of all views, four transverse cross-sectional images (slices) can be produced. Therefore, using 64×64 matrices for both acquisition and reconstruction, 64 transverse slices can be generated. From transverse slices, appropriate pixels are sorted out along the horizontal and vertical long axes, and used to form sagittal and coronal images, respectively. It is a common practice to lump several slices together to increase the count density in the individual slices to reduce statistical fluctuations.

Filtered Backprojection

The simple backprojection has the problem of “star pattern” artifacts (Fig. 12.3C) caused by “shining through” radiations from adjacent areas of increased radioactivity resulting in the blurring of the object. Because the blurring effect decreases with distance (r) from the object of interest, it can be described by a $1/r$ function (Fig. 12.3D). It can be considered as a spillover of certain counts from a pixel of interest into neighboring pixels, and the spillover decreases from the nearest pixels to the farthest pixels. This blurring effect is minimized by applying a “filter” to the acquisition data, and the filtered projections are then backprojected to produce an image that is more representative of the original object. Such methods are called the filtered back projection. There are in general two methods of filtered backprojection: the convolution method in the spatial domain and the Fourier method in the frequency domain, both of which are described below.

The Convolution Method

The blurring of reconstructed images caused by simple backprojection is eliminated by the convolution method in which a function, termed “kernel,” is convolved with the projection data, and the resultant data are then backprojected. The application of a kernel is a mathematical operation that essentially removes the $1/r$ function by taking some counts from the neighboring pixels and putting them back into the central pixel of interest. Mathematically, a convolved image $f'(x, y)$ can be expressed as

$$f'(x, y) = \sum_{i=-N}^N \sum_{j=-N}^N h_{ij} \odot f_{ij}(x-i, y-j) \quad (12.1)$$

where $f_{ij}(x-i, y-j)$ is the pixel count density at the $x-i, y-j$ location in the acquired projection, the h_{ij} values are the weighting factors of the convolution kernel, and \odot indicates the convolution operation. The arrangement of h_{ij} is available in many forms.

A familiar “nine-point smoothing” kernel (i.e., 3×3 size), also called smoothing filter, has been widely used in nuclear medicine to decrease statistical variation. The essence of this technique is primarily to average the counts in each pixel with those of the neighboring pixels in the acquisition matrix. An example of the application of nine-point smoothing to a section of an image is given in Figure 12.5. Let us assume that the thick-lined pixel with value 5 in the acquisition matrix is to be smoothed. First, we assume a 3×3 acquisition matrix (same as 3×3 kernel matrix) centered at the pixel to be convolved. Each pixel datum of this matrix is multiplied by the corresponding weighting factor, followed by the summation of the products. The weighting factors are calculated by dividing the individual pixel values of the kernel matrix by the sum of all pixel values of the matrix. The result of this operation is that the value of the pixel has changed from 5 to 3. Sim-




Fig. 12.5. The smoothing technique in the spatial domain using a 9-point smoothing kernel. The thick-lined pixel with value 5 is smoothed by first assuming a 3×3 acquisition matrix (same size as the smoothing matrix) centered at this pixel and multiplying each pixel value of the matrix by the corresponding weighting factor, followed by summing the products. The weighting factor is calculated by dividing the individual pixel value by the sum of all pixel values of the smoothing matrix. After smoothing the value of the pixel is changed from 5 to 3. Similarly all pixel values of the acquisition matrix are smoothed by the 9-point smoothing kernel, to give a smoothed image.

ilarly, all pixels in the acquisition matrix are smoothed, and the profiles are then backprojected.

The spatial kernel described above with all positive weighting factors reduces noise but degrades spatial resolution of the image. Sharp edges in the original image become blurred in the smoothed image as a result of averaging the counts of the edge pixels with those of the neighboring pixels.

Another filter kernel commonly used in the spatial domain consists of a narrow central peak with both positive and negative values on both sides of the peak, as shown in Figure 12.6. When this so-called edge-sharpening filter is applied centrally to a pixel for correction, the negative values in effect cancel or erase all neighboring pixel count densities, thus creating a corrected central pixel value. This is repeated for all pixels in each projection and the corrected projections are then backprojected. This technique reproduces the original image with better spatial resolution but with increasing noise. Note that blurring due to simple backprojection is removed by this technique but the noise inherent in the data acquisition due to the limitations of the spatial resolution of the imaging device is not removed but rather enhanced.

The Fourier Method

Nuclear medicine data obtained in the spatial domain (Fig. 12.7A) can be expressed in terms of a Fourier series in the frequency domain as the sum




Fig. 12.6. A filter in the spatial domain. The negative side-lobes in the spatial domain cancel out the unwanted contributions that lead to blurring in the reconstructed image.

of a series of sinusoidal waves of different amplitudes, spatial frequencies, and phase shifts running across the image (Fig. 12.7B). This is equivalent to sound waves that are composed of many sound frequencies. Thus, the data for each row and column of the acquisition matrix can be considered as composed of sinusoidal waves of varying amplitudes and frequencies in the frequency domain. The process of determining the amplitudes of sinusoidal waves is called the Fourier transformation (Fig. 12.7C) and the method of changing from the frequency domain to the spatial domain is called the inverse Fourier transformation.




Fig. 12.7. Representation of an object in the spatial and frequency domains. A profile in the spatial domain can be expressed as an infinite sum of sinusoidal functions (the Fourier series). For example, the activity distribution as a function of distance in an organ (A) can be composed by the sum of the four sine functions (B). The Fourier transform of this activity distribution is represented in (C), in which the amplitude of each sine wave is plotted at the corresponding frequency of the sine wave.

The Fourier method of reconstruction can be applied in two ways: either directly or by using filters. In the direct Fourier method, the Fourier transforms of individual acquisition projections are taken in polar coordinates in the frequency domain, which are then used to calculate the values in rectangular coordinates. Inverse Fourier transforms of these profiles are taken to compute the image. The method is not a true backprojection and is rarely used in reconstruction of images in nuclear medicine because of the time-consuming computation.

A more convenient method of reconstruction is the filtered backprojection (FBP) using the Fourier technique. In this method, filters are used to eliminate the blurring function $1/r$ that arises from simple backprojection of the projection data. These filters are analogous to tone controls or equalizers in radios or CD players that act as filters to vary the amplitudes of different frequencies, bass for low-frequency amplitudes and treble for high-frequency amplitudes. In image reconstruction, filters do the same thing, modulating the amplitudes of different frequencies, preserving the broad structures (the image) represented by low frequencies and removing the fine structures (noise) represented by high frequencies.

The Fourier method of filtering the projection data is based on the initial transformation of these data from the spatial domain to the frequency domain, which is symbolically expressed as

$$F'(v_x, v_y) = \mathcal{F}f(x, y) \quad (12.2)$$

where $F'(v_x, v_y)$ is the Fourier transform of $f(x, y)$ and \mathcal{F} denotes the Fourier transformation. Next a Fourier filter, $H(v)$ is applied in the frequency domain; that is,

$$F'(v) = H(v) \cdot F(v) \quad (12.3)$$

where $F'(v)$ is the filtered projection in the frequency domain, which is obtained as the multiplication product of $H(v)$ and $F(v)$. Finally, the inverse Fourier transformation is performed to obtain the filtered projections, which are then backprojected. The results obtained by the Fourier method are identical to those obtained by the convolution method. Although the Fourier method appears to be somewhat cumbersome and difficult to understand, the use of modern computers has made it much easier and faster to compute the reconstruction of images than the convolution method.

Types of Filters

A number of Fourier filters have been designed and used in the reconstruction of images in nuclear medicine. All of them are characterized by a maximum frequency, called the Nyquist frequency, that gives an upper limit to the number of frequencies necessary to describe the sine or cosine curves

representing an image projection. Because the acquisition data are discrete, the maximum number of peaks possible in a projection would be in a situation in which peaks and valleys occur in every alternate pixel (i.e., one cycle per two pixels or 0.5 cycle/pixel), which is the Nyquist frequency. If the pixel size is known for a given matrix, then the Nyquist frequency can be determined. For example, if the pixel size in a 64×64 matrix is 4.5 mm for a given detector, then the Nyquist frequency will be

$$\begin{aligned} \text{Nyquist frequency} &= 0.5 \text{ cycle/pixel} \\ &= 0.5 \text{ cycle}/0.45 \text{ cm} \\ &= 1.11 \text{ cycles/cm} \end{aligned}$$

A common and well-known filter is the ramp filter (name derived from its shape in the frequency domain) shown in Figure 12.8 in the frequency domain. An undesirable characteristic of the ramp filter is that it amplifies the noise associated with high frequencies in the image even though it removes the blurring effect of simple backprojection. To eliminate the high-frequency noise, a window is applied to the ramp filter. A window is a function that is designed to eliminate high-frequency noises and retain the low-frequency patient data. Typical filters that are commonly used in reconstruction are basically the products of a ramp filter that has a sharp cut-off at the Nyquist frequency (0.5 cycle/pixel) and a window with amplitude 1.0 at low frequencies but gradually decreasing at higher frequencies. A few of these windows (named after those who introduced them) are illustrated in Figure 12.9, and the corresponding filters (more correctly, filter-window combinations) are shown in Figure 12.10.

The effect of a decreasing window at higher frequencies is to eliminate the noise associated with them. The frequency above which the noise is




FIG. 12.8. The ramp filter in the frequency domain.




Fig. 12.9. Different windows for reconstruction filters in SPECT. Different windows suppress the higher spatial frequencies to a variable degree with a cutoff Nyquist frequency of 0.5 cycle/pixel.

eliminated is called the cut-off frequency (ν_c). As the cut-off frequency is increased, spatial resolution improves and more image detail can be seen up to a certain frequency. At a too high cut-off value, image detail may be lost due to inclusion of inherent noise. Thus, a filter with an appropriate




Fig. 12.10. Different filters for SPECT that are obtained by multiplying the respective windows by the ramp filter with cutoff at Nyquist frequency of 0.5 cycle/pixel.

cut-off value should be chosen so that primarily noise is removed, and image detail is preserved. Note that the Nyquist frequency is the highest cut-off frequency for a reconstruction filter and typical cut-off frequencies vary from 0.2 to 1.0 times the Nyquist frequency. Filters are selected based on the amplitude and frequency of noise in the data. Normally, a filter with a lower cut-off value is chosen for noisier data as in the case of obese patients and in ^{201}Tl myocardial perfusion studies or other studies with poor count density.

Hann, Hamming, Parzen, and Shepp–Logan filters are all *low-pass* filters because they preserve low-frequency structures, while eliminating high-frequency noise. All of them are defined by a fixed formula with a user-selected cut-off frequency. It is clear from Figure 12.10 that most smoothing is provided by the Parzen filter and the Shepp–Logan filter produces the least smoothing.

An important low-pass filter that is most commonly used in nuclear medicine is the Butterworth filter (Fig. 12.11). This filter has two parameters: the critical frequency (f_c) and the order or power (n). The critical frequency is the frequency at which the filter attenuates the amplitude by 0.707, but not the frequency at which it is reduced to zero as with other filters. The parameter, order or power n , determines how rapidly the attenuation of amplitudes occurs with increasing frequencies. The higher the order, the sharper the fall. Lowering the critical frequency, while maintaining the order, results in more smoothing of the image.

Another class of filters, the Weiner and Metz filters, enhances a specific frequency response.




FIG. 12.11. Butterworth filter with different orders and cutoff frequencies.

Many commercial software packages are available offering a variety of choices for filters and cut-off values. The selection of a cut-off value is important such that noise is reduced and image detail is preserved. Reducing a cut-off value will increase smoothing but will curtail low-frequency patient data and thus degrade image contrast particularly in smaller lesions. No filter is perfect and, therefore, the design, acceptance, and implementation of a filter are normally done by trial and error with the ultimate result of clinical utility.

As already mentioned, filtered backprojection was originally applied only to transverse slices from which vertical and horizontal long axis slices are constructed. Filtering between the adjacent slices is not performed, and this results in distortion of the image in planes other than the transverse plane. With algorithms available in current SPECT systems, filtering can be applied to slices perpendicular to transverse planes or in any plane through the 3-D volume of an object. This process is called volume smoothing. However, because of increased popularity of iterative methods described below, the 3-D volume smoothing is not widely applied.

Iterative Reconstruction

The basic principle of iterative reconstruction involves a comparison between the measured image and an estimated image that is repeated until a satisfactory agreement is achieved. In practice, an initial estimate is made of individual pixels in a projection of a reconstruction matrix of the same size as that of the acquisition matrix, and the projection is then compared with that of the measured image. If the estimated pixel values in the projection are smaller or greater than the measure values, then each pixel value is adjusted in relation to other pixels in the projection to obtain an updated estimated projection, which is then compared with the measured projection. This process is repeated until a satisfactory agreement is obtained between the estimated and actual images. The schematic concept of iterative reconstruction is illustrated in Figure 12.12. The method makes many iterations requiring long computation time and thus discouraging its general use in image reconstruction until recently. However, with the availability of faster computers nowadays, this method is gaining popularity in image reconstruction, particularly in PET imaging.

Initially a uniform image is arbitrarily estimated for comparison (e.g., all pixel counts equal to 0, 1, or a mean value). The image is then unfolded into a set of projections by a process called *forward projection* as opposed to backprojection. It is accomplished by determining the weighted sum of the activities in all pixels in the projection across the estimated image. From Figure 12.13, a projection q_i in the estimated image is given by

$$q_i = \sum_{j=1}^N a_{ij} C_j \quad (12.4)$$




FIG. 12.12. Conceptual scheme of iterative reconstruction.




FIG. 12.13. A projection q_i in the estimated image is the sum of counts in all pixels C_j and is compared with the measured projection p_i .

where C_j is the counts (activity) in the j th pixel and a_{ij} is the probability that an emission from pixel j is recorded in the i th projection. The weight, a_{ij} , is equal to the fraction of activity in the j th pixel out of the total activity along the i th projection. If p_i is the measured projection, then the error is calculated as the difference ($p_i - q_i$), or as the ratio p_i/q_i . The weighting factors are then applied to distribute this error ($p_i - q_i$ or p_i/q_i) into all pixels (N) along the i th projection according to

$$\Delta C_j = \frac{a_{ij}(p_i - q_i)}{\sum_{j=1}^N a_{ij}} \quad \text{or} \quad \Delta C_j = \frac{a_{ij}(p_i/q_i)}{\sum_{j=1}^N a_{ij}} \quad (12.5)$$

where ΔC_j is the error introduced into the j th pixel from all pixels in the i th projection. Note that in error calculation, only pixels belonging to the same projection have been considered. However, in reality, all image pixels have a finite probability of contributing counts to any pixel in any projection and the computation of errors becomes very time consuming.

There are three ways of calculating and applying error corrections. In a point-by-point correction technique, the errors due to all pixels from all projections passing through a particular pixel are calculated and used to correct that pixel before proceeding to the next pixel. In a projection-by-projection correction technique, the error is computed for each projection and the image is updated before proceeding to the next projection. In the simultaneous iteration technique, errors for all projections are computed which are then applied simultaneously to update the image.

The two iterative algorithms widely used in image reconstruction are the *maximum-likelihood expectation maximization* (MLEM) algorithm and the *ordered subset expectation maximization* (OSEM) algorithm. The main feature of the MLEM algorithm is to update the image during each iteration using Eqs. (12.4) and (12.5). This method requires many iterations to achieve a satisfactory agreement between the estimated and measured images, demanding a lengthy computation time. To circumvent this problem, the OSEM algorithm has been introduced, which is a modification of MLEM in that projections are grouped into a number of subsets separated by some fixed angle. For each subset, MLEM is applied and the expected projection values are computed from the estimation of the pixel values in all projections in the subset and compared with the measured image. The variance is entered into the next subset, MLEM is applied, and the image is updated. This is repeated for all subsets. After all subsets are exhausted, a single iteration is considered complete. Such iteration is repeated until an expected agreement is achieved between the estimated and measured images. It has been shown that if there are n subsets, and once all subsets are used in a single iteration of the OSEM, an estimate is produced which is similar to that obtained by n iterations of the MLEM using all projections. It is this property of the OSEM that accelerates the




FIG. 12.14. Comparison of filtered backprojection and iterative OSEM method with attenuation correction.

computation process, and, in general, the computation time decreases with the decreasing number of subsets (i.e., more projections in each subset).

Corrections for detection efficiency variations, noise component, random coincidences, scatter coincidences, and photon attenuation are made prior to reconstruction in the FBP method. In the MLEM or OSEM method, these factors are inherently incorporated a priori in the estimated image and need not be applied separately. In general, iterative reconstruction methods do not produce artifacts observed with the FBP method and provide a better signal-to-noise ratio in regions of low tracer uptake (Fig. 12.14). Overall, iterative methods provide high-quality images and are currently included in image reconstruction in PET and SPECT.

Another algorithm, the row-action maximum likelihood algorithm (RAMLA), has been proposed as a special case of OSEM requiring sequences of orthogonal projections, which lead to faster convergence than the OSEM itself.

SPECT/CT

Accurate medical diagnosis of human disease can be made if both anatomical and functional status of the patient's disease is known. In the interpretation of nuclear medicine studies, physicians always like to have a comparison between high-resolution CT or MR images and low-resolution PET or SPECT images for accurate localization of lesions. In PET and SPECT imaging, *in vivo* measurement of organ physiology, cellular metabolism, and perfusion and other functional status of the organ is made. However, these studies have poor resolution due to poor photon flux and lack anatomical detail. On the other hand, computed tomography (CT) or magnetic resonance (MR) imaging provide excellent spatial resolution with high anatomical detail, but little functional information.




FIG. 12.15. A SPECT/CT camera, Symbia model (Courtesy of Siemens Medical Solutions, Hoffman Estates, IL).

Efforts are made to co-register the two sets of images, in which the matrix size, voxel intensity, and rotation are adjusted to establish one-to-one spatial correspondence between the two images. Various techniques of such alignment are employed, and co-registered images are displayed side by side with a linked cursor indicating spatial correspondence, or may be overlaid or fused using the gray or color scale. The major drawback of these alignment techniques arises from positional variations of the patient scanned on different equipment and at different times. Furthermore, patient motion, voluntary or involuntary, adds to the uncertainty in the co-registration. Even with the sophisticated algorithm, a misalignment of 2 to 3mm is not uncommon.

To overcome the problem of positional variations in alignment of images from different equipment, a dual-modality system has been introduced, in which a SPECT camera and a CT scanner are combined into a single system for imaging the patient in the same clinical setting. Both units are mounted on the same gantry, with the SPECT camera in the front and the CT scanner in the back, and use a common imaging table (Fig. 12.15). The two units are mounted fixed, therefore the centers of the scan fields of SPECT and CT scanners are separated by a fixed distance, called the displacement distance. The axial travel range of the scanning table varies with different designs of the manufacturers. The scan field is limited by the maximum travel range of the table minus the displacement distance.

The details of CT scanners are found in standard reference books on CT and are not given here. The CT scanner consists of an x-ray tube that projects an intense beam of x-rays (of energy ~ 70 to 140 keV) through the

patient's organ of interest, and the transmitted beam is detected by an array of detectors. The acquired data are used to reconstruct images using reconstruction methods described for SPECT images. Currently multislice CT scanners are available, which are fast and provide 6, 16, or 64 slices of the organ in seconds. These scanners have produced high-resolution diagnostic-quality images and reduced the imaging time significantly thus improving the patient throughput.

Either CT or SPECT imaging can be performed first, followed by the other. For example, CT images are taken first with the organ of interest in the CT field of view. Next, the scan table with the patient in the same position is moved to the center of the SPECT FOV and images are taken. Both CT (anatomical) and SPECT (functional) images are reconstructed and then fused together by applying appropriate alignment algorithms. Various vendors provide commercially available fusion software, namely, Syngo by Siemens, Syntegra by Philips Medical, MIM of MIMVISTA, and Volumetrix of GE Healthcare. Because the position of the patient on the table does not change, both CT and SPECT images are aligned very accurately and the overall accuracy is improved by 20 to 25% compared to either modality alone.

A major advantage of including CT in the dual-modality is that the CT data can be utilized in attenuation correction of SPECT data, which is particularly useful in cardiac perfusion imaging. Apparent perfusion defects are often seen in the anterior wall in women due to breast position and in the inferior wall in men, and soft-tissue attenuation also shifts between rest and stress images. As will be described later, attenuation correction using CT transmission data compensates for these artifacts more accurately in a shorter time than using the conventional sealed source transmission data. Such CT transmission attenuation correction can be applied to other organ imaging as well.

GE Healthcare pioneered the first commercial SPECT/CT system integrating its Infinia SPECT camera with the Hawkeye CT scanner. Currently, Symbia Truepoint SPECT/CT by Siemens and Precedence by Philips Medical have been added to the commercial market.

Factors Affecting SPECT

Photon Attenuation

γ -Ray photons are attenuated in body tissue while passing through a patient. Attenuation causes less count density generating artifacts particularly at the center of the image. The degree of attenuation depends on the photon energy, the thickness of tissue, and the linear attenuation coefficient of the photons in the tissue. If I_0 is the number of photons emitted from an organ and I is the number of photons detected by the gamma camera, then

$$I = I_0 e^{-\mu x} \quad (12.6)$$




FIG. 12.16. **A.** Illustration of photons traveling different depths of tissue, thus suffering variable attenuation. **B.** Two photons traversing distances a and b are detected by the two detectors oriented at 180° . Attenuation correction can be applied by taking the geometric mean of the two counts I_a and I_b , and using the total thickness D of the tissue in place of a and b separately.

where μ is the linear attenuation coefficient of the photon in tissue and x is the depth of tissue traversed by the photon (Fig. 12.16). Photons originate from different depths of tissue, which are not exactly known. However, in SPECT imaging, a solution to this problem is to obtain two counts in opposite projections and then calculate the geometric mean of the two counts. This is easily obtained in 360° data acquisition in SPECT. Thus the geometric mean is given by

$$I_g = (I_a \times I_b)^{1/2} \tag{12.7}$$

where I_a and I_b are the measured attenuated counts in the opposite projections. Considering Figure 12.16B and applying Eq. (12.6), Eq. (12.7) becomes

$$\begin{aligned} I_g &= (I_a \times I_b)^{1/2} = (I_{a0} \times I_{b0})^{1/2} e^{-\mu(a+b)/2} \\ &= (I_{a0} \times I_{b0})^{1/2} e^{-\mu D/2} \end{aligned} \tag{12.8}$$

where I_{a0} and I_{b0} are the unattenuated counts detected in opposition and D is the total thickness of the tissue. For parallel collimators, which are most commonly used in SPECT imaging, the photon density does not change with distance; that is, I_{a0} and I_{b0} are approximately equal. Then Eq. (12.8) becomes

$$I_0/I_g = e^{\mu D/2} \tag{12.9}$$

Equation (12.9) is the attenuation correction factor that is applied to the geometric mean counts to obtain the unattenuated counts.

Attenuation Correction

There are two methods of attenuation correction: the Chang method and the transmission method.

Chang Method. The application of Eq. (12.9) is called the Chang method.

In this method, an attenuation map is generated from individual pixel values based on the estimated thickness of an organ of interest and the assumption of a constant μ . Emission data are obtained by Eq. (12.7) and attenuation correction is applied using the factors from the map. This method works reasonably well for organs such as the brain and abdomen, where the attenuating tissue can be considered essentially uniform. However, the situation is complicated in areas such as the thorax, where μ varies due to close proximity of various organs, and the Chang method is difficult to apply.

Transmission method. The most acceptable method for attenuation correction in SPECT is the transmission method. Several SPECT systems currently use a transmission source of a radionuclide that is mounted opposite to the detector such as is an x-ray tube in computed tomography. The detector collects the transmission data to correct for attenuation in emission data. For ^{99m}Tc imaging, common transmission sources are gadolinium-153 (^{153}Gd) (48 keV, 100 keV) and ^{57}Co (122 keV), whereas for ^{201}Tl imaging, americium-241 (^{241}Am) (60 keV) and ^{153}Gd are used in different configurations. In one common configuration, a well-collimated line source is mounted that is translated across the plane parallel to the detector face to collect transmission data. The line source is scanned at each angular stop during the SPECT data collection to apply attenuation correction to each angular projection.

Typically, a blank scan is obtained without the patient in the scanner. The data from this scan are used for all subsequent patients for the day. Then a transmission scan is obtained with the patient in the scanner before the emission scan is acquired. The ratio of counts of each pixel between the blank scan and the transmission scan is the attenuation correction factor for the pixel, which is applied to the emission pixel data obtained next. This is done for each patient for the day. Because the patient is positioned separately in the two scans, error may result in the attenuation correction.

Because the transmission and emission photons have different energy, it is possible that SPECT cameras can be used to collect both transmission and emission data simultaneously using separate discriminator settings. However, one should keep in mind that there is spillover of scattered high-energy photons (i.e., with reduced energy) into the low-energy photopeak window. The transmission data are used to calculate the attenuation factors, which are then applied to the emission data. It should be noted that in the

iterative reconstruction method, the attenuation factor is taken into consideration in the estimated image and it is perhaps the best approach for attenuation correction in SPECT.

In SPECT/CT, the CT transmission data are utilized for attenuation correction, with an advantage of fast data collection in less than a minute thus improving the patient throughput. In the CT technique, typically a blank CT scan is taken without the patient in the scanner at the beginning of the day and is used for subsequent patient studies for the day. Next, the CT transmission scan of the patient is taken and an attenuation correction map is generated from the ratios of counts of each pixel of the blank and the patient transmission scans. After the CT scan, the scanning table with the patient in the same position is moved to the SPECT scanning field and the emission scan is obtained. Factors from the map are then applied to the corresponding pixels in the patient's emission scan for attenuation correction. A typical patient scan is shown in Figure 12.17 with and without attenuation correction.

Note that the attenuation depends on photon energy and, therefore, correction factors derived from ~ 70 keV CT x-ray scans must be scaled to the energy of single photons of the radionuclide used (e.g., 140 keV of ^{99m}Tc) by applying a scaling factor defined by the ratio of mass attenuation coefficient of the photons in tissue to that of 70 keV photons. This factor is assumed to be the same for all tissues except bone, which has a slightly higher mass attenuation coefficient. Because the position of the patient does not change in CT transmission and PET emission scans, the error in positional misalignment of pixels between the two scans is minimized. The CT transmission method provides essentially noiseless images.

Several factors introduce errors in CT attenuation correction factors. One is the respiratory motion of the thorax during scanning that causes mismatch in the fusion of SPECT and CT images. These effects are minimized by multislice CT scanning and a short scanning time (~ 25 s), with a breath hold at end expiration. Also, contrast agents affect the CT attenuation



FIG. 12.17. Illustration of attenuation correction of cardiac SPECT image reconstructed by filtered backprojection. A. Cardiac SPECT images without attenuation showing deficient activity in the inferior wall. B. The same images with attenuation correction using the CT transmission method showing improvement in count density in the inferior wall.

factors because contrast-enhanced pixels overestimate attenuation. Some investigators advocate not using contrast agents and others suggest the use of water-based contrast agents to mitigate this effect.

Partial-Volume Effect

Partial-volume effects are inherent flaws of all imaging devices, because no imaging device has perfect spatial resolution. When a “hot” spot relative to a “cold” background is smaller than twice the spatial resolution of the imaging device, the activity around the object is smeared over a larger area than it occupies in the reconstructed image. Although the total counts are preserved, the object appears to be larger and to have a lower activity concentration than it actually has. Similarly, a small cold spot relative to a hot background would appear smaller as if with higher activity concentration. Such underestimation and overestimation of activities around smaller objects result from what is called the partial-volume effect.

The partial-volume effect is a serious problem for smaller structures in images, and correction needs to be applied for the overestimation or underestimation of the activities in them. A correction factor, called the recovery coefficient, is the ratio of the reconstructed count density to the true count density of the region of interest that is smaller than twice the spatial resolution of the system. The recovery coefficient can be determined by measuring the count densities of different objects containing the same activity but with sizes larger as well as smaller than the spatial resolution of the system. Recovery coefficients are usually measured using phantoms which may not truly be representative of the human body. The measured recovery coefficients are then applied to the image data of the patient to correct for partial volume effect. The recovery coefficient would be one for the larger objects.

Center of Rotation

The center of rotation (COR) parameter is a measure of the alignment of the opposite views (e.g., posterior versus anterior or right lateral versus left lateral) obtained by the SPECT system. The COR must be accurately aligned with the center of the acquisition matrix in the computer. If the COR is misaligned, then a point source would be seen as a “donut” on the image (Fig. 12.17). Thus, an incorrect COR in a SPECT system would result in image degradation. For example, an error of 3 mm in the alignment of COR is likely to cause a loss of resolution of ~30% in a typical SPECT system (Todd-Pokropek, 1983).

The misalignment of COR may arise from improper shifting in camera tuning, mechanics of the rotating gantry, and misaligned attachment of the collimator to the detector. The COR off by more than one pixel may cause degradation in the reconstructed image. It is essential that the COR alignment is assessed routinely in SPECT systems to avoid potential degrada-

tion in spatial resolution. Manufacturers provide detailed methods of determining COR alignment for SPECT systems, which should be included in the routine quality control procedure. Currently, manufacturers include the COR alignment in their maintenance services.

Sampling

It is understood that the larger the number of projections (i.e., small angular increment), the less is the star or streaking effect and hence the image quality is better. However, this requires a longer acquisition time. Ideally, for accurate reconstruction, the number of angular projections should be at least equal to the size of the acquisition matrix (e.g., 64 angular projections for a 64×64 matrix or 128 projections for a 128×128 matrix). A fewer number of projections may not erase the streaking effect. How many angular projections should be taken over 180° or 360° to reconstruct the images accurately depends on the spatial resolution of the camera. As a general rule, 120 to 128 projections (using a 128×128 matrix) are needed for large organs such as lungs and liver, whereas 60 to 64 projections (using a 64×64 matrix) are sufficient for smaller organs such as head and heart. Typically, angular sampling at 6° intervals for 360° acquisition or at 3° intervals for 180° acquisition are commonly used for most SPECT systems.

Scattering

Radiations are scattered in patients, and the scattered photons, depending on the energy and angle of scattering, may strike the detector. Scattering may occur in the detector itself, and also within or outside the FOV. Normally, most of these scattered photons fall outside the photopeak window and are rejected. However, a fraction whose photon energy falls within the photopeak window will be counted, but their (X, Y) positions remain uncertain causing degradation of the image resolution. In SPECT imaging, more than 95% of the 140 keV photons of ^{99m}Tc are scattered in the patient and pose a serious problem. Scatter correction should be applied to improve the spatial resolution of images.

There are a few methods of scatter correction, of which the most common method is the use of two windows: a scatter window and a photopeak window. The scatter window is set at a lower energy than the photopeak window, and it is assumed that scatter in the photopeak window is the same as that in the scatter window. The scatter counts in the scatter window are subtracted from the photopeak counts for each projection to obtain the scatter-corrected projections, which are then used for reconstruction. The scatter spectrum is variable in energy; therefore, to have more accurate scatter corrections, multiple scatter windows can be used. Scatter corrections are made prior to attenuation correction, because the former are amplified during the latter operation.

Performance of SPECT Cameras

Performance of SPECT cameras is assessed by evaluating several parameters such as spatial resolution, sensitivity, and contrast, which are discussed below.

Spatial Resolution

Spatial resolution of a SPECT camera is affected by all the factors discussed in Chapter 10. Typically, it consists of intrinsic resolution, collimator resolution, and scatter resolution. Because of 3-D orientation, SPECT images have both axial and transaxial spatial resolutions. The axial spatial resolution refers to the resolution along the axis of rotation of the camera heads, whereas the transaxial spatial resolution indicates the resolution across the transverse FOV perpendicular to the axis. Normally, planar images have better spatial resolution than SPECT images. Typical values of extrinsic resolution for SPECT cameras with low-energy, high-resolution collimators are 7 to 10mm at a radius of rotation of 10cm. Spatial resolution deteriorates but sensitivity increases with increasing slice thickness. As a trade-off between spatial resolution and sensitivity, an optimum slice thickness should be chosen.

Sensitivity

The sensitivity of an imaging system is always desired to be higher for better image contrast. All factors that affect the conventional cameras also affect the SPECT systems in the same manner, as discussed in Chapter 10. The sensitivity of gamma cameras varies from 200 to 500kcps/ μ Ci. The SPECT systems are designed for greater sensitivity so that high counts can be accumulated for images of thin slices of an organ in a reasonable time. For conventional two-dimensional planar images of good contrast, about 500,000 counts are required. Thus, if each sectional image (i.e., slice) of an organ requires 500,000 counts for the same contrast as in a conventional image, and if there are, for example, 20 sectional images of an organ of interest, then 10 million counts would be needed for the entire organ. For most SPECT systems using low-energy all-purpose collimators, 5 to 20 million counts are acquired. Total counts may be increased by either counting for a longer period or by administering more activity. However, long counting is inconvenient for the patient and administering a larger amount of activity increases the radiation dose to the patient.

Other Parameters

Other important parameters to evaluate include effects of high-count rates, uniformity, and contrast of images. These parameters have been discussed

in detail in Chapter 10 for conventional gamma cameras and are equally applicable to multihead SPECT cameras.

Quality Control Tests for SPECT Cameras

Daily Tests

Photopeaking and Uniformity

The daily photopeaking and uniformity tests for SPECT cameras are the same as for conventional gamma cameras described in Chapter 10 with more stringent requirements. These tests must be done for each head of the system. In SPECT systems, nonuniformities are substantially magnified by the filtered FBP method causing a ring artifact in the image particularly at the center of rotation. To achieve uniformity on SPECT images, UFOV nonuniformity should be less than 1%. Ideally this can be achieved by acquiring at least 30 million counts for 64×64 images or 120 million counts for 128×128 images. For practical reasons, however, 5 million counts for large FOV cameras and 3 million counts for small FOV cameras are appropriate. For Siemens e-Cam cameras, acquisition of 5 million counts in a 1024×1024 matrix is recommended.

Weekly Tests

Spatial Resolution

For single-head or multihead SPECT cameras, spatial resolution for each head is checked by using bar phantoms in the same manner as conventional gamma cameras described in Chapter 10. In determining spatial resolution by the intrinsic method, the two detectors must be kept apart at maximum radius, and a ^{99m}Tc point source is placed in a source holder on the rear bed mechanism provided by the manufacturer. The bar phantom is placed on the detector, and the bed is raised to a maximum height.

Center of Rotation

The COR corrections are performed weekly or bi-weekly using the computer software provided by the manufacturer. To begin COR corrections, the camera face must be parallel to the axis of rotation. Generally, a point or line source is placed in the FOV of the camera and then SPECT scans of the source are obtained for 360° . The software analyzes the scans and determines if the COR is within acceptable limits.

Nowadays, many manufacturers provide a phantom using 5-point sources for low-energy high-resolution collimators or 3-point sources for medium- and high-energy collimators. The phantom with the sources in position is




FIG. 12.18. An illustration of the effect of a misaligned center of rotation. A “donut”-shaped image appears from inaccurately aligned center of rotation. (From Todd-Pokropek A. The mathematics and physics of emission computerized tomography (ECT). In: Esser PD, Westerman BR, eds. *Emission Computed Tomography*. New York: Society of Nuclear Medicine 1983; 3.)

placed on the patient table. SPECT data are collected for 360° in a circular orbit of 20cm, when the detector heads are in 180° configurations. In the case of 90° configurations, the default radius is applied as given by the manufacturer. In a proper COR-corrected system, the point sources should be visible in all projection images.

This method is also used to check the head alignment and noncircular orbit configuration in multihead camera systems with the computer software.

Quality Control Tests for CT Scanners

Most of the CT quality control tests are performed by menu-driven software protocols provided by the manufacturer. First, a checkup automatically executes a set of CT warm-up acquisitions, automatic function checks, and air calibration. This checkup is performed daily and guarantees optimum image quality.

The next daily test is the quality control test that is performed by using a quality phantom provided by the manufacturer. After the phantom is installed in the CT field, data are acquired by a menu-driven protocol provided by the manufacturer to calculate the CT value of water (HU), pixel noise, and tube voltage. This test must be performed prior to the first patient examination.

An important monthly test for CT scanners is the CT constancy test, which is performed by using the CT quality phantom. The following factors are checked by this test: position of the light marker, laser slice thickness, homogeneity and noise of kV values, MTF, and the table position. The manufacturer provides the appropriate software to conduct these tests.

Questions

1. Describe the principles of single photon emission computed tomography (SPECT).
2. Explain how SPECT images are reconstructed by the filtered back-projection technique.
3. Elucidate different factors that affect SPECT images.
4. What is the optimum angular sampling for the SPECT system?
5. How does the center of rotation affect the SPECT image and from what does it originate? What is the ideal location of COR in a 64×64 matrix?
6. With a multihead camera, does the sensitivity of an imaging device increase or decrease?
7. What are the different types of common filters used in the reconstruction of images in SPECT?
8. Which of the following acquisition matrices would give better spatial resolution: a 64×64 matrix or a 128×128 matrix?
9. What are the different factors that affect the uniformity of an imaging device? Describe the methods to improve uniformity.
10. In the filtered backprojection method, the Fourier method is preferred to the convolution method. Why?
11. What is edge packing and how is it rectified?
12. High-frequency data represent noise in SPECT image reconstruction. True or False?
13. A filter with a low cut-off frequency should be used for noisy data and low count density studies in order not to curtail image detail. True or False?
14. A gamma camera has a NaI(Tl) detector of 38-cm diameter. Data of a study are acquired in a 64×64 matrix. What is the Nyquist frequency?
15. Describe the OSEM method of reconstruction of SPECT images.

References and Suggested Readings

- Bacharach SL. Image analysis. In: Wagner HN Jr, Szabo Z, Buchanan JW, eds. *Principles of Nuclear Medicine*. Philadelphia: W.B. Saunders; 1995; 393–404.
- Brooks RA, DiChiro G. Theory of image reconstruction in computed tomography. *Radiology*. 1975; 117:561–572.
- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia; W.B. Saunders: 2003.

- English RT, Brown SE. *Single Photon Emission Computed Tomography: A Primer*. 2nd ed. New York: Society of Nuclear Medicine; 1990.
- Groch MW, Erwin WD. Single-photon emission computed tomography in the year 2001: Instrumentation and quality control. *J Nucl Med Technol*. 2001; 29:12.
- Groch MW, Erwin WD. SPECT in the year 2000: Basic principles. *J Nucl Med Technol*. 2000; 28:233–244.
- Jaszczak RJ, Coleman E. Single photon emission computed tomography (SPECT): Principles and instrumentation. *Invest Radiol*. 1985; 20:897–910.
- Madsen MT. Introduction to emission CT. *Radiographics*. 1995; 15:975–991.
- Rollo FD, ed. *Nuclear Medicine Physics, Instrumentation, and Agents*. St. Louis: Mosby; 1977.
- Todd-Pokropek A. The mathematics and physics of emission computerized tomography (ECT). In: Esser OD, ed. *Emission Computed Tomography*. New York: Society of Nuclear Medicine; 1983:3.

13

Positron Emission Tomography*

Positron emission tomography (PET) is based on the detection in coincidence of the two 511-keV annihilation radiations that originate from β^+ -emitting sources (e.g., the patient containing β^+ -emitting radioactivity). Positrons are annihilated in body tissue and produce two 511-keV annihilation photons that are emitted in opposite directions (180°). Two photons are detected in an electronic time interval, called “coincidence time window,” by two detectors connected in coincidence. Conversion of 511-keV photons to light photons in the detector, formation of a pulse by the PM tube, and pulse-height analysis follow the same principles as in conventional gamma cameras. Detectors are arranged in the array of several rings to have the organ of interest in the field of view. Data collected over 360° simultaneously around the body axis of the patient are used to reconstruct the image of the activity distribution in the slice of interest. Because the two opposite photons are detected in a straight line, no collimator is needed to limit the field of view, and the technique is called the *electronic collimation*.

Positron-Emitting Radionuclides

Various positron-emitting radionuclides are used in preparing radiopharmaceuticals for PET studies, and the production and characteristics of the most common ones are listed in Tables 5.1 and 13.1. Of these, $^{82}\text{Rb-RbCl}$ is commonly used for myocardial perfusion imaging and ^{18}F -fluorodeoxyglucose for metabolic imaging of the heart, brain, and various types of cancers. Readers are referred to reference books on the synthesis of PET radiopharmaceuticals.

* This chapter is adapted from Saha (2005, Chapters 2–5).

TABLE 13.1. Characteristics of PET radionuclides.

Radionuclides	Half-life	Mode of decay (%)	E_{β^+max} (MeV)
^{11}C	20.4 min	$\beta^+(100)$	0.970
^{13}N	10 min	$\beta^+(100)$	1.2
^{15}O	2 min	$\beta^+(100)$	1.74
^{18}F	110 min	$\beta^+(97)$ EC (3)	0.64
^{68}Ga	68 min	$\beta^+(89)$ EC(11)	1.90
^{82}Rb	75 s	$\beta^+(95)$ EC(5)	3.15
^{124}I	4.2d	$\beta^+(23)$ EC(77)	2.14

Detector for PET

Several types of scintillation detectors are available for PET scanners, of which BGO and LSO are most commonly used by commercial manufacturers because of their properties (illustrated in Table 8.1). NaI(Tl) was initially used in earlier PET scanners, but nowadays only one vendor uses this detector (C-PET, marketed by Philips Medical Inc.). Both BGO and LSO are not hygroscopic, and hermetic sealing is not required. Both detectors have high stopping power (high density and linear attenuation coefficient), but LSO has shorter scintillation decay time (40 ns) than BGO (300 ns) and has higher light output (29 photons vs. 6 photons). Currently, LSO has become the preferred detector in PET scanners. However, due to the intrinsic property, its energy resolution is poor. It is to be noted that LSO contains a naturally occurring radioisotope of its own, ^{176}Lu , with an abundance of 2.6% and a half-life of 3.6×10^{10} years. This radionuclide decays by emission of β^- rays and γ -rays of 88 to 400 keV. However, the activity level is low enough to ignore radiation exposure from ^{176}Lu , and its photon energy lower than 511 keV does not interfere in PET imaging.

GSO has been used in PET scanners by some manufacturers, despite its lower light output and stopping power than LSO. These crystals are fragile and require great care in fabrication.

BaF_2 has the shortest decay time of 0.6 ns and is primarily used in time-of-flight PET scanners that are rarely used clinically.

PM Tubes and Pulse-Height Analyzers

Like conventional cameras, PET scanners use PM tubes and pulse-height analyzers (PHA) for determining the locations of the origin of pulses and their amplitudes. As mentioned in Chapter 8, PM tubes convert light

photons arising from the interaction of γ -rays in detectors to pulses, which are used to determine the X -, Y -positions of the two detectors that detect the two 511-keV photons (discussed later), and the PHA is used to check if the pulse height is within the acceptable range, which is normally set at 350 keV to 650 keV for 511-keV annihilation photopeak in PET scanners.

PET Scanners

In the original PET scanners, many detectors (hundreds to thousands), each connected to a photomultiplier tube, were arranged in multiple circular, hexagonal, or orthogonal rings. The number of rings in current scanners (18–32), and the number of detectors per ring vary with the manufacturer. The number of rings and, hence, the width of the array of rings define the axial field of view. Despite the advantage of good resolution with these scanners, the use of many PM tubes and packing them with as many detectors is very costly and impractical.

Block Detectors

In modern PET scanners, the *block detector* is used, in which small detectors are created by making partial cuts in a block of detector crystal and each block detector is coupled to two to four PM tubes. A schematic block detector is shown in Figure 13.1. Typically, each block is about 3-cm deep and grooved into 6×8 , 7×8 , or 8×8 elements by partial cuts through the crystal with a saw. The cuts are made at varying depths, with the deepest cut at the edge of the block. The cuts are filled with opaque reflective materials to prevent spillover of light between elements. The size of each detector varies from 3 mm to 6.5 mm.

The block detectors are arranged in an array of full or partial rings with a diameter of 80 to 90 cm. The full ring array may be in the form of a circular or hexagonal shape. Different arrangements of block detectors adopted by manufacturers are shown in Figure 13.2. Siemens' ECAT ACCEL has 24 rings and 64 detectors/block with a total of 9216 LSO detectors, coupled to 576 PM tubes. The ADVANCE Nxi of General Electric Healthcare has 18 rings and a total of 12,096 BGO detectors with 36 detectors/block, coupled to 672 PM tubes. In the case of partial ring configurations, the blocks have to be rotated around the patient to obtain 360° acquisition of data.

Coincidence Timing Window

In ideal coincidence counting in PET, the two 511-keV annihilation photons should be detected by the two detectors exactly at the same time. In reality, however, one photon may arrive at one detector earlier than the other in




FIG. 13.1. A schematic block detector is segmented into 8×8 elements, and four PM tubes are coupled to the block for pulse formation. The pulses from the four PM tubes (A, B, C, and D signals) determine the location of the element in which 511-keV γ -ray interaction occurs, and the sum of the four pulses determines if it falls within the energy window of 511 keV.

the opposite detector. This uncertainty in detection time is called the timing resolution or coincidence timing window. The timing resolution results from the difference in pulse formation in the detector primarily due to statistical variations in gain and scintillation decay time of the detector. Furthermore, there is a time delay in the arrival of one photon relative to the other, because of the difference in distances traveled by the two photons, partic-




FIG. 13.2. Different configurations of PET scanners. (A) A circular full-ring scanner. (B) A partial ring scanner with a 15° angular shift between two blocks of detectors. (C) Continuous detectors using curve plates of NaI (TI). (D) Hexagonal array of quadrant-sharing panel detectors. (Reprinted with the permission of the Cleveland Clinic Foundation.)

ularly if the annihilation occurs at the edge of the FOV. This also adds to the timing resolution. For a whole-body scanner with a diameter of 1 m, the maximum distance traveled by one photon can be as large as 1 m, and the other photon travels almost no distance, if the annihilation occurs at the edge of the FOV. Because the velocity of light is 3×10^8 m/s, the difference between the arrival times of the two photons is about 3 to 4 ns (time to travel 1 m). This is the limit of timing resolution of a PET scanner with 1-m diameter.

After annihilation, two timing signals A and B are formed with timing width, say τ , depending on the scanner system. Signal B may arrive at one detector just τ ahead of (Fig. 13.3A) or τ behind (Fig. 13.3B) the arrival of signal A, and both signals will be counted in coincidence. This is the extreme




FIG. 13.3. Two time signals with time width τ define the coincidence timing window of 2τ . **(A)** Pulse 2 (dashed line) of time with τ arrives at the opposite detector just after the arrival of Pulse 1 (solid line) of time width τ at one of the detector pair, overlapping only at the edge and two events are counted in coincidence. **(B)** Pulse 2 of time width τ arrives at one of the detector pair exactly prior to the arrival of Pulse 1 of time width τ at the other detector, overlapping only at the edge and two events are counted in coincidence. **(C)** Pulse 2 overlaps Pulse 1 partially or completely depending on the time of arrival of the pulses at the detector pair, and two events are counted in coincidence. Both A and B indicate that the timing window of the coincidence circuit must be at least 2τ to detect all events in coincidence.




FIG. 13.4. The transverse field of view determined by the acceptance angles of individual detectors in a PET scanner. Each detector is connected in coincidence with as many as half the total number of detectors in a ring and the data for each detector are acquired in a “fan beam” projection. All possible fan beam acquisitions are made for all detectors, which define the FOV as shown in the figure. (Reprinted with the permission of the Cleveland Clinic Foundation.)

limit, and all other annihilation events overlapping in their time widths (Fig. 13.3C) will be counted as coincident events. Thus the timing resolution or coincidence timing window has to be a minimum of 2τ . The typical coincidence timing window of conventional PET scanners is set at 6 to 20 ns.

In a PET scanner, each detector element is connected by a coincidence circuit with a time window to a set of opposite detector elements (both in plane and axial). The number of opposite detectors can vary from one to a maximum of half the total number of detectors present in a ring. Each detector element can be connected in coincidence to a maximum of half the total number N of opposite detector elements ($N/2$). Note that less than $N/2$ detectors can also be connected. As shown in Figure 13.4, depending on the number of opposite detectors connected, each detector element has a number of projections that define an *angle of acceptance*. These angles of acceptance for all detector elements in the ring form the transaxial FOV. The angle of acceptance increases with the number of opposite detector elements connected in coincidence and, hence, the FOV.

PET/CT Scanners

Parallel to SPECT/CT systems, PET/CT has been developed, in which functional PET images are fused with anatomical CT images to provide more accurate diagnosis of human diseases. In a PET/CT unit, both scanners are mounted on a common gantry with the CT unit in the front and PET unit in the back attached to the CT unit. Typically, this is opposite to the SPECT/CT arrangement in which the SPECT camera is placed in the front and the CT unit in the back. Both units use the same scanning table. A commercial PET/CT is shown in Figure 13.5. As in SPECT/CT, the centers of the PET FOV and CT FOV are separated by a fixed distance, called the displacement distance. The horizontal travel range of the scanning table varies with the designs of the scanners. The actual scan field is the maximum travel range of the scanning table minus the displacement distance. Typically, a CT scan is taken first of the patient on the scanning table and in the CT scan field. The table with the patient in the same position is moved to the PET scan field and imaging is performed. Note that ^{18}F -FDG was administered 60 to 90 minutes prior to scanning, and low-intensity 511-keV



Fig. 13.5. A commercial PET/CT scanner, Biograph 16. (Courtesy of Siemens Medical Solutions, Hoffman Estates, IL.)

photons in the patient do not interfere with the CT scan, because of the high intensity of the CT photon beam. Fusion of the two sets of images is performed by using commercial software as discussed under SPECT/CT. As described later, CT transmission data is used to calculate the attenuation factors for the PET emission data. The principle of attenuation correction is the same as described in SPECT/CT, based on the ratios of pixel data of a blank and the patient's CT transmission scans. Factors affecting the attenuation corrections are discussed later.

Because the position of the patient on the table does not change, both CT anatomical and PET functional images can be fused with accurate alignment providing better accuracy in detection of diseases. The overall accuracy of diagnosis increases by 20 to 25% with PET/CT compared to either modality alone. Because CT scanning is fast, the total scanning time is significantly reduced, and the patient throughput is increased. Because of the increased reimbursement, whole-body imaging with PET/CT for accurate diagnosis of various cancers has become the standard of practice. The sale of PET/CT units worldwide has outpaced that of PET scanners alone by almost nine to one.

Mobile PET or PET/CT

Largely because of the low patient volume and high cost, many community hospitals cannot afford PET or PET/CT scanners, but can take advantage of mobile PET or PET/CT that provides PET scanning services to different locations. PET or PET/CT scanners and necessary accessories are installed in sturdy vans, along with nuclear pharmacy facilities. The mobile unit moves to different clients' facilities on different days depending on the schedule. The patient's schedule and delivery of PET tracers must be well coordinated to provide efficient services. The owner of the mobile unit must have a license from appropriate authorities to operate the mobile PET/CT and also a letter of agreement between the client and the licensee to provide the service. The van must meet the Department of Transportation's overload regulations, and the rules and regulations of fire safety and security of local authorities.

Micro-PET

For research animal imaging, clinical PET scanners with a large bore give poor spatial resolution. Micro PET scanners with a smaller bore and, hence, smaller in overall size (to be fitted in small rooms) have been developed by several manufacturers. The typical bore diameter is about 16 cm. The spatial resolution can be obtained as small as 1 mm with the use of LSO detectors. These scanners are useful for drug evaluation in animals.

Hybrid Gamma Cameras

Conventional dual-head and triple-head gamma cameras (Fig. 12.2) can be utilized as PET cameras by connecting the appropriate heads with a coincidence circuitry. The typical time window is ~ 12 ns for dual-head and ~ 10 ns for triple-head cameras. In SPECT mode, the cameras are used with collimators, whereas in PET mode, collimators are removed, and therefore they can be used in either mode as needed. These cameras are called hybrid gamma cameras and are very attractive to community hospitals because of their low cost with the scope of PET imaging.

The hybrid cameras suffer from a disadvantage of low sensitivity due to low detection efficiency of NaI(Tl) crystal for 511-keV photons. To improve sensitivity, thicker crystals of sizes 1.6 cm to 2.5 cm have been employed in some cameras, with a resultant increase in coincidence photopeak efficiency of only 3 to 4%, with concomitant degradation of spatial resolution. There is a significant camera dead time loss and pulse pile-up of counts in PET mode in the absence of a collimator. Overall, the spatial resolution of a hybrid camera is poorer than that of a dedicated PET scanner.

Data Acquisition

In PET imaging, two 511-keV annihilation photons are detected in coincidence by two opposite detectors along a straight line, called the *line of response* (LOR). In a full ring system, data are collected in 360° simultaneously, whereas in the partial ring system, the rings are rotated around the patient for 360° data acquisition. There are three steps in PET data acquisition. First, the location of the detector pair in the ring is determined for each coincident event. Next, the pulses are analyzed by PHA to check if they are within the energy window set for 511 keV. Finally, the position of the LOR is determined in polar coordinates to store the data in computer memory.

Because each detector is connected to many opposite detectors in coincidence, which detector pair detected a coincidence event must be determined. As in gamma cameras, the position X , Y of each detector in the ring is determined by

$$X = \frac{(C + D) - (A + B)}{A + B + C + D} \quad (13.1)$$

$$Y = \frac{(A + D) - (B + C)}{A + B + C + D} \quad (13.2)$$

where A, B, C, D are the pulses from the four PM tubes attached to the block, as shown in Figure 13.1.

Next, the four pulses (A, B, C, and D) are summed up to give a Z pulse, which is then checked by the PHA if its amplitude is within the energy window set for the 511-keV photons. If it is outside the window, it is rejected; otherwise, it is accepted for storage.

The last step in data acquisition is the storage of the data in the computer. Unlike conventional planar imaging where individual events are stored in a (X, Y) matrix, the coincidence events in PET imaging are stored in the form of a *sinogram*. Consider an annihilation event occurring at the * position in Figure 13.6A. The coincidence event is detected along the LOR indicated by the arrow between the two detectors. It is not known where along the line of travel of the two photons the event occurred, because they are accepted within the set time window (say, 12 ns) and their exact times of arrival are not compared. The only information we have is the positions of the two detectors in the ring that registered the event, that is, the location of the LOR is established by the (X, Y) positioning of the two detectors. Many coincidence events arise from different locations along the LOR and all are detected by the same detector pair and stored in the same pixel, as described below.

For data storage in sinograms, each LOR is defined by the distance (r) of the LOR from the center of the scan field (i.e., the center of the gantry)




FIG. 13.6. PET data acquisition in the form of a sinogram. **(A)** Each LOR datum is plotted in (r, ϕ) coordinates. **(B)** Data for all r and ϕ values are plotted to yield the sinogram indicated by the shaded area (only a part is shown). (Reprinted with the permission of the Cleveland Clinic Foundation.)




FIG. 13.7. A typical normal sinogram indicating all detectors are working properly.

and the angle of orientation (ϕ) of the LOR (i.e., the angle between r and the vertical axis of the field). A matrix of an appropriate size is chosen defined by the r , ϕ coordinates, rather than by X -, Y -coordinates used in SPECT data acquisition, and counts in each LOR are stored in the corresponding pixel in the matrix. If we plot the distance r on the x -axis and the angle ϕ on the y -axis, then the coincidence event along the LOR (r , ϕ) will be assigned at the cross-point of r and ϕ values (Fig. 13.6B). In a given projection, adjacent detector pairs constitute parallel LORs (at different r values in Fig. 13.6A) at the same angle of orientation. The plot of these LORs will be seen as a horizontal row at angle ϕ . Similarly LORs from different projections (i.e., at different angles ϕ) for the same r values can be plotted, which will give a vertical line. When all projections around the field of view are considered, the plot of the LORs at different projection angles and r values will result in the shaded area in Figure 13.6B, which is called the sinogram. A typical normal sinogram is shown in Figure 13.7.

The sinogram represents a single slice of data for a transverse FOV obtained from a single ring of the PET scanner. PET data are acquired directly into a sinogram in a matrix of appropriate size in the computer. Each pixel corresponds to a particular LOR characterized by (r , ϕ) containing all coincidence counts detected by the detector pair along the LOR. Data can be collected in both static and dynamic imaging using either the frame mode or the list mode, described in Chapter 11.

Because PET scanners are axially fixed, whole-body imaging is accomplished by the use of a computer-controlled bed-table that moves along the axis of the scanner. The whole-body scan of the patient is obtained at different axial positions of the bed.

Two-Dimensional Versus Three-Dimensional Data Acquisition

Coincident counts detected by a detector pair are called the *prompts* which include true, random, and scatter events described later (Fig. 13.8). To eliminate random and scatter events, annular septa (~1 mm thick and radial width of 7 to 10 cm) made of tungsten or lead are inserted between rings in multiring PET scanners (Fig. 13.9A). The septa act as do parallel hole collimators in gamma cameras. They mostly allow direct coincidence events to be recorded from a given ring and prevent random and scatter from other rings. This mode of data collection is called two-dimensional (2-D) acquisition. The use of septa reduces the contribution of scattered photons from 30–40% without septa to 10–15%. To improve sensitivity, detector pairs in two adjacent or nearby rings are also connected in coincidence. Such cross-coincidence connections can be made at most among five adjacent rings. Coincidence events detected by the detectors connected in the same ring are called the *direct plane* events, whereas those detected by detectors interconnected between different rings are called the *cross plane* events. Although the cross plane events increase the sensitivity, they degrade the




FIG. 13.8. **(A)** True coincidence events. **(B)** Random coincidence events detected by two detectors connected in coincidence along the dotted line. The two 511-keV photons originated from different positron annihilations. **(C)** Scattered coincidence events. One or both of the 511 keV photons from the same annihilation event may be scattered with little loss of energy and may fall within the PHA window and also within the coincidence time window to be detected as a coincidence event by two detectors.




FIG. 13.9. (A) 2-D data acquisition with the septa placed between the rings so that true coincidence counts are obtained avoiding random events and scatters. Detectors connected in the same ring give direct plane events. However, detectors are connected in adjacent rings and cross plane data are obtained as shown. (B) When septa are removed, the 3-D data acquisition takes place, which includes random and scatter events along with true events. (Reprinted with the permission of the Cleveland Clinic Foundation.)

spatial resolution as a trade-off. The overall sensitivity in 2-D acquisition is 2 to 3% at best.

To improve further the sensitivity of PET scanners, the three-dimensional (3-D) acquisition is employed in which the septa are retracted, or they are not included in the scanner (Fig. 13.9B). In this mode, all events detected by detectors in coincidence in all rings are counted including random and scatter events, and the sensitivity in the 3-D mode increases four- to eight-fold over 2-D acquisition. The incidences of random and scatter can be reduced by having a smaller angle of acceptance; that is, a detector is connected to a smaller number of opposite detectors.

Image Reconstruction

Image reconstruction of 2-D PET data is accomplished by the same filtered backprojection and iterative methods that have been described in detail under SPECT in Chapter 12. In PET, the LORs in a sinogram are backprojected by the Fourier method. In the iterative method, the projections are estimated by determining the weighted sum of the activities in all pixels along a LOR across the estimated image, and then compared with the measured projection.

The reconstruction of images from 3-D data is complicated by a very large volume of data, especially in a multiring scanner. The direct application of filtered backprojection and the iterative method to these data is difficult, and so the 3-D sinogram data are rebinned into a set of 2-D equivalent projections by assigning axially tilted LORS to transaxial planes intersecting them at their axial midpoints. This method is called the single slice rebinning method (SSRB). In another method, called the Fourier rebinning (FORE) method, rebinning is performed by applying the Fourier method to each oblique sinogram in the frequency domain. This method is more accurate than the SSRB method because of the more accurate estimate of the source axial location. After rebinning of 3-D data into 2-D data, either the filtered backprojection or iterative method is applied.

Factors Affecting PET

As in gamma cameras, PET acquisition data are affected by photon attenuation, variation in detection efficiency of the detectors, scatter coincidences, partial volume effect, and dead time. These factors are already discussed under SPECT and therefore only subtle points pertinent to PET data will be highlighted here. In addition, PET data are affected by some unique factors such as random coincidence and parallax error (radial elongation), which will be discussed below.

Normalization

There are 10,000 to 20,000 detectors arranged in blocks, and coupled to several hundred PM tubes in modern PET scanners. Practically, as in gamma cameras, the detection efficiency varies from detector pair to detector pair due to variation in the gain of PM tubes and location of the detector in the block, resulting in nonuniformity of the PET data. Data are corrected for this factor by using what is called *normalization*. In normalization of PET data, all detectors are exposed uniformly to a 511-keV photon source (e.g., ^{68}Ge source), without an object in the field of view, and data are collected in the 2-D or 3-D mode. The normalization factors F_i are calculated for individual pixels as

$$F_i = \frac{A_{mean}}{A_i} \quad (13.3)$$

where A_{mean} is the mean of all pixel counts and A_i is the counts in the i th pixel. The observed count C_i in the i th pixel from the patient is then normalized by

$$C_{norm,i} = C_i \times F_i \quad (13.4)$$

where $C_{norm,i}$ is the normalized count in the i th pixel. The normalization data collection requires a long time (~6–8 hrs) and is normally carried out overnight. These factors are obtained weekly or monthly and most vendors offer algorithms to obtain them routinely.

Photon Attenuation Correction

The two 511-keV photons in PET can traverse different thicknesses of tissues before detection and are attenuated to a different degree similar to situations discussed under SPECT. If the two photons traverse a and b thicknesses of tissues of an organ, then the attenuation correction P for each pixel (i.e., each LOR) is given by

$$P = e^{-\mu a} \times e^{-\mu b} = e^{-\mu(a+b)} = e^{-\mu D} \quad (13.5)$$

where μ is the linear attenuation coefficient of 511-keV photons in tissue and D is the total thickness of the organ (Fig. 13.10). When photons traverse various organs, differences in linear attenuation coefficients and organ thicknesses must be taken into consideration. As in SPECT, Eq. (13.5) has been employed to correct for attenuation in brain PET imaging, based on the assumption of uniform density of tissue and a constant μ for 511-keV photons in tissue (Chang method). However, the method tends to cause artifacts due to underestimation of attenuation in the thorax area.

The transmission method is widely used for attenuation correction of PET emission scan data. In this method, normally a ^{68}Ge source is used to obtain the transmission scan in dedicated PET imaging. The source is placed in a holder mounted at the edge of the scanner bore and the holder is rotated by a motor so that data detected by all detector pairs can be acquired. Normally, a blank scan is obtained at the beginning of the day without any object or patient in the scanner. These blank scan data are used for all subsequent patients for the day. Next a transmission scan is obtained with the patient in the scanner for each patient. The ratios of the counts in each pixel (i.e., each LOR) between the blank scan and the transmission scan are calculated for each patient. The emission scan is taken with the patient in the same position as in the transmission scan to minimize errors in correction factors, and then each LOR datum is corrected for attenuation by using the corresponding ratio. It takes 20 to 40 min for acquisition of the transmission scan depending on the source strength. Normally the




FIG. 13.10. Two 511-keV annihilation photons traverse thicknesses a and b of tissues of an organ. However, attenuation of the two photons depends on the total thickness D of the organ regardless of a and b .

transmission scan is performed before the emission scan to avoid interference of radiations from the administered radioactivity. Other approaches include post-injection transmission scanning (transmission scan after the emission scan) and simultaneous emission/transmission scanning, but each method suffers from various disadvantages of its own.

In PET/CT, the CT transmission scan is utilized for attenuation correction, which takes less than a minute thus improving the patient throughput. As mentioned in the SPECT/CT technique, typically at the beginning of the day, a blank CT scan without the patient in the scanner is obtained, which is later used for subsequent patient studies for the day. Next, the CT transmission scan of each patient is taken with the patient in the CT scan field before the emission scan, and an attenuation correction map is generated from the ratios of counts of each pixel (i.e., LOR) of the blank scan and the transmission scan. Because the PET and CT units are fixed on the same gantry, the patient remains in the same position on the table, which is then moved to the PET scan field for the emission scan. Factors from the attenuation map are subsequently applied to each LOR in the patient's emission scan. The CT transmission method provides essentially noiseless images.

Attenuation depends on photon energy; therefore correction factors derived from ~ 70 -keV CT x-ray scans must be scaled to the 511-keV

photons of PET by applying a scaling factor defined by the ratio of mass attenuation coefficient of 511-keV photons to that of 70-keV photons. This factor is assumed to be the same for all tissues except bone, which has a slightly higher mass attenuation coefficient. As mentioned in the SPECT/CT section, the respiratory motion of the thorax during scanning and intravenous contrast agents affect the CT attenuation factors. Use of breath hold and water-based contrast agents helps mitigate these effects, respectively.

Random Coincidences

Random coincidence events occur when two unrelated 511-keV photons, arising from two different positron annihilation events, are detected by a detector pair within the same time window (Fig. 13.8B). Random coincidences are largely minimized in 2-D acquisition by septa, whereas in 3-D acquisition in the absence of septa, their contribution is high causing loss of image contrast. They increase with increasing pulse-height window, coincidence time window, and activity [varies as the square of the activity; see Eq. (13.7) below], and can be reduced by decreasing these variables. Also increasing the ring diameter reduces the random coincidences.

Corrections for random coincidences can be made by separately measuring two single count rates, R_1 and R_2 , of a radioactive source by each of the detector pair and by using the following equation,

$$R_c = 2\tau R_1 R_2 \quad (13.7)$$

where τ is the coincidence time window and R_c is the random count rate that is to be subtracted from the prompt count rate to obtain the true coincidence count rate. In another method, two coincidence circuits are employed in which one is set at a standard coincidence timing window (e.g., 12ns) and the other at a delayed time window (e.g., 55ns that is, counts arriving within 55ns and 67ns will be counted.), using the same pulse-height window in both cases. The counts in the standard time window contain true, scatter, and random events, whereas in the delayed time window, only random events and no true or scatter coincidence events are recorded, because true and scatter photons from the same annihilation decay arrive at the detectors within the short coincidence time window. The random counts will be the same in both coincidence and delayed coincidence windows. The true counts are then obtained by subtracting the delayed window counts from the standard window counts. Use of faster electronics and a shorter coincidence time window are some of the physical techniques that are used to minimize random events. In another method, a very high radioactive source is counted by the PET scanner over time until the radioactivity is reduced to a level where no random event is recorded. Random events are calculated by subtracting the low activity counts (true plus scatter) from the high activity counts (true plus scatter plus random).

Scatter Coincidences

Annihilation radiations may undergo Compton scattering while passing through the body tissue and, due to high energy of 511 keV, they are mostly scattered forward without much loss of energy (Hoffman and Phelps, 1986). Such scattering may also occur in the detector material itself. These scattered photons may fall within the coincidence time window (Fig. 13.8C) and be detected by the detector pair. One or both of the 511 keV photons from the same annihilation event may be scattered. Note that coincident counts of scattered photons from two separate annihilation events will be considered as random counts. Background of the image is increased by these radiations with concomitant loss of image contrast. Scattering increases with the density and depth of tissue, the density of the detector material, the activity, and the pulse-height window.

Narrowing the pulse-height window reduces the scattered events significantly. The use of septa in 2-D acquisition reduces the scatter events considerably, but in 3-D acquisition, this becomes a problem. In modern PET scanners, the scatter fraction is around 15% in 2-D acquisition, whereas it can be as high as 40% in 3-D acquisition.

The practical methods of scatter correction in PET are essentially similar to that in SPECT described earlier.

Dead Time

The effects of dead time and pulse pile-up have been discussed in Chapter 8. The effects of high-count rates on the performance of gamma cameras have been discussed in Chapter 10. These effects equally apply to the counting of 511-keV annihilation photons in PET imaging. The correction for dead time loss is made by measuring the observed count rates as a function of increasing concentrations of activity. The dead time is calculated from these data and then applied to actual data obtained in the patient's scanning. Uses of high-speed electronics, buffers, and pulse pile-up rejection circuits are some of the techniques that are employed to improve dead time loss.

Radial Elongation

Radial elongation, also called the parallax error or radial astigmatism, occurs from LORs that are off-centered. As shown in Figure 13.11, the two 511-keV photons originating along the actual LOR (solid line) may strike the detectors tangentially at the back of the detector and form a coincident event. But the X , Y positioning of the detectors is defined by the dashed line some distance d away from the actual LOR. This effect results in some blurring of the image due to unknown depth of interactions, and worsens with the LORs farther away from the center of the FOV and with a thicker




FIG. 13.11. An illustration of radial elongation. An off-center event (solid line) strikes the back of the detector pair tangentially. The X -, Y -positioning of the detectors (dashed line) is a distance d away from the actual location of the positron annihilation, causing the blurring of the image. (Reprinted with the permission of the Cleveland Clinic Foundation.)

detector. However, as can be seen from Figure 13.11, the use of a larger diameter of the ring improves this effect. Correction can be made for this effect by measuring light in the front and back of each detector and using the difference to measure the depth of the photon interaction in the detector.

Performance of PET Scanners

Spatial Resolution

The spatial resolution of a PET scanner is defined by several factors that are discussed below.

Detector size. In the absence of collimators, unlike in SPECT, the intrinsic resolution R_i is the predominant factor in spatial resolution, which is related to the detector size d in multidetector PET scanners. It is normally given by d at the face of the detector and by $d/2$ midway between the two detectors.

Positron range. A positron with energy travels a distance in tissue losing energy before it almost comes to rest and then combines with an electron to produce two 511-keV photons. The locations of positron emission and annihilation are separated by the effective range of the positron (Fig. 13.12), which adds to the uncertainty of X, Y positioning of the detector pair. This error R_p increases with the positron energy and decreases with the tissue density. This value is reported to be 0.2 mm for ^{18}F and 2.6 mm for ^{82}Rb in tissue (Tarantola et al., 2003).

Noncolinearity. Positrons at the end of their range still possess some residual momentum and, therefore, the two annihilation photons are not emitted exactly 180° , but at slight deviation. This deviation from 180° is $\pm 0.25^\circ$ at maximum. Because of this deviation, the LOR is displaced from the point of annihilation (Fig. 13.13) and, thus, an error R_a is introduced in the spatial resolution of the PET scanner. This value deteriorates with the diameter of the ring and is estimated to be

$$R_a = 0.0022D \quad (13.8)$$




FIG. 13.12. Positrons travel a distance before annihilation in the absorber and the distance increases with positron energy. Because positrons with different energies travel in zigzag directions, the effective range is the shortest distance between the nucleus and the direction of 511-keV photons. This effective range degrades the spatial resolution of the PET scanner. (Reprinted with the permission of the Cleveland Clinic Foundation.)




FIG. 13.13. Noncolinearity of 511-keV annihilation photons. Because there is some residual momentum associated with the positron, the two annihilation photons are not emitted exactly at 180° , but at a slight deviation from 180° . Two detectors detect these photons in a straight line, which is slightly deviated from the original annihilation line. The maximum deviation is $\pm 0.25^\circ$. (Reprinted with the permission of the Cleveland Clinic Foundation.)

Reconstruction method. If the filter backprojection method is used for reconstruction, the choice of filter with a selected cut-off frequency degrades spatial resolution. This error (K_r) is reported to be a factor of 1.2 to 1.5 depending on the type of filter used.

Location of the detector. The use of block detectors rather than single detectors introduces an error in the (X , Y) positioning of the detector pair, which causes degradation of the spatial resolution. This value (R_l) is approximately 2.2 mm for BGO detectors, and somewhat less for LSO detectors.

Combining all these factors, the spatial resolution of a PET scanner is given by

$$R = K_r \times \sqrt{R_i^2 + R_p^2 + R_a^2 + R_l^2} \quad (13.9)$$

The transaxial spatial resolution typically ranges from 5 mm to 7 mm at 10 cm and the axial spatial resolution from 5.4 mm to 6.6 mm for commercial PET scanners. The transverse resolution is best at the center of FOV and worsens toward the periphery of the scanner.

Sensitivity

The sensitivity of a PET scanner is defined by the number of counts per unit time for each unit of activity and is normally given in counts per second per microcurie (or megabecquerel) (cps/ μ Ci or cps/MBq). Commercial vendors give this in units of volume sensitivity, kcps/ μ Ci/cc or cps/Bq/cc. Sensitivity depends on the geometric efficiency, detection efficiency of the detector, pulse-height window, and dead time of the detector. These factors have been discussed in detail in Chapter 10 for conventional cameras. However, the geometric efficiency needs further consideration because of the specific configuration of the PET scanners. The geometric efficiency depends on the distance d between the detector and the source, the diameter D of the ring, and the number of rings n in the scanner. Increasing the distance d and the diameter D reduces the solid angle subtended by the detector at the source and, thus, decreases the geometric efficiency. With all these factors taken into consideration, the sensitivity S of a single-ring PET scanner is given by (Budinger, 1998)

$$S = \frac{A \cdot \varepsilon^2 \cdot e^{-\mu t} \cdot 3.7 \cdot 10^4}{\pi D^2} \text{ (cps}/\mu\text{Ci)} \quad (13.10)$$

where A is the detector area seen by a point source to be imaged, ε = detector efficiency, μ = linear attenuation coefficient of 511 keV photons in detector material, t is the thickness of the detector, and D is the diameter of the ring. The sensitivity increases with the number of rings in the scanner.

The sensitivity of the PET scanners in 2-D mode acquisition is about 0.2 to 0.5%, whereas it is 2 to 10% in 3-D acquisition due to the absence of septa between the rings. However, the latter contains a large number of random and scatter coincidences.

Noise Equivalent Count Rate

As discussed in Chapter 10 under contrast, noise degrades the image contrast and primarily arises from the statistical variation of the count rates. Noise is given by $1/\sqrt{N}$, where N is the count density. An important parameter related to noise is the noise equivalent count rate (*NECR*) that is given by

$$NECR = \frac{T^2}{T + S + R} \quad (13.11)$$

where T , S , and R are the true, scatter, and random coincidences, respectively. It serves as a parameter to compare the performances of different PET scanners. Image noise can be minimized by maximizing $NECR$.

Quality Control Tests for PET Scanners

Daily Tests

Sinogram Check

This is normally carried out by using a standard 20-cm long cylindrical phantom containing a positron emitter (normally a 1 to 3 mCi [37–111 MBq] ^{68}Ge source). It is mounted on the patient table and centered both vertically and horizontally in the field of view. This arrangement allows uniform exposure to radiations of all detectors to provide a uniform sinogram. Initially, a reference blank sinogram is obtained during the last setup of the scanner. Subsequently, a blank sinogram is taken daily before the patient study is started on each day. This daily sinogram is then compared with the initial blank sinogram. The difference between the two sinograms is characterized by a quantity called average variance, which is calculated by the square sum of the differences of the relative detector efficiencies between the two scans weighted by the inverse variances of the differences. To obtain the average variance, the sum is then divided by the number of detectors. When this value exceeds 2.5, recalibration is needed, whereas for a value greater than 5.0, manufacturer's service is warranted (Buchert et al., 1999). Normally, in modern PET cameras, this is carried out by the manufacturer's menu-driven software.

Weekly Tests

Normalization

As mentioned before, normalization corrects for nonuniformities in the acquired PET data. It is accomplished by using a standard 20-cm long cylinder containing 1 to 3 mCi (37 to 111 MBq) ^{68}Ge activity accurately positioned at the center of the FOV. All detectors are uniformly exposed to radiations in the absence of any object in the FOV. From the acquired 2-D or 3-D data, the correction factors are calculated for each detector by dividing the average counts of all detector pairs by each individual detector pair count [i.e., along the LOR; (Eq. 13.3)]. These factors are saved and later applied to corresponding detector pairs in the acquired emission data of the patient (Eq. 13.4). Usually, normalization factors need to be taken weekly or monthly, although some manufacturers recommend quarterly. For better statistical accuracy, long hours of counting (6 to 8 hrs) are necessary, and often, overnight data acquisition is made. Data acquisition and

calculation of normalization factors and their storage are carried out by the manufacturer's menu-driven software.

Questions

1. Describe the principles of positron emission tomography.
2. How is the attenuation of annihilation photons corrected for in PET?
3. Attenuation of photons is directly proportional to the thickness and density of the material through which they pass and inversely proportional to photon energy. True or False?
4. How do random and scatter radiations affect PET images? What are the methods for correcting these effects?
5. What are the typical values of spatial resolutions of PET scanners?
6. What is the cause of the partial-volume effect and how would you rectify it?
7. What are the most common radionuclides used in PET?
8. Explain why ^{13}N -ammonia gives better spatial resolution than ^{82}Rb in the PET imaging of myocardium.
9. What are the factors that affect the spatial resolution of a PET scanner?
10. What are the initial steps that are taken in the reconstruction of 3-D data?
11. The overall sensitivity in 3-D acquisition is four to eight times higher than in 2-D acquisition. Why?
12. What is the daily quality control test for a PET scanner?
13. The sensitivity of a PET scanner increases with the size of the detector. True or False?
14. Explain why LSO detectors are preferred to BGO detectors.
15. Transverse resolution is worse at the center of the field of view than away from the center. True or False?

References and Suggested Readings

- Bacharach SL. Image analysis. In: Wagner HN Jr, Szabo Z, Buchanan JW, eds. *Principles of Nuclear Medicine*. Philadelphia: W. B. Saunders; 1995:393–404.
- Budinger TF. PET instrumentation: what are the limits? *Semin Nucl Med*. 1998; 28:247.
- Cherry SR, Dahlbom M. PET: Physics, instrumentation, and scanners. In: Phelps ME. *PET: Molecular Imaging and Its Biological Applications*. New York; Springer: 2004.
- Cherry SR, Sorensen JA, Phelps ME. *Physics in Nuclear Medicine*. 3rd ed. Philadelphia; W. B. Saunders: 2003.
- Hoffman EJ, Phelps ME. Positron emission tomography: Principles and quantitation. In: Phelps ME, Mazziotta J, Schelbert H, eds. *Positron Emission Tomogra-*

- phy and Autoradiography: Principles and Applications for the Brain and Heart.* New York: Raven; 1986:237–286.
- Keim P. An overview of PET quality assurance procedures: Part I. *J Nucl Med Technol.* 1994; 22:27–34.
- Koeppel RA, Hutchins GD. Instrumentation for positron emission tomography: Tomographs and data processing and display systems. *Semin Nucl Med.* 1992; 22:162–181.
- Saha GB. *Basics of PET Imaging.* New York; Springer: 2005.
- Tarantola G, Zito F, Gerundini P. PET instrumentation and reconstruction algorithms in whole-body applications. *J Nucl Med.* 2003; 44:756.

14

Internal Radiation Dosimetry

Radiation can cause detrimental effects on human tissues, and these effects depend on various factors, such as dose, dose rate, time of exposure and so on. This chapter describes the method of calculating absorbed doses in various organs from radionuclides ingested internally either purposely (e.g., medical procedures) or accidentally.

Radiation Units

Three units of measure are related to radiation: the roentgen (R) for exposure, the rad (radiation absorbed dose) for absorbed dose, and the rem (roentgen equivalent man) for dose equivalent.

The *roentgen* is the amount of x- or γ -radiation that produces ionization of one electrostatic unit of either positive or negative charge per cubic centimeter of air at 0°C and 760 mm Hg, standard temperature and pressure (STP). Because 1 cm³ air weighs 0.001293 g at STP and a charge of either sign carries 1.6×10^{-19} C or 4.8×10^{-10} electrostatic units, it can be shown that

$$1 \text{ R} = 2.58 \times 10^{-4} \text{ C/kg} \quad (14.1)$$

It should be noted that the roentgen applies only to air and to x- or γ -radiations. Because of practical limitations of the measuring instruments, the R unit is applicable only to photons of less than 3 MeV energy.

The *rad* is a more universal unit. It is a measure of the energy deposited per unit mass of any material by any type of radiation. The rad is specifically defined as

$$1 \text{ rad} = 100 \text{ ergs/g absorber} \quad (14.2)$$

Since 1 joule (J) = 10^7 ergs,

$$1 \text{ rad} = 10^{-2} \text{ J/kg} \quad (14.3)$$

Another radiation unit is *kerma* (acronym for kinetic energy released in matter) which is defined as the sum of initial kinetic energies of all charged particles liberated by uncharged ionizing radiation per unit mass of material. For all practical purposes, kerma and rad are identical.

In SI units, the *gray* (Gy) is the unit of radiation absorbed dose and kerma and is given by

$$1 \text{ Gy} = 100 \text{ rad} \quad (14.4)$$

$$= 1 \text{ J/kg absorber} \quad (14.5)$$

It can be shown that the energy absorbed per kilogram of air due to an exposure of 1 R is

$$1 \text{ R} = 86.9 \times 10^{-4} \text{ J/kg in air}$$

Therefore,

$$1 \text{ R} = 0.869 \text{ rad in air}$$

or,

$$1 \text{ R} = 0.00869 \text{ Gy in air}$$

The rad is not restricted by the type of radiation or absorber or by the energy or intensity of the radiation. It should be understood that the rad is independent of the weight of the material. This means that a radiation dose of 1 rad (0.01 Gy) is always 1 rad (0.01 Gy) in 1, 2, or 10 g of the material. However, the integral absorbed dose is given in units of gram-rad ($\text{g} \cdot \text{rad}$ or $\text{g} \cdot \text{Gy}$) and calculated by multiplying the rad (Gy) by the mass of material. For example, if the radiation dose to a body of 45 g is 10 rad (0.1 Gy), then the integral radiation dose to the material is $450 \text{ g} \cdot \text{rad}$ (or $4.5 \text{ g} \cdot \text{Gy}$); however, the radiation dose is still 10 rad (0.1 Gy).

The dose equivalent unit, *rem*, has been developed to account for the differences in effectiveness of different types of radiation in causing biological damage. In radiobiology, the rem is defined as

$$\text{rem} = \text{rad} \times \text{RBE} \quad (14.6)$$

where RBE is the relative biological effectiveness of the radiation. It is defined as the ratio of the dose of a standard radiation to produce a particular biological response to the dose of the radiation in question to produce the same biological response. Radiations of 250 kV x-rays are normally chosen as the standard radiation because of their widespread use. RBE varies with the linear energy transfer (LET) of the radiation, radiation dose, dose rate, and the biological system in which RBE is determined.

In radiation protection, RBE is replaced by the *radiation weighting factor*, W_r , to account for differences in effectiveness of various radiations in causing biological damage. The rem is then defined as

TABLE 14.1. Radiation weighting factors.

Type and energy range	Radiation weighting factors, W_r
Photons, all energies	1
Electrons, muons, all energies	1
Neutrons, energy <10 keV	5
10 keV to 100 keV	10
>100 keV to 2 MeV	20
>2 MeV to 20 MeV	10
>20 MeV	5
Protons, other than recoil protons, energy >2 MeV	5
Alpha particles, fission fragments, heavy nuclei	20

Adapted with permission from ICRP Publication 60: *1990 Recommendations of the International Commission on Radiological Protection*. New York: Pergamon Press; 1991.

$$\text{rem} = \text{rad} \times W_r \quad (14.7)$$

The International Commission on Radiological Protection (ICRP) has suggested the W_r values for different radiations, which are listed in Table 14.1 (ICRP 60, 1991). These values depend on the LET of the radiation. When a radiation dose comes from several radiations, the total dose equivalent is calculated by adding the absorbed doses from individual radiations multiplied by the W_r of each radiation.

In the past, the W_r values were called *quality factors*, which are somewhat different from the W_r values. The US Nuclear Regulatory Commission (NRC) still adopts these values for regulatory purposes, and the values are listed in Table 14.2.

In SI units, the dose equivalent is expressed in *sievert*, which is defined as

$$1 \text{ sievert (Sv)} = 100 \text{ rem} \quad (14.8)$$

In practical situations, all these radiation units are often expressed in miliroentgens (mR), millirads (mrad), and millirems (mrem), which are 10^{-3} times the units, roentgen, rad, and rem, respectively. In SI units, the equivalent quantities are milligrays (mGy) and millisieverts (mSv). A rad is also commonly expressed as centigray (cGy), one-hundredth of a gray.

TABLE 14.2. Quality factors for different radiations.

Type of radiation	QF
X-rays, γ -rays, β -particles	1.0
Neutrons and protons	10.0
α -Particles	20.0
Heavy ions	20.0

Dose Calculation

The radiation absorbed dose depends on a number of factors: (1) the amount of radioactivity administered; (2) the physical and biological half-lives of the radioactivity; (3) the fractional abundance of the radiation in question from the radionuclides; (4) the biodistribution of radioactivity in the body; and (5) the fraction of energy released from the source organ that is absorbed in the target volume, which is related to the shape, composition, and location of the target. The physical characteristics of a radionuclide are well established. Information concerning the biodistribution of ingested radioactivity can be obtained from various experimental studies in humans and animals. The factors 4 and 5 are variable from one individual to another and, therefore, they are approximated for a "standard" or "average" 70-kg man.

Radiopharmaceuticals administered to patients are distributed in different regions of the body. A region of interest for which the absorbed dose is to be calculated is considered the "target," whereas all other regions contributing to the radiation dose to the target are considered "sources." The source and the target become the same when the radiation dose due to the radioactivity in the target itself is calculated.

Radiation Dose Rate

Suppose a source volume r contains A μCi of a radiopharmaceutical emitting several radiations. If the i th radiation has energy E_i and a fractional abundance N_i per disintegration, then the energy absorbed per hour (dose rate) by a target of mass m and volume v from the i th radiation emitted by the source volume r is given by

$$\begin{aligned} R_i(\text{rad/hr}) &= A/m(\mu\text{Ci/g})N_iE_i(\text{MeV/disintegration}) \\ &\quad \times [3.7 \times 10^4 \text{ disintegrations}/(\text{s} \cdot \mu\text{Ci})] \\ &\quad \times (1.6 \times 10^{-6} \text{ erg/MeV}) \\ &\quad \times (0.01 \text{ g} \cdot \text{rad/erg}) \\ &\quad \times (3600 \text{ s/hr}) \\ &= 2.13(A/m)N_iE_i \end{aligned}$$

The above equation is valid for nonpenetrating radiations only, meaning all energy is absorbed in the absorber. For penetrating radiations, total or part of the radiation energy may be absorbed in the absorbing material. If the target and the source are not the same, then a factor must be introduced to account for the partial absorption, if any, of the radiation energy. Thus,

$$R_i(\text{rad/hr}) = 2.13(A/m)N_iE_i\phi_i(v \leftarrow r) \quad (14.9)$$

Here $\phi_i(v \leftarrow r)$ is called the *absorbed fraction* and is defined as the ratio of the energy absorbed by the target volume v from the i th radiation to the energy emitted by the i th radiation from the source volume r . This is a critical factor that is difficult to evaluate, because the absorbed fraction ϕ_i depends on the type and energy of the radiation, the shape and size of the source volume, and the shape, composition, and distance of the target volume. However, in the case of β -particles, conversion electrons, α -particles, and x- and γ -rays of energies less than 11 keV, all of the energy emitted by a radionuclide is absorbed in the volume r larger than 1 cm. Then, ϕ_i becomes 0, unless v and r are the same, in which case $\phi_i = 1$. For x- and γ -rays with energies greater than 11 keV, the value of ϕ_i decreases with increasing energy and varies between 0 and 1, depending on the energy. The values of ϕ_i are calculated by statistical Monte Carlo methods on the basis of fundamental mechanisms of interaction of radiation with matter, and are available in standard textbooks on radiation dosimetry, particularly the medical internal radiation dose (MIRD) pamphlets published by the Society of Nuclear Medicine.

The quantity $2.13N_iE_i$ is a constant for the i th radiation and is often denoted by Δ_i . Thus,

$$\Delta_i = 2.13N_iE_i \quad (14.10)$$

The quantity Δ_i is called the *equilibrium dose constant* for the i th radiation and has the unit $\text{g} \cdot \text{rad}/(\mu\text{Ci} \cdot \text{hr})$ based on the units chosen in Eq. (14.9). It should be pointed out that since β -particles are emitted with a distribution of energy, the average energy \bar{E}_β of β -particles is used in the calculation of Δ_i . Thus, Eq. (14.9) becomes

$$R_i(\text{rad/hr}) = (A/m)\Delta_i\phi_i(v \leftarrow r) \quad (14.11)$$

The activity A will change due to the physical decay and biological elimination of the radiopharmaceutical, and therefore the dose rate will also change. If A_o is the initial administered activity, then the activity localized in an organ is a fraction f of A_o . Assuming an effective exponential change in A with time, Eq. (14.11) can be written

$$R_i(\text{rad/hr}) = (f \cdot A_o/m)\Delta_i e^{-\lambda_e t} \phi_i(v \leftarrow r) \quad (14.12)$$

Here λ_e is the effective decay constant of the radiopharmaceutical, and t is the time over which the original activity has decayed.

Cumulative Radiation Dose

The cumulative radiation dose D_i to the target due to the i th radiation of the radionuclide during the period $t = 0$ to t can be obtained by integrating Eq. (14.12). Thus,

$$D_i(\text{rad}) = (f \cdot A_o/m)\Delta_i\phi_i(v \leftarrow r) \int_0^t e^{-\lambda_e t} dt$$

$$\begin{aligned}
 &= (f \cdot A_o / m) \Delta_i \phi_i(v \leftarrow r) \frac{1}{\lambda_e} (1 - e^{-\lambda_e t}) \\
 &= 1.44(f \cdot A_o / m) \Delta_i T_e (1 - e^{-\lambda_e t}) \phi_i(v \leftarrow r) \quad (14.13)
 \end{aligned}$$

Here, T_e is the effective half-life of the radiopharmaceutical in hours (discussed in Chapter 3). If $t = \infty$, that is, the radiopharmaceutical is completely eliminated, then the exponential term $e^{-\lambda_e t}$ approaches zero and the absorbed dose in Eq. (14.13) may be written as

$$D_i(\text{rad}) = 1.44(f \cdot A_o / m) \Delta_i T_e \phi_i(v \leftarrow r) \quad (14.14)$$

If the radionuclide has n radiations with energies E_1, E_2, \dots, E_n and fractional abundances N_1, N_2, \dots, N_n per disintegration, then the total dose D can be obtained by summing Eq. (14.14) over all radiations. Thus,

$$D_i(\text{rad}) = 1.44(f \cdot A_o / m) T_e \sum_{i=1}^n \Delta_i \phi_i(v \leftarrow r) \quad (14.15)$$

This summation can also be applied to Eq. (14.12) for the dose rate R_i . The total dose to the target from different sources of radiations can be calculated by summing Eq. (14.15) over all sources.

In the MIRD pamphlets, the values of Δ_i have been compiled on the basis of various nuclear characteristics of the radionuclide in question. The ϕ_i values have been calculated on the basis of different sizes and compositions of the targets receiving the radiation dose and the radiation characteristics of the radionuclide. In MIRD pamphlet no. 11, Eq. (14.15) has been substituted by

$$D(\text{rad}) = \tilde{A} \cdot S \quad (14.16)$$

where

$$\tilde{A} = 1.44 \times f \cdot A_o \times T_e \quad (14.17)$$

$$S = \sum_{i=1}^n \Delta_i \phi_i / m \quad (14.18)$$

The quantity \tilde{A} is called the *cumulated activity* and has the unit of $\mu\text{Ci} \cdot \text{hr}$. The quantity S is called the *mean absorbed dose per cumulated activity* and has the unit of $\text{rad}/\mu\text{Ci} \cdot \text{hr}$. These two quantities are further discussed next.

Factors Affecting \tilde{A}

The cumulated activity \tilde{A} in Eq. (14.17) is given as

$$\tilde{A} = 1.44 f \cdot A_o T_e$$

This is calculated on the assumption that the radiopharmaceutical localizes in the organs instantaneously and cleared by both physical decay and biological elimination.

There are situations when the uptake of the tracer is gradual and the clearance also is slow. In these cases,

$$\tilde{A} = 1.44f \cdot A_o T_e (T_Q/T_U) \quad (14.19)$$

where T_U is the biological uptake half-time, T_e is the effective excretion half-time [Eq. (3.12)] and T_Q is the effective uptake half-time. T_Q is calculated by Eq. (3.12) using the physical half-life T_P and the biological uptake half-time T_U .

Two other situations can occur when the uptake is instantaneous, but the T_P of the radionuclide is greater than the biological half-life T_B , or T_B is greater than T_P .

When $T_P \gg T_B$, the cumulated activity is given by

$$\tilde{A} = 1.44f \cdot A_o T_B \quad (14.20)$$

If the tracer is excreted by several excretion routes such as urinary excretion, fecal excretion, etc., the fraction of activity excreted and the effective half-time of each mode are used to calculate the fractional cumulated activity of each mode, which are then summed to calculate the total cumulated activity.

When $T_B \gg T_P$, the cumulated activity is calculated as

$$\tilde{A} = 1.44f \cdot A_o T_P \quad (14.21)$$

In this case, there is no biological excretion.

The S Values

The mean absorbed dose per cumulated activity, S , is more appropriately expressed as

$$S(v \leftarrow r) = \frac{1}{m} \sum_{i=1}^n \Delta_i \phi_i(v \leftarrow r) \quad (14.22)$$

where the symbols v and r represent the target and the source, respectively. The calculation of these values is quite laborious. The MIRD Committee of the Society of Nuclear Medicine calculates these values for radiopharmaceuticals commonly used in nuclear medicine and publish them periodically. Table 14.3 includes a partial list of S values for ^{99m}Tc obtained from MIRD pamphlet no. 11.

Problem 14.1

Calculate the absorbed dose to the lungs from the administration of 4 mCi (148 MBq) ^{99m}Tc -MAA particles, assuming that 99% of the particles are trapped in the lungs. The value of S for the lungs is 5.2×10^{-5} rad/ $\mu\text{Ci} \cdot \text{hr}$. Assume that the ^{99m}Tc activity is uniformly distributed in the lungs and 45% of the activity is cleared from the lungs with a biological half-life of 3 hr and 55% with a biological half-life of 7 hr.

TABLE 14.3. S^* , mean absorbed dose per unit cumulated activity ($\text{rad}/\mu\text{Ci}\cdot\text{hr}$)[†] for ^{99m}Tc .

Target organs	Source organs											Total body
	Bladder content	Stomach content	Kidneys	Liver	Lungs	Ovaries	Red marrow	Spleen	Testes	Thyroid	Thyroid	
Adrenals	1.5E-07	2.7E-06	1.1E-05	4.5E-06	2.7E-06	3.3E-07	2.3E-06	6.3E-06	3.2E-08	1.3E-07	1.3E-07	2.3E-06
Bladder wall	1.6E-04	2.7E-07	2.8E-07	1.6E-07	3.6E-08	7.2E-06	9.9E-07	1.2E-07	4.8E-06	2.1E-09	2.1E-09	2.3E-06
Bone (total)	9.2E-07	9.0E-07	1.4E-06	1.1E-06	1.5E-06	1.5E-06	4.0E-06	1.1E-06	9.2E-07	1.0E-06	1.0E-06	2.5E-06
Stomach	2.7E-07	1.3E-04	3.6E-06	1.9E-06	1.8E-06	8.1E-07	9.5E-07	1.0E-05	3.2E-08	4.5E-08	4.5E-08	2.2E-06
Kidneys	2.6E-07	3.5E-06	1.9E-04	3.9E-06	8.4E-07	9.2E-07	2.2E-06	9.1E-06	4.0E-08	3.4E-08	3.4E-08	2.2E-06
Liver	1.7E-07	2.0E-06	3.9E-06	4.6E-05	2.5E-06	5.4E-07	9.2E-07	9.8E-07	3.1E-08	9.3E-08	9.3E-08	2.2E-06
Lungs	2.4E-08	1.7E-06	8.5E-07	2.5E-06	5.2E-05	6.0E-08	1.2E-06	2.3E-06	6.6E-09	9.4E-07	9.4E-07	2.0E-06
Marrow (red)	2.2E-06	1.6E-06	3.8E-06	1.6E-06	1.9E-06	5.5E-06	3.1E-05	1.7E-06	7.3E-07	1.1E-06	1.1E-06	2.9E-06
Ovaries	7.3E-06	5.0E-07	1.1E-06	4.5E-07	9.4E-08	4.2E-03	3.2E-06	4.0E-07	0.0	4.9E-09	4.9E-09	2.4E-06
Skin	5.5E-07	4.4E-07	5.3E-07	4.9E-07	5.3E-07	4.1E-07	5.9E-07	4.7E-07	1.4E-06	7.3E-07	7.3E-07	1.3E-06
Spleen	6.6E-07	1.0E-05	8.6E-06	9.2E-07	2.3E-06	4.9E-07	9.2E-07	3.3E-04	1.7E-08	1.1E-07	1.1E-07	2.2E-06
Testes	4.7E-06	5.1E-08	8.8E-08	6.2E-08	7.9E-09	0.0	4.5E-07	4.8E-08	1.4E-03	5.0E-10	5.0E-10	1.7E-06
Thyroid	2.1E-09	8.7E-08	4.8E-08	1.5E-07	9.2E-07	4.9E-09	6.8E-07	8.7E-08	5.0E-10	2.3E-03	2.3E-03	1.5E-06
Total body	1.9E-06	1.9E-06	2.2E-06	2.2E-06	2.0E-06	2.6E-06	2.2E-06	2.2E-06	1.9E-06	1.8E-06	1.8E-06	2.0E-06

* Adapted by permission of the Society of Nuclear Medicine from Snyder, et al. "S" absorbed dose per unit cumulated activity for selected radionuclides and organs. MIRD pamphlet no. 11. New York: Society of Nuclear Medicine; 1975.

† Divide by 3.7 to convert to SI unit ($\text{Gy}/\text{MBq}\cdot\text{hr}$).

Answer

The half-life of ^{99m}Tc = 6 hr. The effective half-life of two biological clearances are

$$T_{e1} = \frac{3 \times 6}{3 + 6} = 2 \text{ hr}$$

$$T_{e2} = \frac{7 \times 6}{7 + 6} = 3.2 \text{ hr}$$

Using Eq. (14.17)

$$\begin{aligned} \tilde{A} &= 1.44 \times 4000 \times 0.99 \times (0.45 \times 2 + 0.55 \times 3.2) \\ &= 15,200 \mu\text{Ci} \cdot \text{hr} \quad (0.562 \text{ GBq} \cdot \text{hr}) \end{aligned}$$

Using Eq. (14.16)

$$\begin{aligned} D &= \tilde{A} \cdot S \\ &= 15200 \times 5.2 \times 10^{-5} \\ &= 0.79 \text{ rad} \\ &= 790 \text{ mrad} \quad (7.9 \text{ mGy}) \end{aligned}$$

Radiation Dose in SI Units

The radiation dose in System Internationale (SI) units due to the administration of a radiopharmaceutical can be calculated by assuming a source volume r containing A MBq of the radiopharmaceutical that emits several radiations. If the i th radiation has energy E_i and a fractional abundance N_i per disintegration, then the energy absorbed per hour by a target of mass m and volume v from the i th radiation emitted by the source volume r (dose rate) is given by

$$\begin{aligned} R_i(\text{Gy/hr}) &= A/m(\text{MBq/g})N_iE_i(\text{MeV/disintegration}) \\ &\quad \times 10^6 \text{ disintegrations}/(\text{s} \cdot \text{MBq}) \\ &\quad \times (1.6 \times 10^{-6} \text{ erg/MeV}) \\ &\quad \times (1 \times 10^{-4} \text{ g} \cdot \text{Gy/erg}) \\ &\quad \times (3600 \text{ s/hr}) \\ &= 0.576 (A/m)N_iE_i \end{aligned}$$

When the target and the source are not the same, the absorbed fraction $\phi_i(v \leftarrow r)$ must be taken into account. Thus,

$$R_i(\text{Gy/hr}) = 0.576(A/m)N_iE_i\phi_i(v \leftarrow r) \quad (14.23)$$

The quantity $0.576N_iE_i$ is a constant and can be denoted by Δ_i as in Eq. (14.10). Thus

$$\Delta_i = 0.576N_iE_i \quad (14.24)$$

With this value of Δ_i , Eqs. (14.11) to (14.18) are equally applicable to radiation doses in SI units. It should be understood that the equations in SI units contain a constant $\Delta_i = 0.576N_iE_i$ and activities expressed in MBq, whereas the equations in rad units contain the equilibrium dose constant $\Delta_i = 2.13N_iE_i$ and activities expressed in microcuries. Also note that A should be equal to $f \cdot A_o$, if A_o is the initial administered activity.

Table 14.4 presents radiation absorbed doses from various radiopharmaceuticals to different organs in adults. All values have been obtained from package inserts except those noted by footnotes.

TABLE 14.4. Radiation absorbed doses in adults for various radiopharmaceuticals.

Radiopharmaceuticals	Organ	Dose	
		rad/mCi	mGy/GBq
^{99m}Tc -pertechnetate	Thyroid	0.130	35.1
	Upper large intestine	0.120	32.4
	Lower large intestine	0.110	30.0
	Stomach	0.051	13.8
	Ovaries	0.030	8.1
	Testes	0.009	2.4
^{99m}Tc -sulfur colloid	Liver	0.335	91.2
	Spleen	0.210	57.4
	Red marrow	0.028	7.4
^{99m}Tc -DTPA	Bladder wall (2 hr void)	0.115	31.1
	Kidneys	0.090	24.3
^{99m}Tc -MAA	Gonads	0.011	3.0
	Lungs	0.220	59.5
	Kidneys	0.011	3.0
	Liver	0.018	4.9
	Ovaries	0.008	2.2
^{99m}Tc -MDP	Testes	0.006	1.6
	Bone	0.035	9.5
	Bladder wall	0.130	35.1
	Kidneys	0.040	10.8
	Red marrow	0.026	7.0
	Ovaries	0.012	3.2
^{99m}Tc -sestamibi (Cardiolite)	Testes	0.008	2.1
	Gall bladder	0.067	18.1
	Upper large intestine	0.100	27.0
	Lower large intestine	0.180	48.6
	Heart (wall)	0.017	4.6
	Kidneys	0.067	18.1
	Ovaries	0.053	14.3
	Bladder wall	0.140	37.8

TABLE 14.4. *Continued*

Radiopharmaceuticals	Organ	Dose	
		rad/mCi	mGy/GBq
^{99m}Tc -mercaptoacetylglycylglycylglycine (MAG3)	Bladder wall	0.480	129.7
	Lower large intestine	0.033	8.9
	Gall bladder	0.016	4.3
	Kidneys	0.014	3.8
	Ovaries	0.026	7.0
^{99m}Tc -HMPAO (Ceretec)	Brain	0.026	7.0
	Thyroid	0.100	27.0
	Kidneys	0.130	35.1
	Gall bladder	0.190	51.4
	Lachrymal gland	0.258	69.7
^{99m}Tc -tetrofosmin (at rest) (Myoview)	Gall bladder	0.180	48.7
	Upper large intestine	0.113	30.5
	Lower large intestine	0.082	22.2
	Heart (wall)	0.015	4.1
	Kidneys	0.046	12.4
	Ovaries	0.035	9.5
	Bladder wall	0.071	19.2
	Brain	0.020	5.4
^{99m}Tc -ECD (2 hr void) (Neurolite)	Kidneys	0.027	7.3
	Gall bladder	0.091	24.6
	Upper large intestine	0.061	16.5
	Liver	0.020	5.4
	Ovaries	0.022	6.0
	Bladder wall	0.110	29.8
	Testes	0.008	2.2
	Liver	0.047	12.7
^{99m}Tc -mebrofenin (Choletec)	Lower large intestine	0.474	128.1
	Upper large intestine	0.364	98.4
	Gall bladder	0.137	37.0
	Bladder	0.029	7.8
	Red marrow	0.034	9.1
	Ovaries	0.101	27.3
	Bladder wall	0.220	60.0
	Kidneys	0.050	14.0
^{99m}Tc -apcitide (AcuTect)	Upper large intestine	0.038	10.0
	Uterus	0.034	9.2
	Lungs	0.016	4.3
	Red marrow	0.009	2.5
	Kidneys	0.330	90.0
	Spleen	0.160	42.0
	Testes	0.110	31.0
	Red marrow	0.078	21.0
^{99m}Tc -depreotide (NeoTect)	Liver	0.078	21.0
	Bladder wall	0.033	8.9
	Bladder wall	0.070	18.9
	Kidneys	0.630	170.3
	Liver	0.031	8.56
	Red marrow	0.022	5.86
	Ovaries	0.013	3.60
	Testes	0.007	1.80
^{99m}Tc -DMSA			

TABLE 14.4. *Continued*

Radiopharmaceuticals	Organ	Dose	
		rad/mCi	mGy/GBq
¹³¹ I-iodide	Thyroid	1300.00	3.5×10^5
	Liver	0.48	130.00
	Ovaries	0.14	37.8
¹³¹ I-MIBG	Bladder wall	2.960	800.0
	Liver	2.920	789.2
	Spleen	2.180	589.2
	Heart (wall)	1.410	381.1
	Adrenal medulla	0.780	210.8
	Kidneys	0.330	89.2
	Ovaries	0.270	73.0
¹²³ I-iodide	Thyroid	13.00	3513.5
	Testes	0.02	5.4
¹¹¹ In-pentetreotide (OctreoScan)	Kidneys	1.807	488.4
	Liver	0.407	110.0
	Spleen	2.460	664.9
	Bladder wall	1.007	272.2
	Ovaries	0.163	44.1
¹¹¹ In-capromab pendetide (ProstaScint)	Liver	3.700	1000.0
	Spleen	3.260	881.1
	Kidneys	2.480	670.3
	Red marrow	0.860	232.4
	Testes	1.120	339.0
	Prostate	1.640	443.2
¹¹¹ In-WBC	Spleen	26.000	7027.0
	Liver	38.000	10270.0
	Red marrow	26.000	7027.0
	Skeleton	7.280	1967.6
	Ovaries	3.800	1027.0
⁸² Rb-RbCl	Kidneys	0.032	8.6
	Heart (wall)	0.007	1.9
²⁰¹ Tl-thallos chloride	Heart	0.500	135.1
	Kidneys	1.200	324.3
	Liver	0.550	148.6
	Thyroid	0.650	175.7
	Testes	0.500	135.1
¹⁸ F-FDG*	Brain	0.070	18.9
	Heart	0.220	59.5
	Bladder wall	0.700	189.2
	Spleen	0.140	37.8
	Ovaries	0.063	17.0
	Uterus	0.085	23.0
	Liver	0.46	124.3
⁶⁷ Ga-gallium citrate	Red marrow	0.58	156.7
	Spleen	0.53	143.2
	Upper large intestine	0.56	151.4
	Lower large intestine	0.90	243.2
	Gonads	0.26	70.2
	Bone surfaces	25.000	6756.8
	Red marrow	5.700	1540.0
¹⁵³ Sm-lexidronam (Quadramet)	Bladder wall	3.600	973.0

TABLE 14.4. *Continued*

Radiopharmaceuticals	Organ	Dose	
		rad/mCi	mGy/GBq
⁸⁹ Sr-strontium chloride (Metastron)	Kidneys	0.065	17.6
	Ovaries	0.032	8.6
	Liver	0.019	5.1
	Bone surfaces	63.0	17000.0
	Red marrow	40.7	11000.0
	Lower bowel	17.4	4700.0
	Bladder wall	4.8	1300.0
⁹⁰ Y-ibritumomab tiuxetan (Zevalin) [†]	Ovaries	2.9	800.0
	Kidneys	2.9	800.0
	Spleen	27.2	7350.0
	Liver	16.0	4320.0
	Lungs	7.6	2050.0
	Bladder wall	3.3	890.0
	Red marrow	2.2	590.0
	Kidneys	0.8	220.0
	Other organs	1.5	400.0
	¹²³ I-tositumomab (Bexxar)	Thyroid	10.027
Kidneys		7.252	1960.0
Upper large intestine		4.958	1340.0
Heart wall		4.625	1250.0
Red marrow		2.405	650.0
Bladder wall		2.368	640.0
Ovaries		0.925	250.0
Brain		0.481	130.0
Total Body		0.888	240.0
¹³³ Xe-xenon		Lungs	0.008

* From Stabin MG, Stubbs JB, Toohey RE. Radiation dose estimates for radiopharmaceuticals. Radiation Internal Dose Information Center, Oak Ridge Institute for Science and Foundation, 1996.

[†] From Wiseman GA, Kormehl E, Leigh B, et al. Radiation dosimetry results and safety correlations from ⁹⁰Y-ibritumomab tiuxetan radioimmunotherapy for relapsed or refractory non-Hodgkin's lymphoma. Combined data from four (4) clinical trials. *J Nucl Med* 2003; 44:465.

Effective Dose Equivalent and Effective Dose

Historically, the whole-body dose or total body dose was used to evaluate the relative radiation risks of different procedures involving radiations. This quantity is calculated according to the MIRD method by using the *S* factor for the whole body as the source organ as well as the target organ. This value does not take into consideration the effect of tissue sensitivity to radiation.

In 1977 the ICRP introduced the concept of effective dose equivalent (EDE) to take into account the different sensitivity of tissues to radiation (ICRP 26). The tissues weighting factor (W_T) for an organ was defined as the ratio of the whole-body dose, which would cause a certain probability of cancer induction to the absorbed dose in that organ which would cause

the same probability of cancer induction in that organ. For example, a dose of 3 rem to the whole body causes some probability of cancer induction; a dose of 100 rem to the thyroid causes the same numerical probability of thyroid cancer induction. Then the W_T for thyroid is equal to 0.03.

The *effective dose equivalent* (H_E) is defined as the sum of weighted dose equivalents in all tissues and organs, and is calculated as

$$H_E = \sum_T W_T H_T \quad (14.25)$$

where W_T is the tissue weighting factor for an organ and H_T is the dose equivalent (rem) to the organ. H_E can be explicitly written as

$$H_E = \sum_T W_T \sum_r W_r \times (\text{rad})_{T,r} \quad (14.26)$$

where $(\text{rad})_{T,r}$ is the absorbed dose to tissue T from radiation of type r and W_r is the radiation weighting factor discussed earlier.

The effective dose equivalent provides an overall risk estimate for an individual exposed to radiation, which is computed from dose equivalent to each organ that is weighted for tissue sensitivity. For assessment of risk versus benefit, the effective dose equivalent is a more appropriate parameter than the whole-body dose, because it takes into consideration the different tissue sensitivities of the organ. The W_T values are assigned such that their sum equals one. These W_T values from ICRP 26 have been adopted by the NRC (10CFR20).

In 1990, ICRP adopted a different set of W_T values and renamed the effective dose equivalent as simply the effective dose (ED) (ICRP 60). Table 14.5 summarizes the W_T values recommended by ICRP for both EDE

TABLE 14.5. Tissue weighting factors W_T .

Tissue	W_T for EDE*	W_T for ED†
Gonads	0.25	0.20
Breast	0.15	0.05
Thyroid	0.03	0.05
Bone surfaces	0.03	0.01
Bone marrow (red)	0.12	0.12
Lung	0.12	0.12
Colon	not given	0.12
Stomach	not given	0.12
Bladder	not given	0.05
Liver	not given	0.05
Esophagus	not given	0.05
Skin	not given	0.01
Remainder	0.3	0.05
Total Body	1.0	1.0

* W_T values from 10CFR20. NRC adopted these values from ICRP 26 (1977).

† Adapted with permission from ICRP 60. *1990 Recommendations of International Commission of Radiological Protection*. New York: Elsevier, 1991.

TABLE 14.6. Effective doses from various radiopharmaceuticals in nuclear medicine.

Radiopharmaceuticals	Effective dose*	
	rem/mCi	mSv/MBq
^{99m} Tc-pertechnetate	0.048	0.013
^{99m} Tc-sestamibi (exercise)	0.030	0.008
^{99m} Tc-MAA	0.004	0.001
^{99m} Tc-tetrofosmin (exercise)	0.026	0.007
^{99m} Tc-DTPA aerosol	0.022	0.006
^{99m} Tc-MDP	0.022	0.006
^{99m} Tc-red blood cell (RBC)	0.026	0.007
^{99m} Tc-iminodiacetic acid (IDA) derivatives	0.063	0.017
^{99m} Tc-DTPA	0.019	0.005
^{99m} Tc-dimercaptosuccinic acid (DMSA)	0.033	0.009
^{99m} Tc-sulfur colloid	0.033	0.009
^{99m} Tc-white blood cell (WBC)	0.004	0.001
^{99m} Tc-HMPAO	0.033	0.009
^{99m} Tc-ECD	0.041	0.011
^{99m} Tc-glucoheptonate	0.019	0.005
^{99m} Tc-MAG3	0.026	0.007
¹¹¹ In-WBC	0.133	0.036
¹¹¹ In-DTPA	0.078	0.021
¹¹¹ In-pentetreotide	0.185	0.050
¹²³ I-NaI (35% uptake)	0.814	0.220
¹³¹ I-NaI (35% uptake)	88.80	24.00
²⁰¹ Tl-TlCl	0.814	0.22
¹⁸ F-FDG	0.070	0.019
⁶⁷ Ga-citrate	0.370	0.100
¹³¹ I-MIBG	0.052	0.014
⁸² Rb-RbCl	0.013	0.003

* Adapted with permission from ICRP publication no. 80. New York: Pergamon Press; 1999.

(ICRP 26) and ED (ICRP 60). Some W_T values are different and some are the same in the two schema, whereas others are not given in the EDE scheme. Because the radiosensitivity of tissues varies with age, the effective dose is age dependent. Table 14.6 lists the effective doses in adult humans from different nuclear medicine studies using various radiopharmaceuticals (ICRP 80, 1999).

Pediatric Dosages

The metabolism, biodistribution, and excretion of drugs are different in children from those in adults, and therefore radiopharmaceutical dosages for children must be adjusted. Several methods and formulas have been reported on pediatric dosage calculations based on body weight, body

TABLE 14.7. Fraction of adult administered dosages for pediatric administration.

Weight in kg (lb)	Fraction	Weight in kg (lb)	Fraction
3 (6.6)	0.10	28 (61.6)	0.58
4 (8.8)	0.14	30 (66.0)	0.62
8 (17.6)	0.23	32 (70.4)	0.65
10 (22.0)	0.27	34 (74.8)	0.68
12 (26.4)	0.32	36 (79.2)	0.71
14 (30.8)	0.36	38 (83.6)	0.73
16 (35.2)	0.40	40 (88.0)	0.76
18 (39.6)	0.44	42 (92.4)	0.78
20 (44.0)	0.46	44 (96.8)	0.80
22 (48.4)	0.50	46 (101.2)	0.83
24 (52.8)	0.53	48 (105.6)	0.85
26 (57.2)	0.56	50 (110.0)	0.88

Adapted from Paediatric Task Group European Association Nuclear Medicine Members. A radiopharmaceuticals schedule for imaging paediatrics. *Eur J Nucl Med.* 1990;17:127.

surface area, combination of weight and area, and simple ratios of adult dosages. The calculation based on body surface area is more accurate for pediatric dosages. The body surface area of a standard adult is 1.73 m^2 and proportional to the 0.7 power of the body weight. Based on the information, the Paediatric Task Group European Association Nuclear Medicine Members published the fractions of the adult dosages needed for children, which are shown in Table 14.7. For most nuclear studies, however, there is a minimum dosage required for a meaningful scan, which is normally established in each institution based on experience.

Questions

1. Calculate the absorbed dose to the thyroid gland of a hyperthyroid patient from a dosage of $30\text{ mCi }^{131}\text{I}$, assuming 60% uptake, a biological half-life of 4 days for thyroid clearance of ^{131}I , and S equal to $2.2 \times 10^{-2}\text{ rad}/\mu\text{Ci}\cdot\text{hr}$.
2. Calculate the dose in rems and sieverts to a tumor that received 35 rads (0.35 Gy) from neutron therapy (radiation weighting factor = 10 for neutrons).
3. What is the difference between the effective dose equivalent and effective dose?
4. Identify as true or false if the following affect the absorbed fraction of a γ -emitting radionuclide:
 - (a) γ -ray energy
 - (b) Shape of the target organ

- (c) Composition of the target organ
 - (d) Amount of the radioactivity present in the source
 - (e) Shape of the source organ
5. Does the mean absorbed dose per cumulated activity, S , depend on:
 - (a) Absorbed fraction
 - (b) Target mass
 - (c) Photon energy
 - (d) Photon abundance
 6. What is the important parameter that is considered in adjusting the activity to be administered to children compared to adults for a nuclear medicine test?
 7. Calculate the cumulated activity \tilde{A} in a 55-g source organ containing 3 mCi (111 MBq) of ^{99m}Tc ($t_{1/2} = 6$ hr) with a biological $t_{1/2} = 14$ hr.
 8. A target organ has a mass of 35 g and contains 1 mCi (37 MBq) of a radionuclide emitting a β^- -particle with $\Delta_1 = 0.3 \text{ g}\cdot\text{rad}/\mu\text{Ci}\cdot\text{hr}$ and $\phi_1 = 1.0$, and a γ -radiation with $\Delta_2 = 0.2 \text{ g}\cdot\text{rad}/\mu\text{Ci}\cdot\text{hr}$ and $\phi_2 = 0.35$. Calculate the mean absorbed dose per cumulated activity.
 9. An external beam deposits 360 ergs of energy in 3 g of tissue. What is the radiation dose in rad and cGy?
 10. Explain why the concept of effective dose was introduced.
 11. The absorbed doses are 10 rad (10 cGy) to organ A , 5 rad (5 cGy) to organ B , and 6 rad (6 cGy) to organ C from each radiation from a 20 mCi (370 MBq) source containing two radiations having radiation weighting factors, W_r , as 1 and 10. The tissue weighting factors of organs A , B , and C are 0.30, 0.22, and 0.46, respectively. Considering the contribution from other organs negligible, calculate the effective dose.

References and Suggested Readings

- Federal Register. *Code of Federal Regulations*. 10CFR20. Washington, DC: US Government Printing Office; 1996.
- Fourth International Pharmaceutical Dosimetry Symposium, CONF-85113, Oak Ridge, Tenn.; November, 1985.
- International Commission on Radiation Protection. 1990 Recommendations of the International Commission on Radiological Protection; *ICRP 60*. New York: Pergamon Press; 1991.
- International Commission on Radiation Protection. Radiation Doses to Patients from Radiopharmaceuticals. *ICRP 80*. New York: Pergamon Press; 1999.
- International Commission on Radiological Protection. *Radiation Dose to Patients from Radiopharmaceuticals ICRP 53*. New York: Pergamon Press; 1988.
- Kereiakes JG, Rosenstein M. *Handbook of Radiation Doses in Nuclear Medicine and Diagnostic X-ray*. Boca Raton, FL: CRC Press; 1980.

- Snyder WS, et al. "*S*" absorbed dose per unit cumulated activity for selected radionuclides and organs. MIRD pamphlet no. 11. New York: Society of Nuclear Medicine; 1975.
- Snyder WS, Ford MR, Warner GG. *Specific absorbed fractions for radiation sources uniformly distributed in various organs of a heterogeneous phantom*. MIRD pamphlet no. 12. New York: Society of Nuclear Medicine; 1977.

15

Radiation Biology

The subject of radiation biology deals with the effects of ionizing radiations on living systems. During the passage through living matter, radiation loses energy by interaction with atoms and molecules of the matter, thereby causing ionization and excitation. The ultimate effect is the alteration of the living cells. Radiation biology is a vast subject, and it is beyond the scope of this book to include the full details of the subject. The following is only a brief outline of radiation biology, highlighting the mechanism of radiation damage, radiosensitivity of tissues, different types of effect on living matter, and risks of cancer and genetic effects from radiation exposure.

The Cell

The cell is the building unit of living matter and consists of two primary components: the nucleus and the cytoplasm (Fig. 15.1). All metabolic activities are carried out in the cytoplasm under the guidance of the nucleus.

The nucleus contains chromosomes, which have a threadlike structure of two arms connected by a centromere (Fig. 15.2). Chromosomes are formed of genes, which are the basic units of heredity in the cells of all living species. Genes are composed of deoxyribonucleic acid (DNA) molecules. The structural relationship of DNA molecules, genes, and chromosomes is shown in Figure 15.2. The sequence of genes in the chromosome characterizes a particular chromosome. Two categories of cells—namely, germ cells (reproductive cells such as oocytes and spermatozoa) and somatic cells (all other cells)—are based on the number of chromosomes they contain. Whereas germ cells contain n number of individual chromosomes, somatic cells contain $2n$ number of chromosomes in pairs, where n varies with species of the animal. In humans, n is equal to 23; therefore, there are 23 chromosomes in germ cells and 46 chromosomes in somatic cells.

In the cytoplasm of the cell exist four important organelles—ribosomes, endoplasmic reticula, mitochondria, and lysosomes—that carry out the




FIG. 15.1. Structure of a typical mammalian cell.




FIG. 15.2. Structural relationship of chromosomes, genes, and DNA molecules.

cellular metabolic activities. Ribosomes are made up of protein and ribonucleic acid (RNA) and are responsible for protein synthesis in living matter. Endoplasmic reticula are tubular structures mostly responsible for protein synthesis. Mitochondria are ellipsoidal structures with a central cavity and contain specific enzymes to oxidize carbohydrate and lipid to produce energy. Lysosomes are small organelles in the cytoplasm that contain enzymes capable of lysing many nutrients and cells.

The entire cytoplasm is enclosed within a cell membrane made of lipids and proteins. Its primary function is to selectively prohibit or permit the passage of substances into and out of the cell.

The growth of living matter is caused by proliferation of cells by cell division—a process in which a cell divides into two cells. The cell division of somatic cells is called *mitosis* and that of germ cells is called *meiosis*. Both mitosis and meiosis, designated as M, consist of four phases: *prophase*, *metaphase*, *anaphase*, and *telophase*. Each of these phases involves the rearrangement of the number of chromosomes and represents the progression of cell division (Fig. 15.3) and is described below.

In prophase, the chromosome thickens in the shape of dumbbell with a constriction at the center, called centromere. The nuclear membrane breaks open, leading to the mixture of cytoplasm and nuclear material, and spindles made of fibers are formed extending from one end (pole) of the cell to the other. Next in the metaphase, the chromosomes move to and line up at the central (or equatorial) plane of the cell, and the centromeres divide into two, each attaching to the spindle. Anaphase then follows and two chromatids move to the two poles of the cell. The last step of cell division involves the deconvolution of the chromosomes leading to the regeneration of the nuclear membrane and nucleoli around both poles. Division of cytoplasm (cytokinesis) sets in, and ultimately two daughter cells are formed.

Before cell division, each cell undergoes a long period, termed *interphase*, in which DNA molecules are synthesized. In DNA synthesis, two new DNA molecules are produced from each DNA molecule, which are exact replicas of the original DNA molecule. This period of DNA synthesis is designated the “S” phase, which takes place around the middle of the interphase. The period between the telophase and the S phase is termed G_1 , and the period between the S phase and the prophase is termed G_2 (Fig. 15.4). During the G_1 and G_2 periods, no functional activity related to cell division occurs. The period of the entire cell cycle including the M and S phases varies with the types of cells. The S phase normally is the longest and G_1 is the most variable phase in the cell cycle. The duplicate DNA molecules lead to two identical chromosomes during mitosis, which are termed sister chromatids.

One important difference between mitosis and meiosis is that in meiosis, for a given series of cell division, every alternate cell division skips DNA synthesis, thus keeping the number of chromosomes the same in germ cells.





FIG. 15.3. Different phases of mitosis. See text for details.

FIG. 15.4. The cell cycle. S is the DNA synthesis phase. M is the period of mitosis during which the prophase, metaphase, anaphase, and telophase take place. G_1 is the period between the telophase and S, and G_2 is the period between S and the prophase.



Effects of Radiation

DNA Molecule

The nucleus of the cell is the part most sensitive to radiation and this sensitivity has been attributed to the DNA molecule. To understand the effect of radiation on the DNA molecule, a knowledge of its structure is essential. It has a double-helical structure consisting of two strands, which are like the two rails of a ladder (Fig. 15.5A). The strands are composed of sugars interlinked by phosphate bonds. The two strands are connected to each other by rungs made of four bases: thymine (T), adenine (A), guanine (G), and cytosine (C) (Fig. 15.5B). The bases are bonded to the sugar mole-




FIG. 15.5. **A.** Double-helical structure of DNA molecule composed of four bases: adenine (A), guanine (G), thymine (T), and cytosine (C). **B.** Configuration of a DNA molecule: strands are formed by sugar molecules bonded by phosphate groups. The rungs of the ladderlike structure are formed by bases connected to each other by the hydrogen bond (dashed line) and to the sugar molecule on the strands on both sides.

cule on the strands on both sides, and are paired to each other by hydrogen bonds. These four bases are arranged in a very specific manner to form a specific gene in every living species and provide the unique characteristics to these species.

Radiation damage to the DNA molecule can be due to

- (a) Loss of a base
- (b) Cleavage of the hydrogen bond between bases
- (c) Breakage of one strand of the DNA molecule (single strand)
- (d) Breakage of both strands of the DNA molecule (double strand)

These radiation effects on DNA molecules are illustrated in Figure 15.6. These changes result in so-called *mutations*, which have adverse effects on the genetic codes. The number of mutations increases with increasing radiation exposure. At low-dose exposures, the breaks are single stranded and can be repaired by joining the broken components in the original order. At higher exposures, however, double strand breaks occur and the odds for repair decrease. Also, high-LET radiations cause more damage to the DNA molecule because of the double strand breaks. If the cell is not repaired, it may suffer a minor functional impairment or a major consequence (cell death). If DNA damage occurs in germ cells, future offspring may be affected.

Chromosome

Chromosomes are likely to be affected by mutations of the DNA molecules. However, chromosomes themselves can be cleaved by radiation producing single or double breaks in the arms. These structural changes are called *aberrations*, anomalies, or lesions. These aberrations are categorized as chromatid aberrations and chromosome aberrations. In chromatid aberrations, irradiation occurs after DNA synthesis prior to mitosis and thus only one chromatid will be affected. On the other hand, in chromosome aberrations, irradiation occurs after mitosis prior to DNA synthesis and hence the broken chromatids will be duplicated producing daughter cells with damaged chromosomes.

Whether chromosome aberrations are induced by single-strand breaks or double-strand breaks in the structure determines the fate of the cell. In single-strand breaks, the chromosome tends to repair by joining the two fragments in a process called *restitution*, provided sufficient time is allowed. The cell becomes functionally normal and replicates normally (Fig. 15.7A). However, if the fragments are replicated during DNA synthesis prior to restitution, two strands with centromeres and two strands without centromeres will be produced. Random combination of these fragments will then produce acentric and dicentric chromatids as illustrated in Figure 15.7B. Such chromosomes suffer severe consequences due to the mismatch of genetic information.




FIG. 15.6. Illustration of radiation effects on DNA molecules: **(A)** Normal DNA molecule; **(B)** hydrogen bond is broken without the loss of the base; **(C)** hydrogen bond is broken with the loss of the base T; **(D)** single strand break that can repair; **(E)** double strand breaks which are well separated and can repair; **(F)** double strand breaks that are too close to repair.




FIG. 15.7. **(A)** Illustration of restitution, in which the fragments produced by a single-strand break in one arm of the chromosome by radiation join together to produce the original chromosome. **(B)** Formation of dicentric and acentric chromosomes by combination of the fragments, after replication from a single-strand break in a chromosome.

If radiation produces single-strand breaks in two separate chromosomes, then there are four ways of recombining the broken ends as shown in Figure 15.8. The dicentric and acentric combinations (Fig. 15.8A) are similar to those formed after replication of single strands in the same chromosome shown in Figure 15.7B. However, these cells suffer severe consequences because of the mismatch of genetic information from two separate damaged chromosomes. The translocation is a process in which two fragments—one with a centromere from one chromosome and one without a centromere from another chromosome—combine to form a new chromosome (Fig. 15.8B). In another scenario, radiation can cause two breaks in one arm of a chromosome, resulting in three fragments, only two of which combine with the loss of the third. Such a process is called deletion (Fig. 15.9A). Translocation and deletion, although not as harmful to the cell, cause late effects such as carcinogenesis and hereditary effects due to mismatch or loss of genetic material. An alternative to deletion is the combination of all three fragments into a chromosome with changes along the broken line as shown in Figure 15.9B. This process is called inversion, which has all the original genetic material except a change in the sequence of genes and hence is not as detrimental to the cell.

Repair of chromosomes after irradiation depends on the sites of break in the DNA molecule or the chromosome, the total radiation dose, the dose rate, and the LET of the radiation. Chromosome aberrations by double-strand breaks occur more frequently at high-dose rates than at low-dose rates because of less time to repair and fewer chances of combining




FIG. 15.8. Single-strand breaks in one arm of each of two separate chromosomes. Combination of these four fragments leads to dicentric and acentric chromosomes (A) or translocation (B).




FIG. 15.9. Two breaks in one arm of a chromosome producing three fragments. **(A)** In deletion, two of the fragments combine with the loss of the third, or **(B)** in inversion, all three fragments combine with the interchange of positions.

two fragments in correct sequence of genes. High-LET radiations cause more double-strand breaks in chromosomes than low-LET radiations, and thus repair becomes difficult. For example, α -particles, protons, and neutrons will cause more chromosome aberrations than γ -rays.

Direct and Indirect Actions of Radiation

The DNA molecule of a cell is the most sensitive target to radiation. Radiation damage to the cell can be caused by the direct or indirect action of radiation on the DNA molecules. In the direct action, the radiation hits the DNA molecule directly, disrupting the molecular structure (Fig. 15.10). The structural change leads to cell damage or even cell death. Damaged cells that survive may later induce carcinogenesis or other abnormalities. This process becomes predominant with high-LET radiations such as α -particles and neutrons, and high radiation doses.





FIG. 15.10. Illustration of direct and indirect action of radiation on the DNA molecule. In direct action, radiations hit the DNA structure directly, whereas in indirect action, radiations produce free radicals in the cytoplasm, which react adversely with the DNA molecule to cause structural damage.

In the indirect action, the radiation hits the water molecules, the major constituent of the cell, and other organic molecules in the cell, whereby free radicals such as perhydroxyl ($\text{HO}_2\bullet$) and alkoxy ($\text{RO}_2\bullet$) are produced. A variety of reactions that can occur after radiation interacts with water molecules is shown below.



Free radicals are characterized by an unpaired electron in the structure, which is very reactive, and therefore reacts with DNA molecules to cause a molecular structural damage (Fig. 15.10). Hydrogen peroxide, H_2O_2 , is also toxic to the DNA molecule. The result of indirect action of radiation on DNA molecules is the impairment of function or death of the cell. The number of free radicals produced by ionizing radiation depends on the total dose but not on the dose rate. It has been found that the majority of radiation-induced damage results from the indirect action mechanism because water constitutes nearly 70% of the composition of the cell.

Radiosensitivity of Cells

In living matter, there are two types of cells: differentiated and undifferentiated. Undifferentiated cells do not have any specific physiologic function except to develop into mature cells. They undergo mitosis and serve as the precursors for mature cells. In contrast, all mature cells are differentiated and perform specific functions in the living body. For example, red blood cells (RBCs) are mature and differentiated cells performing the function of oxygen carriers, whereas erythroblasts are undifferentiated cells that develop into RBCs through mitosis.

According to the law of Bergonié and Tribondeau, undifferentiated cells that are undergoing active mitosis are most sensitive to radiation, and differentiated or mature cells are least affected by radiation. For example, in a sample of mixed RBCs, erythroblasts are most damaged and mature RBCs are least affected by radiation. Undifferentiated cells that are killed by radiation may be replaced by new cells, but those that survive with defective DNAs can induce late effects, such as cancer (see later). In contrast, the S phase of DNA synthesis in the cell cycle is least radiosensitive. Radiosensitivity is best assessed by cell death. For differentiated cells, it means loss of cellular function, whereas for undifferentiated cells it means loss of reproductivity.

Groups of cells and their relative radiosensitivity are listed in Table 15.1. As can be seen, lymphocytes, though mature cells, are most sensitive to radiation, owing to a large nucleus; nuclear material is more radiosensitive. Nerve cells and muscle cells are totally differentiated cells and therefore

TABLE 15.1. Different types of cells and their radiosensitivity.

Types of cells*		Radiosensitivity
VIM	Mature lymphocytes	Highly sensitive
	Erythroblasts	
	Spermatogonia	
DIM	Myelocytes	Relatively sensitive
	Intestinal crypt cells	
	Basal cells of epidermis	
MCT	Osteoblasts	Intermediate sensitivity
	Spermatocytes	
	Chondroblasts	
	Endothelial cells	
RPM	Spermatozoa	Relatively resistant
	Granulocytes	
	Erythrocytes	
	Osteocytes	
FPM	Nerve cells	Highly resistant
	Muscle cells	
	Fibrocytes	

Adapted from Casarett AP. *Radiation Biology*. Englewood Cliffs, NJ: Prentice-Hall; 1968:168–169.

* VIM—vegetative intermitotic; DIM—Differentiating intermitotic; MCT—multipotential connective tissue; RPM—reverting postmitotic; FPM—fixed postmitotic.

are highly resistant to radiation. The tissue or organ that contains more radiosensitive cells will be highly radiosensitive and vice versa. For example, bone marrow containing radiosensitive erythroblasts is very radiosensitive, whereas nerves and muscles containing radioresistant cells are less radiosensitive. Following irradiation of blood, depressed blood counts are observed as follows: lymphocytes on the same day, granulocytes in 3 days, platelets in 6 days, and RBCs in 10 days.

Cell Survival Curves

When mammalian cells are irradiated, not all cells are affected to the same extent. Different factors such as the total dose, the dose rate, the LET of the radiation, the particular stage of the cell cycle (M, G₁, S, or G₂) and the type of cell will affect the radiation-induced damage. Some cells may die and some will survive. The cellular response to radiation is illustrated by what is called the *cell survival curve*. It is obtained by plotting the dose along the linear *X*-axis and the surviving fraction along the logarithmic *Y*-axis. Surviving cells are those cells that retain all reproductive as well as functional activities after irradiation, whereas the death of cells is indicated by the loss of their function in differentiated cells and by the loss of repro-

ductive activity in undifferentiated cells. It should be noted that thousands of grays are needed to kill differentiated cells, whereas only hundreds of grays are needed for undifferentiated cells.

Typical cell survival curves are shown in Figure 15.11. For high-LET radiations such as α -particles and low-energy neutrons, the survival curves are nearly a straight line starting from the lowest doses. In contrast, for low-LET radiations (e.g., x- and γ -radiations), the survival curve exhibits an initial shoulder, followed by a straight line. This straight line portion on the semilog plot is an exponential curve on a linear plot. This curve based on a *multitarget model* is characterized by three parameters: D_0 (dose at which 37% of cells survive), the extrapolation number n , and the quasithreshold dose D_q , and they are related by the expression

$$\log_e n = D_q/D_0 \quad (15.1)$$

The quasithreshold dose, D_q , is the dose given by the width of the shoulder of the curve. The D_0 indicates that, at low doses, almost all cells repair




FIG. 15.11. Typical cell survival curves. The cell survival curve for low linear energy transfer (LET) radiations shows a shoulder of width D_q , which is called the quasithreshold dose. After D_q , the plot becomes linear on a semilog scale, indicating an exponential dose-response relationship. The extrapolation number n is obtained by extrapolating the linear portion of the curve back to the ordinate. D_0 is the dose obtained from the slope of the linear portion of the curve, at which 37% of the cells survive. The survival curve for high-LET radiations shows no or little shoulder, indicating D_q to be zero and n to be unity.

after irradiation, and cell killing is minimal, which is due to very limited radiation damage to the cell.

D_0 is determined from the slope of the straight line portion of the survival curve. It is the dose that kills 63% of the total number of cells. The value of D_0 is a measure of radiosensitivity of a given type of cell. For example, a large value of D_0 for a type of cell means that the cells are less radiosensitive and vice versa.

The extrapolation number n is obtained by extrapolating the straight line portion of the survival curve back to the Y -axis. Its value depends on the width of the shoulder of the survival curve, that is, the quasithreshold value, D_q . Its value for mammalian cells varies between 1 and 10.

Although Eq. (15.1) has some merit in expressing cell killing by radiation, the *linear-quadratic model* provides a more accurate description of the radiation-induced cell killing. This model is mathematically expressed as

$$S = e^{-\alpha D - \beta D^2} \tag{15.2}$$

where S is the survival fraction of the cells irradiated with dose D and α and β are constants. For low-LET radiations, βD^2 is negligible at low doses, and the cell survival is proportional to the dose only, making the survival curve linear (Fig. 15.12). At higher doses, the cell survival is proportional to the square of the dose, and the curve tends to bend becoming concave downward (Fig. 15.12). For high-LET radiations, β is zero, and so the survival curve becomes linear.




FIG. 15.12. Cell survival curves based on linear-quadratic model. The initial slope of the linear-logarithmic plot gives cell killing proportional to $e^{-\alpha D}$ and the latter part to $e^{-\beta D^2}$ which bends at higher doses. With high-LET radiations, β is zero, and the curve is exponentially expressed simply by $e^{-\alpha D}$.

Factors Affecting Radiosensitivity

As already mentioned, various factors affect the radiation damage in the cell and hence the survival curve. The dose rate, the LET of the radiation, the presence of chemical molecules and the stage of the cell cycle all affect the survival curve.

Dose Rate

The dose rate, that is, the delivery of dose per unit time, is an important factor in cellular damage. The higher the rate of dose delivery, the greater will be the cell damage. At low-dose rates, only single-strand breaks of DNA molecules occur, and so cells have time to repair, whereas at high-dose rates double-strand breaks occur, and so repair is less likely to occur because of the shorter time available to the cells between ionizing events. Figure 15.13 illustrates the effects of two dose rates on the cell survival curve. The dose-rate effect is very important in radiation therapy, because unless an appropriate dose rate is prescribed, intended therapeutic effect may not be achieved. When a total dose is given to a patient in fractions over a period of time, it should be kept in mind that the interval between fractional doses should be short enough to keep repair of damage to abnormal cells to a minimum.




FIG. 15.13. The cell survival curves indicating the effect of dose rates. At high dose rates, the shoulder of the curve is reduced, with smaller values of D_q . The opposite is true at low dose rates.

Linear Energy Transfer

High-LET radiations do not exhibit a dose-rate effect on the survival curve. Also at high-dose rates (above 100 rad/min) of low- and moderate-LET radiations, no dose-rate effects are observed on the survival curve in contrast to low-dose rates. Thus, high-LET radiations exhibit no shoulder (i.e., no D_q) on the survival curve resulting in an extrapolation number of 1. High-LET radiations are densely ionizing radiations causing more double-strand breaks in the DNA molecules, and thus leading to more cell deaths than low-LET radiations, which are sparsely ionizing radiations. Radiation damage by high-LET radiations is so severe that the chances of repair are minimal, and even if repair takes place, the cell is likely to be defective.

Chemicals

Several chemicals, if present during irradiation, have been found to augment or diminish the effects of radiation on cells. Agents that enhance the cell response to radiation are called *radiosensitizers*, and those that protect cells from radiation-induced damage are called *radioprotectors*.

Radiosensitizers

Oxygen

Oxygen is the best-known sensitizer encountered in radiation biology. It has been found that hypoxic cells are very resistant to radiation, whereas oxygenated cells are highly radiosensitive. Such radiosensitization by oxygen is called the *oxygen effect* and is measured by a quantity called the *oxygen enhancement ratio* (OER). The OER is given by the ratio of the dose required to produce a given radiation damage to cells in the absence of oxygen to that required to produce the same damage in the presence of oxygen. The oxygen effect occurs only when oxygen is administered simultaneously with radiation. It increases with O_2 tension up to 30 mm Hg, and remains constant at higher O_2 tension. For mammalian cells, the oxygen concentration required to produce a radiation response midway between hypoxic and aerobic conditions is approximately 0.5%. The OER value reaches a maximum of 3.0 for x- and γ -radiations, whereas it is about unity for high-LET radiations such as α -particles.

Figure 15.14 illustrates the effects of oxygen on the survival curve. The presence of oxygen makes the curve much steeper, indicating the augmentation of cellular damage at smaller doses relative to the situation of no oxygen. The mechanism of the oxygen effect is not clearly understood but is most likely related to DNA strand breaks. It has been postulated, however, that oxygen combines with already formed free radicals, $R\bullet$, to produce peroxidyl group $RO_2\bullet$, which is more damaging to the DNA molecules. While normally $R\bullet$ could recombine with complementary molecu-




FIG. 15.14. The cell survival curve illustrating the effect of oxygen. In the presence of oxygen, the curve becomes steeper, indicating effective killing of the cells by radiation.

lar components to repair the cell, RO_2^{\cdot} is an altered chemical entity and cannot help in cell repair. The oxygen effect is most predominant for γ - and x-rays, and is practically absent for high-LET radiations (e.g., α -particles). Because tumor cells are hypoxic, treatment of tumors with radiation under high-oxygen pressure has been advocated.

It has been found experimentally that the proportion of hypoxic cells in a tumor remains the same before and after fractionated radiation therapy. Logically, radiotherapy should have killed more oxygenated cells and thus raised the proportion of hypoxic cells. Instead, it remains the same and has brought in the argument of reoxygenation of the tumor cells during fractional radiation therapy, provided sufficient time is allowed for this to happen. This phenomenon has an important implication in radiation therapy in that even though the proportion of hypoxic cells remains the same, the total number of hypoxic tumor cells will be killed by radiation over time, thus leading to a successful treatment. The degree of reoxygenation varies with tumor types. The mechanism of reoxygenation has been attributed to the fact that as the tumor shrinks in size, surviving cells that were previously deprived of oxygen diffusion due to distal location

of the blood vessels find themselves closer to the blood supply and so reoxygenate.

Pyrimidines

Halogenated pyrimidines such as 5-chlorodeoxyuridine (CIUDR), 5-bromodeoxyuridine (BUDR), and 5-iododeoxyuridine (IUDR) are useful radiosensitizers. When cells are treated with these drugs for several days before irradiation with x- or γ -rays, cells become highly sensitive to radiation. Potentiation of radiosensitivity is due to the fact that these drugs are similar to the DNA precursor thymidine, and therefore are incorporated into the DNA molecule, making them more susceptible to damage by radiation. For optimal therapeutic gain in radiotherapy, patients should be treated for a period of time extending over several cell cycles to maximize drug incorporation into the cells.

Others

Radiosensitizers such as actinomycin D, puromycin, methotrexate, and 5-fluorouracil have been successfully used in combination with radiation to treat cancer. Whether these agents truly increase radiosensitivity or are simply toxic to the cells is still not clear.

Investigators have been trying to explore radiosensitizing chemicals to substitute for oxygen that requires the use of a high-pressure technique. Metronidazole (Flagyl), having a structure with high electron affinity, is a good radiosensitizer for hypoxic cells. Another useful radiosensitizer for hypoxic cells is misonidazole, which also has high electron affinity. Misonidazole is almost ten times more effective than metronidazole in sensitizing hypoxic cells. However, clinical trials with this agent provided only disappointing results. Another radiosensitizer of this kind is etanidazole, which is less toxic than misonidazole, and has great potential in radiotherapy. Most side effects of these products are related to neurotoxicity. These compounds are described as “oxygen mimics.”

Radioprotectors

The most common radioprotectors—substances that protect cells from radiation damage—include substances containing sulfhydryl groups (-SH), such as cysteine and cysteamine. These agents protect normal cells from radiation damage by combining with free radicals that are produced by radiation and would be toxic to normal cells. However, these compounds cause severe adverse reactions such as nausea and vomiting.

Less toxic compounds have been developed in which the -SH group is protected by a phosphate group. The phosphate group is hydrolyzed in vivo to release the -SH group for radioprotection. Two most effective compounds of this category are WR-638 and WR-2721 developed at Walter Reed Army Hospital, Washington, DC. Experimental evidence showed that

these products concentrate more in normal cells and less in tumor cells. As a result, normal cells are protected better than tumor cells if these agents are administered immediately before the radiation dose is given. WR-2721, also called amifostine, is an aminothial and protects bone marrow. Its most common toxic effects are hypotension and somnolence. Radioprotectors are most effective with low-LET radiations, because they cause minimal damage.

Stage of Cell Cycle

Radiation damage mostly occurs during the period of mitosis, the M phase, whereas least damage occurs during the DNA synthesis, the S phase. Thus, the stage of the cell cycle determines the extent of radiation damage. Exposure of cells to 100 to 1000rad (100 to 1000cGy) causes delay in the G₂ phase to M phase transition. An exposure of 1000rad (1000cGy) inhibits the progression of the S phase cells by 30%, whereas the S phase to G₂ phase transition is not affected by such an exposure (Prasad, 1995).

Classification of Radiation Damage

Cell death is a measure of extreme radiation damage. Therefore, based on the degree of lethality induced by radiation, radiation damage can be classified into three categories: (1) lethal damage, which causes irreversible death; (2) sublethal damage (SLD), which normally repairs in hours, and thus avoids cellular death, unless followed by another sublethal damage; and (3) potentially lethal dose (PLD), which can potentially kill the cell but can be modified to repair under specific physicochemical conditions. All these damages are relevant in clinical radiation therapy as to the effectiveness of treatment. Lethal damage is a definite end point in treatment, whereas SLD and PLD have variable effects in radiation therapy.

Sublethal damage occurs in mammalian cells, when a radiation dose is given in fractions at different time intervals rather than a single dose. There are four mechanisms, the so-called four R's that play a role in the SLD repair (SLDR) mechanism: repair, redistribution, regeneration, and reoxygenation. Repair involves the healing of the radiation-induced damage in the time interval between the two fractions of the dose. If the second dose is applied too soon after the first application, the damage does not have enough time to repair and the cell will die. In fractionated radiotherapy, normal tissues are spared by SLD because of its repair mechanism. In the redistribution process, the cells are desynchronized and sensitized to show increased damage. Following irradiation, the radiosensitive cells will die, and one would expect the proportion of radioresistant cells and hence the surviving fraction to increase. In fact, however, the surviving cells become sensitized and tend to die. This result depends on the fractionated dose and

the time interval between the doses. Regeneration is a mechanism of response to depopulation of a cell cohort due to radiation damage, and depends on the types of tissue and their proliferating capacity. Protracting a fractionated dose should be beneficial to normal tissues and somewhat harmful to regenerating tumor cells. Reoxygenation discussed earlier is an effect that makes the hypoxic cells more radiosensitive in the presence of oxygen in fractionated radiotherapy.

Sublethal damage repair depends very much on the dose rate and in which stage of the cell cycles the cells are. At lower doses, more SLD can be repaired, and at higher doses, the chances of SLD repair diminish. The dose-rate effect varies with the types of tissue and species. For example, the testis of male rats is most radiosensitive, whereas the small intestine seems to be less affected by radiation. Also, SLD repair depends on the LET of radiations. The repair is significant with x-rays and γ -rays and almost nonexistent for neutrons and α -particles. SLD repair is very important in radiation therapy as it provides maximum survival of normal cells, while killing tumor cells.

Potentially lethal damage after a single dose of radiation can potentially kill the cell but it can be repaired (PLDR) under specific physicochemical conditions. For example, the survival of the HeLa cells increased after irradiation, when the cells were treated with excess thymidine or hydroxyurea for a period of 4 hr. However, opposite results were obtained by other investigators. The importance of PLDR in radiotherapy is a matter of debate.

PLDR and SLDR are found with low-LET radiations (e.g., γ -rays and x-rays giving cell survival curves with a broad shoulder), while they are absent for high-LET radiations (neutrons and α -particles).

Stochastic and Deterministic Effects

Two categories of radiation effects on biological systems are encountered: stochastic and deterministic. Stochastic effects are the biological effects that occur randomly, the probability of which increases with increasing dose without a threshold. Radiation-induced hereditary effects and cancer incidences are examples of stochastic effects. The assumption of no threshold is made on the belief that radiation damage to a few cells or a single cell could theoretically induce the genetic disorder or cancer, and the severity of the disease will be the same, if it ever occurs. It should be noted that the basic principle of ALARA (as low as reasonably achievable) in Nuclear Regulatory Commission (NRC) regulations is based on the assumption of risks linearly proportional to the dose without a threshold. Much debate is currently going on regarding the assumption of the linear-no-threshold (LNT) theory (discussed below).

The deterministic or nonstochastic effects are induced by high radiation doses and the severity of the damages, rather than their probability of occurrence, increases with the dose. These effects have a threshold dose below which no damage is evident. Cataracts, skin erythema, sterility, and fibrosis are examples of deterministic effects induced by high radiation doses.

Acute Effects of Total Body Irradiation

Different tissues of the body respond differently to radiation, due to varying degrees of radiosensitivity. When an adult subject is irradiated over the entire body, various syndromes are manifested depending on the dose applied. The effects of radiation are characterized by the survival time of the species and various stages of acute syndromes following the total-body irradiation. These effects are deterministic types and have a threshold dose.

Cell survival time varies with mammal species depending on the individual radiosensitivity. The radiosensitivity of a given species is commonly characterized by the lethal dose, $LD_{50/60}$, which is the dose that kills 50% of the species in 60 days. The $LD_{50/60}$ for humans is 400 to 600 rad (400 to 600 cGy); for dogs, 300 rad (300 cGy); and for mice, 900 rad (900 cGy).

Acute radiation syndromes appear in four stages: prodromal, latent, manifest illness, and recovery or death. Each stage is dose dependent and can last for a few minutes to weeks. A minimum of 200 to 300 rad (200 to 300 cGy) is required for all four stages to be seen and can cause death.

In the prodromal stage, major symptoms are nausea, vomiting, and diarrhea and they occur in the early phase, lasting for only a short period of time depending on the dose. A dose of 50 rad can induce nausea and vomiting. In the latent stage, biological damage slowly builds up without manifestation of any syndromes, again lasting for hours to weeks, depending on the dose. During the manifest illness stage, radiation syndromes appear as a result of the damage to the organs involved after the latent period, and the subject becomes ill. In the last stage, the subject either recovers or dies.

There are three categories of syndromes in the manifest illness stage depending on the dose: hemopoietic or bone marrow, gastrointestinal (GI), and cerebrovascular.

Hemopoietic Syndrome

Hemopoietic or bone marrow syndromes appear at a total body dose of 250 to 500 rad (250 to 500 cGy) following irradiation. At this dose, the precursors for RBCs and white blood cells (WBCs) are greatly affected, so much

so that they lose the ability to reproduce. Also, the number of lymphocytes are greatly depressed, whereby the immune system of the body is suppressed. Loss of blood cell counts can be noticed at a dose as low as 10 to 15 rad (10 to 15 cGy). Thus, the body loses the defense against bacterial and viral infection and becomes susceptible to them. Immunosuppression by radiation occurs at doses as low as 100 rad (100 cGy) and 90% to 95% of immunosuppression can take place in humans at doses of 200 to 400 rad (200 to 400 cGy).

At this dose level, the platelet count is drastically reduced, and therefore bleeding gradually progresses through various orifices owing to a lack of ability of the blood to coagulate. Fever, bleeding, and infection result, followed by ultimate death in 10 to 21 days. However, bone marrow transplantation at the appropriate time may prompt the recovery of the subject. Whereas at doses <100 rad (100 cGy) survival is almost certain, survival is virtually impossible at doses >500 rad (500 cGy).

Gastrointestinal Syndrome

Gastrointestinal (GI) syndromes are expressed at a total body dose of 500 to 1000 rad (500 to 1000 cGy) and include prodromal syndromes such as nausea, vomiting, and diarrhea of more severity that appear within hours after exposure. The primary effect of radiation exposure in this range is that the intestinal crypt cells are destroyed and not replaced, and consequently the mucosal lining (villi) shrinks and hardens whereby the gut becomes nonfunctional. Because of the denudation of the gut, an intestinal ulcer may develop. These GI syndromes are also accompanied by drastic hemopoietic syndromes including immunosuppression, loss of white blood cells, and infection.

Thus, the loss of nutrients through ulcers, in combination with bacterial infection and excessive bleeding, results in GI death in 3 to 10 days after radiation exposure. Only aggressive medical treatment in the early stages of exposure may lead to recovery in cases at the lower end of the dose spectrum.

Cerebrovascular Syndrome

Cerebrovascular syndromes appear in a matter of minutes after radiation exposure at a total body dose of more than 10,000 rad (10,000 cGy). Because the nerve cells are radioresistant, such a large dose is required for cerebrovascular syndromes to appear. The symptoms include severe nausea, vomiting, and burning sensation of the skin that occurs within minutes of exposure, followed by the malfunction of the neuron motor pump giving rise to motor incoordination, intermittent stupor, coma, and ultimately death in two to three days. The cerebrovascular death sequence is so rapid that there is little time for significant changes to appear in other organs in

the body. At this level of radiation dose, death is a certainty and medical help is of no use.

Long-Term Effects of Radiation

The long-term or late effects of radiation cause various syndromes long after the radiation exposure. These may appear after acute radiation syndromes subside following exposure to a single large dose or after exposure to many smaller doses over a period. The late effects may be somatic or genetic, depending on the respective cells involved. Somatic effects are seen in the form of carcinogenesis, life-shortening, cataractogenesis, and embryologic damage. On the other hand, genetic effects result in abnormalities in the offspring.

Somatic Effects

Carcinogenesis

Cancer develops in three stages: initiation, promotion, and progression. Initiation of cancer is caused by various agents such as chemicals, ultraviolet rays, radiation, and viruses. In the case of radiation, cancer is initiated by the action of radiation on the DNA molecule resulting in the mutation of the cell. Cancer promoters are those agents that cannot initiate the cancer but simply promote it once it is started. Examples of tumor promoters are estrogen, phorbol ester, excessive fat, and radiation. Radiation acts as a promoter by inactivating tumor suppressive genes through the interaction of the free radicals produced in the cytoplasm by radiation. In these two stages, mutated cells proliferate at the site of cancer growth. One or more of these cells become aggressive and are likely to spread to other organs. This stage is called the progression or metastasis of the cancer.

At the cellular level, carcinogenesis is thought to be controlled by two types of genes: oncogenes and suppressor genes. There is evidence that oncogenes are responsible for the growth and proliferation of tumor cells, while suppressor genes inhibit the tumor cell growth. Most oncogenes have their counterpart, proto-oncogenes in normal cells, that are present throughout their eukaryotic evolution. Radiation or other carcinogens activate normal proto-oncogenes to several cancer-causing oncogenes and inactivate suppressor genes, resulting in cell proliferation to cause cancer. Chromosome aberrations (deletion or translocation) caused by radiation are responsible for oncogene activation. There are about 100 oncogenes identified that are associated with various human cancers. For example, deletion of a part of the chromosome is responsible for the activation of *N-ras* oncogene associated with neuroblastoma. Similarly, a translocation between chromosomes 8 and 14 in humans activates the *C-myc* oncogene in B-cell lymphoma.

Suppressor genes exist in normal cells to control the cell growth and protect the genome against carcinogenic agents. After radiation damage, suppressor genes stop cell division and repair the damaged gene. Examples of suppressor genes include the p53 gene found in breast cancer, small cell lung cancer, and bladder cancer; the DCC gene in colon cancer; and the p105 Rb gene in retinoblastoma. Radiation can inactivate these suppressor genes and thus cause cell proliferation leading to malignancy.

Epidemiologic Evidence of Carcinogenesis

The latent period of malignancies varies with the type of malignancy and the absorbed dose. Leukemia has an average latent period of about 5 to 10 years, whereas solid tumors in the head, neck, pharynx, and thyroid have a minimum latent period of 10 years with an average of 20 to 30 years.

Malignancies have been observed in individuals who are exposed to radiation from medical treatment, radiation-related occupation (e.g., industrial exposure), and acts of war. Infants and children are more radiosensitive than adults, and the risk of cancer from radiation exposure in the former is greater than that in the latter, almost by a factor of 2.

In the early 1900s, radium-dial painters used to lick the brush bristle soaked with radium-containing paint to make a fine point for painting clock and watch dials. During the procedure, they ingested radium, which, as a chemical analog of calcium, localized in bone, causing bone tumors. In some cases, the quantity of radium ingested was large, and acute effects including death were observed.

Before the 1930s, the enlarged thymus gland of infants with acute respiratory distress syndromes was commonly treated with therapeutic doses of x-rays to reduce the enlargement. During irradiation, however, the thyroid glands also received a massive radiation dose. A statistically significant number of these infants developed thyroid cancer later in life (about 10 years later).

Radiologists who used x-rays in their profession were found to have a higher incidence of leukemia than other medical professionals. Dentists had higher incidence of finger lesions due to exposure to dental x-rays. These incidences occurred mostly before the 1950s, largely because of the lack of knowledge of radiation effects. Now, through better radiation protection practice, these incidences have been curtailed drastically.

From 1935 to 1944, approximately 15,000 patients with ankylosing spondylitis were irradiated with 100 to 2000 R over spine and pelvis. A 2-year follow-up showed an increased incidence of leukemia in this group of patients.

Increased incidences of leukemia, lung cancer, breast cancer, and thyroid cancer have been observed in the Japanese survivors of the atomic bomb attacks on Hiroshima and Nagasaki.

Uranium mine workers inhale a considerable amount of radioactive dust containing radon gas. The decay products of radon settle in the lungs, and radiations from them can cause lung cancer.

Dose–Response Relationship

A meaningful dose–response relationship for carcinogenesis should be based on data with both low and high radiation exposures. However, the low-dose data (below 10 rad or 10 cGy) that have been accumulated thus far are mostly inconclusive, because of the small sample size, lack of appropriate control, incomplete dosimetry, and other related factors. So risks at low doses are primarily estimated by extrapolation of the data from high-dose experiments. Several authoritative committees (international and national) are responsible for establishing the dose-response relationship, and they are the United Nations Scientific Committee of the Effects of Atomic Radiations (UNSCEAR), the Committee on the Biological Effects of Ionizing Radiations (BEIR), the International Commission on Radiological Protection (ICRP), and the National Council on Radiation Protection and Measurements (NCRP) in the United States. These committees base their analysis on the data of the Japanese survivors of the A-bomb attacks on Hiroshima and Nagasaki, data on human exposures mentioned above (see *Epidemiologic Evidence of Carcinogenesis*), and data from *in vitro* cell culture and animal studies.

The risk versus dose relationship has been controversial, particularly about the minimum level of radiation dose that induces cancer (Murphy, 1991). Some experts propose that the dose–response relationship is linear, without a threshold dose, and that a very minimal dose can cause cancer (Fig. 15.15). A threshold dose is a dose below which no radiation damage occurs in an individual. The LNT dose–response relationship has been based on the extrapolation of high-dose data to low-dose data (below 10 rad or cGy) and has drawn a considerable debate as to its validity. In one argument, the opponents of the theory question the validity of such extrapolation, because the mechanism of radiation damage may be quite different at doses that differ by orders of magnitude. Also, this group points out that the people living in high natural background radiation areas (e.g., Rocky Mountains) do not show to have any more apparent adverse health effects than those in low-dose areas (e.g., sea level). On the other hand, the proponents of the LNT theory suggest that a single hit by a radiation can cause the mutation of a cell, and consequently result in carcinogenesis in a later period, thus supporting the theory. The recent BEIR VII report strongly supports the LNT theory, suggesting that even the smallest dose can cause a small risk of cancer in humans.

All intermediate and high-energy data primarily obtained from the Japanese survivors of the A-bomb attacks are fitted by a linear quadratic curve (Fig. 15.15). The curve is linear at lower doses and becomes propor-




FIG. 15.15. Three general shapes of the dose–response curves permit prediction of different incidences of low-dose radiation effects when the curves are fitted to high-dose data. (Adapted from Murphy PH. Acceptable risk as a basis for regulation. *Radiographics* 1991; 11:889–897.)

tional to D^2 (quadratic) at higher doses. Yet, other experts believe that there is no risk of carcinogenesis up to a certain threshold dose, after which the curve becomes linear or quadratic (Fig. 115.15). While the linear response model is preferred for all solid tumors, the linear-quadratic model is more suitable for leukemia.

Risk Estimates of Excess Cancer

The BETR VII Committee (2005) estimated the excess cancer risk based on the LNT theory and by extrapolation of high-dose incidences to the low-dose situations. It is estimated that 1 in 100 individuals is expected to develop solid tumor or leukemia from a radiation dose of 10rem (0.1 Sv), whereas approximately 42 of these 100 persons are expected to develop these cancers from other causes. Lower doses would produce proportionally lower risks. So, approximately 1 in 1000 individuals would expect to develop cancer from an exposure of 1 rem (0.01 Sv). Another estimate of cancer risk by BETR VII is that 1 in 100 persons would likely develop cancer from a lifetime (70yr) exposure to low-LET natural background radiations excluding radon and high-LET radiations. For comparison, the ICRP estimate of cancer risk for the general population is 0.1 per Sv (0.1% per rem) for high doses and dose rates, and 0.05 Sv (0.05% per rem) for low doses and dose rates (ICRP 60, 1991).

Leukemia

Leukemia is one of the most common cancers induced by radiation in humans, accounting for one in five mortalities from radiocarcinogenesis. Risk of leukemia varies with age, with younger persons being more prone

to radiocarcinogenesis than older persons. Based on a relative-risk linear quadratic dose response model, the BEIR V Committee (1990) predicts a risk of excess lifetime leukemia cancer mortality of 10 in 10,000 (0.1%) with an exposure of 10 rad (10 cGy). The BEIR VII (2005) estimate of this value is 1 to 2.6 per 10,000 person years per Gy (1 to 2.6 per 100 person years per rem). Leukemia appears in as early as 2 to 3 years after the exposure, with an average latent period of 5 to 10 years.

Breast Cancer

The annual incidence of breast cancer in women in the United States is 1 in 11, with a high mortality rate. Several factors—age, estrogen level, heredity, demographics, race, number of children given birth to, and breastfeeding—all influence the risk of breast cancer. Women exposed to low-level radiation can develop breast cancer, and the risk is greater with younger ages than with older ages. According to the BEIR V (1990) report, the excess absolute risk of breast cancer in 25-, 35-, and 45-year-old women is 5, 4, and 2 in 10,000, per Gy (5, 4, and 2 in 100 per rem) respectively. The BEIR VII (2005) estimate of this value is 10 per 10,000 person years per Gy (10 per 100 person years per rem) at age 50. The risk of radiogenic breast cancer apparently is reduced if the dose is given in divided exposures.

While in the past, the breast cancer risk from mammography was of major concern, modern mammographic equipment is well designed and properly shielded, resulting in significant reduction in radiation exposure and hence a lower risk of breast cancer. According to the recommendations of the American College of Radiology, women over 40 should have annual mammography screening for breast cancer.

Other Cancers

Radiation-induced cancers in the thyroid, lungs, bone, skin, and other organs have been found in the general population and are influenced by a variety of etiologic factors such as heredity, occupation, age, sex, and hormonal level. For thyroid cancer, the BEIR VII (2005) reports the excess absolute risk of 4.4 per 10,000 person years per Gy (4.4 per 100 person years per rem).

Radiation Damage to Skin

The skin is sensitive to radiation because of the presence of highly radio-sensitive structures such as hair follicles and sebaceous glands. The radiation effect on the skin is deterministic and has a threshold dose of about 100 rad (100 cGy). The primary skin reactions to radiation are erythema and dermatitis. Initial erythema appears in a few hours to a few days after radiation exposure, which is followed by dry desquamation characterized by atrophy of epidermal papillae and vascular changes. In the third or fourth week, erythema reappears with red, warm, edematous, and tender skin. Temporary hair loss (epilation) can occur at this stage. Severe erythema is

followed by acute radiation dermatitis manifested by blister formation, dermal hypoplasia, edema, and permanent depilation. The normal skin reappears in 2 to 3 months.

The above effects depend on the dose, dose rate, LET of the radiation, and the duration of exposure, and vary from individual to individual and with the location in the same individual. Although sex is not a factor in skin reaction, age is an important factor, with the skin of younger people being more sensitive.

A low-level chronic exposure of radiation on the skin causes atrophy, hyperplasia, and hyperkeratosis. In addition, ulceration and deep fibrosis may result from such exposures. In rare instances, skin cancer, mainly squamous cell carcinoma, may develop.

Radiation Damage to Reproductive Organs

Extremely deleterious effects are expected from radiation exposure to the gonads, because of their high radiosensitivity. In males, spermatogonia are most radiosensitive, and spermatozoa and spermatids are radioresistant, whereas in females, the ovarian follicles are most radiosensitive. The radiation effects vary with the dose, dose rate, sex, and age.

Sterility is the important radiation effect that warrants special attention. In males, temporary sterility can be induced with a dose as low as 15 rad (15 cGy), whereas permanent sterility is reported with an acute dose of 500 to 600 rad (500 to 600 cGy) (Prasad, 1995). Male sterility is evident by a reduced sperm count and low motility. In females, permanent sterility occurs with 320 to 625 rad (320 to 625 cGy), which is manifested by the damage to the ovarian follicles. If the dose is lower, then the follicles may recover in 5 to 6 months. Relatively larger doses are needed in younger women than in older women.

Nonspecific Life-Shortening

Studies have shown that exposure to ionizing radiations results in the shortening of the life span of mice (Rotblat and Lindop, 1961). For acute total body exposure, the life span of mice, rats, and dogs is reduced by about 5% per gray (UNSCEAR, 1982). The irradiated group looks much older than the control group, and radiation effects are similar to those of normal aging, e.g., an increase in connective tissue and a decrease in parenchymal cells. During the period 1945–1955, American radiologists were found to have a shorter life-span than other medical professionals. But the issue of life-shortening by radiation is controversial, because in some cases it has been found that life span is rather lengthened by irradiation at low doses. Such observations have led to the concept of *hormesis*, which states that low doses of radiation are beneficial to health and prolong the life span. It is postulated that hormesis is secondary to an enhanced immune responsiveness due to radiation at low doses.

Cataractogenesis

The lens of the eye is sensitive to radiation and develops cataracts on irradiation with ionizing radiations. The incidence of radiation-induced cataracts is a deterministic effect and depends on the dose given. A dose of 10 to 30 rad (10 to 30 cGy) is required to produce cataracts in mice, whereas a threshold dose of about 200 rad (200 cGy) is needed to produce cataracts in humans in a single exposure. A larger dose is required if the dose is given in fractions, and high-LET radiations almost double the incidence of cataractogenesis. A minimum of 1 year is needed for the latent period of cataractogenesis.

Radiation Damage to Embryo and Fetus

The developing mammalian embryo is extremely sensitive to ionizing radiations, because many cells are differentiating at this stage. The degree of damage depends on the developmental stage of the embryo, the dose, and the dose rate. The entire fetal development is divided into three general stages: (1) *preimplantation*, a period of about 8 to 10 days between fertilization of the egg and its attachment to the uterine wall; (2) *major organogenesis*, a period of about 2 to 6 weeks, when major organs are developed; and (3) the *fetal stage*, the remainder of the pregnancy period, when the organs of the fetus grow further to enable the mammal to survive after birth.

The embryo in the preimplantation stage is most sensitive to ionizing radiations and mostly encounters prenatal embryonic death as a result of radiation exposure. In some species, a dose as low as 5 to 15 rad (5 to 15 cGy) is sufficient to cause deleterious effects on the embryo. At a dose of 200 rad (200 cGy) in the preimplantation stage, embryonic death is certain. Almost all embryos that survive the radiation exposure grow normally in utero and afterward, with the exception of a few that develop abnormalities later.

During the period of major organogenesis, embryos exposed to ionizing radiations develop abnormalities mostly related to the central nervous system (CNS) and bone. These abnormalities are too severe for the fetus to survive and ultimately result in neonatal death. At an exposure of 200 rad (200 cGy) to mouse embryos during this period, almost 70% of the embryos later experienced neonatal death. Growth retardation also is noted at doses above 100 rad (100 cGy). Often it is suggested that a therapeutic abortion should be considered if an embryo receives ~10 rad (10 cGy) during the first 6 weeks after conception.

During the fetal period, however, comparable doses do not cause any abnormality or neonatal death, because fetal cells are more radioresistant than embryonic cells. Relatively higher doses are needed to cause death in this period. A few cases of growth retardation have been noted. In utero irradiation with a dose of 1 to 2 rad (1 to 2 cGy) may increase the risk of

childhood leukemia in the first 10 to 15 years by a factor of 1.5 to 2. Mental retardation in this period has been reported with doses as low as 10 to 20rad (10 to 20cGy), if given in the 8 to 15 weeks of gestation.

Because of these radiation effects, radiological procedures are contraindicated in pregnant women, and practitioners must exercise caution in determining the woman's status of pregnancy before these procedures. Before a procedure, it is a common practice to inquire of the patient if she is pregnant or when she had her last menstrual period, and thus unnecessary fetal exposure can be avoided. If the patient is pregnant and the procedure is essential, then the risk versus benefit to the patient from the procedure should be weighed by the practitioner with due consideration to the stage of pregnancy.

Genetic Effects

As mentioned above, ionizing radiations can cause changes in the DNA structure, which ultimately are expressed in gene mutations. Through the affected germ cells, these mutations propagate to future generations. Genetic effects are not expressed in the individual whose germ cells have been affected by radiation, but are expressed in future generations. Genetic effects appear as Down syndromes, achondroplasia, retinoblastoma, cystic fibrosis, sickle cell anemia, Tay-Sachs disease, and other chromosome disorders.

Spontaneous Mutations

In normal cells, genes occasionally undergo natural mutations even without radiation exposure. Such mutations are called *spontaneous mutations*, and their frequency is about 10^{-5} per gene per generation. This means that the chance of spontaneous mutation is 1 in 100,000. This frequency is increased by various mutagens such as chemicals and radiation.

In a given generation, radiation does not produce any new mutations and simply increases the frequency of spontaneous mutations. BEIR VII (2005) estimated the total risk for all classes of genetic diseases to be 3.0×10^{-5} to 4.7×10^{-5} per rem (3.0×10^{-3} to 4.7×10^{-3} per Sv) per generation. It indicates that the genetic risks are relatively small. The dose-response relationship is linear without threshold, indicating that no dose is safe and any dose, however relatively small. Furthermore, genetic damage is a function of the dose rate and the LET of ionizing radiations. High-LET radiations and high-dose rates cause more mutations. Genetic mutations may appear in future generations long after exposure has occurred.

Doubling Dose

The doubling dose is a measure of the increase in genetic mutations by radiation. It is the amount of radiation dose that doubles the spontaneous muta-

tions in one generation in a species. It is calculated as a ratio of the average spontaneous and induced mutation rates in a set of genes. A small doubling dose indicates a large relative mutation risk and vice versa. In humans, it is considered to be of the order of 100 rad (100 cGy), but it depends on the dose rate, the gender, and the type of species (BEIR V, 1990). Based on this doubling dose, the ICRP has given an estimate of the probability of induced hereditary disorders to be 0.6×10^{-2} per sievert (0.6×10^{-4} per rem) for the working population (ICRP 60, 1991).

Genetically Significant Dose

The genetically significant dose (GSD) is the dose that, if received by everyone of the entire population, would cause the same genetic damage as the gonadal dose now being received by a limited number of individuals of the population through medical procedures, natural radiations, TV viewing, flying at high altitudes, and so forth. The GSD is an index of the expected genetic damage on the whole population, and is calculated as an average value from the gonadal doses received from all exposures by the exposed personnel with proper weighting with respect to the chances of their having offspring. Thus, the GSD depends on the total number of individuals irradiated and the relative expectancy of their having children. The weighting factor is needed because older people have lesser probability of having offspring than younger people.

The contributions of various sources of radiation to GSD are given in Table 15.2. The GSD values from natural radiation sources are considered to be equal to the gonadal dose, because natural radiation exposes the entire population of all ages uniformly. Of all medical procedures, diagnostic x-rays contribute most to the GSD. It is, therefore, essential that strict protective measures are taken to avoid unnecessary gonadal exposure.

TABLE 15.2. Annual genetically significant dose (GSD) in the U.S. population about 1980–1982.

Source	Contributions to GSD in mrems (mSv)
Natural sources	
Radon	10 (0.1)
Other	90 (0.9)
Medical	
Diagnostic x-rays	20–30 (0.2–0.3)
Nuclear medicine	2 (0.02)
Consumer products	5 (0.05)
Occupational	~0.6 (0.006)
Nuclear fuel cycle	<0.05 (0.0005)
Miscellaneous environmental sources	<0.1 (0.001)
Total	~132 (1.32)

Adapted with permission from Table 8.2 in NCRP report No. 93, 1987.

Gonadal shields, appropriate collimation of the x-ray beams, and limited and prudent application of repeat procedures definitely lead to a considerable reduction in GSD from medical procedures.

Genetic effects of radiation can be greatly reduced if a time interval is allowed between radiation exposure and conception. This is the consequence of some repair process after irradiation. It is, therefore, recommended for both men and women that conception should be deferred for about 6 months after a significant radiation exposure such as a radiation accident or radiation therapy involving high gonadal exposure. Such delay in conception is not needed for diagnostic procedures.

Risk Versus Benefit in Diagnostic Radiology and Nuclear Medicine

Millions of x-ray, dental x-ray, computed tomography (CT), and nuclear medicine procedures are performed in the United States for the diagnosis of diseases, and the number is increasing steadily over the years. However, with the remarkable improvement in the evolving techniques and equipment, the effective dose to the individual and to the population as a whole is steadily decreasing.

Of all diagnostic radiological procedures, CT scans and fluoroscopic procedures give the highest effective doses, whereas dental and chest x-rays contribute only minimal effective doses. Gonadal doses are higher with fluoroscopic procedures than with head CT, chest x-ray, and dental procedures. This is primarily due to the fact that the gonads are out of the field of the latter procedures. It should be noted that the mammographic procedure contributes only a little to the total body dose compared to the breast. For obvious reasons, the highest gonadal dose comes from the procedures involving hips and pelvis. The GSD is about 9.8 mrad (98 μ Gy) for males compared to 20.9 mrad (209 μ Gy) for females (NCRP 100, 1989).

The doses to different organs from different nuclear medicine procedures are listed in Table 14.4 and the effective doses in Table 14.5 in Chapter 14. Radiation dose is always higher with long-lived and β -emitting radionuclides. The GSD values for females (1.9 mrad or 19 μ Gy) is almost twice those of males (1.1 mrad or 10.9 μ Gy) (NCRP 100, 1989).

Risks from diagnostic procedures include both somatic and genetic effects. Normally, these effects are minimal from diagnostic procedures for humans because doses from these procedures are considered low. Doses from nuclear medicine procedures are even lower than those from diagnostic x-ray procedures. However, based on the LNT model, there is no reason to believe that there is no risk from diagnostic exposures, no matter how small the doses are. There may not be acute effects, but long-term effects such as carcinogenesis, teratogenic effects from fetal exposure, and genetic effects in the future offspring can occur. The probabilities of fatal

cancers, nonfatal cancers, and hereditary effects have been estimated by the ICRP to be 4.0%, 0.8%, and 0.8%, respectively, for adult radiation workers and 5.0%, 1.0%, and 1.3%, respectively, for the whole population (ICRP 60, 1991).

An important quantity in the assessment of risk from radiation exposures is the collective effective dose, which is defined as the sum of the products of the effective dose and the number of persons exposed for each diagnostic procedure (NCRP 100, 1989). The age-weighted value per year for low dose and dose rates is estimated to be 58,000 person-Sv (5.8 million person-rems) for diagnostic radiological procedures and 13,500 person-Sv (1.35 million person-rems) for nuclear medicine procedures in the United States.

Based on a collective effective dose of 58,000 person-Sv (5.8 million person-rems) and the probabilities of cancers given above, Hall (1994) estimated the risks from 1 year of diagnostic radiology practice to be 2320 fatal cancers, 464 nonfatal cancers, and 464 serious heritable defects. Similarly, based on a collective dose of 13,500 person-Sv (1.35 million person-rems), the risk from 1 year of nuclear medicine practice is 540 fatal cancers, 103 nonfatal cancers, and 108 serious heritable defects. These risks are quite low compared to the number of examinations performed annually.

The benefit from diagnostic procedures (both x-ray and nuclear medicine) is the immediate diagnosis of the disease that can lead to the appropriate treatment and its ultimate cure. Argument should prevail in favor of the benefit for the use of radiation for diagnosis over the risks that may appear in later years in the individual himself or the future offspring. However, a judicious use of these procedures is definitely warranted, and a procedure that is not needed should not be done. This argument for the prudent use of radiation also applies to different screening procedures using x-ray, such as mammography, chest x-rays, and dental x-rays. Many individuals are exposed for screening, but only a small number of people benefit from the early diagnosis, while most of the screened people turn out to be negative. For this reason, the American College of Radiology has recommended annual mammography only for women above 40 years of age, excluding younger women who are much more radiosensitive, some of whom may likely develop breast cancer many years after mammography.

Risk to Pregnant Women

Since radiation can cause a devastating effect on the embryo and fetus in pregnant women, diagnostic radiological and nuclear medicine procedures are contraindicated in pregnant women, despite only a small risk involved with the individual exposed from these procedures. This is particularly important in nuclear medicine procedures, because radiopharmaceuticals reside in the body following a biological half-life and are likely to cross the placenta to cause the fetal damage. β -emitting radionuclides are more

damaging than γ -emitting radionuclides. Radioiodine administered orally to pregnant women during the gestation period of 15 to 22 weeks can cross the placenta and localize in the fetal thyroid to the extent of 50% to 75%. The fetal thyroid dose at 6 weeks of gestation is of the order of 2.1 Gy/MBq (7.8 rad/mCi) (Watson, 1991).

In most cases, radiologic procedures are avoided in pregnant women by proper screening such as asking them prior to the procedure if they are pregnant or when they had their last menstrual period. However, at times, it is discovered after the procedure that the woman is pregnant. In such situations, steps should be taken to estimate the dose received by the embryo or fetus based on the dosimetry parameters of the radiopharmaceutical. Depending on the period of pregnancy, the question of therapeutic abortion may be considered if the dose is excessive. Some experts believe that a dose of 10 cGy (10 rad) is a reasonable value above which therapeutic abortion should be considered. However, the decision to abort depends on a number of socio-personal factors.

In radionuclide therapy, pregnant women are absolutely excluded because of the anticipated excessive fetal dose. ^{131}I treatment of pregnant women is almost prohibited unless benefit outweighs the risk of the fetus from therapy. Besides the in utero effects, there is a small probability of thyroid cancer induced by the ^{131}I therapy of hyperthyroidism.

Dirty Bombs

A dirty bomb, also called a Radiological Dispersal Device (RDD), is a mix of explosive, such as dynamite, with radioactive materials. After the explosion, in addition to the immediate devastating effects of the explosive material causing injury and property damage, radioactive dust and smoke spread the radioactive contamination into the surrounding areas. Radioactive dust and smoke, if inhaled, can cause ill health effects. The use of dirty bombs by perpetrators is to spread radioactive contamination and create fear and panic, more than anything else. Subsequent decontamination could involve considerable time and cost.

A dirty bomb is not an atomic bomb and is primarily used to disrupt and not destroy the human life. Another type of RDD might involve a very high level of radioactive source hidden in a bus, train, or subway station, where people passing close to the source might get a significant dose of radiation. Prompt detection of these devices (bomb or radioactive source) is essential in order to take protective measures.

The sources of radioactive materials are the hospitals, research facilities, and industrial and construction sites where radioactivity is used for various purposes (diagnosis and treatment at hospitals, research work, sterilizing equipment, and check of welding). Some of the highly radioactive sources are cobalt-60, strontium-90, cesium-137, and iridium-192 used in industrial

radiographic services. These radionuclides have long half-lives. Many of these sources are mostly in metallic capsule form and the likelihood of dispersion is minimal. However, they can be available in liquid and powder forms and potentially be used in dirty bombs, which can result in widespread contamination in the surrounding areas of explosion. Because one cannot see, taste, or feel radiations, excessive exposure can be received unknowingly by people in the vicinity of the area.

Types of Radiation Exposure

Radiation exposure from radiation accidents can be localized and/or whole-body type. The localized exposure may be caused by direct handling of or close proximity to highly radioactive sources. The local injury includes erythema, epilation, desquamation, ulceration, or blistering depending on the level of exposure. The treatment of choice for localized injuries is the use of antibiotic for infection and control of pain. In severe cases, amputation or plastic surgery is warranted.

The whole-body exposure causes various acute radiation syndromes that have been discussed earlier in this chapter. These syndromes include hemopoietic, gastrointestinal, and cerebrovascular syndromes depending on the absorbed doses. Although cerebrovascular syndromes occur with 10,000 rem (100Sv), and result in death, the hemopoietic and gastrointestinal syndromes may be managed by bone marrow transplantation and other prophylactic treatment.

When a RDD explodes, radioactive material may be airborne and contaminate food and water. Internal contamination can occur from the ingestion of contaminated food and water, inhalation of the contaminated air, and diffusion through the skin or wounds. The principle of the treatment of internal contamination primarily involves dilution, displacement by non-radioactive material, complex formation, and blockage. In the case of internal contamination with radioiodine (e.g., incidences of fallout from a nuclear explosion or a nuclear reactor accident), both the NRC and FDA have approved the use of potassium iodide (KI) as a preventive measure. Such use of KI is intended to block the thyroid from trapping ^{131}I and it should be taken before the exposure or within several hours of exposure. The recommended daily dose is 130mg of KI for adults, 65mg for 3 to 18 yr old, 32mg for children 1 month to 3yr old, and 16mg for infants less than 1 month old (Mettler and Voelz, 2002).

Outer garments such as clothing and shoes can be contaminated by radioactivity from the explosion of a dirty bomb. Such contamination does not constitute a medical emergency and most of it can be removed by taking off these garments. Minor skin contamination can be eliminated by thorough washing with water and detergent, and a shower, if appropriate. Skin should not be abraded by a heavy brush, as this may facilitate internal absorption. If an individual has a life-threatening condition in addition to

the external contamination, the patient must be first managed for the condition before decontamination is carried out. Burns and wounds that are not contaminated should be first covered and then decontamination of the other affected areas carried out.

Protective Measures in Case of Explosion of a Dirty Bomb

According to the advisory of the NRC and the Center for Disease Control (available at www.nrc.gov and www.bt.cdc.gov), the following steps should be taken in case of a dirty bomb explosion.

1. If you are outside and close to the explosion of a dirty bomb, cover the nose and mouth with a mask or cloth to reduce the risk of breathing in radioactive dust or smoke. If possible, immediately go inside the building that is not affected by the explosion and remove the outer layer of clothing and shoes and seal them in a plastic bag, if available. Store the plastic bag in a safe place for decay of the radioactivity. Next, take a shower to remove any dust that remains on the body.
2. If you are inside and close to the incident that has occurred outside, close all the doors and windows and do not leave the building. Turn off heat and air conditioner to stop air circulation from outside.
3. Listen to the local news for further appropriate instructions.
4. Food and water are unlikely to be contaminated. If contaminated, do not consume these items. Local or federal authorities are likely to monitor the food and water in the area of explosion and inform the public of their suitability for consumption.
5. If it is definitely known that there is radioiodine contamination, KI should be orally administered within a short time according to the regimen described earlier.
6. Monitoring is essential and critical to estimate the level of contamination or decontamination using a GM counter and is performed by hospital personnel, police, and firefighters who are specially trained for this purpose.

Verification Card for Radioactive Patients

After the September 11, 2001 attack on the World Trade Center in New York, numerous security measures have been adopted by the U.S. Federal Government. Congress has passed laws to establish the Department of Homeland Security to implement and monitor different aspects of these security measures. Security checks of airline passengers, background scrutiny of many visitors and suspected terrorist groups, and implementation of the Patriot Act are some of the examples of these security actions

that are currently in place. In view of the concerns over the use of dirty bombs by miscreants, the Homeland Security has established checkpoints at various strategic locations such as airports, tunnels, mass transit, bridges, border crossing points, historical monuments, landmarks, and the like, to monitor the transport of dirty bombs containing radioactivity by using radiation detectors.

One pitfall of this measure is that patients who received radioactive materials for diagnostic and therapeutic purposes may trigger the monitors while passing through these checkpoints and undergo undue hassle with authorities to provide proof that the radioactivity was really from medical uses. These incidents have been reported all over the country. The NRC has suggested providing patients having radioactivity a card to carry with information such as the name of the patient, type of radionuclide administered, date of examination, and a contact phone number for verification of the information given on the card. Many hospitals are adopting this policy and instruct the patients to carry the card for the period discussed below.

A question arises as to how long the patient who has undergone nuclear studies should carry the card. It depends on the half-life of the radionuclide, types of radiations the radionuclide emits, and the biological elimination of the radiotracer from the body. Zuckier (2004) in a paper presented at The Radiological Society of North America annual meeting in Chicago suggested the following periods for different radionuclides for the patients to carry the card.

^{18}F	1 day
$^{99\text{m}}\text{Tc}$ and ^{123}I	3 days
^{111}In	14–17 days
^{67}Ga and ^{201}Tl	30 days
^{131}I	95 days

Radiation Phobia

The public in general is unduly fearful of radiation because of several factors. One important factor is the graphic images of the devastating effects of the atomic bombs detonated in Hiroshima and Nagasaki in 1945, and to a lesser extent, the images of the Chernobyl reactor accident in 1986. The most noticeable effects of these incidents are death of living species and destruction of property at the site of the explosion and its immediate vicinity. Because of these images, many people associate radiation exposure with adverse health effects and death. These images are firmly embedded in the minds of the public causing perpetual fear of radiation. However, these incidents are very uncommon.

Another flashpoint in creating radiation phobia in the public's mind is the knowledge of assumption that any level of ionizing radiation is dan-

gerous to health, alluding to the linear no-threshold (LNT) theory of the dose–response relationship. Psychological warfare with anecdotal rhetoric among the rival countries possessing nuclear weapons also creates fear of radiation among the public. Dreading effects of radiation on children and future offspring, and long-term damage to property are major concerns of the public. Furthermore, the media often play a role in exacerbating the problem of exposure from radiation accidents.

Is there a logical justification for this radiation phobia of the public? Definitely, nuclear detonation causes an instantaneous devastating effect on the population and property, and so can be reason for fear and panic. But the long-term effects of low doses of radiations, even from the fallout of the atomic bombs in Japan and Chernobyl accident, have been shown to be relatively small. The average individual lifetime dose from the Chernobyl fallout is estimated to be 0.6 to 6 rem (6–60 mSv) (Jaworowski, 1999). By comparison, the worldwide average annual dose rate of natural radiation an individual receives on earth is 220 mrem (2.2 mSv) and the lifetime dose of about 15 rem (150 mSv). In the United States, an individual receives an annual dose of about 300 mrem (3 mSv) including radon and a lifetime dose of 21 rem (210 mSv). These values are even ten times higher in some regions in India and Brazil, and yet incidence of excess cancer is not shown to be higher in these places.

People face risk of cancer, injuries, and even death from day-to-day living activities, such as driving, smoking cigarettes, drinking alcohol, eating food, and breathing air, in addition to hazardous job-related activities. Thirty-three percent of the population will contract cancer just from these activities without any radiation exposure, and 22% will die from natural causes (ACS, 2003). If the population is exposed to 1 rem (10 mSv) of radiation exposure, the risk of cancer increases only by 0.1% and half of them (0.05%) will die, which is quite negligible (BEIR VII, 2005). Based on these arguments, it can be said that although nuclear explosions can be a cause for grave concern, low-dose radiations from medical facilities, natural background, and the like, are fairly safe relative to the hazards of different living activities, and the risk from such radiation exposure is small.

To allay the fears of radiation in the public's mind is essential and critical. It can be achieved through education of the public. People knowledgeable in radiation should talk to laymen explaining the relatively small risk of low-level radiations compared to many other day-to-day living activities. The media should play an important role in communicating this information to the public. Radiation experts should hold regular public seminars to explain the minimal risk of low-level radiation. Radiation-related professional organizations such as the Society of Nuclear Medicine, The Radiological Society of North America, Health Physics Society, and American Association of Physicists in Medicine should undertake appropriate approaches of communication with the public to shed their concern and fear of radiation.

Questions

1. (a) What are the mechanisms of radiation damage?
 (b) Does the direct action or indirect action contribute more to radiation damage? Why?
 (c) Which are the free radicals that are most damaging to cells?
 (d) Does the presence of oxygen increase or decrease radiation damage?
2. (a) Why are erythroblasts more radiosensitive than red blood cells?
 (b) Which phase of the cell cycle is most radiosensitive?
 (c) Which molecule of the cell structure is most radiosensitive?
 (d) What are the different factors affecting radiation damage?
3. (a) Define D_0 , D_q , and n as illustrated in the cell survival curve.
 (b) D_q is smaller for high-LET radiations than for low-LET radiations. True or false?
 (c) D_0 is smaller for high-LET radiations than for low-LET radiations. True or false?
 (d) What is the value of n for mammalian cells?
4. (a) The cell survival curve is steeper at high radiation doses than at low radiation doses. Explain why and its implication.
 (b) Does the shape of the cell survival curve vary with high-LET radiations and at very high-dose rates?
5. (a) Choose the dose in rad that has been suggested as a practical threshold for radiation-induced abortion: (i) 2; (ii) 5; (iii) 10; (iv) 20; (v) 50.
 (b) How many days after conception can prenatal death occur as a result of in utero irradiation?
 (c) Which one of the following organs is most affected to be malformed by prenatal radiation exposure? (i) heart; (ii) stomach; (iii) head; (iv) gonads; (v) upper extremities.
 (d) What is the period of pregnancy during which the incidence of abnormalities and malformations in human neonates is expected to be the highest?
 (e) What are the effects of radiation on the fetus?
 (f) The incidence of childhood leukemia after in utero irradiation with a few rad of diagnostic x-rays increases by a factor of (i) 1.5–2.0; (ii) 2.5–3.0; (iii) 3.5–6.0.
6. (a) What are the dose ranges and approximate time limits of death for hemopoietic, gastrointestinal, and cerebrovascular syndromes?
 (b) What are the prodromal syndromes and when do they appear?
 (c) What is the dose at which almost total immunosuppression occurs in humans?
7. (a) Define the oxygen enhancement ratio (OER).
 (b) Why is the oxygen effect absent for high-LET radiations?
 (c) What are radiosensitizers? Name some of them.

- (d) What is the maximum OER for γ and x-rays?
 - (e) Is misonidazole a radiosensitizer or radioprotector for hypoxic cells?
 - (f) What is the specific composition of radioprotectors and how do they function?
8. (a) What is the doubling dose and what is its value for humans?
 - (b) What are the doses at which permanent sterility can be induced in (a) males and (b) females?
 - (c) Define the genetically significant dose (GSD).
 - (d) What is the GSD for humans?
 - (e) Which one of all medical radiations contributes most to the GSD?
 - (f) What are the factors that influence the GSD?
9. (a) The mean latent period for radiation-induced leukemia is about (i) 5 to 10 years; (ii) 12 to 20 years; (iii) 21 to 30 years.
 - (b) The mean latent period for radiation-induced solid tumors is about (i) 5 to 10 years; (ii) 12 to 20 years; (iii) 21 to 30 years.
 - (c) Cataract can be induced in humans with (i) 10 to 30rad (10 to 30cGy); (ii) 100 to 110rad (100 to 110cGy); (iii) 200rad (200cGy).
 - (d) What is the risk of cancer in the general population from small doses of low-LET radiation exposure?
10. What are the two most common chromosome aberrations that are responsible for carcinogenesis?
 11. In the linear quadratic model of the cell survival, what is the value of parameter β for high-LET doses?
 12. Elucidate the mechanisms of sublethal damage repair and potentially lethal damage repair. Give an example of the latter.
 13. What is the recent value of cancer death attributable to radiation exposure in Japanese survivors of the atomic bomb?
 14. What are the different dose-response models preferred for solid tumors and leukemia?
 15. What are the risk estimates of fatal cancer by the ICRP and the BEIR VII for the general population?
 16. What is a dirty bomb? How does it differ from an atomic bomb?
 17. What are the common radionuclides used in the radiological dispersal device? What are the common sources of radioactive materials used in dirty bombs?
 18. Describe the types of effects caused by radiation.
 19. Describe the basic principles of decontamination of the contaminated individuals.
 20. What are the recommended steps one should take in the case of the explosion of a dirty bomb?
 21. The U.S. Homeland Security monitors radioactivity for dirty bombs at strategic points of commuting. The patients undergoing nuclear studies are given cards by the hospitals to provide proof of radioactive examinations. How long should the patient normally carry the card for ^{18}F , $^{99\text{m}}\text{Tc}$, ^{123}I , ^{111}In , ^{67}Ga , ^{201}Tl , and ^{201}Tl ?

References and Suggested Readings

- American Cancer Society. *Radiation Exposure and Cancer*. Atlanta, GA; 2003.
- BEIR V Committee. *The Effects on Populations of Exposure to Low Levels of Ionizing Radiations*. Washington DC: National Academy of Sciences/National Research Council; 1990.
- BEIR VII, Phase 2. *Health Risks from Exposure to Low Levels of Ionizing Radiations*. Washington, DC: National Academy of Sciences/National Research Council; 2005.
- Hall EJ. *Radiobiology for the Radiologist*. 5th ed. Philadelphia: JB Lippincott Williams & Wilkins; 2000.
- ICRP report no. 26. *Recommendations of the International Commission on Radiological Protection*. New York: Pergamon; 1977.
- ICRP report no. 60. *1990 Recommendations of the International Commission on Radiological Protection*. New York: Pergamon, 1991.
- Jaworowski Z. Radiation risk and ethics. *Physics Today*. 1999;52:24.
- Mettler FA, Upton AC. *Medical Effects of Ionizing Radiations*. 2nd ed. Philadelphia: W.B. Saunders; 1995.
- Mettler FA, Voelz GL. Major radiation exposure—What to expect and how to respond. *New Eng J Med*. 2002;346:1554.
- Murphy PH. Acceptable risk as a basis for regulation. *Radiographics*. 1991; 11:889–897.
- NCRP report no. 93. *Ionizing Radiation Exposure of the Population of the United States*. Bethesda, MD: NCRP; 1987.
- NCRP report no. 100. *Exposure of the U.S. Population from Diagnostic Medical Radiation*. Bethesda, MD: NCRP; 1989.
- Nias AHW. *Introduction to Radiobiology*. 2nd ed. Hoboken, NJ: Wiley; 1988.
- Pizzarello DJ, Witcofski RL. *Medical Radiation Biology*. 2nd ed. Philadelphia: Lea & Febiger; 1982.
- Prasad KN. *Handbook of Radiobiology*. 2nd ed. Boca Raton, FL: CRC Press; 1995.
- Ring JP. Radiation risks and dirty bombs. *Health Phys*. 2004;86:S42–S47.
- Rotblat J, Lindop P. Long-term effects of a single whole body exposure of mice to ionizing radiation, II. Causes of death. *Proc R Soc Lond (Biol)*. 1961;154:350–368.
- Travis EL. *Primer of Medical Radiobiology*. 2nd ed. Chicago: Year Book Medical Publishers; 1989.
- United Nations Scientific Committee on the Effects of Atomic Radiation (UNSCEAR). *Ionizing Radiations: Sources and Biological Effects*. New York: United Nations, 1982.
- United Nations Scientific Committee on the Effects of Atomic Radiation (UNSCEAR). *Sources, Effects and Risks of Ionizing Radiation*. New York, United Nations, 1988.
- Watson EE. Radiation absorbed dose to the human fetal thyroid. In: *5th International Radiopharmaceutical Dosimetry Symposium*. Oak Ridge, TN, May 7–10, 1991.
- Zuckier L, Stabin M, Garetano G et al. Sensitivity of personal homeland security radiation detectors to medical radionuclides and implications for counseling of nuclear medicine patients. *RSNA Annual Meeting*. 2004; Abstract SSJ19-01.

16

Radiation Regulations and Protection

Radiation hazards to humans are well documented. To minimize their risks, international and national organizations have been established to set guidelines for safe handling of radiations. As mentioned in Chapter 15, the ICRP and NCRP are two such organizations. They make recommendations and guidelines for radiation workers to follow in handling radiations. The Nuclear Regulatory Commission (NRC) and state agencies adopt many of these recommendations into regulations to implement radiation protection programs in the United States. The NRC regulations are published in the Federal Register in the form of the Code of Federal Regulations (CFR). The regulations pertinent to the practices of nuclear medicine are briefly described here.

Currently, in the United States, the NRC regulates all reactor-produced by-product materials as to their use and disposal and the radiation safety of personnel using them as well as the public. On the other hand, naturally occurring and accelerator-produced radionuclides are regulated by individual states. However, Congress has recently authorized the NRC to take control of these radioactive materials as well and currently they are in the process of drafting the rule. For convenience of operations, at present, 33 states have entered into agreement with the NRC, which authorizes each state to regulate the by-product materials as well. These states are called *Agreement States*. The rules and regulations of these states must be as restrictive as those of the NRC, if not stricter.

Sources of Radiation Exposure

The population at large receives radiation exposure from various sources such as natural radioactivity, medical procedures, consumer products, and occupational sources. The estimates of annual effective dose equivalents from different radiation sources to the U.S. population are tabulated in Table 16.1.

The major contribution of the exposure comes from natural sources, particularly from radon from building materials, amounting to 200mrem (2mSv)/year accounting for 82% of the total exposure. Excluding radon exposure, the average exposure from natural background consisting of cosmic radiations, terrestrial radiations, and so on amounts to about 100mrem (1mSv)/year. This exposure varies with the altitude of places above sea level. For example, the annual cosmic ray exposure in cities such as Denver is about 50mrem (0.5mSv) compared to 26mrem (0.26mSv) at sea level. Air travel at a height of 39,000ft (12km) gives 0.5mrem/hr (5 μ Sv/hr), resulting in an annual dose of 1mrem (10 μ Sv) to the population.

Terrestrial radiation exposure arises from radionuclides such as ^{40}K and from decay products of thorium and uranium in soil. It varies from about 16mrem (160 μ Sv)/year in the Atlantic ocean to 63mrem (630 μ Sv)/year in the Rockies with an average of 28mrem (280 μ Sv)/year.

Radionuclides ingested through food, water, or inhalation include ^{40}K and decay products of thorium and uranium, particularly ^{210}Po , and contribute about 39mrem (390 μ Sv) annually.

Man-made exposure constitutes about 18% of the total exposure. Medical procedures contribute the highest exposure of all man-made radiation sources. The most exposure comes from diagnostic radiographic procedures with about 39mrem (390 μ Sv) annually compared to 14mrem (140 μ Sv) for nuclear medicine procedures. Exposure from radiation therapy is relatively small.

Consumer products such as tobacco, water supply, building materials, agricultural products, and television receivers contribute to radiation expo-

TABLE 16.1. Annual effective dose equivalent in the U.S. population from different sources circa 1980 to 1982.

Sources	Average annual effective dose equivalent in mrem (mSv)
Natural sources	
Radon	200 (2.0)
Cosmic rays	27 (0.27)
Terrestrial	28 (0.28)
Ingested radionuclides	39 (0.39)
Medical procedures	
Diagnostic x-rays	39 (0.39)
Nuclear medicine	14 (0.14)
Radiation therapy	<1 (0.01)
Consumer products	5-13 (0.05-0.13)
Occupational	0.9 (0.009)
Nuclear fuel cycle	0.05 (0.0005)
Miscellaneous	0.06 (0.0006)
Total	~360 (3.6)

Adapted with permission from NCRP Report No. 93. *Ionizing Radiation Exposure of the Population of the United States*. Bethesda, MD: NCRP; 1987: Tables 8.1 and 2.4.

sure through consumption. Exposure from smoking has been estimated to be 1.3 mrem ($13 \mu\text{Sv}$)/year, which is not included in Table 16.1, because it is difficult to calculate the collective effective dose equivalent for the entire population. The total exposure from consumer products varies between 5 and 13 mrem (50 and $130 \mu\text{Sv}$)/year.

Occupational exposure is received by the workers in reactor plants, coal mines, and other industries using radionuclides. This value is about 0.9 mrem ($9 \mu\text{Sv}$)/year, which is quite small, because a great deal of precaution is taken to reduce exposure in the workplace.

Nuclear power plants around the country release small amounts of radionuclides to the environment, which cause radiation exposure to the population. This value is of the order of 0.05 mrem ($0.5 \mu\text{Sv}$)/year.

License

Authorization for the use of radioactive materials is granted by issuance of a license by the NRC or the Agreement State. There are two types of licenses:

1. General domestic license: Although the general domestic license is given for the use of by-product material in various devices according to 10CFR31, only the provisions of 10CFR31.11 are applicable for general license for the use of by-product material in certain in vitro clinical laboratory tests. Such general licenses are given to physicians, veterinarians, clinical laboratories, and hospitals only for in vitro tests, not for the use of by-product material in humans or animals. An application must be filed with the NRC using the Form NRC-483, "Registration Certificate—In vitro Testing with By-product Material under General License" and a validated copy must be obtained prior to the use of by-product material. The total amount to be possessed at any one time should not exceed $200 \mu\text{Ci}$ (7.4 MBq) of ^{125}I , ^{131}I , ^{75}Se , and/or ^{59}Fe . The amount of ^{14}C and ^3H can be obtained in units of $10 \mu\text{Ci}$ (370 kBq) and $20 \mu\text{Ci}$ (740 kBq), respectively. The products must be supplied in prepackaged units.

2. Specific licenses: The specific licenses are given in two categories: one to manufacture or transfer for commercial distribution certain items containing by-product material (10CFR32) and the other to possess, use, and transfer by-product material in any chemical or physical form with the limitations of the maximum activity specified (10CFR33). The former types of specific licenses are typically given to commercial manufacturers. The latter type is called the specific license of broad scope or "broad license" and has three categories based on the maximum activity allowed for the receipt, acquisition, ownership, possession, use, and transfer of any chemical or physical form of by-product material (10CFR33.11). The Type A broad license allows specified quantities of activities usually in multicuries; the Type B broad license allows maximum activities of by-product material

specified in 10CFR33.100, Schedule A, Column I; and the Type C license permits maximum activities of by-product material specified in 10CFR33.100, Schedule A, Column II, which are an order of magnitude less than those in the Type B license.

In the Type A license, a radiation safety committee and a radiation safety officer are required to implement and monitor all aspects of radiation safety in the use and disposal of by-product material. Such licenses are mainly offered to large medical institutions with previous experience that are engaged in medical research, and in diagnostic and therapeutic uses of by-product material. Individual users are authorized by the radiation safety committee to conduct specific protocols using by-product materials.

The Type B specific license requires a radiation safety officer, but no radiation safety committee, to implement and monitor all radiation safety regulations. The Type C specific license requires neither the radiation safety officer nor the committee, but a definite statement that the by-product material will be used by the licensee or by persons under his direct supervision who has the training specified in 10CFR33.15.

In all cases of specific licenses, an application must be filed to the NRC using the NRC Form 313 with all information related to the possession, use, and disposal of by-product materials.

Radiation Protection

Rules and regulations pertaining to radiation protection set by the NRC are contained in 10CFR20. Because it is beyond the scope of this book to include the entire 10CFR20, only the relevant highlights are included.

Definition of Terms

Several terms related to absorbed dose as defined in the 10CFR20 are given here.

Committed dose equivalent ($H_{T,50}$) is the dose equivalent to organs or tissues of reference (T) that will be received from an intake of radioactive material by an individual during the 50-year period following the intake.

Deep-dose equivalent (H_d), which applies to the external whole-body exposure, is the dose equivalent at a tissue depth of 1 cm (1000 mg/cm²).

Shallow-dose equivalent (H_s), which applies to the external exposure of the skin or an extremity, is the dose equivalent at a tissue depth of 0.007 cm (7 mg/cm²) averaged over an area of 1 cm².

Tissue weighting factor (W_T) for an organ or tissue is the proportion of the risk of stochastic effects resulting from irradiation of that organ or tissue to the total risk of stochastic effects when the total body is irradiated uniformly.

The values of W_T from the 10CFR20 are given in Table 14.5.

Effective dose equivalent (H_E) is the sum of the products of the committed dose equivalent to each of the body organs and tissues when the total body is irradiated uniformly and the weighting factor of the corresponding organ or tissue ($H_E = \Sigma W_T \cdot H_{T,50}$).

Annual limit on intake (ALI) is the derived limit on the amount of radioactive material allowed to be taken into the body of an adult worker by inhalation or ingestion in a year. These values are given in 10CFR20 (Table 1, Appendix B).

Total effective dose equivalent (TEDE) is the sum of the deep-dose equivalent (for external exposure) and the committed effective dose equivalent (for internal exposure). It is the effective dose defined by Eq. (14.25).

Restricted area is an area where an individual could receive in excess of 5 mrem (0.05 mSv) per hour at 30 cm from a radiation source.

High-radiation area is an area where an individual could receive from a radiation source a dose equivalent in excess of 100 mrem (1 mSv) in 1 hr at 30 cm from the source.

Very high-radiation area is an area where an individual could receive from radiation sources an absorbed dose in excess of 500 rad (5 Gy) in 1 hr at 1 m from the source.

Unrestricted area is an area in which an individual could receive from an external source a dose of 2 mrem ($20 \mu\text{Sv}$)/hr and 50 mrem (0.5 mSv)/yr.

Caution Signs and Labels

The NRC requires that specific signs, symbols, and labels be used to warn people of possible danger from the presence of radiation. These signs use magenta, purple, or black color on yellow background; some typical signs are shown in Figure 16.1.

Caution: Radiation Area. This sign must be posted in radiation areas.

Caution: High Radiation Area or *Danger: High-Radiation Area.* This sign must be posted in high-radiation areas.

Caution: Radioactive Material or *Danger: Radioactive material.* This sign is posted in areas or rooms in which 10 times the quantity of any licensed material specified in Appendix C of 10CFR20 are used or stored. All containers with quantities of licensed materials exceeding those specified in Appendix C of 10CFR20 should be labeled with this sign. These labels must be removed or defaced before disposal of the container in the unrestricted areas.

Caution signs are not required in rooms storing the sealed sources, provided the radiation exposure at 1 foot (30 cm) from the surface of the source reads less than 5 mrem ($50 \mu\text{Sv}$)/hr. Caution signs are not needed in rooms




FIG. 16.1. Various radiation caution signs and labels.

where radioactive materials are handled for less than 8 hr, during which time the materials are constantly attended.

Occupational Dose Limits

The annual limit of the occupational dose to an individual adult is the more limiting of (a) TEDE of 5 rem (0.05 Sv) or (b) the sum of the deep-dose equivalent and the committed dose equivalent to any individual organ or tissue other than the lens of the eye being equal to 50 rem (0.5 Sv). It should be noted that there is no lifetime cumulative dose limit in the revised 10CFR20, although the NCRP recommends a lifetime cumulative dose of 1 rem (10 mSv) \times age in years.

The annual limit on the occupational dose to the lens of the eye is 15 rem (0.15 Sv).

The annual limit of the occupational dose to the skin and other extremities is the shallow-dose equivalent of 50 rem (0.5 Sv).

Depending on the license conditions, both internal and external doses have to be summed to comply with the limits. A licensee may authorize under *planned special procedures* an adult worker to receive additional dose in excess of the prescribed annual limits, provided no alternative procedure is available. The total dose from all planned procedures plus all doses in excess of the limits must not exceed the dose limit (5 rem or 50 mSv) in a given year, nor must it exceed five times the annual dose limits in the individual's lifetime.

The annual occupational dose limits for minors is 10% of the annual dose limits for adults. The dose limit to the fetus/embryo during the entire pregnancy (gestation period) due to occupational exposure of a declared pregnant woman is 0.5 rem (5 mSv).

The total effective dose equivalent to individual members of the public is 0.1 rem (1 mSv) per year. However, this limit can be increased to 0.5 rem (5 mSv) provided the need for such a higher limit is demonstrated.

ALARA Program

The established dose limits are the upper limits for radiation exposure to individuals. The NRC has instituted the ALARA (as low as reasonably achievable) concept to reduce radiation exposure to individuals to a minimum. The ALARA concept calls for a reasonable effort to maintain individual and collective radiation exposure as low as possible. Under this concept, techniques, equipment, and procedures are all critically evaluated. According to NRC Regulatory Guide, under the ALARA concept, when the exposure to a radiation worker exceeds 10% of the occupational exposure limit in a quarter (Action Level I), an investigation is made by the RSO, and the report is reviewed by the RSC. When the exposure exceeds 30% of the occupational exposure limit (Action Level II), corrective actions are taken or the licensee must justify a higher dose level for ALARA in that particular situation, but not to exceed annual occupational dose limit.

Principles of Radiation Protection

Of the various types of radiation, the α -particle is most damaging because of its charge and great mass, followed in order by the β -particle and the γ ray. Heavier particles have shorter ranges and therefore deposit more energy per unit path length in the absorber, causing more damage. On the other hand, γ -rays and x-rays have no charge or mass and therefore have a longer range in matter and cause relatively less damage in tissue. Knowledge of the type and energy of radiations is essential in understanding the principles of radiation protection.

The cardinal principles of radiation protection from external sources are based on four factors: time, distance, shielding, and activity.

Time

The total radiation exposure to an individual is directly proportional to the time of exposure to the radiation source. The longer the exposure, the higher the radiation dose. Therefore, it is wise to spend no more time than necessary near radiation sources.

TABLE 16.2. Exposure rate constants of commonly used radionuclides.

Radionuclides	Γ_{20} (R · cm ² /mCi · hr at 1 cm)	Γ_{20} (μGy · m ² /GBq · hr at 1 m)*
¹³⁷ Cs	3.26	88.11
^{99m} Tc	0.59	15.95
²⁰¹ Tl	0.45	12.16
⁹⁹ Mo	1.46	39.46
⁶⁷ Ga	0.76	20.54
¹²³ I	1.55	41.89
¹¹¹ In	2.05	55.41
¹²⁵ I	1.37	37.03
⁵⁷ Co	0.56	15.16
¹³¹ I	2.17	58.65
¹⁸ F†	5.70	154.05

* R · cm²/mCi · hr is equal to 27.027 μGy · m²/GBq · hr.

† Personal communication with Dr. M. Stabin, Oak Ridge Associated Universities, Inc., Oak Ridge, Tennessee.

Adapted from Goodwin PN: Radiation safety for patients and personnel. In: Freeman LM, ed. *Freeman and Johnson's Clinical Radionuclide Imaging*. 3rd ed. Philadelphia: WB Saunders Co; 1984: 320.

Distance

The intensity of a radiation source, and hence *the radiation exposure, varies inversely as the square of the distance from the source to the point of exposure*. It is recommended that an individual should keep as far away as practically possible from the radiation source. Procedures and radiation areas should be designed so that individuals conducting the procedures or staying in or near the radiation areas receive only minimum exposure.

The radiation exposure from γ -ray and x-ray emitting radionuclides can be estimated from the *exposure rate constant*, Γ , which is defined as the exposure from γ -rays and x-rays in R/hr from 1 mCi (37 MBq) of a radionuclide at a distance of 1 cm. Each γ - and x-ray emitter has a specific value of Γ , which has the unit of R · cm²/mCi · hr at 1 cm or, in System Internationale (SI) units, μGy · m²/GBq · hr at 1 m. The Γ values are derived from the number of γ -ray and x-ray emissions from the radionuclide, their energies, and their mass absorption coefficients in air.^a Because γ -rays or x-rays below some 10 or 20 keV are absorbed by the container and thus do not contribute significantly to radiation exposure, often γ -rays and x-rays above these energies only are included in the calculation of Γ . In these instances, they are denoted by Γ_{10} or Γ_{20} . The values of Γ_{20} for different radionuclides are given in Table 16.2.

The exposure rate X from an n -mCi radionuclide source at a distance d cm is given by

^a The Γ value of photon-emitting radionuclides can be calculated from the expression $\Gamma = 199 \sum N_i E_i \mu_i$, where N_i is the fractional abundance of photons of energy E_i in MeV, and μ_i is the mass absorption coefficient (cm²/g) of photons of energy E_i in air.

$$X = \frac{n\Gamma}{d^2} \quad (16.1)$$

where Γ is the exposure rate constant of the radionuclide.

Problem 16.1

Calculate the radiation exposure at 25 cm from a vial containing 30 mCi (1.11 GBq) of ^{201}Tl .

Answer

The exposure rate constant Γ_{20} of ^{201}Tl is $0.45 \text{ R} \cdot \text{cm}^2/\text{mCi} \cdot \text{hr}$ at 1 cm from Table 16.2. Therefore, using Eq. (16.1), at 25 cm

$$X = \frac{30 \times 0.45}{25^2} = 21.6 \text{ mR/hr}$$

Because Γ_{20} of ^{201}Tl in SI units is $12.16 \mu\text{Gy} \cdot \text{m}^2/\text{GBq} \cdot \text{hr}$ at 1 m, X for 1.11 GBq of ^{201}Tl at 25 cm is

$$\begin{aligned} X &= \frac{1.11 \times 12.16}{(0.25)^2} \\ &= 215.96 \mu\text{Gy/hr} \end{aligned}$$

It should be pointed out that because the patient is not a point source, the exposure rate does not vary exactly as the inverse square of the distance.

Shielding

Various high atomic number (Z) materials that absorb radiations can be used to provide radiation protection. Because the ranges of α - and β -particles are short in matter, the containers themselves act as shields for these radiations. γ -Radiations, however, are highly penetrating. Therefore, highly absorbing material should be used for shielding of γ -emitting sources, although for economic reasons, lead is most commonly used for this purpose. The half-value layer (HVL) of absorbent material for different radiations is an important parameter in radiation protection and is related to linear attenuation coefficient of the photons in the absorbing material. This has been discussed in detail in Chapter 6.

Problem 16.2

Calculate the number of HVLs and the amount of lead necessary to reduce the exposure rate from 100 mCi (3.7 GBq) of ^{131}I to less than 10 mR/hr at 10 cm from the source. ($\Gamma = 2.17 \text{ R} \cdot \text{cm}^2/\text{mCi} \cdot \text{hr}$ at 1 cm and 1 HVL = 3 mm of lead).

Answer

$$\text{Exposure at 10 cm} = \frac{2170 \times 100}{10^2} = 2170 \text{ mR/hr}.$$

A factor of $2170/10 = 217$ or more would be needed to reduce the exposure to less than 10 mR/hr. In terms of HVL, $2^8 = 256$, that is, 8 HVLs would be needed. Since 1 HVL = 3 mm of lead, 8 HVLs would be equal to 24 mm. Therefore, 8 HVLs or 24 mm of lead would be necessary.

Obviously, shielding is an important means of protection from radiation. Radionuclides should be stored in a shielded area. The radiopharmaceutical dosages for patients should be carried in shielded syringes. Radionuclides emitting β -particles should be stored in containers of low- Z material such as aluminum and plastic because in high- Z material, such as lead, they produce highly penetrating bremsstrahlung radiations. For example, ^{32}P is a β^- emitter and should be stored in plastic containers instead of lead containers.

Activity

It should be obvious that the radiation exposure increases with the intensity of the radioactive source. The greater the source strength, the more the radiation exposure. Therefore, one should not work unnecessarily with large quantities of radioactivity.

Personnel Monitoring

According to 10CFR20, personnel monitoring is required under the following conditions:

1. Occupational workers including minors and pregnant women likely to receive in 1 year a dose in excess of 10% of the annual limit of exposure from the external radiation source
2. Individuals entering high or very high radiation areas

Monitoring for occupational intake of radioactive material is also required if the annual intake by an individual is likely to exceed 10% of the ALIs in 10CFR20, Table 1, Appendix B, and if minors and pregnant women are likely to receive a committed effective dose equivalent in excess of 0.05 rem (0.5 mSv) in 1 year.

Three devices are used to measure the exposure of ionizing radiations received by an individual: the pocket dosimeter, the film badge, and the thermoluminescent dosimeter. The pocket dosimeter has been described in Chapter 7.

Film Badge

The film badge is most popular and cost-effective for personnel monitoring and gives reasonably accurate readings of exposures from β -, γ and x-radiations. The film badge consists of a radiation-sensitive film held in a plastic holder (Fig. 16.2). Filters of different metals (aluminum, copper, and cadmium) are attached to the holder in front of the film to differentiate




FIG. 16.2. (A) Film badge holder. (B) Film badge.

exposure from radiations of different types and energies. Filters of metals of different densities stop different energy radiations, thus discriminating exposures from them. After exposure the optical density of the developed film is measured by a densitometer and compared with that of a calibrated film exposed to known radiation. Film badges are usually changed monthly for each radiation worker in most institutions. Film badges provide an integral dose and a permanent record. The main disadvantage of the film badge is the long waiting period (a month) before the exposed personnel know about their exposure. The film badge also tends to develop fog resulting from heat and humidity, particularly when in storage for a long time, and this may obscure the actual exposure reading. The film badges of all workers are normally sent to a commercial firm that develops and reads the density of the films and sends back the report of exposure to the institution. The commercial firm must be approved by the National Voluntary Laboratory Accreditation Program (NVLAP) of the National Institute of Standards and Technology.

Thermoluminescent Dosimeter

A thermoluminescent dosimeter (TLD) consists of inorganic crystals (chips) such as lithium fluoride (LiF) and manganese-activated calcium fluoride ($\text{CaF}_2:\text{Mn}$) held in holders like the film badges and plastic rings. When these crystals are exposed to radiation, electrons from the valence band are excited and trapped by the impurities in the forbidden band. If the radiation-exposed crystal is heated to 300°C to 400°C , the trapped electrons are raised to the conduction band; they then fall back into the valence band, emitting light. The amount of light emitted is proportional to the amount of radiation absorbed in the TLD. The amount of light is measured and read

as the amount of radiation exposure by a TLD reader, a unit that heats the crystal and reads the exposure as well. The TLD gives an accurate exposure reading and can be reused after proper heating (annealing).

It should be noted that exposure resulting from medical procedures and background radiations are not included in occupational dose limits. Therefore, radiation workers should wear film badges or dosimeters only at work. These devices should be taken off during any medical procedures involving radiation such as radiographic procedures and dental examinations, and also when leaving after the day's work. Also radiation workers should not wear these badges for certain period of time after undergoing a diagnostic or therapeutic nuclear medicine procedure or radiation therapy permanent implant procedure.

Dos and Don'ts in Radiation Protection Practice

Do wear laboratory coats and gloves when working with radioactive materials.

Do work in a ventilated fume hood while working with volatile material.

Do cover the trays and workbench with absorbent paper.

Do store and transport radioactive material in lead containers.

Do wear a film badge while working in the radiation laboratory.

Do identify all radionuclides and dates of assay on the containers.

Do survey work areas for contamination as frequently as possible.

Do clean up spills promptly and survey the area after cleaning.

Do not eat, drink, or smoke in the radiation laboratory.

Do not pipette any radioactive material by mouth.

Do monitor hands and feet after the day's work.

Do notify the radiation safety officer (RSO) in the case of any major spill or other emergencies related to radiation.

Bioassay

NRC Regulatory Guide 8.20 gives the details of bioassay requirements for ^{131}I and ^{125}I radionuclides. Bioassays are required when the level of radioiodine activity handled (volatile or dispersible) exceeds the following values:

Open bench: 1 mCi (37 MBq)

Fume hood: 10 mCi (370 MBq)

Glove box: 100 mCi (3.7 GBq)

When the radioiodinated material is nonvolatile, the limits of activity are higher by a factor of 10. Stricter limits may be imposed in the license by the NRC.

For iodine radionuclides, bioassay is performed by the thyroid uptake test within 72 hr and at 14 days after handling the radioactivity. Sometimes urine analysis may also be required soon after the exposure. Bioassays may be

required for other radionuclides, depending on the amount and type of radionuclides.

Receiving and Monitoring of Radioactive Packages

Individual users or institutions are authorized to possess and use radioactive materials on issuance of a radioactive material license by the NRC or the Agreement State. The suppliers require documentation of licensing of the user as to the types and limits of quantities of radioactive material before shipping.

Monitoring of packages is required if the packages are labeled as containing radioactive material to check if the packages are damaged or leaking. A radioactive shipment must be monitored as soon as possible after receipt but no later than 3 hr after delivery if the delivery takes place during normal hours, or not later than 3 hr from the beginning of the next working day if it is received after working hours. Two types of monitoring are performed: survey for external exposure and wipe test for contamination on the surface of the package resulting from potential leakage of liquid. The survey reading of external exposure should not exceed 200mrem/hr (2mSv/hr) on the surface of the container or 10mrem/hr (100 μ Sv/hr) at 1 m from the surface of the container. The wipe test is performed by swabbing an area of 300 cm² of the package and should show less than the limit of 6600 dpm or 111 Bq/300 cm². If the readings exceed these limits, the NRC and the final delivering carrier must be notified by telephone and telegram, mailgram, or facsimile. Advice should be sought from these authorities as to whether the shipment should be returned.

After all surveys are completed, the data must be entered into a receipt book. The information logged in includes the date of the receipt, the manufacturer, the lot number, name and quantity of the product, date and time of calibration, and survey data along with the name of the individual processing the receipt.

Radioactive Waste Disposal

Radioactive waste generated in nuclear medicine or pharmacy (e.g., syringes, needles, vials containing residual activities and contaminated papers, tissues, and liners) are disposed of by the following methods according to the guidelines set forth in 10CFR20 and 10CFR35.

1. Decay in storage
2. Release into a sewerage system
3. Transfer to authorized recipient (commercial land disposal facilities)
4. Other disposal methods approved by the NRC (e.g., incineration of solid waste and atmospheric release of radioactive gases)

The following is a brief description of different methods of radioactive waste disposal, but one should consult 10CFR20 and 10CFR35 for details.

Decay in Storage

Although 10CFR20 does not spell out the conditions of the decay-in-storage method, 10CFR35.92 describes this method in detail. Radionuclides with half-lives less than 120 days usually are disposed of by this method. These radionuclides are allowed to decay in storage and monitored before disposal. If the radioactivity of the waste cannot be distinguished from background, it can be disposed of in the normal trash after removal or defacing of all radiation labels. Note that the requirement for 10 half-lives decay is eliminated from the revised 10CFR35. This method is most appropriate for shortlived radionuclides such as ^{99m}Tc , ^{123}I , ^{201}Tl , ^{111}In , ^{67}Ga and ^{131}I . Radioactivities should be stored separately according to half-lives for convenience of timely disposal of each radionuclide.

Release into Sewerage System

The NRC permits radioactive waste disposal into the sewerage system provided the radioactive material is soluble or dispersible in water and the quantity disposed monthly does not exceed the maximum permissible limits set in 10CFR20. Disposal depends on the total volume and flow rate of water used but is limited to 1 Ci (37 GBq) of ^{14}C , 5 Ci (185 GBq) of ^3H , and 1 Ci (37 GBq) of all other radionuclides annually. Excreta from humans undergoing medical diagnosis or treatment with radioactive material are exempted from these limitations. However, items contaminated with radioactive excreta (e.g., linen, diapers, etc., contaminated with urine or feces) are not exempted from these limitations. To adopt this method of radioactive disposal, one must determine the total volume and the flow of sewer water in the institution and the number of users of a specific radionuclide so that for each individual user, a limit can be set for sewer disposal of the radionuclide in question.

Transfer to Authorized Recipient

This method of transfer to an authorized recipient is adopted for longlived radionuclides and usually involves transfer of radioactive wastes to authorized commercial firms that bury or incinerate at approved sites or facilities.

Although the columns of the ^{99}Mo - ^{99m}Tc generators may be decayed to background for disposal to normal trash, a convenient method of disposing of this generator is to return them to the vendors, who let them decay and later dispose of them. Normally, the used generator is picked up by the authorized carrier when a new one is delivered.

Other Disposal Methods

A licensee may adopt methods of radioactive waste disposal different from those mentioned here, provided regulatory agency approval is obtained. Impact of such disposal methods on environment, nearby facilities, and population is heavily weighed before approval. Incineration of solid radioactive waste and carcasses of research animals containing radioactive materials is allowed by this method. Radioactive gases such as ^{133}Xe and ^{127}Xe are released by venting through the fumehood, as long as their maximum permissible concentration at the effluent side of the exhaust to the atmosphere does not exceed the NRC limits. Radioactive waste containing $0.05\ \mu\text{Ci}$ ($1.85\ \text{kBq}$) or less of ^3H or $^{14}\text{C}/\text{g}$ of medium used for liquid scintillation counting or animal tissue may be disposed of in the regular nonradioactive trash.

Records must be maintained as to the date of storage and the amount and kind of activity stored in a waste disposal log book. The stored packages must be labeled with pertinent information. The date of disposal and the amount of disposed activity must also be recorded in the log book, along with the initials of the individual disposing of the waste.

Radioactive Spill

Accidental spillage of radioactivity can cause unnecessary radiation exposure to personnel and must be treated cautiously and expeditiously. Appropriate procedures must be established for handling radioactive spills. There are two types of spills: major spill and minor spill. No definitive distinction exists between a minor and a major spill. A major spill usually occurs when the spilled activity cannot be contained in a normal way and can cause undue exposure to personnel. In the case of a major spill, the RSO should be notified immediately. In either case, the access to the area should be restricted. Areas, personnel, and equipment must be decontaminated, keeping in mind the principle of containment of radioactivity. Survey and wipe tests must be performed after decontamination. The RSO will investigate the accident and recommend corrective action if a major spill occurs.

Recordkeeping

Records must be maintained for the receipt, storage, and disposal of radioactive materials, as well as for various activities performed in the radiation laboratories. According to the NRC regulations, these records must contain specific information and must be kept for a certain period of time specified by the NRC.

Medical Uses of Radioactive Materials

The NRC regulates the medical use of by-product materials by enforcing 10CFR35; similar regulations are implemented by the states for naturally occurring and accelerator-produced radionuclides. However, the Agreement States regulate both categories of radionuclides. The revised 10CFR35 has been in effect as of October 24, 2002 when all NRC-regulated states were required to adopt all new regulations. However, the Agreement States have been given a three-year grace period (October, 2005) to adopt these regulations, after which time 10CFR35 will be effective for medical use of by-product material equally for all states in the United States.

There are six categories of medical uses of radioactive materials according to 10CFR Part 35. They are: (1) radiopharmaceuticals for uptake, dilution, and excretion (10CFR35.100); (2) radiopharmaceuticals for imaging and localization including generators and kits (10CFR35.200); (3) radiopharmaceuticals for therapy (10CFR35.300); (4) sealed sources for brachytherapy (10CFR35.400); (5) sealed sources for diagnosis such as sources of ^{125}I and ^{153}Gd for bone mineral analysis (10CFR35.500); and (6) sealed sources for teletherapy such as sources of ^{60}Co and ^{137}Cs in teletherapy units or gamma stereotactic radiosurgery units (10CFR35.600).

The regulations for the medical use of all radioactive materials are given in 10CFR35, but radiopharmaceuticals under categories 1, 2, and 3 only are relevant in nuclear medicine. These radiopharmaceuticals must be approved for human clinical use by the FDA under an IND or NDA. The $^{99\text{m}}\text{Tc}$ activity is eluted from the ^{99}Mo - $^{99\text{m}}\text{Tc}$ generator and reagent kits are used to prepare $^{99\text{m}}\text{Tc}$ -labeled radiopharmaceuticals according to instructions given by the manufacturer in the package inserts. Only reagent kits that are approved by the FDA under an IND or NDA may be used for radiopharmaceutical preparation.

Applications, Amendments, and Notifications

As already mentioned, applications for a license and its renewals must be made by the licensee's management for the medical uses of by-product materials. Amendments to the license must be made by the licensee's management for the following:

- (a) Appointment or discontinuation of an authorized user, radiation safety officer, authorized medical physicist, or authorized nuclear pharmacist
- (b) Change of name or address of the licensee
- (c) Change or addition of the use areas
- (d) Use of excess or new by-product materials not permitted before in the license

Notification of the above must be made within 30 days of occurrence. Change or addition of areas of use for uptake and dilution (10CFR35.100) and for localization and imaging (10CFR35.200) need not be amended. Licenses with Type A specific license of broad scope are exempt from these requirements.

Authority and Responsibilities of the Licensee

According to 10CFR35.24, the licensee's management is responsible for the overall implementation of the radiation protection program in the medical uses of by-product material. The licensee's management shall approve in writing all new authorized users, radiation safety officer, or nuclear pharmacist, and ministerial changes in the radiation safety program that do not require license amendment (10CFR35.26).

The licensee's management shall appoint a Radiation Safety Officer (RSO), who accepts in writing responsibilities to implement a radiation protection program. It may appoint one or more temporary RSOs for 60 days in a year, if all conditions of an RSO are met.

The licensee's management also must appoint a Radiation Safety Committee (RSC), if the licensee is authorized for two or more different types of uses of by-product material. Examples are the use of therapeutic quantities of unsealed by-product material (10CFR35.300) and manual brachytherapy (10CFR35.400), or manual brachytherapy and low-dose-rate therapy units (10CFR35.600), or teletherapy units (10CFR35.600) and gamma knife units (10CFR35.600). Use of by-product materials for both uptake and dilution (10CFR35.100) and imaging and localization (10CFR35.200) does not require an RSC. The RSC must include as a minimum an authorized user of each type of use permitted in the license, the RSO, a representative of the nursing service, and a representative of management, and in addition, other members, if appropriate. The NRC does not prescribe any definite frequencies of the RSC meetings nor record-keeping of the minutes.

Supervision

According to 10CFR35.27, a licensee that permits an individual to work under an authorized user or authorized nuclear pharmacist using by-product material must instruct the supervised individual to follow strictly all regulations and conditions of the license and all procedures involving by-product material. There is no requirement for periodic review of the supervised individual's work and records. The licensee is responsible for the acts and omissions of the supervised individuals.

Mobile Nuclear Medicine Service

According to 10CFR35.80, a licensee providing mobile nuclear medicine service to a client must

- (a) Have a letter signed by the licensee and the management of each client spelling out the details of the responsibility and authority of the client and the licensee,
- (b) Calibrate and check the instruments for measuring dosages and surveying,
- (c) Measure dosages and perform surveys of the area of uses at the client address, and
- (d) The client must have a license for receiving and using by-product material.

Written Directives

According to 10CFR35.40, a written directive is required when a dosage greater than $30\mu\text{Ci}$ (1.11 MBq) of $^{131}\text{I-NaI}$ or a therapeutic dosage of an unsealed by-product material other than $^{131}\text{I-NaI}$ is administered to a patient or human research subject. The written directive must be dated and signed by an authorized user and must contain the patient's name, the dosage, the name of the drug, and route of administration. A revision of the written directive can be made, if necessary, provided it is signed and dated by the authorized user before administration. In case of an emergency, an oral revision to an existing written directive is acceptable, which must be followed by a written directive within 48 hours.

According to 10CFR35.41, the licensee shall develop and maintain a copy of the written procedures for the written directive that include specific verifications of the identity of the patient before each administration, and that the administration is in accordance with the written directive. The identity of the patient may be verified by the name, driver's license, birthday, any hospital's I.D. number, and so on.

Measurement of Dosages

According to 10CFR35.63, all dosages for patient administration must be measured in an instrument (dose calibrator) that is calibrated with nationally recognized standards or the manufacturer's instructions (10CFR35.60). Although the methods of calibration are not specifically prescribed in 10CFR35, the constancy, accuracy, linearity, and geometry of the dose calibrator must be checked as described in Chapter 7.

For unit dosages, the activity can be determined by direct measurement or by the decay correction of the activity provided by the licensed manufacturer. For dosages other than unit dosages, the activity must be deter-

mined by direct measurement, a combination of measurement of radioactivity and mathematical calculations, or a combination of volumetric measurements and mathematical calculations based on the activity provided by the manufacturer.

Unless otherwise directed by the authorized user, the licensee may not use a dosage if it does not fall within the prescribed dosage range, or if it differs from the prescribed dosage by more than 20%. The licensees who use only unit dosages supplied by the manufacturer may not need to have a dose calibrator.

Calibration, Transmission, and Reference Sources

The following sources of by-product material are permitted for check, calibration, transmission, and reference use (10CFR35.65):

- (a) Sealed sources not exceeding 30mCi (1.11GBq) manufactured and supplied by a licensed manufacturer, or a licensee authorized to redistribute such sources
- (b) Any by-product material with a half-life not longer than 120 days in individual amounts not exceeding 15mCi (0.56GBq)
- (c) Any by-product material with a half-life longer than 120 days in individual amounts not to exceed the smaller of 200 μ Ci (7.4MBq) or 1000 times the quantities in Appendix B of 10CFR30
- (d) ^{99m}Tc in amounts as needed

A licensee may only use sealed sources for medical use, manufactured by a licensed manufacturer (10CFR49).

Requirement for Possession of Sealed Sources

According to 10CFR35.67, sealed sources of radionuclides with a half-life greater than 30 days and containing more than 100 μ Ci (3.7MBq) of γ -emitting material or more than 10 μ Ci (370kBq) α -emitting material must be leak tested and inventoried semiannually. If a source shows a leak of 0.005 μ Ci (185Bq) or more of removable contamination, it must be immediately removed from use and stored, disposed of, or repaired according to regulations, and a report must be filed with the NRC within five days of the leak test describing the source involved, the test results, and the action taken.

Labeling of Vials and Syringes

Each syringe and vial containing radioactivity must be labeled to identify the radioactive drug (10CFR35.69). Each syringe or vial shield also must

be labeled, unless the label on the syringe or vial is visible through the shield. Although syringe shields are not required by the NRC regulations for administration of radiopharmaceuticals, they should be used to maintain ALARA exposures.

Surveys of Ambient Radiation Exposure Rate

According to 10CFR35.70, the NRC requires that the licensee shall survey all areas where unsealed by-product material requiring a written directive was prepared for use or administered. The survey must be performed at the end of each day of use with a radiation detection instrument. In addition, according to 10CFR20, surveys must be conducted in areas of radiation use to maintain ALARA principles.

Calibration of Survey Instruments

According to 10CFR35.61, the survey meter must be calibrated before its first use, annually, and after repairs that affect calibration. Calibration must be made in all scales with readings up to 1000mrem (10mSv) per hour with a radiation source, and two separated readings must be calibrated on each scale or decade (digital) that is used to show compliance. The date and the results of calibration must be noted on the instrument. The licensee may not use the survey instruments if the difference between the indicated exposure rate and the calculated exposure rate is more than 20%.

The requirement for wipe testing of various areas of use for removable contamination has been eliminated in new 10CFR35 regulations. However, it is advisable to adopt wipe testing for removable contamination as part of the ALARA principle.

Training and Experience Requirements for Medical Uses of By-Product Materials

Authorized users, radiation safety officers, and nuclear pharmacists are required to have appropriate training and experience for medical uses of by-product materials. Normally there are two methods of approval: (1) certification by a specific medical specialty board, and (2) training and work experience in radionuclide handling techniques applicable to specific medical use of by-product material. The names of the boards recognized by the NRC are posted on the NRC Web site. For recognition, each board must meet all requirements of the training and work experience in a specific category described below and be approved by the NRC.

The training part includes a specified period of didactic classroom and laboratory training in the areas of (a) radiation physics and instrumenta-

tion, (b) radiation protection, (c) mathematics pertinent to radioactivity, (d) chemistry of by-product material, and (e) radiation biology and radiation dosimetry (for radiation safety officer).

The work experience must be under an authorized user, radiation safety officer, or nuclear pharmacist depending on the specific authorization of by-product material requested and must include (a) ordering, receiving, and unpacking radioactive materials, and surveying; (b) calibration of dose calibrators and survey meters; (c) calculating, measuring, and preparing dosages for patients; (d) procedures for spill management; (e) safely administering dosages to patients (for authorized users only); and (f) elution of radioactive generators (for localization and imaging studies).

In addition, approval by the training and experience method requires a written certification by a preceptor that the individual has acquired competence in the techniques to function independently for a specified use of by-product material.

The required hours of training and experience vary for different types of uses of radioactive material and are listed below.

Radiation Safety Officer (10CFR35.50)	200 hrs of classroom and laboratory training in radiation safety plus 1 yr work experience under a radiation safety officer
Nuclear Pharmacist (10CFR35.55)	700 hrs of classroom and laboratory training and supervised work experience in nuclear pharmacy
Authorized User (10CFR35.190) (Uptake, dilution, and excretion)	60 hrs of training and work experience, including 8 hrs of classroom and laboratory training in radionuclide handling
Authorized User (10CFR25.290) (Localization and imaging)	700 hrs of training and experience including 80 hrs of classroom training in basic radionuclide handling
Authorized User (10CFR35.390) (Therapeutic use)	700 hrs of training and work experience, including 200 hrs of classroom training in basic radionuclide handling, plus 3 cases in each therapeutic use of by-product material
Authorized User (10CFR35.392) (Hyperthyroidism using less than 33 mCi [1.22 GBq] $^{131}\text{I-NaI}$)	80 hrs of classroom and laboratory training in ^{131}I therapy procedures plus 3 cases of hyperthyroid treatments
Authorized User (10CFR35.394) (Thyroid cancer using greater than 33 mCi [1.22 GBq] $^{131}\text{I-NaI}$)	80 hrs of classroom and laboratory training in ^{131}I therapy procedures plus 3 cases of thyroid cancer treatments

Report and Notification of a Medical Event

The term *medical event* has substituted for “misadministration” and “recordable event” under the revised 10CFR35.3045. A medical event

occurs when a dose exceeds 5 rem (0.05 Sv) effective dose equivalent, or 50 rem (0.5 Sv) to an organ or tissue or skin from any of the following situations.

- (a) The total dose delivered differs from the prescribed dose by 20% or more,
- (b) The total dosage delivered differs from the prescribed dosage by 20% or more, or falls outside the prescribed dosage range,
- (c) Administration of a wrong radioactive drug containing by-product material,
- (d) Administration by wrong route,
- (e) Administration to a wrong individual.

The licensee must notify by telephone a medical event to the NRC Operation Center no later than 1 calendar day after discovery of the event, followed by a written report to the NRC Regional Office within 15 days. The report must include the licensee's name, prescribing physician's name, brief description of the event, cause of the event, effect of the event, if any, on the individual, corrective action taken, if any, and whether the affected individual or his or her relative or guardian has been notified. The individual's name or identification number shall not be included in the report.

The licensee shall notify the individual and the referring physician of the event no later than 24 hours after the discovery, unless the referring physician personally takes the responsibility of informing or not informing the individual based on medical judgment. If a verbal notification is made, the licensee shall inform the individual of the availability of a written description of the event, which the licensee will provide upon request.

In addition, the licensee shall annotate a copy of the report filed with the NRC with the name and social security number or other identification number of the affected individual and provide a copy of the annotated report to the referring physician, if other than the licensee, within 15 days of occurrence of the event. Recordkeeping of medical events is not required because the reports are provided to the NRC.

Report and Notification of a Dose to an Embryo/Fetus or a Nursing Child

The licensee shall report to the NRC an event in which an embryo/fetus receives more than 5 rem (50 mSv) dose equivalent due to the administration of by-product material to a pregnant individual, unless such a dose was specifically approved in advance by the authorized user. Also, a report must be made to the NRC if the dose to a nursing child, from the administration of by-product material to a breast-feeding individual, exceeds

5 rem (50 mSv) total effective dose equivalent, or has resulted in unintended permanent functional damage to an organ or biological system of the child.

The conditions, timing, and descriptions of the report are identical to those of the medical events described above.

Release of Patients Administered with Radiopharmaceuticals

According to 10CFR35.75, a licensee can release a patient administered with a radiopharmaceutical or a permanent radioactive implant, provided the TEDE to any other individual from exposure to the released patient is not likely to exceed 500 mrem (5 mSv). Practically in nuclear medicine, patients treated with $^{131}\text{I-NaI}$ are commonly considered under these regulations. In these cases, when the activity in the patient is less than 33 mCi (1.2 GBq) or the measured exposure rate is less than 7 mrem/hr (0.07 mSv/hr) at 1 meter, then the patient can be released. However, patients administered with higher ^{131}I -activities, as high as 200 mCi (7.4 GBq), may be released provided the dose calculations using patient-specific parameters show that the potential TEDE to any other individual would be no greater than 0.5 rem (5 mSv; NRC Regulatory Guide 8.39). The patient-specific calculations depend on the choice of the occupancy factor and the physical or effective half-life. An occupancy factor of 0.75 is chosen for $t_{1/2}$ of less than one day and a value of 0.25 for $t_{1/2}$ greater than one day. A value of 0.25 for the occupancy factor would be valid if the patient follows the instructions, such as, for the first two days, (1) maintain the distances from others; (2) sleep alone or, better yet, live alone; (3) do not travel by airplane or mass transportation; (4) do not travel in automobiles with others; (5) have the sole use of the bathroom; (6) drink plenty of water; and (7) limit visits by others. These instructions must be given in writing to the patients to follow after release.

The released patient must be given instructions, including written instructions, to maintain the dose as low as reasonably achievable if the TEDE to any other individual is likely to exceed 100 mrem (1 mSv). In the case of ^{131}I treatment, instructions must be given to the patient, when the activity in the patient is more than 7 mCi (259 MBq) or when the measured exposure rate exceeds 2 mrem/hr (0.02 mSv/hr) at 1 meter. If the dose to a breast-fed infant or child could exceed 100 mrem (1 mSv) assuming continuous breast-feeding by a patient administered with a radiopharmaceutical, then instructions on discontinuation of breast-feeding and consequences of failure to follow the guidance must also be given. Table 16.3 lists the activity limits for giving instructions to the breast-feeding patients and activity limits for cessation of breast-feeding.

TABLE 16.3. Limits of activities that require instructions to breast-feeding patients and recordkeeping.*

Radiopharmaceutical	Activity above which instructions are needed mCi (MBq)		Activity above which record is needed mCi (MBq)		Recommended duration of cessation of breast-feeding
¹³¹ I-NaI	0.0004	(0.01)	0.002	(0.07)	complete cessation
¹²³ I-NaI	0.5	(20)	3	(100)	—
¹³¹ I-MIBG	2	(70)	10	(400)	12 hr (4 mCi/150 MBq)
^{99m} Tc-DTPA	30	(1000)	150	(6000)	—
^{99m} Tc-MAA	1.3	(50)	6.5	(200)	12.6 hr (4 mCi/150 MBq)
^{99m} Tc-Perchnetate	3	(100)	15	(600)	12 hr (12 mCi/440 MBq)
^{99m} Tc-DISIDA	30	(1000)	150	(6000)	—
^{99m} Tc-Glucoheptonate	30	(1000)	170	(6000)	—
^{99m} Tc-Sestamibi	30	(1000)	150	(6000)	—
^{99m} Tc-MDP	30	(1000)	150	(6000)	—
^{99m} Tc-PYP	25	(900)	120	(4000)	—
^{99m} Tc-RBC in vivo	10	(400)	50	(2000)	6 hr (20 mCi/740 MBq)
^{99m} Tc-RBC in vitro	30	(1000)	150	(6000)	—
^{99m} Tc-Sulfur colloid	7	(300)	35	(1000)	6 hr (12 mCi/440 MBq)
^{99m} Tc-MAG3	30	(1000)	150	(6000)	—
^{99m} Tc-WBC	4	(100)	15	(600)	12 hr (12 mCi/440 MBq)
⁶⁷ Ga-Citrate	0.04	(1)	0.2	(7)	1 wk (0.2 mCi/7 MBq)
¹¹¹ In-WBC	0.2	(10)	1	(40)	1 wk (0.5 mCi/20 MBq)

* NRC regulatory Guide 8.39.

Records of release of patients are required, if the TEDE is calculated by using the retained activity rather than the administered activity, using an occupancy factor less than 0.25 at 1 meter, using a biological or effective $t_{1/2}$ or considering the shielding by tissue. Records are also required if instructions are given to a breast-feeding woman who may give a TEDE exceeding 500 mrem (5 mSv) to the infant from continuous breast-feeding (10CFR35.2075). Table 16.3 gives the activity limits that require record-keeping in the case of breast-feeding.

Recordkeeping

Records must be maintained for the receipt, storage, and disposal of radioactive materials, and also for various activities performed in the radiation laboratories. According to the NRC regulations, these records must contain specific information and be kept for a certain period of time. Table 16.4 lists different records that are required by the NRC and the period of time to be kept.

TABLE 16.4. Recordkeeping of various activities related to radioactive materials.

Type of operation	Information needed	Time to maintain the records
Written directives (10CFR35.2040)	Copy of the written directives	3 years
Procedures requiring written directives (10CFR35.2041)	Copy of the procedures	Duration of the license
Dosage of radiopharmaceuticals dispensed (10CFR35.2063)	Name, lot number, expiration date, patient's name or identification number, prescribed dosage and dispensed dosage, date and time of administration, and name of the individual	3 years
Calibration of dose calibrator (10CFR35.2060)	Model, serial number of the dose calibrator, date and results of test, and name of the individual	3 years
Calibration of survey meters (10CFR35.2061)	Model and serial number of the instrument, date and results of calibration, and name of the individual	3 years
Semiannual leak tests and inventory of sealed sources (10CFR35.2067)	Model and serial number of each source and its radionuclide, estimated activity, measured activity in μCi (Bq), date of test, location of source (inventory), and name of the individual	3 years
Moly breakthrough (10CFR35.2204)	μCi (kBq) of ^{99}Mo per mCi (MBq) of ^{99m}Tc , date and time of measurement, name of the individual	3 years
Thyroid bioassay and whole body counting (10CFR20.2106)	Name of the individual having the bioassay, date of reading, and the individual taking the measurement	Until the NRC terminates the license
Personnel exposure monitoring records (10CFR20.2106)	Must be on NRC-5 form according to items described in the form	Until the NRC terminates the license
Radioactive waste disposal by decay-in-storage (10CFR35.2092)	Date of disposal, instrument used, background reading, and surface reading of the waste container, and the name of the individual	3 years
Planned special procedures (10CFR20.2105)	Circumstances, name of authorizing individual, doses expected	Until the NRC terminates the license
Surveys (10CFR35.2070)	Date, area, trigger level (mR/hr), survey data, instrument used, and name of the individual	3 years
Release of patients with unsealed by-product material (10CFR35.2075)	Basis of calculation to release the patient, such as retained activity, occupancy factor less than 0.25 at 1 meter, using T_p or T_e , or considering shielding by tissue	3 years
Instructions given to breast-feeding female (10CFR25.2075)	Instruction given if dose to the infant exceeds 0.5 rem (5 mSv)	3 years

Transportation of Radioactive Materials

The transportation of radioactive materials is governed by the U.S. Department of Transportation (DOT), which establishes the guidelines for packaging, types of packaging material, limits of radioactivity in a package, and exposure limits. Title 49 of the *Code of Federal Regulations* (49CFR) and 10CFR71 contain all these regulations related to packaging and transportation of radioactive materials.

There are two types of packaging:

Type A: This type of packaging is used primarily for most radiopharmaceuticals. Such packaging is sufficient to prevent loss of radioactive material with proper shielding to maintain the prescribed exposure during normal transportation. The limits of radioactivities of various radionuclides under this category are specified in 49CFR and 10CFR71.

Type B: When the radioactivity exceeds the limits specified in Type A, Type B packaging must be used. Such packaging is considerably more accident resistant and is required for very large quantities of radioactive material.

The packages must pass certain tests such as the drop test, corner drop test, compression test, and 30-min water spray test.

The radioactive packages must be labeled properly before transportation. There are three types of labels (Fig. 16.3) according to the exposure reading in mR/hr at 1 m from the surface of the package (*transport index*). The criteria for the three labels are given in Table 16.3. The transport index (TI) must be indicated on the label and the sign “RADIOACTIVE” must be placed on the package. The maximum permissible TI value is 10, although it is limited to three for passenger-carrying aircrafts. For liquids, the label “THIS SIDE UP” must be placed on the package. Each package must be labeled on opposite sides with the appropriate warning label (one of the labels in Table 16.5). The label must identify the content and amount of radionuclide in curies or becquerels. The package must contain shipping



FIG. 16.3. The three types of U.S. Department of Transportation labels required for transportation of radioactive materials.

TABLE 16.5. Labeling categories for packages containing radioactive materials.

Type of label	Exposure (mR/hr)	
	At surface	At 1 m
White-I	<0.5	—
Yellow-II	>0.5 ≤ 50	<1
Yellow-III	>50 ≤ 200	>1 ≤ 10

Note: No package shall exceed 200 mR/hr at the surface of the package or 10 mR/hr at 1 m. Transport index is the reading in mR/hr at 1 m from the package surface (10CFR71).

documents inside, bearing the identity, amount, and chemical form of the radioactive material and the TI.

Placards are necessary on the transport vehicles carrying yellow-III-labeled packages and must be put on four sides of the vehicle.

According to 49CFR173.421, radionuclides are exempt from the packaging and labeling requirements if only a limited quantity of these radionuclides is shipped. The surface exposure readings should not exceed 0.5 mR/hr at all points of the package surface, and the wipe test indicates no removable contamination in excess of 6600 dpm/300 cm². The outside of the inner packaging or, if there is no inner packaging, the outside of the packaging itself bears the marking, “Radioactive”. The outside of the package must be marked with UN2910. If the inner packaging materials read less than 0.5 mR/hr for exposure and 6600 dpm/300 cm² for contamination, then the shipment is considered “Empty” and should be labeled “UN2908” on the outside. No shipping paper is required for both limited quantity or empty packages, if the material is not hazardous. The limited quantities for some important radionuclides are given in Table 16.6 based on 10⁻⁴ A₂ for liquids and 10⁻³ A₂ for solids and gases, as stated in 49CFR173.425. The values of A₂ are obtained from 49CFR173.435.

TABLE 16.6. Limited quantities of several radionuclides that are exempt from shipping and labeling requirements, according to 49CFR 173.425.

Radionuclides*	Quantity		Radionuclides*	Quantity	
	(mCi)	(MBq)		(mCi)	(MBq)
⁵⁷ Co (s)	270	800	¹¹¹ In (l)	8.1	300
⁶⁷ Ga (l)	8.1	600	³² P (l)	1.4	50
¹²³ I (l)	81	3000	^{99m} Tc (l)	11	400
¹²⁵ I (l)	8.1	300	²⁰¹ Tl (l)	11	400
¹³¹ I (l)	1.9	70	¹³³ Xe (g)	270	10,000
⁸⁹ Sr (l)	1.6	60	¹⁸ F (l)	1.6	60

* s = solid; l = liquid; g = gas.

Employees who ship hazardous material including radioactive material must have hazmat training to be able to recognize and identify hazardous material, to conduct their specific function, and to enforce safety procedures to protect the public. The training must be given to a new employee within 90 days of employment and then repeated every three years. The training is provided by the hazmat employer or other public or private sources, and a record of training must be made.

European Regulations Governing Radiopharmaceuticals

The European Union (EU) is a union of 25 member countries in Europe, which adopt various rules and regulations that are uniformly applicable to all member states. Regulation governing the use of ionizing radiations varied among member states until 1989 when the EU applied uniform regulations for radiopharmaceuticals to be implemented by each member state. The European Regulatory Organizations of the EU have three main instruments: directives, guidelines, and regulations. Directives are mandatory to be translated into national legislation and implemented in each member country. Guidelines are recommendations (not mandatory) for implementation of the directives by each member country. Regulations are mandatory for all member countries without adoption into individual national legislation.

Similar to the US Codes of Federal Register (CFR), the European Economic Council (EEC) directives are issued for regulation of all food and drugs. Directive 65/65 EEC defines medicinal products and details the requirements for registration for commercialization of the medicinal products by each member state. This directive has since been amended by 83/570 EEC, 87/21 EEC, and 89/381 EEC to include medicinal products derived from human blood or plasma. Directive 89/341 EEC exempts certain drugs from the registration requirements, which include those compounded by the pharmacy for patients and those for research and clinical trials.

Directive 75/319 EEC establishes the requirements for application for authorization to market a medicinal product, which is drawn up by qualified experts and approved by a Committee for Proprietary Medicinal Products. This directive facilitates the free movement of medicinal drugs among the member states, and sets guidelines for production of quality medicinal products using good manufacturing methods supervised by qualified personnel. This directive also controls the importation of medicinal products from a third country, particularly regarding the quality and integrity of the products. Until 1989, radiopharmaceuticals were exempt from these regulations, and then in 1990, Directive 89/343 EEC was adopted regulating all radiopharmaceuticals as to their production, quality control, packaging, and labeling.

A drug can be approved for marketing in either a centralized or decentralized way. For the centralized procedure of approval, the European Medicines Agency (EMA) was established in 1995, which controls the evaluation and supervision of medicines for human and veterinary use. Once a drug is considered to be safe, of high quality, and efficacious, a single market authorization is granted for the product for the entire EU. In the decentralized procedure, an application for drug marketing is made to one member state that approves or disapproves it after review. The applicant state can then present the authorization to other member states for registration to market the drug in those states.

Directive 91/356 EEC establishes the principles of good manufacturing practices (GMP), which apply to drug manufacturing at the industrial level as well as compounding in hospitals. These guidelines are parallel to those of the CGMP of the USFDA. For each approved radiopharmaceutical for marketing, a summary of product characteristics (SPC), equivalent to package inserts in the United States, must be included.

The European Pharmacopoeia specifies the characteristics of all radiopharmaceuticals in monograph forms, regarding radionuclidic and radiochemical purity, pH, sterility, pyrogenicity, and so on. It is mandatory in drug product manufacturing in the entire EU and equivalent to the USP in the United States.

Within the EU, it is the responsibility of the radiopharmacist to ensure the quality, safety, and efficacy of a radiopharmaceutical if it is intended for human administration. Similarly, the nuclear physician is responsible for the administration of a radiopharmaceutical to the patient and the clinical care of the patient for any adverse reactions thereof. Claims may be made against any adverse reaction for up to 10 years after the event, therefore the patient's and preparation's records must be maintained for this length of time.

Clinical trials are conducted by qualified clinicians based on protocols approved by an ethics committee in each institution. Data are collected on pharmacokinetic characteristics, clinical efficacy, safety, and the like for a particular indication of a disease, which are then submitted for marketing authorization. Any modification in a product or its administration requires a new clinical trial.

The radiation aspects of radiopharmaceuticals are regulated by directives from the European Atomic Energy Community (EURATOM). Initial Directives 84/466 EURATOM, and 84/467 EURATOM were based on the recommendations of the ICRP, and mandate regulations for radiation protection for patients, radiation workers, and the public. Because the ICRP revised the basic standards for radiation protection in 1996, Directive 84/467 EURATOM was repealed and substituted with Directive 96/29/EURATOM. Upon further revision of the basic standards by the ICRP, Directive 96/29/EURATOM has been amended by 97/43 EURATOM, and finally Directive 84/466/EURATOM was repealed. These

directives are similar to NRC 10CFR20 and 10CFR35 regulations in the United States and address all aspects of radiation protection to individuals involved with radiation. They include exposure limits to workers, patients, minors, pregnant women, nursing mothers, and members of the public, monitoring of radiation areas and working personnel, designation of controlled and supervised (restricted) areas, optimization of radiation exposure (similar to ALARA in the United States), survey of work areas, training and experience of the radiation workers and physicians, monitoring in the case of accidental or emergency exposure, recordkeeping, reporting of incidents, and so on. Note that the dose limits in these directives are the same as those adopted by the NRC in the United States. The disposal, recycling, or reuse of radioactive substances is required to have prior authorization, but may be exempt from these requirements if the clearance levels of radioactivity established by the member state comply with those of the EURATOM directive. Directive 90/642 EURATOM provides regulations for monitoring exposure to outside workers. Interstate shipment of radioactive materials between member states is governed by Directive 1493/93 EURATOM. Similar to an RSO in the United States, a radiation protection advisor (RPA) is an expert in radiation protection principles, who implements and supervises radiation safety regulations in institutions.

Although all EU regulations and directives are equally applicable to all member states, the actual situation differs from country to country, because of the lack of effective implementation of the rules in many states. So in some member states, these directives are effectively implemented, and in others they are leniently applied, and in some cases, there may be a breach of these community laws.

It should be noted that detailed information of different directives given above may be available from the EU Web site.

Questions

1. Define committed dose equivalent, deep-dose equivalent, total effective dose equivalent, radiation area, and high radiation area.
2. What are the annual dose limits for radiation workers for:
 - (a) Whole body
 - (b) Lens
 - (c) Extremities
3. What is the dose limit in the unrestricted area and for the individual members of the public?
4. (a) Calculate the exposure rate at 10 inches from a 150-mCi (5.55-GBq) ^{131}I source (Γ_{20} of ^{131}I = $2.17 \text{ R} \cdot \text{cm}^2/\text{mCi} \cdot \text{hr}$ at 1 cm).
 - (b) If the half-value layer (HVL) of lead for ^{131}I is 3 mm, how much lead is needed to reduce the exposure to 10% of the calculated value at 10 inches?

5. Why is ^{32}P stored in plastic and not in lead containers?
6. What is the approximate amount of lead necessary to reduce the exposure rate from a 200-mCi- $^{99\text{m}}\text{Tc}$ source to less than 5 mR/hr at 20 cm from the source? (Γ_{20} of $^{99\text{m}}\text{Tc} = 0.59 \text{ R} \cdot \text{cm}^2/\text{mCi} \cdot \text{hr}$ at 1 cm = $15.95 \mu\text{Gy} \cdot \text{m}^2/\text{GBq} \cdot \text{hr}$ at 1 m; HVL of Pb for $^{99\text{m}}\text{Tc} = 0.3 \text{ mm}$).
7. If 1% of the primary beam exits through a patient, calculate the exposure at the midline of the patient.
8. (a) Who are required to wear personnel monitoring devices?
 (b) Film badges can discriminate different types of radiation. True or false?
 (c) Film badges can discriminate radiations of different energies. True or false?
 (d) Why are filters used in film badges?
 (e) Filters convert radiation energies into visible light. True or false?
 (f) Filters protect the individual from radiation exposure. True or false?
 (g) Describe how thermoluminescent dosimeters work.
9. (a) What is the ALARA program?
 (b) What is an Agreement State?
 (c) How often should area surveys and wipe tests be performed in nuclear medicine?
 (d) When does one take a bioassay?
 (e) What are the NRC requirements for survey of the packages on receipt?
 (f) Describe different methods of disposal of radioactive waste.
 (g) What are the general principles of handling radioactive spillage?
 (h) What is a transportation index (TI), and how is it used in the transportation of radioactive material?
10. What are the criteria for the release of patients administered with radiopharmaceutical?

References and Suggested Readings

- Cox PH. European legislation and its effects on the production of radiopharmaceuticals. In: Sampson CB, ed. *Textbook of Radiopharmacy* 3rd ed. Amsterdam: Gordon and Breach Science Publishers; 1999.
- Federal Register. Code of Federal Regulations.* 10CFR20. Washington, DC: U.S. Government Printing Office; 1996.
- Federal Register. Code of Federal Regulations.* 10CFR31. Washington, DC: U.S. Government Printing Office; 1993.
- Federal Register. Code of Federal Regulations.* 10CFR33. Washington, DC: U.S. Government Printing Office; 1992.
- Federal Register. Code of Federal Regulations.* 10CFR35. Washington, DC: U.S. Government Printing Office; 2002.
- Federal Register. Code of Federal Regulations.* 10CFR71. Washington, DC: U.S. Government Printing Office; 1992.

- Federal Registers. Code of Federal Regulations.* 49CFR170 to 49CFR189. Washington, DC: U.S. Government Printing Office; 1989, 2004.
- Martin JE. *Physics of Radiation Protection.* Hoboken, NJ: Wiley Interscience; 2000.
- National Council on Radiation Protection and Measurements. *Basic Radiation Protection Criteria.* Bethesda, MD: NCRP Publication 39; 1971.
- National Council on Radiation Protection and Measurements. *Nuclear Medicine—Factors Influencing the Choice and Use of Radionuclides Diagnosis and Therapy.* Bethesda, MD: NCRP Publication 70; 1982.
- National Council on Radiation Protection and Measurements. *Ionizing Radiation Exposure of the Population of the United States.* Bethesda, MD: NCRP Publication 90; 1987.
- National Council on Radiation Protection and Measurements. *Radiation Protection and Allied Health Personnel.* Bethesda, MD: NCRP Publication 105; 1989.
- Shapiro J. *Radiation Protection.* 3rd ed. Cambridge, MA: Harvard University Press; 1990.
- U.S. NRC NUREG-1556, vol 9. *Consolidated Guidance about Materials Licences,* U.S. Government Printing Office; 2002.
- U.S. NRC Regulatory Guide 8.39. Washington, DC: U.S. Government Printing Office; 1997.

Appendix A

Units and Constants

Energy

1 electron volt (eV)	= 1.602×10^{-12} erg
1 kiloelectron volt (keV)	= 1.602×10^{-9} erg
1 million electron volts (MeV)	= 1.602×10^{-6} erg
1 joule (J)	= 10^7 ergs
1 watt (W)	= 10^7 ergs/s
	= 1 J/s
1 rad	= 1×10^{-2} J/kg
	= 100 ergs/g
1 gray (Gy)	= 100 rad
	= 1 J/kg
1 sievert (Sv)	= 100 rem
	= 1 J/kg
1 horsepower (HP)	= 746 W
1 calorie (cal)	= 4.184 J

Charge

1 electronic charge	= 4.8×10^{-10} electrostatic unit
	= 1.6×10^{-19} C
1 coulomb (C)	= 6.28×10^{18} charges
1 ampere (A)	= 1 C/s

Mass and Energy

1 atomic mass unit (amu)	= 1.66×10^{-24} g
	= 1/12 the atomic weight of ^{12}C
	= 931 MeV
1 electron rest mass	= 0.511 MeV
1 proton rest mass	= 938.78 MeV
1 neutron rest mass	= 939.07 MeV
1 pound	= 453.6 g

Length

1 micrometer, or micron (μm)	= 10^{-6} m
	= 10^4 Å

1 nanometer (nm)	= 10^{-9} m
1 angstrom (Å)	= 10^{-8} cm
1 fermi (F)	= 10^{-13} cm
1 inch	= 2.54 cm
<i>Activity</i>	
1 curie (Ci)	= 3.7×10^{10} disintegrations per second (dps) = 2.22×10^{12} disintegrations per minute (dpm)
1 millicurie (mCi)	= 3.7×10^7 dps = 2.22×10^9 dpm
1 microcurie (μ Ci)	= 3.7×10^4 dps = 2.22×10^6 dpm
1 becquerel (Bq)	= 1 dps = 2.703×10^{-11} Ci
1 kilobecquerel (kBq)	= 10^3 dps = 2.703×10^{-8} Ci
1 megabecquerel (MBq)	= 10^6 dps = 2.703×10^{-5} Ci
1 gigabecquerel (GBq)	= 10^9 dps = 2.703×10^{-2} Ci
1 terabecquerel (TBq)	= 10^{12} dps = 27.03 Ci
<i>Constants</i>	
Avogadro's number	= 6.02×10^{23} atoms/g · atom = 6.02×10^{23} molecules/g · mole
Planck's constant (h)	= 6.625×10^{-27} erg · s/cycle
Velocity of light	= 3×10^{10} cm/sec
π	= 3.1416
e	= 2.7183

Appendix B

Terms Used in Text

- Absorption.* A process by which the total energy of a radiation is removed by an absorber through which it passes.
- Accelerator.* A machine to accelerate charged particles linearly or in circular paths by means of an electromagnetic field. The accelerated particles such as α -particles, protons, deuterons, and heavy ions possess high energies and can cause nuclear reactions in target atoms by irradiation.
- Accuracy.* A term used to indicate how close a measurement of a quantity is to its true value.
- Annihilation radiation.* γ -Radiations of 511 keV energy emitted at 180° after a β^+ -particle is annihilated by combining with an electron in matter.
- Atomic mass unit (amu).* By definition, one twelfth of the mass of $^{12}_6\text{C}$, equal to 1.66×10^{-24} g or 931 MeV.
- Atomic number (Z).* The number of protons in the nucleus of an atom.
- Attenuation.* A process by which the intensity of radiation is reduced by absorption and/or scattering during its passage through matter.
- Attenuation coefficient.* The fraction of γ -ray energy attenuated (absorbed plus scattered) per unit length of an absorber (linear attenuation coefficient, μ) or per gram of an absorber (mass attenuation coefficient, μ_m).
- Auger electron.* An electron ejected from an energy shell, instead of a characteristic x-ray emission, carrying the energy equal to that of the x-ray minus its binding energy.
- Average life (τ).* See Mean life.
- Avogadro's number.* The number of molecules in 1 g · mole of any substance or the number of atoms in 1 g · atom of any element. It is equal to 6.02×10^{23} .
- Becquerel (Bq).* A unit of radioactivity. One becquerel is equal to 1 disintegration per second.
- Binding energy.* The energy to bind two entities together. In a nucleus, it is the energy needed to separate a nucleon completely from other nucleons in the nucleus. In a chemical bond, it is the energy necessary to separate two binding partners an infinite distance.

- Biological half-life (T_b)*. The time by which one half of an administered dosage of a substance is eliminated by biological processes such as urinary and fecal excretions.
- Bremsstrahlung*. γ -Ray photons produced by deceleration of charged particles near the nucleus of an absorber atom.
- Carrier*. A stable element that is added in detectable quantities to a radionuclide of the same element, usually to facilitate chemical processing of the radionuclide.
- Carrier-free*. A term used to indicate the absence of any stable atoms in a radionuclide sample.
- Collimator*. A device to confine a beam of radiation within a specific field of view. Collimators may be converging, pinhole, diverging, and parallel-hole types.
- Collimator efficiency*. The number of photons passing through the collimator for each unit of activity present in a source.
- Collimator resolution*. A component of spatial resolution of an imaging system contributed by the collimator. It is also called *geometric resolution*.
- Committed dose equivalent ($H_{T,50}$)*. The dose equivalent to organs or tissues of reference (T) that will be received from an intake of radioactive material by an individual during the 50-year period following intake.
- Compton scattering*. In this process, a γ -ray transfers only a partial amount of energy to an outer orbital electron of an absorber, and the photon itself is deflected with less energy.
- Conversion electron (e^-)*. See Internal conversion.
- Critical organ*. See Organ, critical.
- Cross section (σ)*. The probability of occurrence of a nuclear reaction or the formation of a radionuclide in a nuclear reaction. It is expressed in a unit termed *barn*; 1 barn = 10^{-24} cm².
- Curie (Ci)*. A unit of activity. A curie is defined as 3.7×10^{10} disintegrations per second.
- Dead time*. The period of time that a counter remains insensitive to count the next after an event.
- Decay constant (λ)*. The fraction of atoms of a radioactive element decaying per unit time. It is expressed as $\lambda = 0.693/t_{1/2}$, where $t_{1/2}$ is the half-life of the radionuclide.
- Deep-dose equivalent (H_d)*. Dose equivalent at a tissue depth of 1 cm (1000 mg/cm²) resulting from external whole-body exposure.
- Dose*. The energy of radiation absorbed by any matter.
- Dosimeter*. An instrument to measure the cumulative dose of radiation received during a period of radiation exposure.
- Dosimetry*. The calculation or measurement of radiation absorbed doses.
- Effective half-life (T_e)*. Time required for an initial administered dose to be reduced to one half as a result of both physical decay and biological elimination of a radionuclide. It is given by $T_e = (T_p \times T_b)/(T_p + T_b)$, where T_e

- is the effective half-life, and T_p and T_b are the physical and biological half-lives, respectively.
- Electron* (e^-). A negatively charged particle rotating around the atomic nucleus. It has a charge of 4.8×10^{-10} electrostatic units and a mass of 9.1×10^{-28} g, equivalent to 0.511 MeV, or equal to 1/1836 of the mass of a proton.
- Electron capture* (EC). A mode of decay of a proton-rich radionuclide in which an orbital electron is captured by the nucleus, accompanied by emission of a neutrino and characteristic x-rays or Auger electrons.
- Electron volt* (eV). The kinetic energy gained by an electron when accelerated through a potential difference of 1 V.
- Energy resolution*. Capability of a detecting system to separate two γ -ray peaks of different energies. It is given by the full width at half maximum (FWHM) of a given photopeak.
- Erg*. The unit of energy or work done by a force of 1 dyne through a distance of 1 cm.
- Fission* (f). A nuclear process by which a nucleus divides into two nearly equal smaller nuclei, along with the emission of two to three neutrons.
- Free radical*. A highly reactive chemical species that has one or more unpaired electrons.
- Generator, radionuclide*. A device in which a short-lived daughter is separated chemically and periodically from a long-lived parent adsorbed on adsorbent material. For example, ^{99m}Tc is separated from ^{99}Mo from the Moly generator with saline.
- Gray* (Gy). The unit of absorbed radiation dose in SI units. One gray is equal to 100 rad.
- Half-life* ($t_{1/2}$). A unique characteristic of a radionuclide, defined by the time during which an initial activity of a radionuclide is reduced to one half. It is related to the decay constant λ by $t_{1/2} = 0.693/\lambda$.
- Half-value layer* (HVL). The thickness of an absorbing material required to reduce the intensity or exposure of a radiation beam to one half of the initial value when placed in the path of the beam.
- Internal conversion*. An alternative mode to γ -ray decay in which nuclear excitation energy is transferred to an orbital electron, which is then ejected from the orbit.
- Intrinsic efficiency*. The number of radiations detected divided by the number of radiations striking the detector.
- Intrinsic resolution*. A component of the spatial resolution of an imaging system that is contributed by the detector and associated electronics and depends on the photon energy, detector thickness, and the number of PM tubes.
- Ion*. An atom or group of atoms with a positive charge (cation) or a negative charge (anion).
- Isobars*. Nuclides having the same mass number, that is, the same total number of neutrons and protons. Examples are $^{57}_{26}\text{Fe}$ and $^{57}_{27}\text{Co}$.

Isomeric transition (IT). Decay of the excited state of an isomer of a nuclide to a lower excited state or the ground state.

Isomers. Nuclides having the same atomic and mass numbers but differing in energy and spin of the nuclei. For example, ^{99}Tc and $^{99\text{m}}\text{Tc}$ are isomers.

Isotones. Nuclides have the same number of neutrons in the nucleus. For example, $^{131}_{53}\text{I}$ and $^{132}_{54}\text{Xe}$ are isotones.

Isotopes. Nuclides having the same atomic number, that is, the same number of protons in the nucleus. Examples are $^{14}_6\text{C}$ and $^{12}_6\text{C}$.

LD_{50/60}. A quantity of a substance that, when administered or applied to a group of any living species, kills 50% of the group in 60 days.

Linear energy transfer (LET). Energy deposited by radiation per unit length of the matter through which the radiation passes. Its usual unit is keV/ μm .

Mass defect. The difference between the mass of the nucleus and the combined masses of individual nucleons of the nucleus of a nuclide.

Mass number (A). The total number of protons and neutrons in a nucleus of a nuclide.

Mean life (τ). The average expected lifetime of a group of radionuclides before disintegration. It is related to the half-life and decay constant by $\tau = 1/\lambda = 1.44t_{1/2}$.

Metastable state (m). An excited state of a nuclide that decays to a lower excited or the ground state by isomeric transition with a measurable half-life.

Modulation transfer function. A quantitative value of the spatial resolution of an imaging system.

Neutrino (ν). A particle of no charge and mass emitted with variable energy during β^+ , and electron capture decays of radionuclides. An antineutrino ($\bar{\nu}$) is emitted in β^- decay.

No carrier added (NCA). A term used to characterize the state of a radioactive material to which no stable isotope of the compound has been added purposely.

Nucleon. A common term for neutrons or protons in the nucleus of a nuclide.

Organ, critical. The organ that is functionally essential for the body and receives the highest radiation dose after administration of radioactivity.

Organ, target. The organ intended to be imaged and expected to receive the greatest concentration of administered radioactivity.

Pair production. γ -Rays with energy greater than 1.02 MeV interact with the nucleus of an absorber atom, and a positron and an electron are produced at the expense of the photon.

Photoelectric effect. A process in which a γ -ray, while passing through an absorber, transfers all its energy to an orbital electron, primarily the K-shell electron of an absorber, and the photoelectron is ejected from the shell.

Photofraction. The fraction of all detected γ -rays that contributes to the photopeak.

Physical half-life (T_p). See Half-life.

- Precision.* A term used to indicate the reproducibility of the measurement of a quantity when measurements are made repeatedly.
- Quality factor (QF).* A factor dependent on linear energy transfer that is multiplied by absorbed doses to calculate the dose equivalents in rem. It is used in radiation protection to take into account the relative radiation damage caused by different radiations. It is 1 for x-, γ -, and β -rays and 10 for neutrons and protons.
- Rad.* The unit of radiation-absorbed dose. One rad is equal to 100 ergs of radiation energy deposited per gram of any matter, or 10^{-2} J/kg of any matter.
- Radiation weighting factor (W_r).* A factor that depends on the types of radiation and is used to convert rad to rem in radiation protection. $\text{Rem} = \text{rad} \times W_r$.
- Range.* The straight line distance traversed by a charged particle in an absorber.
- Relative biologic effectiveness (RBE).* A factor used to calculate the dose equivalent in rem from rad. It is defined as the ratio of the amount of a standard radiation that causes certain biological damage to the amount of radiation in question that causes the same biological damage.
- Roentgen.* The quantity of x- or γ -radiations that produces one electrostatic unit of positive or negative charge in 1 cm^3 of air at 0°C and 760-mm Hg pressure (standard temperature and pressure, STP). It is equal to $2.58 \times 10^{-4} \text{ C/kg}$ air.
- Roentgen equivalent man (rem).* A dose equivalent defined by the absorbed dose (rad) times the relative biological effectiveness or quality factor of the radiation in question.
- Sensitivity.* The number of counts per unit time detected by an imaging device for each unit of activity present in a source. It is expressed in $\text{cps}/\mu\text{Ci}$.
- Shallow-dose equivalent (H_s).* Dose equivalent at a tissue depth of 0.007 cm (7 mg/cm^2) averaged over an area of 1 cm^2 from external exposure to the skin.
- Sievert (Sv).* The SI unit of dose equivalent and equal to 100 rem.
- Spatial resolution.* A measure of the ability of an imaging device to faithfully reproduce the image of an object. It is given by the modulation transfer function (MTF) and is determined by the Fourier transform of the line spread function.
- Specific activity.* The amount of radioactivity per unit mass of a radionuclide or labeled compound.
- Specific ionization.* The number of primary and secondary ion pairs produced by an incident radiation per unit path length in an absorber.
- Thermal neutron.* Neutrons of thermal energy 0.025 eV.
- Tissue weighting factor (W_T).* The weighting factor of an organ or tissue is the proportion of risk of stochastic effects resulting from irradiation of that organ or tissue to the total risk of stochastic effects when the total body is irradiated uniformly.

Appendix C

Answers to Questions

Chapter 2

3. 81.3%
7. 130 keV

Chapter 3

1. (a) 1.11×10^{15} atoms
(b) $0.24 \mu\text{g}$
2. (a) 4.75×10^{14} dpm
(b) 216 Ci or 7.99×10^{12} Bq
3. 6.97 hr
4. (a) 429 mCi (15.9 GBq)
(b) 120.7 mCi (4.46 GBq)
5. 25.5 hr
6. 4.03 days
7. 6.4 mCi (237 MBq)
9. 330 min
10. 63%
11. 1.32 hr
12. N/2
13. 143.6 mCi (5.3 GBq)
14. 11 hr

Chapter 4

3. (a) 1707 ± 13.8 cpm
(b) 1647 ± 14.9 cpm
4. 40,000 counts
5. 3 standard deviations
6. 1111 counts
7. 90%

Chapter 5

8. 570.6 Ci (21.1 GBq)
9. 8.92 mCi (330 MBq)

Chapter 6

15. (a) 7.32 HVLs
(b) 8 HVLs
17. 10 HVLs
18. 2.31 cm
19. 2 mm

Chapter 8

8. (a) 25%
(b) 50%
14. 61.4%

Chapter 9

5. (c) 1911 counts/cm²

Chapter 10

6. (e) 0.35

Chapter 12

14. 0.84 cycles/cm

Chapter 14

1. 36,541 rad (365.4 Gy)
2. 350 rem (3.5 Sv)
7. 18,144 $\mu\text{Ci}\cdot\text{hr}$
8. 1.06×10^{-2} rad/ $\mu\text{Ci}\cdot\text{hr}$
9. 1.2 rad (1.2 cGy)
11. 75.5 rem (0.76 Sv)

Chapter 16

4. (a) 0.5 R/hr
(b) 6.96 mm Pb
6. 1.77 mm Pb
7. 10%

Index

A

Aberration, chromosome, 231
Absorbed dose, 208–222, 271–272
 annual limit on intake (ALI), 272
 committed dose equivalent,
 271, 303
 cumulated activity, 213
 cumulative dose, 212–216
 deep-dose equivalent, 271
 dose rate, 211–212
 effective dose, 220–222
 effective dose equivalent, 220–222
 shallow-dose equivalent, 271, 306
 SI unit, 216–217
 total effective dose equivalent,
 272
 radiation weighting factor, 209–210,
 306
 tissue weighting factor, 271
absorbed fraction, 212
Acceptance angle in PET, 188
Accuracy, 34, 302
Activity, 21, 26
 radioactivity, 21
 units of, 26
Acute effects of total body irradiation,
 247–249
Agreement states, 268
ALARA, program, 274
Alpha (α) decay, 14–15
Aluminum breakthrough test, 54
Analog-to-digital converter (ADC),
 142–143
Anger scintillation camera, see Gamma
 cameras

Annihilation radiations, 18, 92, 302
 coincidence detections in PET, 182,
 191
 escape peaks, 92
Annual limit on intake (ALI), 272
Antineutrino, 15
Atom, 3
 binding energy of electron of, 5
 composition of, 3
 electronic structure of, 3–5
 structure of the nucleus of, 6–7
Atomic mass unit, 3, 302
Atomic number, 6, 302
Attenuation of γ -radiations, 64–68
 half value layer, 65–68
 linear attenuation coefficient, 64–67
 mass attenuation coefficient, 65
 in PET, 197–199
 in SPECT, 171–175
 tenth value layer, 67
Auger electron, 14, 302
Auger process, 14
Authorized user, 288
Avalanche ionization, 73
Average value, 35
Avogadro's number, 24, 50, 302

B

Backprojection in tomography, 156–166
 filtered backprojection, 159–166
 in PET, 196
 in SPECT, 156–166
 simple backprojection, 156–158
Backscatter peak, 91
Backscattering of γ -rays, 62

- Bar phantom, 122–123
 Barium fluoride detector, 82–83
 Becquerel (Bq), 26, 302
 Beta (β^-) particle, 15
 average energy of, 15
 decay, 15
 energy spectrum of, 16
 range of, 58–59
 Binding energy, 5–8, 302
 of electron, 5
 nuclear, 7–8
 Bioassay, 279–280
 Biological half-life, 25, 303
 Bismuth germanate (BGO) detector, 82–83, 183
 Bit, 139
 Block detectors in PET, 184
 Bragg ionization peak, 57
 Bremsstrahlung, 60, 303
 Byte, 140
- C**
 ^{11}C (Carbon-11), 48
 Cadmium-zinc-tellurium detector, 85
 Calibration
 of dose calibrator, 75–77
 of high voltage or energy, in well counter, 101–102
 of survey instruments, 287
 of sources, 286
 Carcinogenesis by radiation, 249–253
 breast cancer, 253
 dose-response relationship, 251–253
 leukemia, 252–253
 other cancers, 253
 risk estimate, 252
 Carrier, 45, 303
 Carrier-free, 45, 303
 Cataractogenesis, 255
 Caution signs and labels, 272–273
 Cell survival curve, 238–245
 D_0 , 239–240
 D_q , 239–240
 effects of dose rate on, 241
 effects of LET of radiations on, 242
 effects of radioprotectors on, 244–245
 effects of radiosensitizers on, 242–244
 effects of stage of cell cycle, 245
 extrapolation number n , 239
 oxygen effect on, 242–244
 Cells, 226–229
 chromosome, 226
 composition of, 226
 cytoplasm, 226
 DNA synthesis, 228
 gene, 226
 meiosis, 228
 mitosis, 228
 nucleus, 226
 survival curve, 238–245
 Center of rotation in SPECT, 175–176, 178–179
 Central processing unit (CPU), 140–141
 Cerebrovascular syndrome, 248–249
 Cesium iodide (Cs(Tl) detector, 85
 Chain reaction, nuclear, 46
 Characteristic x rays, 14, 90–91
 peaks, 90–91
 Chart of the nuclides, 9
 Chi-square test, 39–41
 Chromosome aberration
 acentric fragment, 231–234
 deletion, 234–235
 dicentric fragment, 231–234
 inversion, 234–235
 restitution, 231
 translocation, 234
 Chromosome in cells, 226
 Coincidence circuit in liquid scintillation, 93
 Coincidence detection in PET, 182–188
 Collective effective dose in radiation biology, 259
 Collimators, 111–112, 303
 converging, 111
 diverging, 111
 efficiency of, 128–129, 303
 fan beam, 112
 holes, 111
 parallel hole, 111
 pinhole, 111
 resolution of, 119, 303
 septal thickness, 120–121
 for thyroid probe, 104
 Color quenching, 94

- Committed dose equivalent, 271, 303
 Compton edge, 90
 Compton electron, 62
 Compton plateau, 90
 Compton scattering, 62–63, 303
 Compton valley, 90
 Computers, 139–152
 bit, 139–140
 byte, 140
 central processing unit, 140–141
 computer memory, 141
 digital data acquisition, 144–146
 frame mode, 145
 list mode, 145
 digital images, 143–144
 digital-to-analog conversion, 143
 digitization of analog data, 142–143
 display, 148–149
 dynamic study, 146–147
 external storage devices, 141
 gated study, 147
 input/output devices, 141–142
 PACS, 150–152
 polar images, 149
 software, 149–150
 static study, 146
 word, 140
 Confidence level in statistics, 36
 Contrast, image, 132–133
 effect of noise, 133
 patient motion and, 133
 scatter radiations and, 133
 Control rods, reactor, 46
 Converging collimator, 111
 Conversion electron, *see* Internal conversion
 Convolution method in image reconstruction, 159–160
 Counting, statistics of, *see* Statistics of counting
 Counting of radioactivity, 88
 differential, 88
 integral, 88
 Cross section in nuclear reaction, 50, 303
 CT scanners, 170–171
 quality control tests of, 179–180
 Cumulated activity in dosimetry, 213
 Curie (Ci), 26, 303
 Cyclotron, 44–46
 cyclotron-produced radionuclides, 44–46
 equation for production of radionuclides, 49–51
 targets, 49
D
 Dead time, 99–101, 303
 GM counters, 99
 loss, 99
 in nonparalyzable systems, 100–101
 in paralyzable systems, 100–101
 pulse pileup, 99, 132
 gamma cameras, 132
 Decay constant, 21, 303
 Decay of radionuclides, 11–20
 alpha (α), 14–15
 beta minus (β^-), 15–17
 electron capture, 19–20
 equations, 21–25, 29–32
 general, 21–25
 successive, 29–32
 half-life, 22–25
 isomeric transition (IT), 12–14
 mean life, 25
 positron (β^+), 17–18
 scheme, 15
 secular equilibrium, 32
 spontaneous fission, 11
 successive decay, 29–32
 transient equilibrium, 30–31
 Decay scheme of radionuclide, 15
 Deep-dose equivalent, 271, 303
 Delta rays, 56
 Deoxyribonucleic acid, 226–231
 Department of Transportation, 293
 Detection efficiency, 96–99
 geometric efficiency, 97–99
 intrinsic efficiency, 97
 photopeak efficiency or photofraction, 97
 Detectors, 81–85
 bismuth germanate (BGO), 82
 cadmium-zinc-telluride (CZT), 85
 cesium iodide (CsI), 85
 gadolinium oxythosilicate (GSO), 84
 gas-filled, 71–80

- germanium and silicon, 84–85
 - in PET scanners, 183
 - liquid scintillation, 93–95
 - lutetium oxyorthosilicate (LSO), 84
 - of gamma cameras, 110
 - semiconductor, 84–85
 - sodium iodide (NaI(Tl)), 82, 86,
 - thyroid probe, 104–105
 - well type, 101–104
 - yttrium oxyorthosilicate (YSO), 84
 - Deterministic effects in radiation
 - biology, 246–247
 - Diagnostic tests, evaluation of, 41–43
 - negative predictive value, 42–43
 - positive predictive value, 42–43
 - sensitivity, 42–43
 - specificity, 42–43
 - Differentiated cells, 237
 - Digital camera, 115–116
 - Digital computers, *see* Computers
 - Digital image, 143–144
 - matrix sizes in, 143–144
 - zoom factor, 144
 - Dilution quenching, 94
 - Direct and indirect action of radiation, 235–237
 - Dirty bomb, 260–262
 - protective measure, 262
 - Disintegration rate, 21
 - Display, 114, 148–149
 - Distance in radiation protection, 275–276
 - Diverging collimator, 111
 - Dos and don'ts in radiation protection, 279
 - Dose
 - calculation, 211
 - calibrators, 74–77
 - accuracy, 75–76
 - constancy, 75–76
 - geometry, 77
 - linearity, 75–76
 - effect of, on cell survival, 241
 - equivalent, 220–222
 - limits, occupational, 273–274
 - rate, 211
 - Dose-response relationship in radiation
 - biology, 251–252
 - Dosimeter, 277–279, 303
 - film badge, 277–278
 - pocket, 77,
 - thermoluminescent, 278–279
 - Dosimetry, internal radiation, *see* Internal radiation dosimetry
 - Double-escape peak, 92
 - Double-strand break in chromosome, 231–235
 - Doubling dose, 256–257
 - Dynode of photomultiplier tube, 86–87
- E**
- EC, *see* Electron capture
 - Edge packing, 131
 - Effective dose, 220–222
 - Effective dose equivalent, 220–222
 - Effective half-life, 25, 303
 - Effects of radiation, 230–237
 - chromosome, 231–235
 - direct and indirect actions of radiation, 235–237
 - DNA molecule, 230–231
 - Electromagnetic radiations, 2
 - frequency, 2
 - wavelength, 2
 - Electron, 3, 304
 - Electron capture (EC) decay, 19–20, 304
 - Electron volt, 2, 304
 - Electronic collimation in PET, 182
 - Electronic structure of atom, 3–6
 - Electrons, 3
 - auger, 14
 - binding energies, 5
 - conversion, 13
 - Embryonic damage by radiation, 255–256
 - Emission computed tomography, 153
 - PET, 182–207
 - SPECT, 153–181
 - Energy and mass, 1
 - Energy resolution, 95–96, 304
 - Equilibrium dose constant, 212
 - Erg, 2, 304
 - Error
 - propagation of, 37–38
 - random, 34
 - systematic, 34

Escape peak
 double, 92
 iodine, 91–92
 single, 92
 Excitation of atoms, 5, 56
 Excited states, 12
 Exponential decay, 22
 Exposure rate constants, 275
 Exposure rate, survey of, 287
 External storage devices in computers,
 141

F

¹⁸F (Fluorine-18), 48
 Fetus, effects of radiation on, 255–256
 Film badge for personnel monitoring,
 277–278
 Filter, ramp, in computed tomography,
 163
 Filtered backprojection, 159–166
 Fission, 46–47
 and radionuclide production, 47
 spontaneous, 11
 Fluorescence yield, 14
 Forces, nuclear, 7
 Fourier method in image
 reconstruction, 160–166
 Frame mode of data acquisition,
 144–145
 Free radical, 236–237, 304
 Frequency of electromagnetic
 radiation, 2
 Frequency domain, 160
 Full width at half maximum (FWHM),
 95–96, 121, 124

G

⁶⁷Ga (Gallium-67), 48
⁶⁸Ga (Gallium-68), 48
 Gain, amplification, 87
 Gamma cameras, 108–117
 digital camera, 115–116
 performance parameters, 118–138
 quality control tests, 133–136
 solid state digital camera, 116
 tomography with, 153
 Gamma camera tuning, 131
 Gamma ray emission, 12
 Gamma well counter, 101–104

Gas-filled detectors, 71–80
 amplification, 71
 ionization chambers, 74–77
 Cutie pie survey meter, 74
 dose calibrator, 74–77
 pocket dosimeter, 77
 Geiger-Müller counter, 77–79
 Geiger region, 73
 proportional region, 73
 region of continuous discharge, 74
 region of limited proportionality,
 73
 region of recombination, 71
 region of saturation, 71
 Gastrointestinal syndrome, 248
 Gaussian distribution, 35–36
 Geiger-Müller (GM) counters, 77–79
 avalanche, 73, 77
 counting efficiency, 79
 dead time, 79
 quenching in, 77–78
 General domestic licenses, 270
 Generators, radionuclide, 51–54, 304
 aluminum breakthrough, 54
⁹⁹Mo-^{99m}Tc generator, 53–54
 molybdenum breakthrough, 54
 Genetic effects of radiation, 256–258
 doubling dose, 256–257
 genetically significant dose,
 257–258
 spontaneous mutations, 256
 Geometric efficiency, 97–99, 128–129
 in photopeak detection, 97–99
 in scintillation counter, 128–129
 Germanium and silicon detectors,
 84–85
 Gray (Gy), 209, 304

H

³H (Tritium), 48
 Half-life, in radioactive decay, 22–25,
 304
 biological, 25, 303
 definition, 22
 effective, 25, 303
 physical, 25, 305
 Half-value layer (HVL), 65–67, 304
 Hemopoietic syndrome, 247–248
 High count rates, effects of, 132

Hospital Information System (HIS),
151

Hybrid gamma camera, 191

I

¹²³I (iodine-123), 48

¹²⁴I (iodine-124), 48

¹²⁵I (iodine-125), 48

¹³¹I (iodine-131), 48

decay scheme of, 16

Image distortion by converging and
diverging collimators, 111

Image reconstruction, 156–169

filtered backprojection, 159–166

iterative method, 166–169

simple backprojection, 156–158

¹¹¹In (indium-111), 48

decay scheme of, 19

Indirect action of radiation, 235–237

Input/output devices, 141–142

Interaction of radiations with matter,
56–70

attenuation of γ -radiations, 64–69

interaction of charged particles,
56–60

interaction of γ -radiations, 60–64

interaction of neutrons, 68

Internal conversion, 13–14, 304

characteristic x-ray, 14

conversion electron, 13

Internal radiation dosimetry, 208–225

absorbed fraction, 211–212

cumulated activity, 213

dose calculation, 211

dose rate, 211–212

effective dose, 220–222

effective dose equivalent, 220–222

equilibrium dose constant, 212

mean absorbed dose, 213

radiation dose in SI units, 216–217

radiation weighting factor, 209

tissue weighting factors, 221

Intrinsic efficiency, 97, 304

Intrinsic resolution of gamma camera,
118–119, 304

Iodine escape peak, 91–92

Ion, 5, 304

Ionization, 5–6, 56

by charged particles, 56

Ionization chamber, 74–77

Cutie pie survey meter, 74

dose calibrator, 74–77

pocket dosimeter, 77

Ionizing radiations, 56

Isobars, 8, 304

Isomeric transition, 12–14, 305

gamma ray emission, 12

internal conversion, 13–14

Isomers, 8, 305

Isotones, 8, 305

Isotopes, 8, 305

Iterative method in image

reconstruction, 166–169

MLEM, 168–169

OSEM, 168–169

K

K absorption edges, 61

K shell, 3–5, 61

K shell binding energies, 5

Kerma, 209

Kinetics of radioactive decay, 21–33

L

Labeling of vials and syringes, 286–287

LD_{50/60}, 247, 305

Lead shielding, 276–277

Lead x-ray peak, 90

Lethal damage in radiation biology,
245–246

Licenses, 270–271

general domestic license, 270

specific licenses, 270–271

Life-shortening effect of radiation, 254

Line of response in PET, 191

Line of stability of nuclides, 7

Line spread function (LSF), 124

Linear amplifier, 87

Linear attenuation coefficient of
photons in absorber, 65

Linear energy transfer (LET), 57–58,
305

definition, 57–58

effects of, in radiation biology, 242

Linear no threshold theory, 251–252

Liquid scintillation counters, 93–95

counting efficiency, 95

primary solute, 93–94

Liquid scintillation counters (*cont.*):

- quenching, 94
- secondary solute, 94
- solubilizing agent, 94
- solvent, 94
- wavelength shifter, 94

List mode data acquisition, 144–145

Lithium-drifted detectors, 84

Long-term effects of radiation, 249–258

- genetic effects, 256–258
- somatic effects, 249–256

Lutetium oxyorthosilicate detector, 83–84

M

Magic numbers, 6

Major organogenesis, effects of radiation, 255

Mass attenuation coefficient, 65

Mass defect in nucleus, 8, 305

Mass and energy, 1

Mass number, 6, 305

Matrix in computers, 143, 144

Mean absorbed dose per cumulated activity, 213

Mean life of radionuclide, 25, 305

Mean range, 59

Mean value, 35

Measurement of dosages, 285–286

Medical cyclotron, 45

Medical event, 288–289

Medical uses of radioactive materials, 283–292

Meiosis, 228

Metastable states of nucleus, 12, 305

Micro-PET, 190

Minimum detectable activity, 41

Mitosis, 228

Mobile nuclear medicine service, 285

Mobile PET or PET/CT, 190

⁹⁹Mo-^{99m}Tc generators, 53–54

aluminum breakthrough, 54

⁹⁹Mo breakthrough, 53–54

Moderators in reactors, 46

Modulation transfer function (MTF), 125–127, 305

⁹⁹Mo (molybdenum-99), 48

breakthrough, 53–54

decay scheme of, 17

fission product, 48

-^{99m}Tc generator, 53–54

Monitoring, personnel, 277–279

Multichannel analyzer (MCA), 88

Mutations, 231, 256–257

and doubling dose, 256–257

genetic, 256

in DNA molecule, 231

spontaneous, 256

N¹³N (nitrogen-13), 48

NaI(Tl) detector, 82–83, 86, 101, 104, 110

properties of, 83

in gamma cameras, 110

in gamma well counters, 101

in thyroid probes, 104

NaI(Tl) well counter, 101–104

detection efficiency of, 104

effect of sample volume, 103

energy calibration, 101–102

NCA (no carrier added), 45

Negative predictive value, 42

Neutrino, 17, 19, 305

Neutrons, 6

characteristics of, 3

fast, 46

interaction of, in matter, 68

mass of, 3

thermal, 46

Neutron activation analysis, *see*

Neutron capture reaction

Neutron capture reaction, 47

No carrier added (NCA), 45

Noise

background in LS counting, 95

effect of, on image contrast, 133

Noise equivalent count rate, 204–205

Nonparalyzable counting systems, 100–101

Nonpenetrating radiations, 56

Normalization in PET, 196–197, 205–206

NRC (Nuclear Regulatory Commission), 268

Nuclear binding energy, 7

Nuclear chain reaction, 46

Nuclear fission, 46

- Nuclear forces, 7
- Nuclear reactors, 46
 - equation for production of radionuclides, 49–51
 - operation of, 46
 - production of radionuclides in, 46–47
- Nuclear Regulatory Commission, 268
- Nucleons, 3, 305
- Nucleus, 3, 6–8
 - binding energy of, 8
 - line of stability, 7
 - liquid drop model of, 6
 - notation for, 6
 - shell model of, 6
 - size of, 3
- Nuclides, 8
 - chart of, 8–9

- O**
- ¹⁵O (oxygen-15), 48
- Occupational dose limits, 273–274
- Orbital electrons, 3–4
- Organ, critical, 305
- Organ, target, 305
- Oxygen effect, 242–244
- Oxygen enhancement ratio, 242

- P**
- ³²P (phosphorus-32), 48
- Packaging in transportation of radioactive material, 293–294
- PACS, 150–152
- Pair production, 63–64, 305
- Parallel hole collimators, 111–112
 - classification of, 112
 - resolution, 119–122
 - sensitivity, 127
- Paralyzable dead time, 100–101
- Partial volume effect, 175
- Particulate radiations, definition of, 2
- Patient motion, 133
- Penetrating radiations, 56
- Personnel monitoring, 277–279
- PET/CT scanners, 189–190
- PET scanners, 184–188
 - block detectors, 184
 - coincidence timing window, 184–188
 - quality control tests, 205–206
- Photodisintegration, 64
- Photoelectric effect, 61–62
- Photofraction, 97, 305
- Photomultiplier (PM) tubes
 - in gamma cameras, 112
 - in LS counting, 93–94
 - in PET scanner, 183–184
 - in scintillation counters, 86–87
- Photons
 - annihilations, 60, 92
 - attenuation of, 64–68
 - Compton scattering of, 62–63
 - definition of, 2
 - pair production of, 63, 305
 - photoelectric effect of, 61–62, 305
- Photopeak
 - detection efficiency, 96–99
 - energy resolution of, 95–96
- Physical half-life, 22–25, 305
- Pinhole collimator, 111
- Pixels, 143
- Placards in transportation of radioactive material, 294
- Planck's constant, 2
- Pocket dosimeter, 77
- POPOP, in liquid scintillation, 94
- Positive predictive value, 42
- Positrons (β^+)
 - annihilation of, 17–18
 - decay, 17–18, 60
 - range, 202
- Positron emission tomography (PET), 182–206
 - angle of acceptance, 188
 - block detectors in, 184
 - coincidence detection in, 182, 191
 - coincidence timing window, 184–188
 - data acquisition, 191–193
 - dead time, 200
 - detectors in, 182
 - electronic collimation, 182
 - image, reconstruction, 196
 - line of response, 191
 - noise equivalent count rate, 204–205
 - normalization, 196–197
 - photon attenuation correction, 197–199
 - positron range, 202
 - quality control tests, 205
 - radial elongation, 200–201

Positron emission tomography (PET)

(cont.):

- random coincidences, 199
- scatter coincidences, 200
- sensitivity, 204
- sinogram, 192
- spatial resolution, 201–204
- Potentially lethal dose in radiation
 - biology, 245–246
- Preamplifiers, 87
- Precision, 34, 306
- Predictive value, 42
- Preimplantation stage, effects of
 - radiation, 255
- Prodromal stage, 247
- Propagation of errors, 37–39
- Proportional counters, 73
- Protons, 3, 6
 - atomic number, 6
 - properties of, 3
- Pulse height analyzers (PHA), 87–88
 - discriminator settings, 88
 - gamma cameras, 114
 - liquid scintillation counter, 93
 - multichannel, 88
 - single channel, 88
- Pulse pileup, 99

Q

Quality control tests

- CT scanners, 179–180
- dose calibrators, 75–77
- gamma cameras, 133–136
- extrinsic method, 134
- intrinsic method, 134
- PET, 205–206
- positioning of photopeak, 134–135
- SPECT, 178–179
- uniformity, 135–136
- Quality factor (QF), 210, 306
- Quantum number, 3–4
- Quenching
 - in GM counters, 77–78
 - in liquid scintillation counting, 94

R

- R (roentgen), 208
- ⁸²Rb (rubidium-82), 48
- Rad, 208, 306

Radiation(s)

- definition of, 1
- electromagnetic, 2
- genetic effects of, 256–258
- ionizing, 56
- nonpenetrating, 56
- particulate, 2
- penetrating, 56
- somatic effects of, 249–256
- Radiation area
 - high, 272
 - very high, 272
- Radiation damage to embryo and fetus, 255–256
- Radiation damage to reproductive organs, 254
- Radiation damage to skin, 253–254
- Radiation detectors, *see* Detectors
- Radiation dose, *see* Internal radiation dosimetry
- Radiation dosimetry, internal, *see* Internal radiation dosimetry
- Radiation effects
 - acute, 247–249
 - carcinogenesis, 249–253
 - cataractogenesis, 255
 - cell survival curves, 238
 - cerebrovascular syndrome, 248–249
 - chemicals, effects of, 242–245
 - chromosome aberration, 231
 - differentiated cells and, 237
 - direct action, 235–237
 - dose rate and, 241
 - doubling dose, 256–257
 - gastrointestinal syndrome, 248
 - genetic effects, 256–258
 - genetically significant dose, 257–258
 - hemopoietic syndrome, 247–248
 - indirect action, 235–237
 - in utero, 255
 - linear energy transfer and, 242
 - long-term, 249–258
 - oxygen effect, 242–244
 - radioprotectors, 244–245
 - radiosensitizers, 242–244
 - somatic effects, 249–256
 - spontaneous mutation, 256
 - stage of cell cycle, 245

- tissue sensitivities, 237–238
- undifferentiated cells, 245
- Radiation exposure
 - sources, 268–270
 - types of, in dirty bomb, 261–262
- Radiation phobia, 263–264
- Radiation protection, 271–282
 - activity, 277
 - ALARA, 274
 - definition of terms, 271–272
 - distance, 275–276
 - dos and don'ts in, 279
 - film badge, 277–278
 - personnel monitoring, 277–279
 - principles of, 274–277
 - shielding, 276–277
 - sources of radiation exposure, 268–270
 - thermoluminescent dosimeter, 278–279
 - time, 274
- Radiation regulations
 - agreement state, 268
 - ALARA, 274
 - bioassay, 279–280
 - caution signs and labels, 272–273
 - Department of Transportation, 293–295
 - limited quantity, 294
 - training, 295
 - high radiation area, 272
 - license, 270–271
 - general domestic license, 270
 - specific license, 270–271
 - medical uses of radionuclides, 283–292
 - Nuclear Regulatory Commission (NRC), 268
 - personnel monitoring, 277–279
 - principles of radiation protection, 274–277
 - radiation safety committee, 271, 284
 - radiation safety officer, 271, 284, 288
 - radioactive spill, 282
 - radioactive waste disposal, 280–282
 - receiving and monitoring of
 - radioactive packages, 280
 - recordkeeping, 291–292
 - release of patients administered with radiopharmaceuticals, 290–291
 - report and notification of a dose to an embryo/fetus or a nursing child, 289–290
 - report and notification of a medical event, 288–289
 - requirement for possession of sealed sources, 286
 - surveys for ambient radiation
 - exposure rate, 287
 - survey instruments, calibration, 287
 - training and experience
 - requirements for medical uses of byproduct materials, 287–288
 - transportation of radioactive material, 293–295
- Radiation Safety Committee, 271, 284
- Radiation Safety Officer, 271, 284, 288
- Radiation units, 208–210
- Radiation weighting factor, 209–210, 306
- Radioactive decay, 11–20, 21–33
 - alpha particle (α), 14–15
 - annihilation radiation in, 18
 - antineutrino in, 15
 - auger electron, 14
 - auger process, 14
 - beta minus (β^-), 15–17
 - characteristic x-ray, 14, 20
 - conversion electron, 13
 - decay constant, 21
 - decay equations, 21–22
 - effective half-life, 25
 - electron capture (EC), 19–20
 - fluorescence yield, 14
 - half-life, 22–25
 - internal conversion, 13
 - isomeric transition, 12–14
 - mean life, 25
 - of mixed radionuclide sample, 24
 - neutrino in, 17–19
 - positron (β^+), 17–18
 - secular equilibrium, 32
 - specific activity, 26–27
 - spontaneous fission, 11
 - successive decay, 29–32
 - transient equilibrium, 30–31
 - transition energy, 15

- Radioactive packages, receiving and monitoring, 280
- Radioactive spill, 282
- Radioactive waste disposal, 280–282
 decay in storage, 281
 incineration, 282
 other disposal methods, 282
 release into sewerage system, 281
 transfer to authorized recipient, 281
- Radiology Information System (RIS), 150
- Radionuclide generators, 51–54
⁹⁹Mo–^{99m}Tc generator, 53–54
- Radioprotectors, 244–245
- Radiosensitivity of cells and tissues, 237–238
- Radiosensitizer, 243–244
- Ramp filter in computed tomography, 163
- Random coincidences in PET, 199
- Random error, 34
- Ranges of charged particles in absorber, 58–59, 306
- Reactors, nuclear, 46–47
 chain reaction, 46
 control rod, 46
 equation for production of radionuclides, 49–51
 fission, 46
 moderators, 46
 neutron capture reaction, 47
 targets, 49
- Receiving of radioactive packages, 280
- Reconstruction, image, 156–169
 backprojection, 156–166
 iterative, 166–169
- Recordkeeping in NRC regulations, 282, 291–292
- Region of recombination in gas ionization, 71
- Relative biological effectiveness, 209, 306
- Release of patients administered with radiopharmaceuticals, 209–291
- Rem, 209, 306
- Report and notification of a dose to an embryo/fetus or a nursing child, 289–290
- Report and notification of a medical event, 288–289
- Requirement for possession of sealed sources, 286
- Restitution, 231
- Restricted areas, 272
- Risk of pregnant women, 259–260
- Roentgen (R), 208, 306
- S**
- S, mean absorbed dose, 213
- ⁸²Sr (strontium-82), 48
- Sampling in SPECT, 176
- Saturation factor in production of radionuclides, 50
- Scaled sources, requirement for possession of, 286
- Scatter radiation
 Compton, 62–63
 effect of, on spatial resolution, 121
 in PET, 200
 in SPECT, 176
- Scatter resolution, 121
- Scintillation detectors, 81–85
 detection efficiency, 96–99
 energy resolution, 95–96
 geometric efficiency, 97–99
 intrinsic efficiency, 99
 intrinsic resolution, gamma camera, 118–119
 photofraction, 97
 photopeak efficiency, 97
 for gamma camera, 128–129
 for thyroid probe, 104
 for well counter, 104
- Secular equilibrium in successive decay, 32
- Semiconductor detectors, 84–85
- Sensitivity, 306
 in PET, 204
 in scintillation camera, 128–129
 in SPECT, 177
 of diagnostic tests, 42
- Septal penetration, 120–121
- Septal thickness, 120
- Sewerage disposal of radioactive waste, 281
- Shallow dose equivalent, 271, 306
- Shell model of nucleus, 6

- Shielding in radiation protection, 276–277
- Sievert (Sv), 210, 306
- Single-strand break in chromosome, 231
- Single channel analyzer (SCA), 88
- Single escape peak, 92
- Single photon emission computed tomography (SPECT), *see* SPECT
- Sinogram in PET, 192
- SI units, *see* System Internationale Unit
- Sodium iodide detectors, *see* Detectors
- Software and DICOM, 149–150
- Solid state digital camera, 116
- Somatic effects of radiations, 249–256
- Sources of radiation exposure, 268–270
- Spatial resolution, 118–127, 306
 - collimator resolution, 119–121
 - evaluation of, 122–127
 - bar phantoms in, 122–123
 - line spread function in, 124
 - modulation transfer function in, 125–127
 - in PET, 201–204
 - in SPECT, 177
 - intrinsic resolution, 118–119
 - scatter resolution, 121
- Specific activity, 26–27, 306
- Specificity of diagnostic tests, 42
- Specific ionization, 57, 306
- Specific license, 270–271
- SPECT
 - attenuation correction, 173–175
 - center of rotation, 175–176
 - data acquisition, 155–156
 - factors affecting, 171–176
 - partial volume effect, 175
 - performance of SPECT camera, 177–178
 - photon attenuation, 171–175
 - quality control tests for SPECT cameras, 178–179
 - reconstruction of image, 156–169
 - filtered backprojection, 159–166
 - iterative reconstruction, 166–169
 - simple backprojection, 156–158
 - scattering, 176
 - sensitivity, 177
 - spatial resolution, 177
- SPECT camera, 153–154
- SPECT/CT, 169–171
- Spectrum of γ -rays, 88–89
 - ideal, 88–89
 - actual, 89
- Spontaneous discharge in gas ionization, 74
- Spontaneous fission, 11
- Spontaneous mutations, 256
- Standard deviation, 35
- Statistics of counting, 34–41
 - accuracy, 34
 - Chi-square test, 39–41
 - errors, 34
 - Gaussian distribution, 35–36
 - minimum detectable activity, 41
 - percent standard deviation, 36
 - precision, 34
 - propagation of errors, 37–39
 - standard deviation of counts, 35
 - standard deviation of count rates, 37
- Stochastic effects in radiation biology, 246
- Straggling of range, 58
- Structure of the nucleus, 6–7
- Sublethal damage in radiation biology, 245
- Successive decay of radionuclides, 29–32
- Surveys of ambient radiation exposure rate, 287
- Survey meters
 - Cutie pie, 74
 - Geiger-Müller, 77–79
- Survival curves, cell, *see* Cell survival curve
- System Internationale (SI) unit
 - becquerel, 26
 - gray, 209, 306
 - sievert, 210, 306
- Systematic errors, 34
- T**
- ^{99m}Tc (Technetium-99m), 48
 - decay scheme of, 12
 - exposure rate constant of, 275
 - generator, 53–54
 - half-value layer in lead, 67, 304

Target in production of radionuclides, 49

Target, organ, 305

Tenth value layer (TVL), 67

Terms used in the text, 302–306

Thermal neutrons, 46, 306

Thermoluminescent dosimeter (TLD), 278–279

Thyroid probe, 104–105

Thyroid uptake, 105

Tissue weighting factors, 220–221, 271–272, 306

²⁰¹Tl (thallium-201), 49

exposure rate constant, 275

half-value layer in lead, 67

production, 49

Tomographic imaging, 153

emission computed tomography, 153

PET, 182–207

SPECT, 153–181

transmission tomography, 153

Total effective dose equivalent, 272

Training and experience requirements for medical uses of byproduct materials, 287–288

Transient equilibrium in radioactive decay, 30–31

Transition energy in radioactive decay, 15

Transport index, 293–294

Transportation of radioactive materials, 293–295
limited quantity, 294

U

Undifferentiated cells, 237

Uniformity, gamma camera, 129–131
edge packing, 131

nonlinearity, 130

pulse height variation, 129–130,

Units and constants, 300–301

Unrestricted area, 272

Uranium-235, fission, 47

V

Verification cards for radioactive patients, 262–263

W

Waste disposal, radioactive, 280–282

Wavelength of electromagnetic radiation, 2

Wavelength shifter in liquid scintillation, 94

Well counter, gamma, 101–104
energy calibration of, 101–102
photopeak efficiency for, 104
sample volume effects, 103

Window setting, in pulse height analysis, 87–88

Wipe test, 280, 282, 287

Written directives, 285

X

X rays, characteristic, 14, 20

X, Y positioning circuit, 112–114

Y

Yttrium oxyorthosilicate detector, 83–84

Z

Zoom factor, 144

Z-pulse in gamma camera, 112, 114